



# COVERING NOTE

---

GENERAL SECRETARIAT INTERNATIONAL TELECOMMUNICATION UNION

---

Geneva, 27 January 2005

ITU – TELECOMMUNICATION  
STANDARDIZATION SECTOR

**Subject: Corrigendum 1 (01/2005) to**

ITU-T Recommendation P.57 (07/2002), *Artificial ears*

*In clause 5.3.3, replace the following text:*

"The pinna simulator shall be made from a high-quality elastomer, the shore-A hardness of which, measured at the surface 15 mm forward to the ear canal opening should be  $55 \pm 10^\circ$  Shore-00 (reference ASTM D2240 / DIN53505)."

*by:*

"The pinna simulator shall be made from a high-quality elastomer, the hardness of which, measured at the surface 15 mm forward to the ear canal opening should be  $55 \pm 10$  Shore 00 (reference ASTM D2240 / DIN53505)."