

International Telecommunication Union

ITU-T

TELECOMMUNICATION
STANDARDIZATION SECTOR
OF ITU

H.830.8

(01/2015)

SERIES H: AUDIOVISUAL AND MULTIMEDIA SYSTEMS

E-health multimedia services and applications –
Interoperability compliance testing of personal health
systems (HRN, PAN, LAN, TAN and WAN)

**Conformance of ITU-T H.810 personal health
devices: WAN interface Part 8: Consent
Management: Receiver**

Recommendation ITU-T H.830.8



ITU-T H-SERIES RECOMMENDATIONS
AUDIOVISUAL AND MULTIMEDIA SYSTEMS

| | |
|----------------------------------------------------------------------------------------------------|--------------------|
| CHARACTERISTICS OF VISUAL TELEPHONE SYSTEMS | H.100–H.199 |
| INFRASTRUCTURE OF AUDIOVISUAL SERVICES | |
| General | H.200–H.219 |
| Transmission multiplexing and synchronization | H.220–H.229 |
| Systems aspects | H.230–H.239 |
| Communication procedures | H.240–H.259 |
| Coding of moving video | H.260–H.279 |
| Related systems aspects | H.280–H.299 |
| Systems and terminal equipment for audiovisual services | H.300–H.349 |
| Directory services architecture for audiovisual and multimedia services | H.350–H.359 |
| Quality of service architecture for audiovisual and multimedia services | H.360–H.369 |
| Telepresence | H.420–H.429 |
| Supplementary services for multimedia | H.450–H.499 |
| MOBILITY AND COLLABORATION PROCEDURES | |
| Overview of Mobility and Collaboration, definitions, protocols and procedures | H.500–H.509 |
| Mobility for H-Series multimedia systems and services | H.510–H.519 |
| Mobile multimedia collaboration applications and services | H.520–H.529 |
| Security for mobile multimedia systems and services | H.530–H.539 |
| Security for mobile multimedia collaboration applications and services | H.540–H.549 |
| Mobility interworking procedures | H.550–H.559 |
| Mobile multimedia collaboration inter-working procedures | H.560–H.569 |
| BROADBAND, TRIPLE-PLAY AND ADVANCED MULTIMEDIA SERVICES | |
| Broadband multimedia services over VDSL | H.610–H.619 |
| Advanced multimedia services and applications | H.620–H.629 |
| Ubiquitous sensor network applications and Internet of Things | H.640–H.649 |
| IPTV MULTIMEDIA SERVICES AND APPLICATIONS FOR IPTV | |
| General aspects | H.700–H.719 |
| IPTV terminal devices | H.720–H.729 |
| IPTV middleware | H.730–H.739 |
| IPTV application event handling | H.740–H.749 |
| IPTV metadata | H.750–H.759 |
| IPTV multimedia application frameworks | H.760–H.769 |
| IPTV service discovery up to consumption | H.770–H.779 |
| Digital Signage | H.780–H.789 |
| E-HEALTH MULTIMEDIA SERVICES AND APPLICATIONS | |
| Interoperability compliance testing of personal health systems (HRN, PAN, LAN, TAN and WAN) | H.820–H.859 |
| Multimedia e-health data exchange services | H.860–H.869 |

For further details, please refer to the list of ITU-T Recommendations.

Recommendation ITU-T H.830.8

Conformance of ITU-T H.810 personal health devices: WAN interface Part 8: Consent Management: Receiver

Summary

Recommendation ITU-T H.830.8 is a transposition of Continua Health Alliance Test Tool DG2013, Test Suite Structure & Test Purposes, WAN Interface; Part 8: Consent Management. Receiver (Version 1.0, 2014-01-24), that was developed by the Continua Health Alliance. A number of versions of this specification existed before transposition.

This Recommendation includes an electronic attachment with the protocol implementation conformance statements (PICS) and the protocol implementation extra information for testing (PIXIT) required for the implementation of Annex A.

This Recommendation was initially approved as ITU-T H.838 (01/2015) and later renumbered, without further modifications, as ITU-T H.830.8 (01/2015) for consistency with the numbering of new WAN interface conformance testing specifications.

History

| Edition | Recommendation | Approval | Study Group | Unique ID* |
|---------|----------------|------------|-------------|---------------------------------------------------------------------------|
| 1.0 | ITU-T H.838 | 2015-01-13 | 16 | 11.1002/1000/12256 |
| 1.0 | ITU-T H.830.8 | 2015-01-13 | 16 | 11.1002/1000/12594 |

* To access the Recommendation, type the URL <http://handle.itu.int/> in the address field of your web browser, followed by the Recommendation's unique ID. For example, <http://handle.itu.int/11.1002/1000/11830-en>.

FOREWORD

The International Telecommunication Union (ITU) is the United Nations specialized agency in the field of telecommunications, information and communication technologies (ICTs). The ITU Telecommunication Standardization Sector (ITU-T) is a permanent organ of ITU. ITU-T is responsible for studying technical, operating and tariff questions and issuing Recommendations on them with a view to standardizing telecommunications on a worldwide basis.

The World Telecommunication Standardization Assembly (WTSA), which meets every four years, establishes the topics for study by the ITU-T study groups which, in turn, produce Recommendations on these topics.

The approval of ITU-T Recommendations is covered by the procedure laid down in WTSA Resolution 1.

In some areas of information technology which fall within ITU-T's purview, the necessary standards are prepared on a collaborative basis with ISO and IEC.

NOTE

In this Recommendation, the expression "Administration" is used for conciseness to indicate both a telecommunication administration and a recognized operating agency.

Compliance with this Recommendation is voluntary. However, the Recommendation may contain certain mandatory provisions (to ensure, e.g., interoperability or applicability) and compliance with the Recommendation is achieved when all of these mandatory provisions are met. The words "shall" or some other obligatory language such as "must" and the negative equivalents are used to express requirements. The use of such words does not suggest that compliance with the Recommendation is required of any party.

INTELLECTUAL PROPERTY RIGHTS

ITU draws attention to the possibility that the practice or implementation of this Recommendation may involve the use of a claimed Intellectual Property Right. ITU takes no position concerning the evidence, validity or applicability of claimed Intellectual Property Rights, whether asserted by ITU members or others outside of the Recommendation development process.

As of the date of approval of this Recommendation, ITU had not received notice of intellectual property, protected by patents, which may be required to implement this Recommendation. However, implementers are cautioned that this may not represent the latest information and are therefore strongly urged to consult the TSB patent database at <http://www.itu.int/ITU-T/ipr/>.

© ITU 2015

All rights reserved. No part of this publication may be reproduced, by any means whatsoever, without the prior written permission of ITU.

Table of Contents

| | Page |
|-------------------------------------------------------|-------------|
| 1 Scope..... | 1 |
| 2 References..... | 1 |
| 3 Definitions | 2 |
| 3.1 Terms defined elsewhere | 2 |
| 3.2 Terms defined in this Recommendation..... | 2 |
| 4 Abbreviations and acronyms | 2 |
| 5 Conventions | 3 |
| 6 Test suite structure (TSS) | 4 |
| 7 Electronic attachment | 6 |
| Annex A – Test purposes | 7 |
| A.1 TP definition conventions..... | 7 |
| A.2 Subgroup 2.5.1: WAN XDR Transaction (TRANS)..... | 8 |
| A.3 Subgroup 2.5.2: WAN Service Validation (SER)..... | 9 |
| Bibliography..... | 11 |

Electronic attachment: Protocol implementation conformance statements (PICS) and the protocol implementation extra information for testing (PIXIT) required for the implementation of Annex A.

Introduction

This Recommendation is a transposition of Continua Health Alliance Test Tool DG2013, Test Suite Structure & Test Purposes, WAN Interface; Part 8: Consent Management. Receiver (Version 1.0, 2014-01-24), that was developed by the Continua Health Alliance. A number of versions of this specification existed before transposition and these can be found in the table below.

| Version | Date | Revision history |
|----------------|-------------|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| 1.0 | 2012-10-05 | Initial release for Test Tool DG2011 |
| 1.1 | 2013-05-24 | Initial release for Test Tool DG2012. This uses "TSS&TP_DG2012_WAN_PART_8 (REC CM) v1.0.doc" as a baseline as new features included in [b-CDG 2012] do not affect the test procedures specified in this document. |
| 1.2 | 2014-01-24 | Initial release for Test Tool DG2013. This uses "TSS&TP_DG2012_WAN_PART_8 (REC CM) v1.0.doc" as a baseline as new features included in [ITU-T H.810] do not affect the test procedures specified in this document. |

Recommendation ITU-T H.830.8

Conformance of ITU-T H.810 personal health devices: WAN interface Part 8: Consent Management: Receiver

1 Scope

The scope of this Recommendation¹ is to provide a test suite structure and the test purposes (TSS & TP) for the WAN interface based on the requirements defined in the Continua Design Guidelines (CDG) [ITU-T H.810]. The objective of this test specification is to provide a high probability of air interface interoperability between different devices.

The TSS & TP for the WAN interface document have been divided into the eight parts specified below. This Recommendation covers Part 8.

- **Part 1:** Web services interoperability Sender
- **Part 2:** Web services interoperability Receiver
- **Part 3:** SOAP/ATNA. Sender
- **Part 4:** SOAP/ATNA. Receiver
- **Part 5:** PCD-01 HL7 messages. Sender
- **Part 6:** PCD-01 HL7 messages. Receiver
- **Part 7:** Consent management Sender
- **Part 8:** Consent management Receiver

2 References

The following ITU-T Recommendations and other references contain provisions which, through reference in this text, constitute provisions of this Recommendation. At the time of publication, the editions indicated were valid. All Recommendations and other references are subject to revision; users of this Recommendation are therefore encouraged to investigate the possibility of applying the most recent edition of the Recommendations and other references listed below. A list of the currently valid ITU-T Recommendations is regularly published. The reference to a document within this Recommendation does not give it, as a stand-alone document, the status of a Recommendation.

- [ITU-T H.810] Recommendation ITU-T H.810 (2013), *Interoperability design guidelines for personal health systems*.
- [IEEE 11073-20601A] IEEE 11073-20601A-2010, *IEEE Health informatics – Personal health device communication – Part 20601: Application profile – Optimized Exchange Protocol Amendment 1*.
<<http://standards.ieee.org/findstds/standard/11073-20601a-2010.html>>
- [IHE ITI TF-2] IHE ITI TF 2 (2009), *IHE IT Infrastructure Technical Framework, Volume 2* (ITI TF-2), Revision 6.0. It comprises three sub-volumes: 2a (Transactions Part A), 2b (Transactions Part B) and 2x (Appendices).
<http://www.ihe.net/Technical_Framework/upload/IHE_ITI_TF_6-0_Vol2a_FT_2009-08-10.pdf>
<http://www.ihe.net/Technical_Framework/upload/IHE_ITI_TF_6-0_Vol2b_FT_2009-08-10.pdf>
<http://www.ihe.net/Technical_Framework/upload/IHE_ITI_TF_6-0_Vol2x_FT_2009-08-10.pdf>

¹ This Recommendation includes an electronic attachment with the protocol implementation extra information for testing (PIXIT) required for the implementation of Annex A.

3 Definitions

3.1 Terms defined elsewhere

3.1.1 agent [IEEE 11073-20601A]: A node that collects and transmits personal health data to an associated manager.

3.1.2 manager [IEEE 11073-20601A]: A node receiving data from one or more agent systems. Some examples of managers include a cellular phone, health appliance, set top box, or a computer system.

3.2 Terms defined in this Recommendation

None.

4 Abbreviations and acronyms

This Recommendation uses the following abbreviations and acronyms:

| | |
|--------|-------------------------------------------------------|
| ATNA | Audit Trail and Node Authentication |
| ATS | Abstract Test Suite |
| CDA | Clinical Document Architecture |
| CDG | Continua Design Guidelines |
| DUT | Device Under Test |
| ebXML | electronic business using extensible Markup Language |
| EHR | Electronic Health Record |
| INR | International Normalized Ratio |
| IUT | Implementation Under Test |
| MDS | Medical Device System |
| MTOM | Message Transmission Optimization Mechanism |
| NFC | Near Field Communication |
| PCD | Patient Care Device |
| PCO | Point of Control and Observation |
| PCT | Protocol Conformance Testing |
| PHD | Personal Healthcare Device |
| PHDC | Personal Healthcare Device Class |
| PHM | Personal Healthcare Monitoring (report) |
| PICS | Protocol Implementation Conformance Statement |
| PIXIT | Protocol Implementation extra Information for Testing |
| S/MIME | Secure/Multipurpose Internet Mail Extensions |
| SDP | Service Discovery Protocol |
| SOAP | Simple Object Access Protocol |
| TCRL | Test Case Reference List |
| TCWG | Test and Certification Working Group |

| | |
|-------|------------------------------------------------|
| TP | Test Purpose |
| TSS | Test Suite Structure |
| URI | Uniform Resource Identifier |
| USB | Universal Serial Bus |
| WAN | Wide Area Network |
| WDM | Windows Driver Model |
| WS | Web Service |
| WSDL | Web Service Description Language |
| XDR | cross-enterprise Document Reliable interchange |
| XDS.b | cross-enterprise Document Sharing-b |
| XML | extensible Markup Language |

5 Conventions

The key words "SHALL", "SHALL NOT", "SHOULD", "SHOULD NOT", "MAY", "MAY NOT" in this document are to be interpreted as in [b-ETSI SR 001 262].

- SHALL is equivalent to 'must' or 'it is required to'.
- SHALL NOT is equivalent to 'must not' or 'it is not allowed'.
- SHOULD is equivalent to 'it is recommended to'.
- SHOULD NOT is equivalent to 'it is not recommended to'.
- MAY is equivalent to 'is permitted'.
- MAY NOT is equivalent to 'it is not required that'.

NOTE – The above-mentioned key words are capitalized for illustrative purposes only and they do not appear capitalized within this Recommendation.

Reference is made in the ITU-T H.800-series of Recommendations to different versions of the Continua Design Guidelines (CDG) by a specific designation. The list of terms that may be used in this Recommendation is provided in Table 1. Furthermore, the 2013 edition of the Continua design guidelines, which is published as [ITU-T H.810], is designated by "CDG 2013" as an extension of the designations indicated in the bibliography.

Table 1 – List of designations associated with the various versions of the CDG

| CDG name | Transposed as | Version | Description | Designation |
|------------------|---------------|---------|-----------------------------------------------------------------------------------------------------------------------------|-------------|
| 2013 plus errata | ITU-T H.810 | 4.1 | CDG 2013 plus errata noting all ratified bugs. | – |
| 2013 | – | 4.0 | Release 2013 of CDG including maintenance updates of the CDG 2012 and additional guidelines that cover new functionalities. | Endorphin |
| 2012 plus errata | – | 3.1 | CDG 2012 plus errata noting all ratified bugs [CDG 2012]. | – |
| 2012 | – | 3.0 | Release 2012 of the CDG including maintenance updates of CDG 2011 and additional guidelines that cover new functionalities. | Catalyst |

Table 1 – List of designations associated with the various versions of the CDG

| CDG name | Transposed as | Version | Description | Designation |
|------------------|---------------|---------|------------------------------------------------------------------------------------------------------------------------------------------|-------------|
| 2011 plus errata | – | 2.1 | CDG 2011 integrated with identified errata. | – |
| 2011 | – | 2.0 | Release 2011 of CDG including maintenance updates of CDG 2010 and additional guidelines that cover new functionalities [b-CDG 2011]. | Adrenaline |
| 2010 plus errata | – | 1.6 | CDG 2010 integrated with identified errata | – |
| 2010 | – | 1.5 | Release 2010 of the CDG with maintenance updates of CDG Version 1 and additional guidelines that cover new functionalities [b-CDG 2010]. | 1.5 |
| 1.0 | – | 1.0 | First released version of the CDG [b-CDG 1.0]. | – |

6 Test suite structure (TSS)

The test purposes (TPs) for the WAN interface have been divided into the main subgroups specified below. Annex A describes the TPs for subgroups 2.5.1 and 2.5.2 (shown in bold).

- Group 1: Sender (SEN)
 - Group 1.1: Web services interoperability (WSI)
 - Subgroup 1.1.1: Basic profile (BP)
 - Subgroup 1.1.2: Basic security profile (BSP)
 - Subgroup 1.1.3: Reliable messaging (RM)
 - Group 1.2: SOAP (SOAP)
 - Subgroup 1.2.1: SOAP headers (HEAD)
 - Group 1.3: Audit (ATNA)
 - Subgroup 1.3.1: General (GEN)
 - Subgroup 1.3.2: PCD-01 (PCD-01)
 - Subgroup 1.3.3: Consent management (CM)
 - Group 1.4: PCD-01 HL7 messages (PCD-01-DATA)
 - Subgroup 1.4.1: General (GEN)
 - Subgroup 1.4.2: Design guidelines (DG)
 - Subgroup 1.4.3: Pulse oximeter (PO)
 - Subgroup 1.4.4: Blood pressure monitor (BPM)
 - Subgroup 1.4.5: Thermometer (TH)
 - Subgroup 1.4.6: Weighing scales (WEG)
 - Subgroup 1.4.7: Glucose meter (GL)
 - Subgroup 1.4.8: Cardiovascular fitness and activity monitor (CV)
 - Subgroup 1.4.9: Strength fitness equipment (ST)
 - Subgroup 1.4.10: Independent living activity hub (HUB)
 - Subgroup 1.4.11: Adherence monitor (AM)
 - Subgroup 1.4.12: Peak expiratory flow monitor (PF)

- Subgroup 1.4.13: Body composition analyser (BCA)
- Subgroup 1.4.14: Basic electrocardiograph (ECG)
- Group 1.5: Consent management (CM)
 - Subgroup 1.5.1: WAN XDR transaction (TRANS)
 - Subgroup 1.5.2: WAN metadata validation (META)
 - Subgroup 1.5.3: WAN consent directive validation (CDV)
- Group 2: Receiver (REC)
 - Group 2.1: Web service interoperability (WSI)
 - Subgroup 2.1.1: Basic profile (BP)
 - Subgroup 2.1.2: Basic security profile (BSP)
 - Subgroup 2.1.3: Reliable messaging (RM)
 - Group 2.2: SOAP (SOAP)
 - Subgroup 2.2.1: SOAP headers (HEAD)
 - Group 2.3: Audit (ATNA)
 - Subgroup 2.3.1: General (GEN)
 - Subgroup 2.3.2: PCD-01 (PCD-01)
 - Subgroup 2.3.3: Consent management (CM)
 - Group 2.4: PCD-01 HL7 messages (PCD-01-DATA)
 - Subgroup 2.4.1: General (GEN)
 - Subgroup 2.4.2: Design guidelines (DG)
 - Subgroup 2.4.3: Pulse oximeter (PO)
 - Subgroup 2.4.4: Blood pressure monitor (BPM)
 - Subgroup 2.4.5: Thermometer (TH)
 - Subgroup 2.4.6: Weighing scales (WEG)
 - Subgroup 2.4.7: Glucose meter (GL)
 - Subgroup 2.4.8: Cardiovascular fitness and activity monitor (CV)
 - Subgroup 2.4.9: Strength fitness equipment (ST)
 - Subgroup 2.4.10: Independent living activity hub (HUB)
 - Subgroup 2.4.11: Adherence monitor (AM)
 - Subgroup 2.4.12: Peak expiratory flow monitor (PF)
 - Subgroup 2.4.13: Body composition analyser (BCA)
 - Subgroup 2.4.14: Basic electrocardiograph (ECG)
 - Group 2.5: Consent management (CM)
 - **Subgroup 2.5.1: WAN XDR transaction (TRANS)**
 - **Subgroup 2.5.2: WAN service validation (SER)**

7 Electronic attachment

The protocol implementation conformance statements (PICS) and the protocol implementation extra information for testing (PIXIT) required for the implementation of Annex A can be downloaded from <http://handle.itu.int/11.1002/2000/12067>.

In the electronic attachment, letters "C" and "I" in the column labelled "Mandatory" are used to distinguish between "PICS" and "PIXIT" respectively during testing. If the cell is empty, the corresponding PICS is "independent". If the field contains a "C2, the corresponding PICS is dependent on other PICS, and the logical expression is detailed in the "SCR_Expression" field. The static conformance review (SCR) is used in the test tool to assert whether the PICS selection is consistent.

Annex A

Test purposes

(This annex forms an integral part of this Recommendation.)

A.1 TP definition conventions

The test purposes (TP) are defined according to the following rules:

- **TP Id:** This is a unique identifier (TP/<TT>/<DUT>/<GR>/<SGR>/<XX> – <NNN>). It is specified according to the naming convention defined below:
 - Each test purpose identifier is introduced by the prefix "TP".
 - <TT>: This is the test tool that will be used in the test case.
 - WAN: Wide area network
 - <DUT>: This is the device under test.
 - SEN: WAN observation sender
 - REC: WAN observation receiver
 - <GR>: This identifies a group of test cases.
 - <SGR>: This identifies a subgroup of test cases.
 - <XX>: This identifies the type of testing.
 - BV: Valid behaviour test
 - BI: Invalid behaviour test
 - <NNN>: This is a sequential number that identifies the test purpose (TP).
- **TP label:** This is the title of the TP.
- **Coverage:** This contains the specification reference and clause to be checked by the TP.
 - Spec: This indicates the earliest version of the specification from which the testable items to be checked by the TP are included.
 - Testable Item: This contains testable items to be checked by the TP.
- **Test purpose:** This is a description of the requirements to be tested.
- **Applicability:** This contains the PICS items that define if the test case is applicable or not for a specific device. When a TP contains an "ALL" in this field it means that it applies to the device under test within that scope of the test (specialization, transport used, etc.).
- **Initial condition:** This indicates the state to which the DUT needs to be moved at the beginning of TC execution.
- **Test procedure:** This describes the steps to be followed in order to execute the test case.
- **Pass/Fail criteria:** This provides criteria to decide whether the DUT passes or fails the test case.

A.2 Subgroup 2.5.1: WAN XDR Transaction (TRANS)

| | | | | |
|---------------------------|-----------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|----------------------|---------------------|
| TP Id | | TP/WAN/REC/CM/TRANS/BV-000 | | |
| TP label | | Provide and Register Document Set-b Transaction Response | | |
| Coverage | Spec | [b-CDG 2012], WAN interface requirements for consent management | | |
| | Testable items | ConsentReceiver1; M | | |
| | Spec | [IHE ITI TF-2], Volume 2b | | |
| | Testable items | ProvideResponse1; M | ProvideResponse2; M | ProvideProtocol2; M |
| | | ProvideProtocol10; M | ProvideProtocol11; M | |
| Applicability | | C_REC_000 AND C_REC_GEN_002 | | |
| Initial condition | | The receiver under test has published the document recipient service and is ready to receive a "Provide and Register Document Set-b Request" sent by the simulated sender. | | |
| Test procedure | | <p>The simulated sender sends a "Provide and Register Document Set-b Request" with a clinical document architecture (CDA) using MTOM/XOP.</p> <p>The receiver under test responds with a "Provide and Register Document Set-b Response" with status = "urn:oasis:names:tc:ebxml-regrep:ResponseStatusType:Success" using SOAP12 [b-SOAP 1.2].</p> | | |
| Pass/Fail criteria | | All steps are as specified within the test procedure above. | | |
| Notes | | | | |

A.3 Subgroup 2.5.2: WAN Service Validation (SER)

| | | | | |
|---------------------------|-----------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|---------------------|---------------------|
| TP Id | | TP/WAN/REC/CM/SER/BV-000 | | |
| TP label | | Service WSDL | | |
| Coverage | Spec | [IHE ITI TF-2], Volume 2b | | |
| | Testable items | ProvideProtocol3; M | ProvideProtocol4; M | ProvideProtocol5; M |
| | | ProvideProtocol6; M | ProvideProtocol7; M | ProvideProtocol8; M |
| Applicability | | C_REC_000 AND C_REC_GEN_002 | | |
| Initial condition | | The receiver under test has published its WSDL description for a consent management service | | |
| Test procedure | | <p>Look up the WSDL description using the corresponding URL given by the receiver under test (I_REC_CM_001). Check that:</p> <p>There are two xsd:import elements in the definitions/types section for which the namespaces are:</p> <p>namespace="urn:oasis:names:tc:ebxml-regrep:xsd:rs:3.0"</p> <p>namespace="urn:ihe:iti:xds-b:2007"</p> <p>/definitions/message/part/@element = "ihe:ProvideAndRegisterDocumentSetRequest" for the "Provide and Register Document Set-b Request"</p> <p>/definitions/message/part/@element = " ihe:RegistryResponse " for the "Provide and Register Document Set-b Response"</p> <p>/definitions/portType/operation/input/@wsaw:Action = "urn:ihe:iti:2007:ProvideAndRegisterDocumentSet-b" for the "Provide and Register Document Set-b Request"</p> <p>/definitions/portType/operation/input/@wsaw:Action = "urn:ihe:iti:2007:ProvideAndRegisterDocumentSet-bResponse" for the "Provide and Register Document Set-b Response"</p> <p>/definitions/portType/operation/soap12:operation/@soapAction = "urn:ihe:iti:2007:ProvideAndRegisterDocumentSet-b"</p> | | |
| Pass/Fail criteria | | All steps are as specified within the test procedure above. | | |
| Notes | | | | |

| | | | | |
|--------------------------|-----------------------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------|----------------------|----------------------|
| TP Id | | TP/WAN/REC/CM/SER/BV-001 | | |
| TP label | | Service Metadata Validation | | |
| Coverage | Spec | [IHE ITI TF-2], Volume 2b | | |
| | Testable items | ProvideRecipient3; M | ProvideRecipient5; O | ProvideRecipient7; C |
| Applicability | | C_REC_000 AND C_REC_GEN_002 | | |
| Initial condition | | The receiver under test has published the document recipient service and is ready to receive a "Provide and Register Document Set-b Request" sent by the simulated sender. | | |

| | |
|---------------------------|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Test procedure | <p>The Simulated Sender sends a "Provide and Register Document Set-b Request" message including XDSSubmissionSet.sourceId, XDSDocumentEntry.uniqueId, XDSDocumentEntry.hash and XDSDocumentEntry.size elements with correct values.</p> <p>The Receiver under test responds with a "Provide and Register Document Set-b Response" with status Success</p> <p>The simulated sender sends a "Provide and Register Document Set-b Request" message including an XDSDocumentEntry.hash element with a wrong value</p> <p>The receiver under test responds with a "Provide and Register Document Set-b Response" with status Success</p> <p>The simulated sender sends a "Provide and Register Document Set-b Request" message including an XDSDocumentEntry.size element with a wrong value</p> <p>The receiver under test responds with a "Provide and Register Document Set-b Response" with status Success</p> <p>The simulated sender sends a "Provide and Register Document Set-b Request" message including a different XDSSubmissionSet.sourceId element</p> <p>The receiver under test may accept or may not accept the message.</p> |
| Pass/Fail criteria | All steps are as specified within the test procedure above. |
| Notes | |

| | | | |
|---------------------------|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|---------------------------|----------------------|
| TP Id | TP/WAN/REC/CM/SER/BV-002 | | |
| TP label | Multiple Documents and Errors | | |
| | Spec | [IHE ITI TF-2], Volume 2b | |
| | Testable items | ProvideReqActions3; M | ProvideRecipient1; M |
| Applicability | C_REC_000 AND C_REC_GEN_002 | | |
| Initial condition | The receiver under test has published the document recipient service and is ready to receive a "Provide and Register Document Set-b Request" sent by the simulated sender. | | |
| Test procedure | <p>The simulated sender sends a "Provide and Register Document Set-b Request" with two documents using MTOM/XOP</p> <p>The receiver under test responds with a "Provide and Register Document Set-b Response" with status = "urn:oasis:names:tc:ebxml-regrep:ResponseStatusType:Success"</p> <p>The simulated sender sends a "Provide and Register Document Set-b Request" without attaching a document using MTOM/XOP</p> <p>The receiver under test responds with a "Provide and Register Document Set-b Response" with status = "urn:oasis:names:tc:ebxml-regrep:ResponseStatusType:Failure".</p> | | |
| Pass/Fail criteria | All steps are as specified within the test procedure above. | | |
| Notes | | | |

Bibliography

- [b-CDG 1.0] Continua Health Alliance Continua Design Guidelines v1.0. (2008), *Continua Design Guidelines*.
- [b-CDG 2010] Continua Health Alliance Continua Design Guidelines v1.5 (2010), *Continua Design Guidelines*.
- [b-CDG 2011] Continua Health Alliance Continua Design Guidelines (2011) "Adrenaline" *Continua Design Guidelines*.
- [b-CDG 2012] Continua Health Alliance CDG, Continua Design Guidelines (2012) "Catalyst", *Continua Design Guidelines*.
- [b-ETSI SR 001 262] ETSI SR 001 262 v1.8.1 (2003): *ETSI drafting rules*.
- [b-HL7 CDA IG] Health Level Seven (2011), *HL7 Implementation Guide for Clinical Document Architecture, Release 2: Consent Directives, Release 1, HL7 Draft Standard for Trial Use*.
<http://www.hl7.org/documentcenter/public/standards/dstu/CDAR2_IG%20CONSENTDIR_DSTU_2011JAN.pdf>
- [b-SOAP 1.2] W3C SOAP 1.2 (2007), SOAP Version 1.2 (Second Edition).
<<http://www.w3.org/TR/soap/>>

SERIES OF ITU-T RECOMMENDATIONS

| | |
|-----------------|-----------------------------------------------------------------------------------------------------------------------------------------------------------|
| Series A | Organization of the work of ITU-T |
| Series D | General tariff principles |
| Series E | Overall network operation, telephone service, service operation and human factors |
| Series F | Non-telephone telecommunication services |
| Series G | Transmission systems and media, digital systems and networks |
| Series H | Audiovisual and multimedia systems |
| Series I | Integrated services digital network |
| Series J | Cable networks and transmission of television, sound programme and other multimedia signals |
| Series K | Protection against interference |
| Series L | Environment and ICTs, climate change, e-waste, energy efficiency; construction, installation and protection of cables and other elements of outside plant |
| Series M | Telecommunication management, including TMN and network maintenance |
| Series N | Maintenance: international sound programme and television transmission circuits |
| Series O | Specifications of measuring equipment |
| Series P | Terminals and subjective and objective assessment methods |
| Series Q | Switching and signalling |
| Series R | Telegraph transmission |
| Series S | Telegraph services terminal equipment |
| Series T | Terminals for telematic services |
| Series U | Telegraph switching |
| Series V | Data communication over the telephone network |
| Series X | Data networks, open system communications and security |
| Series Y | Global information infrastructure, Internet protocol aspects and next-generation networks |
| Series Z | Languages and general software aspects for telecommunication systems |