ITU-T

H.845.2

TELECOMMUNICATION STANDARDIZATION SECTOR OF ITU (01/2015)

SERIES H: AUDIOVISUAL AND MULTIMEDIA SYSTEMS

E-health multimedia services and applications – Interoperability compliance testing of personal health systems (HRN, PAN, LAN, TAN and WAN)

Conformance of ITU-T H.810 personal health devices: PAN/LAN/TAN interface Part 5B:

**Glucose meter: Agent** 

Recommendation ITU-T H.845.2



# ITU-T H-SERIES RECOMMENDATIONS

# AUDIOVISUAL AND MULTIMEDIA SYSTEMS

CHARACTERISTICS OF VISUAL TELEPHONE SYSTEMS	H.100-H.199
INFRASTRUCTURE OF AUDIOVISUAL SERVICES	
General	H.200-H.219
Transmission multiplexing and synchronization	H.220-H.229
Systems aspects	H.230-H.239
Communication procedures	H.240-H.259
Coding of moving video	H.260-H.279
Related systems aspects	H.280-H.299
Systems and terminal equipment for audiovisual services	H.300-H.349
Directory services architecture for audiovisual and multimedia services	H.350-H.359
Quality of service architecture for audiovisual and multimedia services	H.360-H.369
Telepresence	H.420-H.429
Supplementary services for multimedia	H.450-H.499
MOBILITY AND COLLABORATION PROCEDURES	
Overview of Mobility and Collaboration, definitions, protocols and procedures	H.500-H.509
Mobility for H-Series multimedia systems and services	H.510-H.519
Mobile multimedia collaboration applications and services	H.520-H.529
Security for mobile multimedia systems and services	H.530-H.539
Security for mobile multimedia collaboration applications and services	H.540-H.549
Mobility interworking procedures	H.550-H.559
Mobile multimedia collaboration inter-working procedures	H.560-H.569
BROADBAND, TRIPLE-PLAY AND ADVANCED MULTIMEDIA SERVICES	
Broadband multimedia services over VDSL	H.610-H.619
Advanced multimedia services and applications	H.620-H.629
Ubiquitous sensor network applications and Internet of Things	H.640-H.649
IPTV MULTIMEDIA SERVICES AND APPLICATIONS FOR IPTV	
General aspects	H.700-H.719
IPTV terminal devices	H.720-H.729
IPTV middleware	H.730-H.739
IPTV application event handling	H.740-H.749
IPTV metadata	H.750-H.759
IPTV multimedia application frameworks	H.760-H.769
IPTV service discovery up to consumption	H.770-H.779
Digital Signage	H.780-H.789
E-HEALTH MULTIMEDIA SERVICES AND APPLICATIONS	
Interoperability compliance testing of personal health systems (HRN, PAN, LAN, TAN	H.820-H.859
and WAN)	XX 0.60 XX 0.60
Multimedia e-health data exchange services	H.860–H.869

For further details, please refer to the list of ITU-T Recommendations.

# **Recommendation ITU-T H.845.2**

# Conformance of ITU-T H.810 personal health devices: PAN/LAN/TAN interface Part 5B: Glucose meter: Agent

#### **Summary**

Recommendation ITU-T H.845.2 is a transposition of Continua Test Tool DG2013, Test Suite Structure & Test Purposes, PAN-LAN-TAN Interface; Part 5B: Device Specializations. Agent (Glucose Meter) (Version 1.4, 2014-01-24), that was developed by the Continua Health Alliance. A number of versions of this specification existed before transposition.

This Recommendation includes an electronic attachment with the protocol implementation conformance statements (PICS) and the protocol implementation extra information for testing (PIXIT) required for the implementation of Annex A.

#### **History**

Edition	Recommendation	Approval	Study Group	Unique ID*
1.0	ITU-T H.845.2	2015-01-13	16	11.1002/1000/12263

<sup>\*</sup> To access the Recommendation, type the URL http://handle.itu.int/ in the address field of your web browser, followed by the Recommendation's unique ID. For example, <a href="http://handle.itu.int/11.1002/1000/11830-en">http://handle.itu.int/11.1002/1000/11830-en</a>.

#### **FOREWORD**

The International Telecommunication Union (ITU) is the United Nations specialized agency in the field of telecommunications, information and communication technologies (ICTs). The ITU Telecommunication Standardization Sector (ITU-T) is a permanent organ of ITU. ITU-T is responsible for studying technical, operating and tariff questions and issuing Recommendations on them with a view to standardizing telecommunications on a worldwide basis.

The World Telecommunication Standardization Assembly (WTSA), which meets every four years, establishes the topics for study by the ITU-T study groups which, in turn, produce Recommendations on these topics.

The approval of ITU-T Recommendations is covered by the procedure laid down in WTSA Resolution 1.

In some areas of information technology which fall within ITU-T's purview, the necessary standards are prepared on a collaborative basis with ISO and IEC.

#### **NOTE**

In this Recommendation, the expression "Administration" is used for conciseness to indicate both a telecommunication administration and a recognized operating agency.

Compliance with this Recommendation is voluntary. However, the Recommendation may contain certain mandatory provisions (to ensure, e.g., interoperability or applicability) and compliance with the Recommendation is achieved when all of these mandatory provisions are met. The words "shall" or some other obligatory language such as "must" and the negative equivalents are used to express requirements. The use of such words does not suggest that compliance with the Recommendation is required of any party.

### INTELLECTUAL PROPERTY RIGHTS

ITU draws attention to the possibility that the practice or implementation of this Recommendation may involve the use of a claimed Intellectual Property Right. ITU takes no position concerning the evidence, validity or applicability of claimed Intellectual Property Rights, whether asserted by ITU members or others outside of the Recommendation development process.

As of the date of approval of this Recommendation, ITU had not received notice of intellectual property, protected by patents, which may be required to implement this Recommendation. However, implementers are cautioned that this may not represent the latest information and are therefore strongly urged to consult the TSB patent database at <a href="http://www.itu.int/ITU-T/ipr/">http://www.itu.int/ITU-T/ipr/</a>.

#### © ITU 2015

All rights reserved. No part of this publication may be reproduced, by any means whatsoever, without the prior written permission of ITU.

# **Table of Contents**

			Page
1	Scope	·	1
2	Refer	ences	2
3	Defin	itions	2
	3.1	Terms defined elsewhere	2
	3.2	Terms defined in this Recommendation	2
4	Abbre	eviations and acronyms	2
5	Conve	entions	3
6	Test s	uite structure (TSS)	4
7	Electr	onic attachment	6
Anne	ex A – T	est purposes (TPs)	7
	<b>A.</b> 1	TP definition conventions	7
	A.2	Subgroup 1.3.2: Glucose meter (GL)	8
Bibli	ography	7	57

**Electronic attachment:** Protocol implementation conformance statements (PICS) and protocol implementation extra information for testing (PIXIT) required for the implementation of Annex A.

#### Introduction

This Recommendation is a transposition of Continua Test Tool DG2013, Test Suite Structure & Test Purposes, PAN-LAN-TAN Interface; Part 5B: Device Specializations. Agent (Glucose meter) (Version 1.4, 2014-01-24), that was developed by the Continua Health Alliance. A number of versions of this specification existed before transposition and these can be found in the table below.

Version	Date	Revision history
1.2	2012-10-05	Initial release for Test Tool DG2011. This is the same version as "TSS&TP_1.5_PAN-LAN_PART_5B_v1.2.doc" because new features included in [b-CDG 2011] do not affect the test procedures specified in this document
1.3	2013-05-24	Initial release for Test Tool DG2012. This uses "TSS&TP_DG2011_PAN-LAN_PART_5B_v1.2.doc" as a baseline and adds new features included in [b-CDG 2012]: New GM spec version Max APDU size for GM, BCA and ECG
1.4	2014-01-24	Initial release for Test Tool DG2013. This uses "TSS&TP_DG2012_PAN-LAN_PART_5B_v1.3.doc" as a baseline and adds new features included in [ITU-T H.810]:  • Adds glucose meter BLE  • Adds BLE SSP support  • Adds NFC new transport  • Adds INR device specialization

## **Recommendation ITU-T H.845.2**

# Conformance of ITU-T H.810 personal health devices: PAN/LAN/TAN interface Part 5B: Glucose meter: Agent

## 1 Scope

The scope of this Recommendation<sup>1</sup> is to provide a test suite structure and the test purposes (TSS & TP) for the PAN/LAN/TAN interface based on the requirements defined in the Continua Design Guidelines (CDG) [ITU-T H.810]. The objective of this test specification is to provide a high probability of air interface interoperability between different devices.

The TSS and TP for PAN/LAN/TAN interface document have been divided into ten parts. Each part is listed below:

- **Part 1:** Optimized exchange protocol [ISO/IEEE 11073-20601A] Agent
- Part 2: Optimized exchange protocol [ISO/IEEE 11073-20601A] Manager
- Part 3: Continua design guidelines. Agent
- Part 4: Continua design guidelines. Manager
- **Part 5:** Device specializations. Agent. This document is divided in 12 subparts:
  - Part 5A: Weighing scales
  - Part 5B: Glucose meter
  - Part 5C: Pulse oximeter
  - Part 5D: Blood pressure monitor
  - **Part 5E:** Thermometer
  - Part 5F: Cardiovascular fitness and activity monitor
  - Part 5G: Strength fitness equipment
  - **Part 5H:** Independent living activity hub
  - Part 5I: Adherence monitor
  - **Part 5J:** Insulin pump (Future development)
  - Part 5K: Peak flow
  - Part 5L: Body composition analyser
  - Part 5M: Basic electrocardiograph
  - Part 5N: International normalized ratio monitor
- Part 6: Device specializations. Manager
- Part 7: Continua design guidelines. Agent BLE
- Part 8: Continua design guidelines. Manager BLE
- Part 9: Personal health devices transcoding white paper. Agent
- Part 10: Personal health devices transcoding white paper. Manager

<sup>&</sup>lt;sup>1</sup> This Recommendation includes an electronic attachment with the protocol implementation conformance statements (PICS) and the protocol implementation extra information for testing (PIXIT) required for the implementation Annex A.

#### 2 References

The following ITU-T Recommendations and other references contain provisions which, through reference in this text, constitute provisions of this Recommendation. At the time of publication, the editions indicated were valid. All Recommendations and other references are subject to revision; users of this Recommendation are therefore encouraged to investigate the possibility of applying the most recent edition of the Recommendations and other references listed below. A list of the currently valid ITU-T Recommendations is regularly published. The reference to a document within this Recommendation does not give it, as a stand-alone document, the status of a Recommendation.

[ITU-T H.810] Recommendation ITU-T H.810 (2013), Interoperability design

guidelines for personal health systems.

device communication – Part 10417: Device specialization –

Glucose meter.

<a href="http://standards.ieee.org/findstds/standard/11073-10417-2011.html">http://standards.ieee.org/findstds/standard/11073-10417-2011.html</a>

[ISO/IEEE 11073-20601A] ISO/IEEE 11073-20601:2010, *Health informatics – Personal health* 

device communication – Part 20601: Application profile – Optimized exchange protocol, including ISO/IEEE 11073-

20601:2010 Amd 1:2015.

<a href="http://www.iso.org/iso/home/store/catalogue\_tc/catalogue\_detail.htm?csnumber=54331">http://www.iso.org/iso/home/store/catalogue\_tc/catalogue\_detail.htm?csnumber=54331</a>

with

<a href="http://www.iso.org/iso/home/store/catalogue\_tc/catalogue\_detail.htm?csnumber=63972">http://www.iso.org/iso/home/store/catalogue\_tc/catalogue\_detail.htm?csnumber=63972</a>

[ISO/IEEE 11073-104xx] ISO/IEEE 11073-104xx (in force), *Health informatics – Personal* 

health device communication – Device specialization.

NOTE – This is shorthand used to refer to the collection of device specialization standards that utilize [ISO/IEEE 11073-20601A], where xx

can be any number from 01 to 99, inclusive.

#### 3 Definitions

### 3.1 Terms defined elsewhere

This Recommendation uses the following terms defined elsewhere:

- **3.1.1** agent [ISO/IEEE 11073-20601A]: A node that collects and transmits personal health data to an associated manager.
- **3.1.2** manager [ISO/IEEE 11073-20601A]: A node receiving data from one or more agent systems. Some examples of managers include a cellular phone, health appliance, set top box, or a computer system.

#### 3.2 Terms defined in this Recommendation

None.

#### 4 Abbreviations and acronyms

This Recommendation uses the following abbreviations and acronyms:

ATS Abstract Test Suite

DUT Device Under Test

CDG Continua Design Guidelines

GUI Graphical User Interface

INR International Normalized Ratio

IUT Implementation Under Test

MDS Medical Device System

NFC Near Field Communication

PAN Personal Area Network

PCT Protocol Conformance Testing

PCO Point of Control and Observation

PHD Personal Healthcare Device

PHDC Personal Healthcare Device Class

PHM Personal Health Manager

PICS Protocol Implementation Conformance Statement

PIXIT Protocol Implementation extra Information for Testing

SDP Service Discovery Protocol

SOAP Simple Object Access Protocol

TCRL Test Case Reference List

TCWG Test and Certification Working Group

TP Test Purpose

TSS Test Suite Structure
USB Universal Serial Bus

WDM Windows Driver Model

#### 5 Conventions

The key words "SHALL", "SHALL NOT", "SHOULD", "SHOULD NOT", "MAY", "MAY NOT" in this document are to be interpreted as in [b-ETSI SR 001 262].

- SHALL is equivalent to 'must' or 'it is required to'.
- SHALL NOT is equivalent to 'must not' or 'it is not allowed'.
- SHOULD is equivalent to 'it is recommended to'.
- SHOULD NOT is equivalent to 'it is not recommended to'.
- MAY is equivalent to 'is permitted'.
- MAY NOT is equivalent to 'it is not required that'.

NOTE – The above-mentioned key words are capitalized for illustrative purposes only and they do not appear capitalized within this Recommendation.

Reference is made in the ITU-T H.800-series of Recommendations to different versions of the Continua Design Guidelines (CDG) by a specific designation. The list of terms that may be used in this Recommendation is provided in Table 1.

Table 1 – List of designations associated with the various versions of the CDG

CDG name	Transposed as	Version	Description	Designation
2013 plus errata	ITU-T H.810	4.1	CDG 2013 plus errata noting all ratified bugs.	_
2013	-	4.0	Release 2013 of the CDG including maintenance updates of CDG 2012 and additional guidelines that cover new functionalities.	Endorphin
2012 plus errata	_	3.1	CDG 2012 plus errata noting all ratified bugs [b-CDG 2012].	_
2012	-	3.0	Release 2012 of the CDG including maintenance updates of CDG 2011 and additional guidelines that cover new functionalities.	Catalyst
2011 plus errata	_	2.1	CDG 2011 integrated with identified errata.	_
2011	-	2.0	Release 2011 of the CDG including maintenance updates of CDG 2010 and additional guidelines that cover new functionalities [b-CDG 2011].	Adrenaline
2010 plus errata	_	1.6	CDG 2010 integrated with identified errata	-
2010	-	1.5	Release 2010 of the CDG with maintenance updates of CDG Version 1 and additional guidelines that cover new functionalities [b-CDG 2010].	1.5
1.0	-	1.0	First released version of the CDG [b-CDG 1.0].	_

#### 6 **Test suite structure (TSS)**

The test purposes (TPs) for the PAN/LAN/TAN interface have been divided into the main subgroups specified below. Annex A describes the TPs for subgroup 1.3.2: (shown in bold).

- Group 1: Agent (AG)
  - Group 1.1: Transport (TR)
    - Subgroup 1.1.1: Design guidelines: Common (DGC)
    - Subgroup 1.1.2: USB design guidelines (UDG)
    - Subgroup 1.1.3: Bluetooth design guidelines (BDG)
    - Subgroup 1.1.4: Pulse oximeter design guidelines (PODG)
    - Subgroup 1.1.5: Cardiovascular design guidelines (CVDG)
    - Subgroup 1.1.6: Activity hub design guidelines (HUBDG)
    - Subgroup 1.1.7: ZigBee design guidelines (ZDG)
    - Subgroup 1.1.8: Glucose meter design guidelines (GLDG)
    - Subgroup 1.1.9: Bluetooth low energy design guidelines (BLEDG)
    - Subgroup 1.1.10: Basic electrocardiograph design guidelines (ECGDG)
    - Subgroup 1.1.11: NFC design guidelines (NDG)
  - Group 1.2: Optimized exchange protocol (OXP)
    - Subgroup 1.2.1: PHD domain information model (DIM)

- Subgroup 1.2.2: PHD service model (SER)
- Subgroup 1.2.3: PHD communication model (COM)
- Group 1.3: Devices class specializations (CLASS)
  - Subgroup 1.3.1: Weighing scales (WEG)
  - Subgroup 1.3.2: Glucose meter (GL)
  - Subgroup 1.3.3: Pulse oximeter (PO)
  - Subgroup 1.3.4: Blood pressure monitor (BPM)
  - O Subgroup 1.3.5: Thermometer (TH)
  - Subgroup 1.3.6: Cardiovascular (CV)
  - Subgroup 1.3.7: Strength (ST)
  - Subgroup 1.3.8: Activity hub (HUB)
  - Subgroup 1.3.9: Adherence monitor (AM)
  - Subgroup 1.3.10: Insulin pump (IP) (Future development)
  - Subgroup 1.3.11: Peak flow (PF)
  - Subgroup 1.3.12: Body composition analyser (BCA)
  - Subgroup 1.3.13: Basic electrocardiograph (ECG)
  - Subgroup 1.3.14: International normalized ratio (INR)
- Group 1.4: Personal health device transcoding whitepaper (PHDTW)
  - Subgroup 1.4.1: Whitepaper general requirements (GEN)
  - Subgroup 1.4.2: Whitepaper thermometer requirements (TH)
  - Subgroup 1.4.3: Whitepaper blood pressure requirements (BPM)
  - Subgroup 1.4.4: Whitepaper heart rate requirements (HR)
  - Subgroup 1.4.5: Whitepaper glucose meter requirements (GL)
- Group 2: Manager (MAN)
  - Group 2.1: Transport (TR)
    - Subgroup 2.1.1: Design guidelines: Common (DGC)
    - Subgroup 2.1.2: USB design guidelines (UDG)
    - Subgroup 2.1.3: Bluetooth design guidelines (BDG)
    - Subgroup 2.1.4: Cardiovascular design guidelines (CVDG)
    - Subgroup 2.1.5: Activity hub design guidelines (HUBDG)
    - Subgroup 2.1.6: ZigBee design guidelines (ZDG)
    - Subgroup 2.1.7: Bluetooth low energy design guidelines (BLEDG)
    - Subgroup 2.1.8: NFC design guidelines (NDG)
  - Group 2.2: 20601: Optimized exchange protocol (OXP)
    - Subgroup 2.2.1: General (GEN)
    - Subgroup 2.2.2: PHD domain information model (DIM)
    - Subgroup 2.2.3: PHD service model (SER)
    - Subgroup 2.2.4: PHD communication model (COM)
  - Group 2.3: Devices class specializations (CLASS)
    - Subgroup 2.3.1: Weighing scales (WEG)
    - Subgroup 2.3.2: Glucose meter (GL)

- Subgroup 2.3.3: Pulse oximeter (PO)
- Subgroup 2.3.4: Blood pressure monitor (BPM)
- O Subgroup 2.3.5: Thermometer (TH)
- Subgroup 2.3.6: Cardiovascular (CV)
- Subgroup 2.3.7: Strength (ST)
- Subgroup 2.3.8: Activity hub (HUB)
- Subgroup 2.3.9: Adherence monitor (AM)
- Subgroup 2.3.10: Insulin pump (IP) (Future development)
- Subgroup 2.3.11: Peak flow (PF)
- Subgroup 2.3.12: Body composition analyser (BCA)
- Subgroup 2.3.13: Basic electrocardiograph (ECG)
- Subgroup 2.3.14: International normalized ratio (INR)
- Group 2.4: Personal health device transcoding whitepaper (PHDTW)
  - Subgroup 2.4.1: Whitepaper general requirements (GEN)
  - Subgroup 2.4.2: Whitepaper thermometer requirements (TH)
  - Subgroup 2.4.3: Whitepaper blood pressure measurement requirements (BPM)
  - Subgroup 2.4.4: Whitepaper heart rate requirements (HR)
  - Subgroup 2.4.5: Whitepaper glucose meter requirements (GL)

#### **7** Electronic attachment

The protocol implementation conformance statements (PICS) and the protocol implementation extra information for testing (PIXIT) required for the implementation of Annex A can be downloaded from <a href="http://handle.itu.int/11.1002/2000/12067">http://handle.itu.int/11.1002/2000/12067</a>.

In the electronic attachment, letters "C" and "I" in the column labelled "Mandatory" are used to distinguish between "PICS" and "PIXIT" respectively during testing. If the cell is empty, the corresponding PICS is "independent". If the field contains a "C", the corresponding PICS is dependent on other PICS, and the logical expression is detailed in the "SCR\_Expression" field. The static conformance review (SCR) is used in the test tool to assert whether the PICS selection is consistent.

#### Annex A

# **Test purposes (TPs)**

(This annex forms an integral part of this Recommendation.)

#### A.1 TP definition conventions

The test purposes are defined according to the following rules:

- **TP Id:** This is a unique identifier (TP/<TT>/<DUT>/<GR>/<SGR>/<XX> <NNN>). It is specified according to the naming convention defined below:
  - Each test purpose identifier is introduced by the prefix "TP".
  - <TT>: This is the test tool that will be used in the test case.
    - PAN: Personal area network (Bluetooth or USB)
    - LAN: Local area network (ZigBee)
    - PAN-LAN: Personal area network (Bluetooth or USB) Local area network (ZigBee)
    - LP-PAN: Low power personal area network (Bluetooth Low Energy)
    - TAN: Touch area network (NFC)
    - PLT: Personal area network (Bluetooth or USB) Local area network (ZigBee) Touch area network (NFC)
  - O <DUT>: This is the device under test
    - AG: PAN/LAN Agent
    - MAN: PAN/LAN Manager
  - <GR>: This identifies a group of test cases.
  - <SGR>: This identifies a subgroup of test cases.
  - <XX>: This identifies the type of testing
    - BV: Valid behaviour test
    - BI: Invalid behaviour test
  - <NNN>: This is a sequential number that identifies the test purpose.
- **TP label:** This is the TP's title.
- **Coverage:** This contains the specification reference and clause to be checked by the TP.
  - Spec: This indicates the earliest version of the specification from which the testable items to be checked by the TP were included.
  - Testable item: This contains testable items to be checked by the TP.
- **Test purpose:** This is a description of the requirements to be tested.
- **Applicability:** This contains the PICS items that define if the test case is applicable or not for a specific device. When a TP contains an "ALL" in this field it means that it applies to the device under test within that scope of the test (specialization, transport used, etc.).
- **Initial condition:** This indicates the state to which the DUT needs to be moved at the beginning of TC execution.
- **Test procedure:** This describes the steps to be followed in order to execute the test case.
- **Pass/Fail criteria:** This provides criteria to decide whether the DUT passes or fails the test case.

# A.2 Subgroup 1.3.2: Glucose meter (GL)

A.2 Sul	group 1	3.2: Glucose meter (GL)			
TP ld		TP/PLT/AG/CLASS/GL/BV-000	D_A		
TP label		Get MDS Object for Glucose meter specialization: Mandatory, Conditional and Optional Attributes.			
Coverage	Spec	[IEEE 11073-10417]	[IEEE 11073-10417]		
	Testable items	MDSGL Atr 1; M MDSGL Atr 5; M	MDSGL Atr 2; M	MDSGL Atr 4; M	
Applicability			OVD 170		
Applicability		C_AG_OXP_000 AND C_AG_	UXP_176		
Initial condit	ion	The simulated manager and the	e agent under test are in the op	erating state.	
Test procedu	ıre		sues "roiv-cmip-get" command on the attribute-id-list set to 0 to		
			a "rors-cmip-get" service messa nented attributes of the MDS ob		
		MDS attributes:			
		a. Attribute System-Type must not be present.			
		b. Mandatory attribute System-Type-Spec_List			
		□ attribute-id = MDC_ATTR_SYS_TYPE_SPEC_LIST			
		□ attribute-type = TypeVerList			
		□ attribute-value.length = 4 bytes for each configuration supported			
		attribute-value = {MDC_DEV_SPEC_PROFILE_GLUCOSE, 2} must be found in the list			
		c. Mandatory attribute Sy	stem-model		
		□ attribute-id = MDC_ATTR_ID_MODEL (0x09 0x28)			
		□ attribute-type = SystemModel			
		□ atribute-value.length = <variable></variable>			
		□ attribute-value =			
		<ul> <li>Manufacturer = Check against PIXIT I_AG_OXP_003</li> </ul>			
		Model = Check against PIXIT I_AG_OXP_004			
		d. Mandatory attribute Dev-Configuration-Id			
		☐ IF C_AG_GL_02	3 THEN attribute-value = 0x06A	4 (1700)	
		☐ IF C_AG_GL_02	4 THEN attribute-value = 0x06A	5 (1701)	
		☐ IF C_AG_OXP_1	81 THEN attribute-value = < be	tween 0x4000 and 0x7FFF >	
Pass/Fail cri	teria	All checked values are as spec	sified in the test procedure.		
Notes					
		1			

TP ld		TP/PLT/AG/CLASS/GL/BV-000_B			
TP label		MDS Configuration objects events for Glucose meter specialization.			
Coverage	Spec	[IEEE 11073-10417]			
	Testable items	MDSEvents 1; M			
Applicability		C_AG_OXP_000 AND C_AG_OXP_178			
Initial condit	ion	The simulated manager and the agent under test are in the unassociated state.			
Test procedu	ire	1. The simulated manager receives an association request from the agent under test.  2. The simulated manager responds with a result = accepted-unknown-config.  3. The agent responds with a "Remote Operation Invoke   Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager:  a. APDU Type    field-type = PrstApdu   field-length =2 bytes   field-value =0xE7 0x00   field-type = InvokeIDType   field-length =INT-U16   field-value =     field-value = Not relevant for this test>   c. message   field-type = roiv-cmip-confirmed-event-report     field-type = roiv-cmip-confirmed-event-report     field-type = roiv-cmip-confirmed-event-report     field-type = Relative two bytes     field-type = HANDLE     field-type = HANDLE     field-type = HANDLE     field-length =INT-U16     e. event-time (EventReportArgumentSimple)     field-length =INT-U32     field-length =INT-U32     field-type = ColD-Type     field-type = OID-Type     field-length =INT-U16     field-value=0x0D 0x1C (MDC_NOTI_CONFIG)   g. config-report-id (ConfigReport)     field-length = INT-U16     field-length = INT-U16     field-value = <it configuration="" matches="" tested="" the="">     F C_AG_GL_023 THEN attribute-value = 0x 06A4 (1700)     IF C_AG_GL_024 THEN attribute-value = 0x 06A5 (1701)</it>			

Notes	
Pass/Fail criteria	All checked values are as specified in the test procedure.
	☐ field- value = At least one MDC_MOC_VMO_METRIC_NU
	☐ field-length = INT-U16
	☐ field- type = OID-Type
	<ul> <li>h. obj-class ( ConfigReport → ConfigObjectList (ConfigObject))</li> </ul>
	<ul> <li>IF C_AG_OXP_181 THEN <between 0x00="" 0x40="" 0x7f="" 0xff="" and=""> for extended configuration.</between></li> </ul>

TP ld		TP/PLT/AG/CLASS/GL/BV-000_C		
TP label		MDS objects events for Glucose meter specialization.		
Coverage	Spec	[IEEE 11073-10417]		
	Testable	MDSEvents 3; M	MDSEvents 4; M	MDSEvents 5; M
	items	MDSEvents 6; M	MDSEvents 7; M	MDSEvents 8; M
		MDSEvents 9; M	MDSEvents 10; M	PMStoreObj 4; M
Applicability		C_AG_OXP_000 AND C_AG_ C_AG_OXP_184 OR C_AG_O	OXP_178 AND (C_AG_OXP_1 0XP_189)	82 OR C_AG_OXP_183 OR
Initial conditi	ion	The simulated manager and th	e agent under test are in the op	perating state.
Test procedure		<ol> <li>Take measurements for every supported object in the agent under test.</li> <li>Wait to receive every event report and check:         <ul> <li>a. APDU Type</li> <li>dield-type = Event Report</li> <li>dield-length = 2 bytes</li> <li>dield-value=0x01 0x01 (EventReportArgumentSimple, confirmed)</li> </ul> </li> <li>This field identifies the type of message sent by the agent, for the confirmed event configuration, roiv-cmip-confirmed-event-report.</li> </ol>		
Pass/Fail criteria		MDC_NOTI_SCAN_	REPORT_MP_FIXED REPORT_VAR	med Data APDU
Notes				

TP ld		TP/PLT/AG/CLASS/GL/BV-001			
TP label		Objects for Glucose meter specialization - Standard Configuration (1700 or 1701)			
Coverage	Spec	[IEEE 11073-10417]			
	Testable	BloodGL 1; M	BloodGL 4; M	BloodGL 4b; M	
	items	CtrlSol 1; M			

Applicability	C_AG_OXP_000 AND C_AG_OXP_178 AND (NOT_C_AG_OXP_181)		
Initial condition	The simulated manager and the agent are in the unassociated state.		
Test procedure	The simulated manager receives an association request from the agent under test.		
	2. The simulated manager responds with a result = accepted-unknown-config		
	The agent responds with a "Remote Operation Invoke   Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager.		
	4. Check that the field Dev-Config-Id is set to 0x06A4 (1700) OR 0x06A5 (1701), if it is not, the manager responds with a "unsupported-config" and waits for a new configuration.		
	5. Once the agent under test sends a standard configuration, check that:		
	IF Dev-Config-Id = 0x06A4) THEN Attribute-List:		
	<ul> <li>a. attribute-value (ConfigReport → ConfigObjectList (ConfigObject) → Attribute List), this value depends on the attribute Type. The values we have to check are:</li> </ul>		
	□ Blood Glucose Object is present → MDC_PART_SCADA (0x00 0x02), MDC_CONC_GLU_CAPILLARY_WHOLEBLOOD (0x4A 0x04)		
	IF Dev-Config-Id = 0x06A5) THEN Attribute-List:		
	<ul> <li>a. attribute-value (ConfigReport → ConfigObjectList (ConfigObject) → Attribute List), this value depends on the attribute Type. The values we have to check are:</li> </ul>		
	□ Blood Glucose Object is present → MDC_PART_SCADA (0x00 0x02), MDC_CONC_GLU_UNDETERMINED_PLASMA (0x72 0x70)		
	□ Control solution Object is present → MDC_PART_SCADA (0x00 0x02), MDC_CONC_GLU_CONTROL (0x71 0xD0)		
Pass/Fail criteria	All checked values are as specified in the test procedure and no other object listed.		
Notes			

TP Id TP label		TP/PLT/AG/CLASS/GL/BV-002		
		Objects for Glucose meter specialization - Extended Configuration		
Coverage	Spec	[IEEE 11073-10417]		
	Testable items	BloodGL 1; M	DevSenAn 3; R	BloodGL 28; M
Applicability	1	C_AG_OXP_000 AND	C_AG_OXP_178 AND C_AG_0	OXP_181
Initial condi	tion	The simulated manage	er and the agent are in the unass	sociated state.
Test procedure		<ol> <li>The simulated ma</li> <li>The agent respondence message with an</li> <li>Check that the fieresponds with a "to respond with a</li></ol>	anager responds with a result = and swith a "Remote Operation Involution MDC_NOTI_CONFIG event to sold Dev-Config-Id is in the extendunsupported-config" and waits founder test sends an extended correct (ConfigReport → ConfigObjectLs on the attribute type. The value	woke   Confirmed Event Report" send its configuration to the manager. led range; if it is not, the manager or a new configuration.  Infiguration, check that:  List (ConfigObject)→Attribute List), this

MDC CONC GLU CAPILLARY WHOLEBLOOD (0x71 0xB8)

- IF C\_AG\_GL\_015 THEN MDC\_CONC\_GLU\_CAPILLARY\_PLASMA (0x71 0xBC)
- IF C\_AG\_GL\_016 THEN MDC\_CONC\_GLU\_VENOUS\_WHOLEBLOOD (0x71 0xC0)
- IF C\_AG\_GL\_017 THEN MDC\_CONC\_GLU\_VENOUS\_PLASMA (0x71 0xC4)
- IF C\_AG\_GL\_018 THEN MDC CONC GLU ARTERIAL WHOLEBLOOD (0x71 0xC8)
- IF C\_AG\_GL\_019 THEN MDC\_CONC\_GLU\_ARTERIAL\_PLASMA (0x71 0xCC)
- IF C\_AG\_GL\_012 THEN MDC\_CONC\_GLU\_UNDETERMINED\_WHOLEBLOOD (0x72 0x6C)
- IF C\_AG\_GL\_013 THEN MDC\_CONC\_GLU\_UNDETERMINED\_PLASMA (0x72 0x70)
- IF C\_AG\_GL\_021 THEN MDC\_CONC\_GLU\_ISF (0x71 0xD4)
- ☐ Any of these objects may be present:
  - IF C\_AG\_GL\_001 THEN Control Solution numeric Object is present → MDC\_PART\_SCADA (0x00 0x02), MDC\_CONC\_GLU\_CONTROL (0x71 0xD0)
  - IF C\_AG\_GL\_002 THEN HbA1c numeric Object is present → MDC\_PART\_SCADA (0x00 0x02), MDC\_CONC\_HBA1C (0x71 0xDC)
  - IF C\_AG\_GL\_003 THEN Context Exercise numeric Object is present → MDC\_PART\_PHD\_DM (0X00 0X80), MDC\_CTXT\_GLU\_EXERCISE (0x71 0xE0)
  - IF C\_AG\_GL\_004 THEN Context Medication numeric Object is present → MDC\_PART\_ PHD\_DM (0x00 0x80), MDC\_CTXT\_MEDICATION (0x72 0x04)
  - IF C\_AG\_GL\_005 THEN Context Carbohydrates numeric Object is present → MDC\_PART\_ PHD\_DM (0x00 0x80), MDC\_CTXT\_GLU\_CARB (0x71 0xE4)
  - IF C\_AG\_GL\_007 THEN Device and Sensor annunciation status Enumeration Object is present → MDC\_PART\_ PHD\_DM (0x00 0x80), MDC\_GLU\_METER\_DEV\_STATUS (0x71 0xD8)
  - IF C\_AG\_GL\_008 THEN Context Meal enumeration Object is present → MDC\_PART\_ PHD\_DM (0x00 0x80), MDC\_CTXT\_GLU\_MEAL (0x72 0x48)
  - IF C\_AG\_GL\_009 THEN Context Sample Location enumeration Object is present → MDC\_PART\_ PHD\_DM (0x00 0x80), MDC\_CTXT\_GLU\_SAMPLELOCATION (0x72 0x34)
  - IF C\_AG\_GL\_010 THEN Context Tester enumeration Object is present → MDC\_PART\_ PHD\_DM (0x00 0x80), MDC\_CTXT\_GLU\_TESTER (0x72 0x5C)
  - IF C\_AG\_GL\_011 THEN Context Health enumeration Object is present → MDC\_PART\_ PHD\_DM (0x00 0x80), MDC\_CTXT\_GLU\_HEALTH (0x72 0x1C)

Pass/Fail criteria All checked values are as specified in the test procedure.

Notes

TP ld		TP/PLT/AG/CLASS/GL/BV-004				
TP Label		Blood Glucose Numeric Object - Standard configuration (1700 or 1701)				
Coverage Spec		[IEEE 11073-10417]				
J	Testable	BloodGL 2; M	BloodGL 4; M	BloodGL 6; M		
	items			·		
		BloodGL 8; M	BloodGL 10; M	BloodGL 12; C		
		BloodGL 14; R	BloodGL 18; R	BloodGL 20; M		
		BloodGL 22; C	BloodGL 24; R	BloodGL 26; R		
		MeasDatTx 6;M	BloodGL 31; M			
Applicability		C_AG_OXP_000 AND C_AG_	OXP_178 AND (NOT C_AG_O	XP_181)		
Initial conditi	ion	The simulated manager and th	e agent under are in the unasso	ociated state.		
Test procedu	ıre	The simulated manager re	eceives an association request f	rom the agent under test.		
		responds with a "Remote	esponds with a result = accepted Operation Invoke   Confirmed E ent to send its configuration to the	vent Report" message with an		
		3. Check that the field Dev-Config-Id is set to 0x06A4 (1700) or Dev-Config-Id is set to 0x06A5 (1701); if it is not, the manager responds with a "unsupported-config" and waits for a new configuration.				
		Once the agent under test sends a standard configuration, check that Blood Glucose     Object attributes are:				
		a. Mandatory attribute Handle				
		attribute-id = MDC_ATTR_ID_HANDLE				
		☐ attribute-type = F	IANDLE			
		☐ attribute-value =	0x00 0x01			
		b. Mandatory attribute Ty				
		☐ IF Dev-Config-Id				
			MDC_ATTR_ID_TYPE			
		attribute-type				
		<ul> <li>attribute-value = MDC_PART_SCADA (0x00 0x02), MDC_CONC_GLU_CAPILLARY_WHOLEBLOOD (0x71 0xB8).</li> </ul>				
		☐ IF Dev-Config-Id = 0x06A5:				
		<ul> <li>attribute-id = MDC_ATTR_ID_TYPE</li> </ul>				
		attribute-type	= TYPE			
		<ul> <li>attribute-value = MDC_PART_SCADA (0x00 0x02), MDC_CONC_GLU_UNDETERMINED_PLASMA (0x72 0x70).</li> </ul>				
		c. Mandatory attribute Metric-Spec-Small				
		☐ attribute-id = MDC_ATTR_METRIC_SPEC_SMALL				
		☐ attribute-type = N	MetricSpecSmall (BITS-16)			
		☐ attribute-value.length = 2 bytes				
		□ attribute-value ≠	0x00 0x00			
		Bit 0 (mss-ava)	ail-intermittent(0)), must be set			
		Bit 1 (mss-ava)	ail-stored-data(1)), must be set			
		Bit 2 (mss-upd-aperiodic(2)), must be set				

Notes	
Pass/Fail criteria	All checked values are as specified in the test procedure.
	f. No other attribute shall be present at configuration.
	□ attribute-value= MDC_ATTR_NU_VAL_OBS_BASIC MDC_ATTR_TIME_STAMP_ABS
	☐ attribute-value.length= <variable></variable>
	attribute-type = AttrValMap (sequence of attribute-id(OID-Type) and attribute-length(INT-U16))
	☐ attribute-id = MDC_ATTR_ATRIBUTE_VAL_MAP
	e. Mandatory attribute Attribute-Value-Map
	□ attribute-value= MDC_DIM_MILLI_G_PER_DL
	☐ attribute-value.length = 2 bytes
	☐ attribute-type = OID-Type(INT-U16)
	☐ attribute-id = MDC_ATTR_UNIT_CODE
	d. Mandatory attribute Unit-Code
	The other bits have to be 0.
	Bit 9 (mss-acc-agent-initiated(9)), must be set
	Bit 3 (mss-msmt-aperiodic(3)), must be set

TP Id		TP/PLT/AG/CLASS/GL/BV-005			
TP label		Blood Glucose Numeric Object- Extended configuration			
Coverage Spec		[IEEE 11073-10417]			
Testable		NumObj 3; C	NumObj 5; R	NumObj 6; R	
	items	NumObj 7; R	NumObj 8; R	NumObj 9; R	
		NumObj 12; R	NumObj 22; R	NumObj 23; R	
		NumObj 24; R	BloodGL 5; M	BloodGL 7; M	
		BloodGL 9; M	BloodGL 15; R	BloodGL 19; R	
		BloodGL 25; R	BloodGL 27; R	NumObj 25; R	
		NumObj 2;M			
Applicability	i	C_AG_OXP_000 AND C_AG_OXP_178 AND C_AG_OXP_181			
Initial condit	ion	The simulated manager and the agent under test are in the unassociated state.			
Test procedu	ure	The simulated manager receives an association request from the agent under test.			
		2. The simulated manager responds with a result = accepted-unknown-config. The agent responds with a "Remote Operation Invoke   Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager.			
		3. Check that the field Dev-Config-Id is set in the extended range; if it is not, the manager responds with a "unsupported-config" and waits for a new configuration.			
		4. Once the agent under test sends an extended configuration, check that Blood Glucose Object attributes are:			
		a. Mandatory attribute Type			

	☐ attribute-id = MDC_ATTR_ID_TYPE				
	attri	bute-type = TYPE			
	☐ attri next	bute-value = MDC_PART_SCADA (0x00 0x02), followed by one of the t:			
	•	IF C_AG_GL_014 THEN MDC_CONC_GLU_CAPILLARY_WHOLEBLOOD (0x71 0xB8)			
	•	IF C_AG_GL_015 THENMDC_CONC_GLU_CAPILLARY_PLASMA (0x71 0xBC)			
	•	IF C_AG_GL_016 THEN MDC_CONC_GLU_VENOUS_WHOLEBLOOD (0x71 0xC0)			
	•	IF C_AG_GL_017 THEN MDC_CONC_GLU_VENOUS_PLASMA (0x71 0xC4)			
	•	IF C_AG_GL_018 THEN MDC_CONC_GLU_ARTERIAL_WHOLEBLOOD (0x71 0xC8)			
	•	IF C_AG_GL_019 THEN MDC_CONC_GLU_ARTERIAL_PLASMA (0x71 0xCC)			
	•	IF C_AG_GL_012 THEN MDC_CONC_GLU_UNDETERMINED_WHOLEBLOOD (0x72 0x6C)			
	•	IF C_AG_GL_013 THEN MDC_CONC_GLU_UNDETERMINED_PLASMA (0x72 0x70)			
	•	IF C_AG_GL_021 THEN MDC_CONC_GLU_ISF (0x71 0xD4)			
b.	Not recor	mmended Supplemental -Types Attribute			
	attri	bute-id = MDC_ATTR_SPPLEMENTAL_TYPES			
	attri	bute-type = SupplementalTypeList			
	attri	bute-value.length = <variable> (Sequence of TYPE (TYPE.length= 4 bytes</variable>			
c.	Mandator	ry attribute Metric-Spec-Small			
	attri	bute-id = MDC_ATTR_METRIC_SPEC_SMALL			
	attri	bute-type = MetricSpecSmall (BITS-16)			
	attri	bute-value.length = 2 bytes			
	attri	bute-value ≠ 0x00 0x00			
	•	Bit 0 must be set (mss-avail-intermittent(0))			
	•	Bit 1 must be set (mss-avail-stored-data(1))			
	•	Bit 2 must be set (mss-upd-aperiodic(2))			
	•	Bit 3 must be set (mss-msmt-aperiodic(3))			
	•	Bit 9 must be set (mss-acc-agent-initiated(9))			
	•	Bit 12 may be set (mss-cat-manual(12)) if the reading is entered manually			
d.	IF Not red	commended attribute Metric-Structure-Small is present			
	attri	bute-id = MDC_ATTR_METRIC_STRUCTURE_SMALL			
	attri	bute-type = MetricStructureSmall			
		bute-value.length = <variable>(Sequence of (ms-struct.length =1byte(INT- + ms-comp-no =1byte(INT-U8)))</variable>			
e.	IF Not red	commended attribute Measurement-Status is present			
	attri	bute-id = MDC_ATTR_MSMT_STAT			
	☐ attri	bute-type = MeasurementStatus (BITS-16)			
	attri	bute-value.length =2 bytes			
f.	Condition	nal attribute Metric-Id is present			
	attri	bute-id = MDC_ATTR_ID_PHYSIO			

```
☐ attribute-type = OID-Type (INT-U16)
     ■ attribute-value.length= 2 bytes
g. IF Not recommended attribute Metric-Id-List is present
     □ attribute-id = MDC_ATTR_ID_PHYSIO_LIST
     ■ attribute-type = MetricIdList
     ☐ attribute-value.length= SEQUENCE OF OID-Type (INT-U16)
h. IF Not recommended attribute Metric-Id-Partition is present
     ☐ attribute-id = MDC ATTR METRIC ID PART
     ☐ attribute-type = NomPartition (INT-U16)
     ☐ attribute-value.length = 2 bytes
i. Mandatory attribute Unit-Code
     ☐ attribute-id = MDC_ATTR_UNIT_CODE
     ☐ attribute-type = OID-Type(INT-U16)

☐ attribute-value.length = 2 bytes

     ■ attribute-value= MDC_DIM_MILLI_G_PER_DL OR
        MDC_DIM_MILLI_MOLE_PER_L
j. IF Not recommended attribute Source-Handle-Reference is present
     ☐ attribute-id = MDC_ATTR_SOURCE_HANDLE_REF
     ☐ attribute-type = HANDLE (INT-U16)
     ☐ attribute-value.length = 2 bytes
k. IF Not recommended attribute Relative-Time-Stamp is present
     ☐ attribute-id = MDC_ATTR_TIME_STAMP_REL
     ☐ attribute-type = RelativeTime(INT-U32)

☐ attribute-value.length = 4 bytes

I. IF Not recommended attribute Measure-Active-Period
     ■ attribute-id = MDC_ATTR_TIME_PD_MSMT_ACTIVE
     ■ attribute-type = FLOAT type
     ☐ attribute-value.length = 4 bytes
m. IF Not recommended Compound-Simple-Nu-Observed-Value is present
     □ attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_SIMP
     ■ attribute-type = SimpleNuObsValueCmp
     ■ attribute-value.length =<variable>
n. IF Not recommended attribute Compound-Basic-Nu-Observed-Value is present
     ☐ attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_BASIC

☐ attribute-type = BasicNuObsValueCmp

     ■ attribute-value.length = <variable>
o. IF Not recommended attribute Compound-Nu-Observed-Value is present
     ☐ attribute-id = MDC_ATTR_NU_VAL_OBS
     ■ attribute-type = NuObsValue
     ■ attribute-value.length = <variable>
p. Not recommended attribute Compound-Nu-Observed-Value
     ■ attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_SIMP
     ☐ attribute-type = NuObsValueCmp
     ■ attribute-value.length = <variable>
```

Notes	
Pass/Fail criteria	All checked values are as specified in the test procedure.
	☐ attribute-value.length = 4 bytes
	☐ attribute-type = FLOAT-Type (INT-U32)
	☐ attribute-id = MDC_ATTR_NU_ACCUR_MSMT
	q. IF Recommended attribute Accuracy is present

	TD/DLT/40/GL400/GL/DV 000				
TP Id		TP/PLT/AG/CLASS/GL/BV-006			
TP label		HbA1c Numeric Object -	Extended configuration		
Coverage Spec		[IEEE 11073-10417]		1	
	Testable items	NumObj 3; C	NumObj 4; M	NumObj 5; R	
	items	NumObj 6; R	NumObj 7; R	NumObj 8; R	
		NumObj 9; R	NumObj 12; R	NumObj 16; O	
		NumObj 17; O	NumObj 20; R	NumObj 22; R	
		NumObj 23; R	NumObj 24; R	NumObj 25; R	
		HbA1c 1; M	HbA1c 2; M	HbA1c 3; M	
		HbA1c 4; M	HbA1c 5; M	NumObj 2; M	
Applicability		C_AG_OXP_000 AND C	_AG_OXP_178 AND C_AG_0	OXP_181 AND C_AG_GL_002	
Initial condit	ion	The simulated manager and the agent under test are in the unassociated state.			
Test procedu	ıre	<ol> <li>The simulated mana</li> <li>The agent responds message with an ME</li> <li>Check that the field I manager must response to the agent under attributes are:         <ol> <li>Mandatory attribute-id attribute-type attribute-va (0x71 0xDC)</li> <li>Not recommended attribute-type attribute-type attribute-type attribute-type attribute-type attribute-type attribute-type attribute-type attribute-type attribute-va</li> <li>Mandatory attribute-va</li> <li>Mandatory attribute-id attribute-id</li> </ol> </li> </ol>	ger responds with a result = a with a "Remote Operation Inv DC_NOTI_CONFIG event to s Dev-Config-Id is in the extend and with an "unsupported-conf er test sends an extended conf ute Type = MDC_ATTR_ID_TYPE De = TYPE Iue = MDC_PART_SCADA (0 C) ed Supplemental -Types Attrib = MDC_ATTR_SPPLEMENT De = SupplementalTypeList	voke   Confirmed Event Report" end its configuration to the manager. ed range, if it is not, the simulated fig" and wait for a new configuration. Infiguration, check that HbA1c object exception (CONC_HBA1C) bute AL_TYPES ence of TYPE (TYPE.length= 4 bytes EC_SMALL	

		attribute-value ≠ 0x00 0x00
		Bit 0 must be set (mss-avail-intermittent(0))
		Bit 1 must be set (mss-avail-stored-data(1))
		Bit 2 must be set (mss-upd-aperiodic(2))
		Bit 3 must be set (mss-msmt-aperiodic(3))
		Bit 9 must be set (mss-msm-apenduc(9))     Bit 9 must be set (mss-acc-agent-initiated(9))
٦	IE NI	<ul> <li>Bit 12 may be set (mss-cat-manual(12)) if the reading is entered manually of recommended attribute Metric-Structure-Small is present</li> </ul>
u.		attribute-id = MDC_ATTR_METRIC_STRUCTURE_SMALL
		attribute-type = MetricStructureSmall
		attribute-value.length = <variable>(Sequence of (ms-struct.length =1byte(INT-</variable>
		U8) + ms-comp-no =1byte(INT-U8)))
e.	IF No	ot recommended attribute Measurement-Status is present
		attribute-id = MDC_ATTR_MSMT_STAT
		attribute-type = MeasurementStatus (BITS-16)
		attribute-value.length =2 bytes
f.	Cond	ditional attribute Metric-Id is present
		attribute-id = MDC_ATTR_ID_PHYSIO
		attribute-type = OID-Type (INT-U16)
		attribute-value.length= 2 bytes
		The [Metric-Id-List] attribute shall be used if a compound observed value is used, which does not incorporate the Metric-Id directly. The order of the Metric-Id-List shall correspond to the order of the elements in the compound observed value. Only one attribute of Metric-Id and Metric-Id-List shall be present.
g.	IF N	ot recommended attribute Metric-Id-List is present
		attribute-id = MDC_ATTR_ID_PHYSIO_LIST
		attribute-type = MetricIdList
		attribute-value.length= SEQUENCE OF OID-Type (INT-U16)
		The [Metric-Id-List] attribute shall be used if a compound observed value is used, which does not incorporate the Metric-Id directly. The order of the Metric-Id-List shall correspond to the order of the elements in the compound observed value. Only one attribute of Metric-Id and Metric-Id-List shall be present.
h.	IF No	ot recommended attribute Metric-Id-Partition is present
		attribute-id = MDC_ATTR_METRIC_ID_PART
		attribute-type = NomPartition (INT-U16)
		attribute-value.length = 2 bytes
i.	Man	datory attribute Unit-Code
		attribute-id = MDC_ATTR_UNIT_CODE
		attribute-type = OID-Type (INT-U16)
		attribute-value.length = 2 bytes
		attribute-value= MDC_DIM_PERCENT
j.	IF No	ot recommended attribute Source-Handle-Reference is present
		attribute-id = MDC_ATTR_SOURCE_HANDLE_REF
		attribute-type = HANDLE (INT-U16)
		attribute-value.length = 2 bytes

		Agent supports fixed or variable format MDS event report and it does not support M-Store THEN Mandatory attribute Absolute-Time-Stamp
		☐ attribute-id = MDC_ATTR_TIME_STAMP_ABS
		☐ attribute-type = AbsoluteTime
		□ attribute-value.length = 8 bytes
	I. C	optional attribute Relative-Time-Stamp
		☐ attribute-id = MDC_ATTR_TIME_STAMP_REL
		☐ attribute-type = RelativeTime(INT-U32)
		☐ attribute-value.length = 4 bytes
	m. C	conditional attribute HiRes-Time-Stamp
		☐ attribute-id = MDC_ATTR_TIME_STAMP_REL_HI_RES
		☐ attribute-type = HighResRelativeTime
		☐ attribute-value.length = OCTET STRING (SIZE(8))
	n. IF	Not recommended Compound-Simple-Nu-Observed-Value is present
		☐ attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_SIMP
		☐ attribute-type = SimpleNuObsValueCmp
		☐ attribute-value.length = <variable></variable>
		Agent supports fixed or variable format MDS event report and it does not support M-Store THEN Mandatory attribute Basic-Nu-Observed-Value
		☐ attribute-id = MDC_ATTR_NU_VAL_OBS_BASIC
		☐ attribute-type = BasicNuObsValue
		☐ attribute-value.length = SFLOAT-Type (INT-U16)
	p. IF	Not recommended attribute Compound-Basic-Nu-Observed-Value is present
		☐ attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_BASIC
		☐ attribute-type = BasicNuObsValueCmp
		☐ attribute-value.length = <variable></variable>
	q. IF	Not recommended attribute Compound-Nu-Observed-Value is present
		☐ attribute-id = MDC_ATTR_NU_VAL_OBS
		□ attribute-type = NuObsValue
		☐ attribute-value.length = <variable></variable>
	r. N	ot recommended attribute Compound-Nu-Observed-Value
		☐ attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_SIMP
		□ attribute-type = NuObsValueCmp
		☐ attribute-value.length = <variable></variable>
	s. R	ecommended attribute Accuracy
		☐ attribute-id = MDC_ATTR_NU_ACCUR_MSMT
		□ attribute-type = FLOAT-Type (INT-U32)
		☐ attribute-value.length = 4 bytes
Pass/Fail criteria	All checke	ed values are as specified in the test procedure.
		·
Notes		

TP ld		TP/PLT/AG/CLASS/GL/BV-007			
TP label		Context Exercise Numeric Object - Extended configuration			
Coverage Spec		[IEEE 11073-10417]			
	Testable	NumObj 3; C	NumObj 4; M	NumObj 5; R	
	items	NumObj 6; R	NumObj 7; R	NumObj 8; R	
		NumObj 9; R	NumObj 12; R	NumObj 16; O	
		NumObj 17; O	NumObj 20; R	NumObj 22; R	
		NumObj 23; R	NumObj 24; R	NumObj 25; R	
		ContextEx 1; M	ContextEx 2; M	ContextEx 3; M	
		ContextEx 4; M	ContextEx 5; M	NumObj 2; M	
		ContextEx 6; M			
Applicability		C_AG_OXP_000 AND C_AG_	OXP_178 AND C_AG_OXP_18	31 AND C_AG_GL_003	
Initial conditi	ion	C_AG_OXP_000 AND C_AG_OXP_178 AND C_AG_OXP_181 AND C_AG_GL_003  The simulated manager and the agent under test are in the unassociated state.			
Initial condition  Test procedure		2. The simulated manager responds with a message with an MDC_N  4. Check that the field Dev-Comanager must respond with a mager must respond with a	rype  C_ATTR_ID_TYPE  TYPE  MDC_PART_PHD_DM (0x00 0x00 0x00 0x00 0x00 0x00 0x00 0x	d-unknown-config confirmed Event Report" configuration to the manager. ge, if it is not, the simulated wait for a new configuration. on, check that Context  (02),  PES TYPE (TYPE.length= 4 bytes  ALL	

		<ul> <li>Bit 12 must be set (mss-cat-manual(12)) if the reading is entered manually</li> </ul>
d.	IF N	ot recommended attribute Metric-Structure-Small is present
		attribute-id = MDC_ATTR_METRIC_STRUCTURE_SMALL
		attribute-type = MetricStructureSmall
		attribute-value.length = <variable>(Sequence of (ms-struct.length =1byte(INT-U8) + ms-comp-no =1byte(INT-U8)))</variable>
e.	IF N	ot recommended attribute Measurement-Status is present
		attribute-id = MDC_ATTR_MSMT_STAT
		attribute-type = MeasurementStatus(BITS-16)
		attribute-value.length = 2 bytes
f.	Con	ditional attribute Metric-Id is present
		attribute-id = MDC_ATTR_ID_PHYSIO
		attribute-type = OID-Type(INT-U16)
		attribute-value.length= 2 bytes
		The [Metric-Id-List] attribute shall be used if a compound observed value is used, which does not incorporate the Metric-Id directly. The order of the Metric-Id-List shall correspond to the order of the elements in the compound observed value. Only one attribute of Metric-Id and Metric-Id-List shall be present.
g.	IF N	ot recommended attribute Metric-Id-List is present
		attribute-id = MDC_ATTR_ID_PHYSIO_LIS
		attribute-type = MetricIdList
		attribute-value.length= SEQUENCE OF OID-Type (INT-U16)
		The [Metric-Id-List] attribute shall be used if a compound observed value is used, which does not incorporate the Metric-Id directly. The order of the Metric-Id-List shall correspond to the order of the elements in the compound observed value. Only one attribute of Metric-Id and Metric-Id-List shall be present.
h.	IF N	ot recommended attribute Metric-Id-Partition is present
		attribute-id = MDC_ATTR_METRIC_ID_PART
		attribute-type = NomPartition (INT-U16)
		attribute-value.length = 2 bytes
i.	Man	datory attribute Unit-Code
		attribute-id = MDC_ATTR_UNIT_CODE
		attribute-type = OID-Type(INT-U16)
		attribute-value.length = 2 bytes
		attribute-value= MDC_DIM_PERCENT
j.	IF N	ot recommended attribute Source-Handle-Reference is present
		attribute-id = MDC_ATTR_SOURCE_HANDLE_REF
		attribute-type = HANDLE (INT-U16)
		attribute-value.length = 2 bytes
k.		e agent supports fixed or variable format MDS event report and it does not port PM-Store THEN Mandatory attribute Absolute-Time-Stamp
		attribute-id = MDC_ATTR_TIME_STAMP_ABS
		attribute-type = AbsoluteTime
		attribute-value.length = 8 bytes
I.	Opti	onal attribute Relative-Time-Stamp
		attribute-id = MDC ATTR TIME STAMP REL

	☐ attribute-type = RelativeTime (INT-U32)
	☐ attribute-value.length = 4 bytes
	m. Conditional attribute HiRes-Time-Stamp
	□ attribute-id = MDC_ATTR_TIME_STAMP_REL_HI_RES
	□ attribute-type = HighResRelativeTime
	☐ attribute-value.length = OCTET STRING (SIZE(8))
	n. Mandatory attribute Measure-Active-Period
	□ attribute-id = MDC_ATTR_TIME_PD_MSMT_ACTIVE
	□ attribute-type = FLOAT type
	☐ attribute-value.length = 4 bytes
	o. IF Not recommended Compound-Simple-Nu-Observed-Value is present
	□ attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_SIMP
	□ attribute-type = SimpleNuObsValueCmp
	☐ attribute-value.length = <variable></variable>
	<ul> <li>p. IF the agent supports fixed or variable format MDS event report and it does not support PM-Store THEN Mandatory attribute Basic-Nu-Observed-Value</li> </ul>
	☐ attribute-id = MDC_ATTR_NU_VAL_OBS_BASIC
	☐ attribute-type = BasicNuObsValue
	☐ attribute-value.length = SFLOAT-Type (INT-U16)
	☐ attribute-value= value within the range:[0,100]
	q. IF Not recommended attribute Compound-Basic-Nu-Observed-Value is present
	☐ attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_BASIC
	☐ attribute-type = BasicNuObsValueCmp
	☐ attribute-value.length = <variable></variable>
	r. IF Not recommended attribute Compound-Nu-Observed-Value is present
	☐ attribute-id = MDC_ATTR_NU_VAL_OBS
	☐ attribute-type = NuObsValue
	☐ attribute-value.length = <variable></variable>
	s. Not recommended attribute Compound-Nu-Observed-Value
	☐ attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_SIMP
	☐ attribute-type = NuObsValueCmp
	☐ attribute-value.length = <variable></variable>
	t. Recommended attribute Accuracy
	☐ attribute-id = MDC_ATTR_NU_ACCUR_MSMT
	☐ attribute-type = FLOAT-Type (INT-U32)
	☐ attribute-value.length = 4 bytes
Pass/Fail criteria	All checked values are as specified in the test procedure.
	·
Notes	

TP ld		TP/PLT/AG/CLASS/GL/BV-008				
TP label		Context Medication Numeric Object - Extended configuration				
Coverage Spec		[IEEE 11073-10417]				
	Testable	NumObj 3; C	NumObj 4; M	NumObj 5; R		
	items	NumObj 6; R	NumObj 8; R	NumObj 9; R		
		NumObj 12; R	NumObj 16; O	NumObj 17; O		
		NumObj 20; R	NumObj 22; R	NumObj 23; R		
		NumObj 24; R	NumObj 25; R	ContextMed 1; M		
		ContextMed 2; M	ContextMed 3; M	ContextMed 5; M		
		ContextMed 6; M	NumObj 2; M	ContextMed 7; M		
Applicability	y	C_AG_OXP_000 AND C_AG_	OXP_178 AND C_AG_OXP_18	31 AND C_AG_GL_004		
Initial condi	tion	The simulated manager and the	e agent under test are in the ur	associated state.		
Test proced	lure	The simulated manager re	eceives an association request f	rom the agent under test.		
		2. The simulated manager responds with a result = accepted-unknown-config  2. The simulated manager responds with a result = accepted-unknown-config				
		The agent responds with a "Remote Operation Invoke   Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager.				
		Check that the field Dev-Config-Id is in the extended range, if it is not, the simulated manager must respond with an "unsupported-config" and wait for a new configuration.				
		5. Once the agent under test sends an extended configuration, check that Context Exercise object attributes are:				
		a. Mandatory attribute Type				
		☐ attribute-id = MDC_ATTR_ID_TYPE				
		□ attribute-type = TYPE				
		☐ attribute-value =MDC_PART_PHD_DM (0x00 0x80), MDC_CTXT_MEDICATION (0x72 0x04)				
		b. Not recommended Supplemental –Types Attribute				
		☐ attribute-id = MDC_ATTR_SUPPLEMENTAL_TYPES				
		☐ attribute-type = SupplementalTypeList				
		☐ attribute-value.length = <variable> (Sequence of TYPE (TYPE.length= 4 bytes</variable>				
		c. Mandatory attribute Metric-Spec-Small				
		☐ attribute-id = MDC_ATTR_METRIC_SPEC_SMALL				
		☐ attribute-type = MetricSpecSmall (BITS-16)				
		☐ attribute-value.length = 2 bytes				
		☐ attribute-value ≠ 0x00 0x00				
			nust be set (mss-avail-intermittent(0))			
		• Bit 1 mu	st be set (mss-avail-stored-data	avail-stored-data(1))		
		Bit 2 mu	ust be set (mss-upd-aperiodic(2)	))		
		Bit 3 mu	st be set (mss-msmt-aperiodic)	(3))		
		Bit 9 mu	st be set (mss-acc-agent-initiat	ed(9))		
		• Bit 12 m	nay be set (mss-cat-manual(12)	) if the reading is entered		

	manually					
d.	IF Not recommended attribute Metric-Structure-Small is present					
	□ attribute-id = MDC_ATTR_METRIC_STRUCTURE_SMALL					
	☐ attribute-type = MetricStructureSmall					
	attribute-value.length = <variable>(Sequence of (ms-struct.length =1byte(IN U8) + ms-comp-no =1byte(INT-U8)))</variable>					
e.	Not recommended attribute Measurement-Status					
	☐ attribute-id = MDC_ATTR_MSMT_STAT					
	□ attribute-type = MeasurementStatus(BITS-16)					
	☐ attribute-value.length = 2 bytes					
f.	Mandatory attribute Metric-Id					
	□ attribute-id = MDC_ATTR_ID_PHYSIO					
	☐ attribute-type = OID-Type(INT-U16)					
	☐ attribute-value.length= 2 bytes					
	□ attrbute-value = One of the following					
	<ul> <li>MDC_CTXT_MEDICATION_RAPIDACTING (0x72 0x08)</li> </ul>					
	<ul> <li>MDC_CTXT_MEDICATION_SHORTACTING (0x72 0x0C)</li> </ul>					
	<ul> <li>MDC_CTXT_MEDICATION_INTERMEDIATEACTING (0x72 0x10)</li> </ul>					
	<ul> <li>MDC_CTXT_MEDICATION_LONGACTING (0x72 0x14)</li> </ul>					
	<ul> <li>MDC_CTXT_MEDICATION_PREMIX (0x72 0x18)</li> </ul>					
g.	IF Not recommended attribute Metric-Id-Partition is present					
	☐ attribute-id = MDC_ATTR_METRIC_ID_PART					
	☐ attribute-type = NomPartition (INT-U16)					
	☐ attribute-value.length = 2 bytes					
h.	Mandatory attribute Unit-Code					
	☐ attribute-id = MDC_ATTR_UNIT_CODE					
	☐ attribute-type = OID-Type(INT-U16)					
	☐ attribute-value.length = 2 bytes					
	■ attribute-value= MDC_DIM_MILLI_G OR MDC_DIM_MILLI_L					
i.	IF Not recommended attribute Source-Handle-Reference is present					
	☐ attribute-id = MDC_ATTR_SOURCE_HANDLE_REF					
	☐ attribute-type = HANDLE (INT-U16)					
	☐ attribute-value.length = 2 bytes					
j.	IF the agent supports fixed or variable format MDS event report and it does not support PM-Store THEN Mandatory attribute Absolute-Time-Stamp					
	□ attribute-id = MDC_ATTR_TIME_STAMP_ABS					
	☐ attribute-type = AbsoluteTime					
	☐ attribute-value.length = 8 bytes					
k.	Optional attribute Relative-Time-Stamp					
	□ attribute-id = MDC_ATTR_TIME_STAMP_REL					
	☐ attribute-type = RelativeTime (INT-U32)					
	☐ attribute-value.length = 4 bytes					
I.	Conditional attribute HiRes-Time-Stamp					
	□ attribute-id = MDC_ATTR_TIME_STAMP_REL_HI_RES					
	☐ attribute-type = HighResRelativeTime					

	☐ attribute-value.length = OCTET STRING (SIZE(8))
	m. Not recommended attribute Measure-Active-Period
	☐ attribute-id = MDC_ATTR_TIME_PD_MSMT_ACTIVE
	☐ attribute-type = FLOAT type
	☐ attribute-value.length = 4 bytes
	n. IF Not recommended Compound-Simple-Nu-Observed-Value is present
	☐ attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_SIMP
	□ attribute-type = SimpleNuObsValueCmp
	☐ attribute-value.length = <variable></variable>
	<ul> <li>IF the agent supports fixed or variable format MDS event report and it does not support PM-Store THEN Mandatory attribute Basic-Nu-Observed-Value</li> </ul>
	□ attribute-id = MDC_ATTR_NU_VAL_OBS_BASIC
	☐ attribute-type = BasicNuObsValue
	☐ attribute-value.length = SFLOAT-Type (INT-U16)
	p. IF Not recommended attribute Compound-Basic-Nu-Observed-Value is present
	□ attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_BASIC
	□ attribute-type = BasicNuObsValueCmp
	☐ attribute-value.length = <variable></variable>
	q. IF Not recommended attribute Compound-Nu-Observed-Value is present
	☐ attribute-id = MDC_ATTR_NU_VAL_OBS
	☐ attribute-type = NuObsValue
	☐ attribute-value.length = <variable></variable>
	r. Not recommended attribute Compound-Nu-Observed-Value
	□ attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_SIMP
	□ attribute-type = NuObsValueCmp
	☐ attribute-value.length = <variable></variable>
	s. Recommended attribute Accuracy
	☐ attribute-id = MDC_ATTR_NU_ACCUR_MSMT
	☐ attribute-type = FLOAT-Type (INT-U32)
	☐ attribute-value.length = 4 bytes
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id		TP/PLT/AG/CLASS/GL/BV-009			
TP label		Context Carbohydrates Numeric Object - Extended configuration			
Coverage	Spec	[IEEE 11073-10417]			
	Testable items	NumObj 3; C	NumObj 4; M	NumObj 5; R	
		NumObj 6; R	NumObj 8; R	NumObj 9; R	
		NumObj 12; R	NumObj 16; O	NumObj 17; O	
		NumObj 20; R	NumObj 22; R	NumObj 23; R	
		NumObj 24; R	NumObj 25; R	ContextCarb 1; M	
		ContextCarb 2; M	ContextCarb 3; M	ContextCarb 4; M	
		ContextCarb 5; M	NumObj 2; M	ContextCarb 6; M	
Applicability		C_AG_OXP_000 AND C_AG_OXP_178 AND C_AG_OXP_181 AND C_AG_GL_005			
Initial condit	ion	The simulated manager and the agent under test are in the unassociated state.			
Test procedu	ıre	The simulated manager receives an association request from the agent under test.			
		The simulated manager responds with a result = accepted-unknown-config			
		3. The agent responds with a "Remote Operation Invoke   Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager.			
		4. Check that the field Dev-Config-Id is in the extended range, if it is not, the simulated manager must respond with an "unsupported-config" and wait for a new configuration.			
		<ol><li>Once the agent under test sends an extended configuration, check that Context Carbohydrates object attributes are:</li></ol>			
		a. Mandatory attribute Type			
		☐ attribute-id = MDC_ATTR_ID_TYPE			
		☐ attribute-type = TYPE			
		□ attribute-value =MDC_PART_PHD_DM (0x00 0x80), MDC_CTXT_GLU_CARB (0x71 0xE4)			
		b. Not recommended Supplemental-Types Attribute			
		☐ attribute-id = MDC_ATTR_SPPLEMENTAL_TYPES			
		☐ attribute-type = SupplementalTypeList			
		☐ attribute-value.length = <variable> (Sequence of TYPE (TYPE.length= 4 bytes</variable>			
		c. Mandatory attribute Metric-Spec-Small			
		☐ attribute-id = MDC_ATTR_METRIC_SPEC_SMALL			
		☐ attribute-type = MetricSpecSmall (BITS-16)			
		☐ attribute-value ± 0x00 0x00			
		<ul> <li>□ attribute-value ≠ 0x00 0x00</li> <li>• Bit 0 must be set (mss-avail-intermittent(0))</li> </ul>			
		Bit 0 must be set (mss-avail-intermittent(0))      Bit 1 must be(mss-avail-stored-data(1))			
			pe set (mss-upd-aperiodic(2))		
		Bit 3 must be set (mss-msmt-aperiodic(3))      Bit 9 must be set (mss-acc-agent-initiated(9))			
		Bit 9 must be set (mss-acc-agent-initiated(9))			

		Dit 10 may be not (man not manual(10)) if the reading is entered manually
٦	IE NI	Bit 12 may be set (mss-cat-manual(12)) if the reading is entered manually at recommended attribute Matrix Structure Small in present.
u.		ot recommended attribute Metric-Structure-Small is present
	_	attribute-id = MDC_ATTR_METRIC_STRUCTURE_SMALL attribute-type = MetricStructureSmall
		attribute-value.length = <variable>(Sequence of (ms-struct.length =1byte(INT-</variable>
	_	U8) + ms-comp-no =1byte(INT-U8)))
e.	Not r	recommended attribute Measurement-Status
		attribute-id = MDC_ATTR_MSMT_STAT
		attribute-type = MeasurementStatus(BITS-16)
		attribute-value.length = 2 bytes
f.	Man	datory attribute Metric-Id
		attribute-id = MDC_ATTR_ID_PHYSIO
		attribute-type = OID-Type(INT-U16)
		attribute-value.length= 2 bytes
		attrbute-value = One of the following
		<ul> <li>MDC_CTXT_GLU_CARB_BREAKFAST (0x71 0xE8)</li> </ul>
		<ul> <li>MDC_CTXT_GLU_CARB_LUNCH (0x71 0xEC)</li> </ul>
		<ul> <li>MDC_CTXT_GLU_CARB_DINNER (0x71 0xF0)</li> </ul>
		<ul> <li>MDC_CTXT_GLU_CARB_SNACK (0x71 0xF4)</li> </ul>
		<ul> <li>MDC_CTXT_GLU_CARB_DRINK (0x71 0xF8)</li> </ul>
		<ul> <li>MDC_CTXT_GLU_CARB_SUPPER (0x71 0xFC)</li> </ul>
		<ul> <li>MDC_CTXT_GLU_CARB_BRUNCH (0x72 0x00)</li> </ul>
g.	IF No	ot recommended attribute Metric-Id-Partition is present
		attribute-id = MDC_ATTR_METRIC_ID_PART
		attribute-type = NomPartition (INT-U16)
		attribute-value.length = 2 bytes
h.	Man	datory attribute Unit-Code
		attribute-id = MDC_ATTR_UNIT_CODE
		attribute-type = OID-Type(INT-U16)
		attribute-value.length = 2 bytes
		attribute-value= MDC_DIM_ G
i.	IF No	ot recommended attribute Source-Handle-Reference is present
		attribute-id = MDC_ATTR_SOURCE_HANDLE_REF
		attribute-type = HANDLE (INT-U16)
		attribute-value.length = 2 bytes
j.		e agent supports fixed or variable format MDS event report and it does not ort PM-Store THEN Mandatory attribute Absolute-Time-Stamp
		attribute-id = MDC_ATTR_TIME_STAMP_ABS
		attribute-type = AbsoluteTime
		attribute-value.length = 8 bytes
k.	Optio	onal attribute Relative-Time-Stamp
		attribute-id = MDC_ATTR_TIME_STAMP_REL
		attribute-type = RelativeTime (INT-U32)
		attribute-value.length = 4 bytes

	Conditional attribute HiRes-Time-Stamp
	☐ attribute-id = MDC_ATTR_TIME_STAMP_REL_HI_RES
	☐ attribute-type = HighResRelativeTime
	☐ attribute-value.length = OCTET STRING (SIZE(8))
	m. Not recommended attribute Measure-Active-Period
	☐ attribute-id = MDC_ATTR_TIME_PD_MSMT_ACTIVE
	☐ attribute-type = FLOAT type
	☐ attribute-value.length = 4 bytes
	n. IF Not recommended Compound-Simple-Nu-Observed-Value is present
	☐ attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_SIMP
	□ attribute-type = SimpleNuObsValueCmp
	☐ attribute-value.length = <variable></variable>
	<ul> <li>IF the agent supports fixed or variable format MDS event report and it does not support PM-Store THEN Mandatory attribute Basic-Nu-Observed-Value</li> </ul>
	☐ attribute-id = MDC_ATTR_NU_VAL_OBS_BASIC
	□ attribute-type = BasicNuObsValue
	☐ attribute-value.length = SFLOAT-Type (INT-U16)
	p. IF Not recommended attribute Compound-Basic-Nu-Observed-Value is present
	□ attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_BASIC
	□ attribute-type = BasicNuObsValueCmp
	☐ attribute-value.length = <variable></variable>
	q. IF Not recommended attribute Compound-Nu-Observed-Value is present
	□ attribute-id = MDC_ATTR_NU_VAL_OBS
	☐ attribute-type = NuObsValue
	☐ attribute-value.length = <variable></variable>
	r. IF Not recommended attribute Compound-Nu-Observed-Value is present
	□ attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_SIMP
	☐ attribute-type = NuObsValueCmp
	☐ attribute-value.length = <variable></variable>
	s. Recommended attribute Accuracy
	□ attribute-id = MDC_ATTR_NU_ACCUR_MSMT
	☐ attribute-type = FLOAT-Type (INT-U32)
	☐ attribute-value.length = 4 bytes
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP ld		TP/PLT/AG/CLASS/GL/BV-010			
TP label		Control Solution Numeric Object - Standard configuration 1701			
Coverage	Spec	[IEEE 11073-10417]			
	Testable items	CtrlSol 2; M	CtrlSol 4; M	CtrlSol 5; M	
		CtrlSol 6; M	CtrlSol 8; M	CtrlSol 10; M	
		CtrlSol 12; M			
Applicability		C_AG_OXP_000 AND C_AG_OXP_178 AND C_AG_GL_024 AND (NOT C_AG_OXP_181)			
Initial condit	ion	The simulated manager and the agent under test are in the unassociated state.			
Test procedure		<ol> <li>The simulated manager receives an association request from the agent under test.</li> <li>The simulated manager responds with a result = accepted-unknown-config. The agent responds with a "Remote Operation Invoke   Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager.</li> <li>Check that the field Dev-Config-Id is set to 0x06A5 (1701); if it is not, the manager responds with a "unsupported-config" and waits for a new configuration.</li> </ol>			
		Once the agent under test sends a standard configuration, check that Control Solution     Object attributes are:			
		a. Mandatory attribute Handle			
		☐ attribute-id = MDC_ATTR_ID_HANDLE			
		☐ attribute-type = HANDLE			
		☐ attribute-value = 0x00 0x02			
		b. Mandatory attribute Type			
		☐ attribute-id = MDC_ATTR_ID_TYPE			
		☐ attribute-type = TYPE			
		□ attribute-value = MDC_PART_SCADA (0x00 0x02), MDC_CONC_GLU_CONTROL (0x71 0xD0).			
		c. Mandatory attribute Metric-Spec-Small			
		☐ attribute-id = MD	C_ATTR_METRIC_SPEC_SMA	ALL	
		☐ attribute-type = I	MetricSpecSmall (BITS-16)		
		☐ attribute-value.le	ength = 2 bytes		
		□ attribute-value ≠	0x00 0x00		
		Bit 0 (mss-av	vail-intermittent(0)), must be set		
		Bit 1 (mss-av	vail-stored-data(1)), must be set		
		Bit 2 (mss-up	od-aperiodic(2)), must be set		
		Bit 3 (mss-m	smt-aperiodic(3)), must be set		
		Bit 9 (mss-ad	acc-agent-initiated(9)), must be set		
		The other bits have to be 0.			
		d. Mandatory attribute U	nit-Code		
		☐ attribute-id = MD	C_ATTR_UNIT_CODE		
		☐ attribute-type = OID-Type(INT-U16)			
		□ attribute-value.le	ength = 2 bytes		
		☐ attribute-value= MDC_DIM_MILLI_G_PER_DL			

	e. Mandatory attribute Attribute-Value-Map			
	☐ attribute-id = MDC_ATTR_ATRIBUTE_VAL_MAP			
	attribute-type = AttrValMap (sequence of attribute-id(OID-Type) and attribute-length(INT-U16))			
	☐ attribute-value.length= <variable></variable>			
	<ul><li>attribute-value= MDC_ATTR_NU_VAL_OBS_BASIC MDC_ATTR_TIME_STAMP_ABS</li></ul>			
	f. No other attribute shall be present at configuration.			
Pass/Fail criteria All checked values are specified in the test procedure.				
Notes				

TP ld		TD/DLT/AG/CLASS/GL/B\/-010_A			
		TP/PLT/AG/CLASS/GL/BV-010_A			
TP label		Control Solution Numeric Object - Extended configuration			
Coverage	Spec	[IEEE 11073-10417]			
	Testable items	CtrlSol 4; M		CtrlSol 5; M	CtrlSol 7; M
Applicability		C_AG_OXP_000 AND C_AG_OXP_178 AND C_AG_GL_001 AND C_AG_OXP_181			
Initial condition		The simulated manager and the agent under test are in the unassociated state.			
Test procedu	ure	The simulated manager receives an association request from the agent under test.			
		2. The simulated manager responds with a result = accepted-unknown-config. The agent responds with a "Remote Operation Invoke   Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager.			
		3. Check that the field Dev-Config-Id is set in the extended range; if it is not, the manager responds with a "unsupported-config" and waits for a new configuration.			
		4. Once the agent under test sends an extended configuration, check that Control Solution Object attributes are:			
		a. Mandatory attribute Type			
		☐ attribute-id = MDC_ATTR_ID_TYPE			
		☐ attribute-type = TYPE			
		□ attribute-value = MDC_PART_SCADA (0x00 0x02), MDC_CONC_GLU_CONTROL (0x71 0xD0).			
		b. Not recommended Supplemental –Types Attribute			
		☐ attribute-id = MDC_ATTR_SPPLEMENTAL_TYPES			
		☐ attribute-type = SupplementalTypeList			
		☐ attribute-value.length = <variable> (Sequence of TYPE (TYPE.length= 4 bytes</variable>			
		c. Mandatory attribute Metric-Spec-Small			
		☐ attribute-id = MDC_ATTR_METRIC_SPEC_SMALL			
		☐ attribute-type = MetricSpecSmall (BITS-16)			
		☐ attribute-value.length = 2 bytes			
		☐ attribute-value ≠ 0x00 0x00			
		Bit 0 must be set (mss-avail-intermittent(0))			
		Bit 1 must be set (mss-avail-stored-data(1))			
		Bit 2 must be set (mss-upd-aperiodic(2))			

Bit 3 must be set (mss-msmt-aperiodic(3)) Bit 9 must be set (mss-acc-agent-initiated(9)) d. IF Not recommended attribute Metric-Structure-Small is present ☐ attribute-id = MDC\_ATTR\_METRIC\_STRUCTURE\_SMALL ☐ attribute-type = MetricStructureSmall ☐ attribute-value.length = <variable>(Sequence of (ms-struct.length =1byte(INT-U8) + ms-comp-no =1byte(INT-U8))) e. IF Not recommended attribute Measurement-Status is present ☐ attribute-id = MDC\_ATTR\_MSMT\_STAT □ attribute-type = MeasurementStatus (BITS-16) ■ attribute-value.length =2 bytes f. Conditional attribute Metric-Id is present ☐ attribute-id = MDC\_ATTR\_ID\_PHYSIO ☐ attribute-type = OID-Type (INT-U16) ■ attribute-value.length= 2 bytes g. IF Not recommended attribute Metric-Id-List is present □ attribute-id = MDC\_ATTR\_ID\_PHYSIO\_LIST ☐ attribute-type = MetricIdList ☐ attribute-value.length= SEQUENCE OF OID-Type (INT-U16) h. IF Not recommended attribute Metric-Id-Partition is present ☐ attribute-id = MDC ATTR METRIC ID PART ☐ attribute-type = NomPartition (INT-U16) ☐ attribute-value.length = 2 bytes i. Mandatory attribute Unit-Code ☐ attribute-id = MDC\_ATTR\_UNIT\_CODE ☐ attribute-type = OID-Type(INT-U16) ☐ attribute-value.length = 2 bytes ■ attribute-value= MDC\_DIM\_MILLI\_G\_PER\_DL OR MDC\_DIM\_MILLI\_MOLE\_PER\_L j. IF Not recommended attribute Source-Handle-Reference is present ☐ attribute-id = MDC\_ATTR\_SOURCE\_HANDLE\_REF ☐ attribute-type = HANDLE (INT-U16) ☐ attribute-value.length = 2 bytes k. IF Not recommended attribute Relative-Time-Stamp is present ☐ attribute-id = MDC\_ATTR\_TIME\_STAMP\_REL ☐ attribute-type = RelativeTime(INT-U32) ☐ attribute-value.length = 4 bytes I. IF Not recommended attribute Measure-Active-Period ■ attribute-id = MDC\_ATTR\_TIME\_PD\_MSMT\_ACTIVE ■ attribute-type = FLOAT type ☐ attribute-value.length = 4 bytes m. IF Not recommended Compound-Simple-Nu-Observed-Value is present □ attribute-id = MDC\_ATTR\_NU\_CMPD\_VAL\_OBS\_SIMP ☐ attribute-type = SimpleNuObsValueCmp

	☐ attribute-value.length = <variable></variable>
	n. IF Not recommended attribute Compound-Basic-Nu-Observed-Value is present
	☐ attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_BASIC
	☐ attribute-type = BasicNuObsValueCmp
	☐ attribute-value.length = <variable></variable>
	o. IF Not recommended attribute Compound-Nu-Observed-Value is present
	☐ attribute-id = MDC_ATTR_NU_VAL_OBS
	☐ attribute-type = NuObsValue
	☐ attribute-value.length = <variable></variable>
	p. Not recommended attribute Compound-Nu-Observed-Value
	□ attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_SIMP
	☐ attribute-type = NuObsValueCmp
	☐ attribute-value.length = <variable></variable>
	q. IF Recommended attribute Accuracy is present
	☐ attribute-id = MDC_ATTR_NU_ACCUR_MSMT
	☐ attribute-type = FLOAT-Type (INT-U32)
	☐ attribute-value.length = 4 bytes
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP ld		TP/PLT/AG/CLASS/GL/BV-011			
TP label		Device and Sensor annunciation status Enumeration Object - Extended configuration			
Coverage	Spec	[IEEE 11073-10417]			
	Testable	EnumObj 3; R	DevSenAn 1; M	DevSenAn 5; M	
	items	DevSenAn 6; M	DevSenAn 7; O	DevSenAn 8; R	
		DevSenAn 11; R	DevSenAn 12; R	DevSenAn 13; R	
		DevSenAn 15; R	DevSenAn 21; O	DevSenAn 22; R	
		DevSenAn 23; R	DevSenAn 24; R	DevSenAn 25; R	
		DevSenAn 26; O	DevSenAn 28; O	DevSenAn 9; O	
		DevSenAn 10; R	EnumObj 2; R	EnumObj 4; R	
		EnumObj 5; R	EnumObj 19; O	DevSenAn 30; R	
Applicability		C_AG_OXP_000 AND C_AG_OXP_178 AND C_AG_OXP_181 AND C_AG_GL_007			
Initial condition		The simulated manager and the agent under test are in the unassociated state.			
Test procedure		The simulated manager receives an association request from the agent under test.			
		2. The simulated manager responds with a result = accepted-unknown-config			
		The agent responds with a "Remote Operation Invoke   Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager.			
			Config-Id is in the extended ran- with an "unsupported-config" and		

Once the agent under test sends an extended configuration, check that all Device and Sensor annunciation status Objects have:
a. Mandatory attribute Type
☐ attribute-id = MDC_ATTR_ID_TYPE
☐ attribute-type = TYPE
□ attribute-value = MDC_PART_PHD_DM (0x00 0x80), MDC_GLU_METER_DEV_STATUS (0x71 0xD8)
b. IF Not recommended attribute Supplemental-Types is present
☐ attribute-id = MDC_ATTR_SUPPLEMENTAL_TYPES
☐ attribute-type = SupplementalTypeList
☐ attribute-value.length = <variable>(Sequence of TYPE (TYPE.length= 4 bytes))</variable>
c. Mandatory attribute Metric-Spec_Small
☐ attribute-id = MDC_ATTR_METRIC_SPEC_SMALL
☐ attribute-type = MetricSpecSmall (BITS-16)
☐ attribute-value.length =2 bytes
☐ attribute-value ≠ 0x00 0x00
<ul> <li>Bit 0 must be set (mss-avail-intermittent(0))</li> </ul>
Bit 1 must be set (mss-avail-stored-data(1))
Bit 2 must be set (mss-upd-aperiodic(2))
<ul> <li>Bit 3 must be set (mss-msmt-aperiodic(3))</li> </ul>
Bit 9 must be set (mss-acc-agent-initiated(9))
d. IF Not recommended attribute Metric-Structure-Small is present
☐ attribute-id = MDC_ATTR_METRIC_STRUCTURE_SMALL
☐ attribute-type = MetricStructureSmall
□ attribute-value.length = <variable>(Sequence of (ms-struct.length =1byte(INT-U8) + ms-comp-no =1byte(INT-U8)))</variable>
e. Optional attribute Measurement-Status
☐ attribute-id = MDC_ATTR_MSMT_STAT
☐ attribute-type = MeasurementStatus(BITS-16)
☐ attribute-value.length =2 bytes
f. IF Not recommended attribute Metric-Id is present
☐ attribute-id = MDC_ATTR_ID_PHYSIO
☐ attribute-type = OID-Type (INT-U16)
☐ attribute-value.length = 2 bytes
attribute-value = Only one attribute of Metric-Id and Metric-Id-List shall be present.
g. IF Not recommended attribute Metric-Id is present-List
□ attribute-id = MDC_ATTR_ID_PHYSIO_LIS
☐ attribute-type = MetricIdList
☐ attribute-value.length= <variable>(SEQUENCE OF OID-Type (INT-U16))</variable>
h. IF Not recommended attribute Metric-Id-Partition is present
☐ attribute-id = MDC_ATTR_METRIC_ID_PART
☐ attribute-type = NomPartition (INT-U16)
☐ attribute-value.length = 2 bytes
i. IF Not recommended attribute Unit-Code is present

5.

	D. attribute id. MDC ATTR LINIT CORE
	attribute-id = MDC_ATTR_UNIT_CODE
	□ attribute-type = OID-Type (INT-U16)
	☐ attribute-value.length = 2 bytes
	j. IF Not recommended attribute Source-Handle-Reference is present
	☐ attribute-id = MDC_ATTR_SOURCE_HANDLE_REF
	☐ attribute-type = HANDLE (INT-U16)
	☐ attribute-value.length = 2 bytes
	k. Optional attribute Enum-Observed-Value-Simple-OID
	□ attribute-id= MDC_ATTR_ENUM_OBS_VAL_SIMP_OID
	☐ attribute-type = OID-Type (INT-U16)
	☐ attribute-value.length = 2 bytes
	I. IF Not recommended attribute Enum-Observed-Value-Simple-Bit-Str
	□ attribute-id= MDC_ATTR_ENUM_OBS_VAL_SIMP_BIT_STR
	☐ attribute-type = BITS-32
	☐ attribute-value.length = 4 bytes
	m. IF recommended attribute Enum-Observed-Value-Basic-Bit-Str is present
	□ attribute-id= MDC_ATTR_ENUM_OBS_VAL_BASIC_BIT_STR
	☐ attribute-type = BITS-16
	☐ attribute-value.length = 2 bytes
	n. IF Not recommended attribute Enum-Observed-Value-Simple-Str is rpesent
	□ attribute-id= MDC_ATTR_ENUM_OBS_VAL_SIM_STR
	☐ attribute-type = EnumPrintableString
	☐ attribute-value.length = <variable></variable>
	o. IF Not recommended attribute Enum-Observed-Value is present
	□ attribute-id= MDC_ATTR_VAL_ENUM_OBS
	☐ attribute-type = EnumObsValue
	☐ attribute-value.length = <variable></variable>
	p. Optional attribute Enum-Observed-Value-Partition
	☐ attribute-id= MDC_ATTR_ENUM_OBS_VAL_PART
	☐ attribute-type = NomPartition(INT-U16)
	☐ attribute-value.length = 2 bytes
Pass/Fail criteria	All checked values are as specified in the test procedure.
. acorr an oritoria	7 in oncomed values are as specified in the test procedure.
Notes	

TP Id		TP/PLT/AG/CLASS/GL/BV-01	2		
TP label		Context Meal Enumeration Object - Extended configuration			
Coverage	Spec	[IEEE 11073-10417]			
	Testable	EnumObj 3; R	EnumObj 4; M	EnumObj 5 R	
	items	EnumObj 6; R	EnumObj 7; R	EnumObj 8; R	
		EnumObj 9; R	EnumObj 2; M	EnumObj 12; R	
		EnumObj 16; O	EnumObj 17; O	EnumObj 20; O	
		EnumObj 21; O	EnumObj 22; O	EnumObj 23; O	
		EnumObj 24; O	ContextMeal 1; M	ContextMeal 2; M	
		ContextMeal 3; M			
Applicability		C_AG_OXP_000 AND C_AG_	OXP_178 AND C_AG_OXP_18	31 AND C_AG_GL_008	
Initial conditi	on	The simulated manager and the	e agent under test are in the un	associated state.	
Test procedu	ıre	The simulated manager re	eceives an association request f	rom the agent under test.	
		2. The simulated manager re	esponds with a result = accepted	d-unknown-config	
			a "Remote Operation Invoke   C OTI_CONFIG event to send its		
		Check that the field Dev-Config-Id is in the extended range, if it is not, the simulated manager must respond with an "unsupported-config" and wait for a new configuration.			
		<ul> <li>5. Once the agent under test sends an extended configuration, check that all Context Meal Enumeration Objects have:</li> </ul>			
		a. Mandatory attribute Type			
		□ attribute-id = MD			
		□ attribute-type = T			
			MDC_PART_PHD_DM (0x00 0 U_MEAL (0X72 0X48)	x80),	
			attribute Supplemental-Types is	present	
		☐ attribute-id = MD	C_ATTR_SUPPLEMENTAL_T	/PES	
		☐ attribute-type = S	SupplementalTypeList		
		□ attribute-value.le	ngth = <variable> (Sequence of</variable>	TYPE (TYPE.length= 4 bytes))	
		c. Mandatory attribute M	etric-Spec-Small		
			C_ATTR_METRIC_SPEC_SMA	ALL	
			MetricSpecSmall (BITS-16)		
		□ attribute-value.le	•		
		□ attribute-value ≠			
			pe set (mss-avail-intermittent(0))		
			pe set (mss-avail-stored-data(1))	)	
			pe set (mss-upd-aperiodic(2))		
			pe set (mss-msmt-aperiodic(3))	2))	
			be set (mss-acc-agent-initiated(	יויס	
		Bit 12 may	be set (mss-cat-manual(12))		

d.	IF Not recommended attribute Metric-Structure-Small is present
	□ attribute-id = MDC_ATTR_METRIC_STRUCTURE_SMALL
	☐ attribute-type = MetricStructureSmall
	□ attribute-value.length = <variable>(Sequence of (ms-struct.length =1byte(INT-U8) + ms-comp-no =1byte(INT-U8)))</variable>
e.	IF Not recommended attribute Measurement-Status is present
	☐ attribute-id = MDC_ATTR_MSMT_STAT
	☐ attribute-type = MeasurementStatus (BITS-16)
	☐ attribute-value.length = 2 bytes
f.	IF Not recommended attribute Metric-Id is present
	□ attribute-id = MDC_ATTR_ID_PHYSIO
	☐ attribute-type = OID-Type (INT-U16)
	☐ attribute-value.length = 2 bytes
	attribute-value = Only one attribute of Metric-Id and Metric-Id-List shall be present.
g.	IF Not recommended attribute Metric-Id is present-List is present
	☐ attribute-id = MDC_ATTR_ID_PHYSIO_LIS
	☐ attribute-type = MetricIdList
	☐ attribute-value.length= <variable>(SEQUENCE OF OID-Type (INT-U16))</variable>
h.	IF Not recommended attribute Metric-Id-Partition is present
	☐ attribute-id = MDC_ATTR_METRIC_ID_PART
	☐ attribute-type = NomPartition (INT-U16)
	☐ attribute-value.length = 2 bytes
i.	IF Not recommended attribute Unit-Code is present
	☐ attribute-id = MDC_ATTR_UNIT_CODE
	☐ attribute-type = OID-Type(INT-U16)
	☐ attribute-value.length = 2 bytes
j.	IF Not recommended attribute Source-Handle-Reference is present
	□ attribute-id = MDC_ATTR_SOURCE_HANDLE_REF
	☐ attribute-type = HANDLE (INT-U16)
	☐ attribute-value.length = 2 bytes
k.	IF the agent supports fixed or variable format MDS event report and it does not support PM-Store THEN Mandatory attribute Absolute-Time-Stamp
	☐ attribute-id = MDC_ATTR_TIME_STAMP_ABS
	☐ attribute-type = AbsoluteTime
	☐ attribute-value.length = 8 bytes
I.	IF Optional attribute Relative-Time-Stamp is present
	□ attribute-id = MDC_ATTR_TIME_STAMP_REL
	□ attribute-type = RelativeTime (INT-U32)
	☐ attribute-value.length = 4 bytes
m.	IF Optional attribute HiRes-Time-Stamp is present
	☐ attribute-id = MDC_ATTR_TIME_STAMP_REL_HI_RES
	☐ attribute-type = HighResRelativeTime
	☐ attribute-value.length = OCTET STRING (SIZE(8))
n.	IF the agent supports fixed or variable format MDS event report and it does not

	support PM-Store then mandatory attribute Enum-Observed-Value-Simple-OID
	□ attribute-id= MDC_ATTR_ENUM_OBS_VAL_SIMP_OID
	☐ attribute-type = OID-Type(INT-U16)
	☐ attribute-value.length = 2 bytes
	☐ attribute.value= One of the following nomenclature value will be used:
	<ul> <li>MDC_CTXT_GLU_MEAL_PREPRANDIAL (0x72 0x4C) OR</li> </ul>
	<ul> <li>MDC_CTXT_GLU_MEAL_POSTPRANDIAL (0x72 0x50) OR</li> </ul>
	<ul> <li>MDC_CTXT_GLU_MEAL_FASTING (0x72 0x54) OR MDC_CTXT_GLU_MEAL_BEDTIME (0x72 0x74) OR</li> </ul>
	<ul> <li>MDC_CTXT_GLU_MEAL_CASUAL (0x72 0x58)</li> </ul>
	o. IF Optional attribute Enum-Observed-Value-Simple-Bit-Str is present
	□ attribute-id= MDC_ATTR_ENUM_OBS_VAL_SIM_BIT_STR
	☐ attribute-type = BITS-32
	☐ attribute-value.length = 4 bytes
	p. IF Optional attribute Enum-Observed-Value-Basic-Bit-Str is present
	□ attribute-id= MDC_ATTR_ENUM_OBS_VAL_BASIC_BIT_STR
	☐ attribute-type = BITS-16
	☐ attribute-value.length = 2 bytes
	q. IF Optional attribute Enum-Observed-Value-Simple-Str is present
	☐ attribute-id= MDC_ATTR_ENUM_OBS_VAL_SIM_STR
	☐ attribute-type = EnumPrintableString
	☐ attribute-value.length= <variable></variable>
	r. IF Optional attribute Enum-Observed-Value is present
	☐ attribute-id= MDC_ATTR_VAL_ENUM_OBS
	☐ attribute-type = EnumObsValue
	☐ attribute-value.length = <variable></variable>
	s. IF Optional attribute Enum-Observed-Value-Partition is present
	□ attribute-id= MDC_ATTR_VAL_ENUM_OBS_VAL_PART
	☐ attribute-type = NomPartition (INT-U16)
	☐ attribute-value.length = 2 bytes
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP ld		TP/PLT/AG/CLASS/GL/BV-013				
TP label		Context Sample Location Enui	meration Object - Extended con	figuration		
Coverage	Spec	[IEEE 11073-10417]				
	Testable	EnumObj 3; R	EnumObj 4; M	EnumObj 5 R		
	items	EnumObj 6; R	EnumObj 7; R	EnumObj 8; R		
		EnumObj 9; R	EnumObj 2; M	EnumObj 12; R		
		EnumObj 16; O	EnumObj 17; O	EnumObj 20; O		
		EnumObj 21; O	EnumObj 22; O	EnumObj 23; O		
		EnumObj 24; O	ContxtSamLoc 1; M	ContxtSamLoc 2; M		
			CONIXIOANILOC 1, W	CONCIOANTECC 2, IVI		
		ContxtSamLoc 3; M				
Applicability	/	C_AG_OXP_000 AND C_AG_	OXP_178 AND C_AG_OXP_18	81 AND C_AG_GL_009		
Initial condi	tion	The simulated manager and the	ne agent under test are in the ur	nassociated state.		
Test proced	ure	The simulated manager receives an association request from the agent under test.				
		2. The simulated manager re	esponds with a result = accepte	d-unknown-config		
			a "Remote Operation Invoke   C OTI_CONFIG event to send its			
		4. Check that the field Dev-Config-Id is in the extended range, if it is not, the simulated manager must respond with an "unsupported-config" and wait for a new configuration.				
		5. Once the agent under test	t sends an extended configurati	on, check that all Context Meal		
		Enumeration Objects have a. Mandatory attribute Ty				
		a. Mandatory attribute ry  □ attribute-id = MD				
		□ attribute-type = T				
		□ attribute-value =	MDC_PART_PHD_DM (0x00 0 U_SAMPLELOCATION (0x72 0			
			o_Sampletocarron (0x/2 c attribute Supplemental-Types is	,		
			C_ATTR_SUPPLEMENTAL_T			
		☐ attribute-type = 5	SupplementalTypeList			
		□ attribute-value.le	ngth = <variable> (Sequence of</variable>	TYPE (TYPE.length= 4 bytes))		
		c. Mandatory attribute M	etric-Spec-Small			
		☐ attribute-id = MD	C_ATTR_METRIC_SPEC_SM/	ALL		
		☐ attribute-type = N	MetricSpecSmall (BITS-16)			
		□ attribute-value.le				
		□ attribute-value ≠				
			pe set (mss-avail-intermittent(0)			
			pe set (mss-avail-stored-data(1)	)		
			pe set (mss-upd-aperiodic(2))			
			pe set (mss-msmt-aperiodic(3))	2))		
			pe set (mss-acc-agent-initiated(	שון		
		Bit 12 may	be set (mss-cat-manual(12))			

d.	IF N	ot recommended attribute Metric-Structure-Small is present
		attribute-id = MDC_ATTR_METRIC_STRUCTURE_SMALL
		attribute-type = MetricStructureSmall
		attribute-value.length = <variable>(Sequence of (ms-struct.length =1byte(INT-U8) + ms-comp-no =1byte(INT-U8)))</variable>
e.	IF N	ot recommended attribute Measurement-Status is present
		attribute-id = MDC_ATTR_MSMT_STAT
		attribute-type = MeasurementStatus (BITS-16)
		attribute-value.length = 2 bytes
f.	IF N	ot recommended attribute Metric-Id is present
		attribute-id = MDC_ATTR_ID_PHYSIO
		attribute-type = OID-Type (INT-U16)
		attribute-value.length = 2 bytes
		attribute-value = Only one attribute of Metric-Id and Metric-Id-List shall be present.
g.	IF N	ot recommended attribute Metric-Id-List is present
		attribute-id = MDC_ATTR_ID_PHYSIO_LIS
		attribute-type = MetricIdList
		attribute-value.length= <variable>(SEQUENCE OF OID-Type (INT-U16))</variable>
h.	IF N	ot recommended attribute Metric-Id-Partition is present
		attribute-id = MDC_ATTR_METRIC_ID_PART
		attribute-type = NomPartition (INT-U16)
		attribute-value.length = 2 bytes
i.	IF N	ot recommended attribute Unit-Code is present
		attribute-id = MDC_ATTR_UNIT_CODE
		attribute-type = OID-Type(INT-U16)
		attribute-value.length = 2 bytes
j.	IF N	ot recommended attribute Source-Handle-Reference is present
		attribute-id = MDC_ATTR_SOURCE_HANDLE_REF
		attribute-type = HANDLE (INT-U16)
		attribute-value.length = 2 bytes
k.		e agent supports fixed or variable format MDS event report and it does not oort PM-Store THEN Mandatory attribute Absolute-Time-Stamp
		attribute-id = MDC_ATTR_TIME_STAMP_ABS
		attribute-type = AbsoluteTime
		attribute-value.length = 8 bytes
I.	IF O	ptional attribute Relative-Time-Stamp is present
		attribute-id = MDC_ATTR_TIME_STAMP_REL
		attribute-type = RelativeTime (INT-U32)
		attribute-value.length = 4 bytes
m.	IF O	ptional attribute HiRes-Time-Stamp is present
		attribute-id = MDC_ATTR_TIME_STAMP_REL_HI_RES
		attribute-type = HighResRelativeTime
		attribute-value.length = OCTET STRING (SIZE(8))
n.	IF th	e agent supports fixed or variable format MDS event report and it does not

	support PM-Store then mandatory attribute Enum-Observed-Value-Simple-OID
	□ attribute-id= MDC_ATTR_ENUM_OBS_VAL_SIM_OID
	□ attribute-type = OID-Type(INT-U16)
	☐ attribute-value.length = 2 bytes
	☐ attribute.value= One of the following nomenclature value will be used:
	MDC_CTXT_GLU_SAMPLE_LOCATION_FINGER (0x72 0x38) OR
	MDC_CTXT_GLU_SAMPLE_LOCATION_AST (0x72 0x3C) OR
	MDC_CTXT_GLU_SAMPLE_LOCATION_EARLOBE (0x72 0x40) OR
	MDC_CTXT_GLU_SAMPLE_LOCATION_CTRLSOLUTION (0x72 0x44)
	o. IF Optional attribute Enum-Observed-Value-Simple-Bit-Str is present
	☐ attribute-id= MDC_ATTR_ENUM_OBS_VAL_SIM_BIT_STR
	☐ attribute-type = BITS-32
	☐ attribute-value.length = 4 bytes
	p. IF Optional attribute Enum-Observed-Value-Basic-Bit-Str is present
	☐ attribute-id= MDC_ATTR_ENUM_OBS_VAL_BASIC_BIT_STR
	☐ attribute-type = BITS-16
	☐ attribute-value.length = 2 bytes
	q. IF Optional attribute Enum-Observed-Value-Simple-Str is present
	☐ attribute-id= MDC_ATTR_ENUM_OBS_VAL_SIM_STR
	☐ attribute-type = EnumPrintableString
	☐ attribute-value.length= <variable></variable>
	r. IF Optional attribute Enum-Observed-Value is present
	☐ attribute-id= MDC_ATTR_VAL_ENUM_OBS
	☐ attribute-type = EnumObsValue
	☐ attribute-value.length = <variable></variable>
	s. IF Optional attribute Enum-Observed-Value-Partition is present
	☐ attribute-id= MDC_ATTR_VAL_ENUM_OBS_VAL_PART
	□ attribute-type = NomPartition (INT-U16)
	☐ attribute-value.length = 2 bytes
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP ld		TP/PLT/AG/CLASS/GL/BV-01	4		
TP label		Context Tester Enumeration Object - Extended configuration			
Coverage	Spec	[IEEE 11073-10417]			
	Testable	EnumObj 3; R	EnumObj 4; M	EnumObj 5 R	
	items	EnumObj 6; R	EnumObj 7; R	EnumObj 8; R	
		EnumObj 9; R	EnumObj 2; M	EnumObj 12; R	
		EnumObj 16; O	EnumObj 17; O	EnumObj 20; O	
		EnumObj 21; O	EnumObj 22; O	EnumObj 23; O	
		EnumObj 24; O	ContextTester 1; M	ContextTester 2; M	
		ContextTester 3; M			
Applicability		C_AG_OXP_000 AND C_AG_	OXP_178 AND C_AG_OXP_18	31 AND C_AG_GL_010	
Initial conditi	on	The simulated manager and the	e agent under test are in the un	associated state.	
Test procedu	ire	The simulated manager re	eceives an association request f	rom the agent under test.	
		2. The simulated manager re	esponds with a result = accepted	d-unknown-config	
			a "Remote Operation Invoke   C OTI_CONFIG event to send its		
		Check that the field Dev-Config-Id is in the extended range, if it is not, the simulated manager must respond with an "unsupported-config" and wait for a new configuration.			
		<ol> <li>Once the agent under test sends an extended configuration, check that all Context Meal Enumeration Objects have:</li> </ol>			
		a. Mandatory attribute Type			
		□ attribute-id = MD	•		
		☐ attribute-type = 7	YPE		
		☐ attribute-value = MDC_CTXT_GL	MDC_PART_PHD_DM (0x00 0 U_TESTER (0x72 0x5C)	x80),	
			attribute Supplemental-Types is	present	
		☐ attribute-id = MD	C_ATTR_SUPPLEMENTAL_T	/PES	
		☐ attribute-type = S	SupplementalTypeList		
		□ attribute-value.le	ngth = <variable> (Sequence of</variable>	TYPE (TYPE.length= 4 bytes))	
		c. Mandatory attribute M	etric-Spec-Small		
			C_ATTR_METRIC_SPEC_SMA	ALL	
			MetricSpecSmall (BITS-16)		
		□ attribute-value.le	•		
		□ attribute-value ≠			
			pe set (mss-avail-intermittent(0))		
			pe set (mss-avail-stored-data(1))	)	
			pe set (mss-upd-aperiodic(2))		
			pe set (mss-msmt-aperiodic(3))	2))	
			pe set (mss-acc-agent-initiated(	71)	
		Bit 12 may	be set (mss-cat-manual(12))		

d.	IF N	ot recommended attribute Metric-Structure-Small is present
		attribute-id = MDC_ATTR_METRIC_STRUCTURE_SMALL
		attribute-type = MetricStructureSmall
		attribute-value.length = < variable > (Sequence of (ms-struct.length = 1 byte(INT-U8) + ms-comp-no = 1 byte(INT-U8)))
e.	IF N	ot recommended attribute Measurement-Status is present
		attribute-id = MDC_ATTR_MSMT_STAT
		attribute-type = MeasurementStatus (BITS-16)
		attribute-value.length = 2 bytes
f.	IF N	ot recommended attribute Metric-Id is present
		attribute-id = MDC_ATTR_ID_PHYSIO
		attribute-type = OID-Type (INT-U16)
		attribute-value.length = 2 bytes
		attribute-value = Only one attribute of Metric-Id and Metric-Id-List shall be present.
g.	IF N	ot recommended attribute Metric-Id-List is present
		attribute-id = MDC_ATTR_ID_PHYSIO_LIS
		attribute-type = MetricIdList
		attribute-value.length= <variable>(SEQUENCE OF OID-Type (INT-U16))</variable>
h.	IF N	ot recommended attribute Metric-Id-Partition is present
		attribute-id = MDC_ATTR_METRIC_ID_PART
		attribute-type = NomPartition (INT-U16)
		attribute-value.length = 2 bytes
i.	IF N	ot recommended attribute Unit-Code is present
		attribute-id = MDC_ATTR_UNIT_CODE
		attribute-type = OID-Type(INT-U16)
		attribute-value.length = 2 bytes
j.	IF N	ot recommended attribute Source-Handle-Reference is present
		attribute-id = MDC_ATTR_SOURCE_HANDLE_REF
		attribute-type = HANDLE (INT-U16)
		attribute-value.length = 2 bytes
k.		gent supports fixed or variable format MDS event report and it does not support Store THEN Mandatory attribute Absolute-Time-Stamp
		attribute-id = MDC_ATTR_TIME_STAMP_ABS
		attribute-type = AbsoluteTime
		attribute-value.length = 8 bytes
I.	IF O	ptional attribute Relative-Time-Stamp is present
		attribute-id = MDC_ATTR_TIME_STAMP_REL
		attribute-type = RelativeTime (INT-U32)
		attribute-value.length = 4 bytes
m.	IF O	ptional attribute HiRes-Time-Stamp is present
		attribute-id = MDC_ATTR_TIME_STAMP_REL_HI_RES
		attribute-type = HighResRelativeTime
		attribute-value.length = OCTET STRING (SIZE(8))

	<ul> <li>IF the agent supports fixed or variable format MDS event report and it does not support PM-Store then mandatory attribute Enum-Observed-Value-Simple-OID</li> </ul>
	□ attribute-id= MDC_ATTR_ENUM_OBS_VAL_SIM_OID
	□ attribute-type = OID-Type(INT-U16)
	☐ attribute-value.length = 2 bytes
	☐ attribute.value= One of the following nomenclature value will be used:
	<ul> <li>MDC_CTXT_GLU_TESTER_SELF (0x72 0x60) OR</li> </ul>
	<ul> <li>MDC_CTXT_GLU_TESTER_HCP (0x72 0x64) OR</li> </ul>
	<ul> <li>MDC_CTXT_GLU_TESTER_LAB (0x72 0x68)</li> </ul>
	o. IF Optional attribute Enum-Observed-Value-Simple-Bit-Str is present
	□ attribute-id= MDC_ATTR_ENUM_OBS_VAL_SIM_BIT_STR
	☐ attribute-type = BITS-32
	☐ attribute-value.length = 4 bytes
	p. IF Optional attribute Enum-Observed-Value-Basic-Bit-Str is present
	□ attribute-id= MDC_ATTR_ENUM_OBS_VAL_BASIC_BIT_STR
	☐ attribute-type = BITS-16
	☐ attribute-value.length = 2 bytes
	q. IF Optional attribute Enum-Observed-Value-Simple-Str is present
	□ attribute-id= MDC_ATTR_ENUM_OBS_VAL_SIM_STR
	□ attribute-type = EnumPrintableString
	☐ attribute-value.length= <variable></variable>
	r. IF Optional attribute Enum-Observed-Value is present
	□ attribute-id= MDC_ATTR_VAL_ENUM_OBS
	□ attribute-type = EnumObsValue
	☐ attribute-value.length = <variable></variable>
	s. IF Optional attribute Enum-Observed-Value-Partition is present
	□ attribute-id= MDC_ATTR_VAL_ENUM_OBS_VAL_PART
	□ attribute-type = NomPartition (INT-U16)
	☐ attribute-value.length = 2 bytes
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP ld		TP/PLT/AG/CLASS/GL/BV-015				
TP label		Context Health Enumeration	Object - Extended configuration			
Coverage	Spec	[IEEE 11073-10417]				
	Testable	EnumObj 3; R	EnumObj 4; M	EnumObj 5 R		
	items	EnumObj 6; R	EnumObj 7; R	EnumObj 8; R		
		EnumObj 9; R	EnumObj 2; M	EnumObj 12; R		
		EnumObj 16; O	EnumObj 17; O	EnumObj 20; O		
		EnumObj 21; O	EnumObj 22; O	EnumObj 23; O		
		EnumObj 24; O	ContextHealth 1; M	ContextHealth 2; M		
		ContextHealth 3; M				
Applicability	I .		OXP_178 AND C_AG_OXP_1	81 AND C AG GL 011		
Initial conditi			he agent under test are in the u			
Test procedu			receives an association request			
rest procedi	116		•	•		
		The agent responds with a "Remote Operation Invoke   Confirmed Event Report"				
		message with an MDC_NOTI_CONFIG event to send its configuration to the manager.  4. Check that the field Dev-Config-Id is in the extended range, if it is not, the simulated				
		manager must respond with an "unsupported-config" and wait for a new configuration.				
		5. Once the agent under test sends an extended configuration, check that all Context Meal Enumeration Objects have:				
		a. Mandatory attribute Type				
		☐ attribute-id = MDC_ATTR_ID_TYPE				
		☐ attribute-type = TYPE				
		□ attribute-value = MDC_PART_PHD_DM (0x00 0x80), MDC_CTXT_GLU_HEALTH (0x72 0x5C)				
		b. IF Not recommended attribute Supplemental-Types is present				
		☐ attribute-id = MDC_ATTR_SUPPLEMENTAL_TYPES				
		☐ attribute-type = SupplementalTypeList				
☐ attribute-value.length = <variable> (Sequence of TYPE (TYPE)</variable>			TYPE (TYPE.length= 4 bytes))			
		c. Mandatory attribute N	•			
		☐ attribute-id = MDC_ATTR_METRIC_SPEC_SMALL				
		☐ attribute-type = MetricSpecSmall (BITS-16)				
		☐ attribute-value.length =2 bytes				
		attribute-value ≠ 0x00 0x00				
		<ul> <li>Bit 0 must be set (mss-avail-intermittent(0))</li> <li>Bit 1 must be set (mss-avail-stored-data(1))</li> </ul>				
			be set (mss-avail-stored-data( ) be set (mss-upd-aperiodic(2))	<i>II</i>		
			be set (mss-msmt-aperiodic(3))			
			be set (mss-acc-agent-initiated)			
			be set (mss-cat-manual(12))	(-1)		

d.	IF N	ot recommended attribute Metric-Structure-Small is present
		attribute-id = MDC_ATTR_METRIC_STRUCTURE_SMALL
		attribute-type = MetricStructureSmall
		attribute-value.length = <variable>(Sequence of (ms-struct.length =1byte(INT-U8) + ms-comp-no =1byte(INT-U8)))</variable>
e.	IF N	ot recommended attribute Measurement-Status is present
		attribute-id = MDC_ATTR_MSMT_STAT
		attribute-type = MeasurementStatus (BITS-16)
		attribute-value.length = 2 bytes
f.	IF N	ot recommended attribute Metric-Id is present
		attribute-id = MDC_ATTR_ID_PHYSIO
		attribute-type = OID-Type (INT-U16)
		attribute-value.length = 2 bytes
		attribute-value = Only one attribute of Metric-Id and Metric-Id-List shall be present.
g.	IF N	ot recommended attribute Metric-Id-List is present
		attribute-id = MDC_ATTR_ID_PHYSIO_LIS
		attribute-type = MetricIdList
		attribute-value.length= <variable>(SEQUENCE OF OID-Type (INT-U16))</variable>
h.	IF N	ot recommended attribute Metric-Id-Partition is present
		attribute-id = MDC_ATTR_METRIC_ID_PART
		attribute-type = NomPartition (INT-U16)
		attribute-value.length = 2 bytes
i.	IF N	ot recommended attribute Unit-Code is present
		attribute-id = MDC_ATTR_UNIT_CODE
		attribute-type = OID-Type(INT-U16)
		attribute-value.length = 2 bytes
j.	IF N	ot recommended attribute Source-Handle-Reference is present
		attribute-id = MDC_ATTR_SOURCE_HANDLE_REF
		attribute-type = HANDLE (INT-U16)
		attribute-value.length = 2 bytes
k.		e agent supports fixed or variable format MDS event report and it does not port PM-Store THEN Mandatory attribute Absolute-Time-Stamp
		attribute-id = MDC_ATTR_TIME_STAMP_ABS
		attribute-type = AbsoluteTime
		attribute-value.length = 8 bytes
I.	IF O	ptional attribute Relative-Time-Stamp is present
		attribute-id = MDC_ATTR_TIME_STAMP_REL
		attribute-type = RelativeTime (INT-U32)
		attribute-value.length = 4 bytes
m.	IF O	ptional attribute HiRes-Time-Stamp is present
		attribute-id = MDC_ATTR_TIME_STAMP_REL_HI_RES
		attribute-type = HighResRelativeTime
		attribute-value.length = OCTET STRING (SIZE(8))

	<ul> <li>IF the agent supports fixed or variable format MDS event report and it does not support PM-Store then mandatory attribute Enum-Observed-Value-Simple-OID</li> </ul>
	□ attribute-id= MDC_ATTR_ENUM_OBS_VAL_SIM_OID
	☐ attribute-type = OID-Type(INT-U16)
	☐ attribute-value.length = 2 bytes
	☐ attribute.value= One of the following nomenclature value will be used:
	MDC_CTXT_GLU_HEALTH_MINOR (0x72 0x20) OR
	MDC_CTXT_GLU_HEALTH_MAJOR (0x72 0x24) OR
	MDC_CTXT_GLU_HEALTH_MENSES (0x72 0x28) OR
	MDC_CTXT_GLU_HEALTH_STRESS (0x72 0x2C) OR
	MDC_CTXT_GLU_HEALTH_NONE (0x72 0x30)
	o. IF Optional attribute Enum-Observed-Value-Simple-Bit-Str is present
	□ attribute-id= MDC_ATTR_ENUM_OBS_VAL_SIM_BIT_STR
	□ attribute-type = BITS-32
	☐ attribute-value.length = 4 bytes
	p. IF Optional attribute Enum-Observed-Value-Basic-Bit-Str is present
	□ attribute-id= MDC_ATTR_ENUM_OBS_VAL_BASIC_BIT_STR
	☐ attribute-type = BITS-16
	☐ attribute-value.length = 2 bytes
	q. IF Optional attribute Enum-Observed-Value-Simple-Str is present
	□ attribute-id= MDC_ATTR_ENUM_OBS_VAL_SIM_STR
	☐ attribute-type = EnumPrintableString
	☐ attribute-value.length= <variable></variable>
	r. IF Optional attribute Enum-Observed-Value is present
	☐ attribute-id= MDC_ATTR_VAL_ENUM_OBS
	☐ attribute-type = EnumObsValue
	☐ attribute-value.length = <variable></variable>
	s. IF Optional attribute Enum-Observed-Value-Partition is present
	☐ attribute-id= MDC_ATTR_VAL_ENUM_OBS_VAL_PART
	□ attribute-type = NomPartition (INT-U16)
	☐ attribute-value.length = 2 bytes
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	
	1

TP ld		TP/PLT/AG/CLASS/GL/BV-016				
TP label		PM-Store Attributes for Extended Configuration				
Coverage	Spec	[IEEE 11073-10417]				
Coverage						
	Testable items	PMStrObjAtt 1; M	PMStrObjAtt 5; M	PMStrObjAtt 6; M		
		PMStrObjAtt 8; M	PMStrObjAtt 9; R	PMStrObjAtt 12; M		
Applicability C_AG_OXP_000 AND C_AG_OXP_178 AND C_AG_OXP_041 AND C_AG_OX			1 AND C_AG_OXP_181			
Initial conditi	ion	The simulated manager and th	e agent under test are in the un	associated state.		
Applicability Initial condition Test procedure		1. The simulated manager recommendate and the simulated manager recommendates and the simulated manager recommendates are simulated manager and the simulated manager shattribute-value. Let a ttribute-value and ttribute-id-list set to 0 to in the agent issues a GET recommendates and tribute-id-list set to 0 to in the agent issues a GET recommendates and tribute-id and tribute-id and tribute-id and tribute-value. Let a ttribute-value attribute-value attribute-id attribute-value. Let attribute-value. Let attribute-value. Let attribute-value attribute-val	eceives an association request for esponds with a result = accepted a "Remote Operation Invoke   COTI_CONFIG event to send its above attribute must be: andle IANDLE angth = 2 bytes must be unique and non-zero. A ecialization.  Inall send a Get request for the Findicate all PM-Store attributes. It is esponse with the PM-Store attribute. It is esponse with the PM-S	rom the agent under test. d-unknown-config confirmed Event Report" configuration to the manager.  Actual value may be specified PM-Store object with an butes it supports:  APAC_CNT  GAGE_CNT		
			attribute Sample-Period is pres C_ATTR_TIME_PD_SAMP	ent		
		□ attribute-type = R				
		□ attribute-value.ler				
		☐ attribute-value = <not in="" relevant="" test="" this=""></not>				

Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP ld		TP/P	TP/PLT/AG/CLASS/GL/BV-017			
TP label		PM S	Segment Object	for Exten	ded Configuration	
Coverage	Spec	[IEEE	E 11073-10417]			
	Testable items	PMS	toreObj 8; M		PMStoreObj 9; O	PMStoreObj 10; M
	, i.e.i.e	PMS	toreObj 11; O		PMSegObj 6; M	PMSegObj 7; M
		PMS	egObj 8; M		PMSegObj 10; M	
Applicability	/	C_A	G_OXP_000 AN	ID C_AG_	OXP_178 AND C_AG_OXF	P_041 AND C_AG_OXP_181
Initial condi	tion	The	simulated mana	ger and th	ne agent under test are in the	e operating state.
Test proced	ure	1.	The simulated nattribute-id-list s	nanager s et to 0 to	hall send a Get request for t indicate all PM-Store attribute	he PM-Store object with an tes.
		2. The simulated manager shall send a Get-Segment-Info object action for the PM-Segment object with SegmSelection = all-segments to indicate the PM-Segments attributes of all available PM-Segments.				
		3.	3. The agent issues a response with the PM-Segment attributes it supports:			
		a. Mandatory attribute Segment-Label				
		☐ attribute-id = MDC_ATTR_PM_SEG_LABEL_STRING				
		☐ attribute-type = OCTET STRING				
		☐ attribute-value.length = consistent with value				
		☐ attribute-value = <printable ascii=""></printable>				
		b. Mandatory attribute Segment-Start-Abs-Time				
		☐ attribute-id = MDC_ATTR_TIME_START_SEG				
		☐ attribute-type = AbsoluteTime				
		☐ attribute-value.length = 8 bytes				
			attribut	e-value =		
			• C6	entury =		
			• ye	ear ≤ 99		
			• m	onth ≤ 12		
			• da	ay ≤ 31		
			• ho	our ≤ 24		
			• m	inute ≤ 60	)	
			• S6	econd ≤ 6	0	
			• S6	ec-fraction	ns ≤ 100	
			c. Mandatory a	ttribute S	egment-End-Abs-Time	
			_		C_ATTR_TIME_END_SEG	
					AbsoluteTime	
					ength = 8 bytes	
				e-value =	-	
		• century =				

Í			
	<ul> <li>year ≤ 99</li> </ul>		
	<ul> <li>month ≤ 12</li> </ul>		
	<ul> <li>day ≤ 31</li> </ul>		
	<ul> <li>hour ≤ 24</li> </ul>		
	<ul> <li>minute ≤ 60</li> </ul>		
	<ul> <li>second ≤ 60</li> </ul>		
	<ul> <li>sec-fractions ≤ 100</li> </ul>		
	d. Mandatory attribute Segment-Usage-Count		
	☐ attribute-id = MDC_ATTR_SEG_USAGE_CNT		
	☐ attribute-type = INT-U32		
	☐ attribute-value.length = 4 bytes		
	☐ attribute-value = <not in="" relevant="" test="" this=""></not>		
	e. Mandatory attribute PM-Segment-Entry-Map		
	☐ SegmentEntryHeader.value = One of the next must be set:		
	<ul> <li>seg-elem-hdr-absolute-time(0)</li> </ul>		
	<ul> <li>seg-elem-hdr-relative-time(1)</li> </ul>		
	<ul> <li>seg-elem-hdr-hires-relative-time(2)</li> </ul>		
	☐ SegmEntryElem: < Record the fields for later comparison>		
	4. Repeat step 3 and 4 for every Segment.		
Pass/Fail criteria	All checked values are as specified in the test procedure		
	Every segm-entry-header must contain one of the time formats		
	At least one PM-Segment must reference the Common Glucose in its PM-Segm-Entry- Map		
	If there are more than one PM-Segment, the rest of them must reference one of the objects defined in the spec in its PM-Segm-Entry-Map		
Notes			

TP ld		TP/PLT/AG/CLASS/GL/BV-017_A				
TP label		PM-Segment Object for Extended Configuration.MDS Event Reports				
Coverage	Spec	[IEEE 11073-10417]	-			
	Testable Items	PMStoreObj 5; M	PMStoreObj 5; M PMStoreObj 6; M PMStoreObj 7; M			
Applicability	1	C_AG_OXP_000 AND C_AG_	OXP_178 AND C_AG_OXP_04	11 AND C_AG_OXP_181		
Initial condition The simulated manager and the agent under test are in the operating state.		erating state.				
Test procedure		attribute-id-list set to 0 to 2. The simulated manager s	hall send a Get request for the Findicate all PM-Store attributes.	oject action for the PM-		
		Segment object with SegmSelection = all-segments to indicate the PM-Segments attributes of all available PM-Segments.				
		3. The simulated manager asks for a measurement.				
		4. Check event reports that are sent by the agent.				
Pass/Fail cri	iteria	In step 4, the agent shall not send the data with MDS event reports.				

Notes	

TP ld		TP/PLT/AG/CLASS/GL/BV-018				
TP label		Communication Model:	Association Procedure			
Coverage	Spec	[IEEE 11073-10417]				
	Testable	AgProcAs 1; M	AgProcAs 2; M	AgProcAs 4; M		
	items	AgProcAs 5; M	AgProcAs 6; M	AgProcAs 7; M		
		AgProcAs 8; M	AgProcAs 9; M	AgProcAs 10; M		
		AgProcAs 11; M	AgProcAs 12; M	MDSMethods 3;M		
		AgProcAs 13; M	G .	,		
Applicabilit	<b>y</b>	C_AG_OXP_000 AND	C AG OXP 178			
Initial cond	<u>-                                      </u>		r and the agent under test are in	n the unassociated state		
			<u> </u>			
Test proced	dure	<ol> <li>The agent sends a sent by the Agent</li> </ol>		mulated manager, the expected fields		
		a. APDU Type				
		☐ field- type = AarqApdu				
		☐ field-length =2 bytes				
		☐ field-value =0xE2 0x00.				
		b. assoc-version				
		☐ field- type = AssociationVersion				
		☐ field-length =BITS-32				
		☐ field- value=0x80 0x00 0x00 0x00				
		c. data-proto-id				
		☐ field- type = DataProtoId(INT-U16)				
		☐ field-length =2 bytes				
		☐ field- value=0x50 0x79 (20601)				
		d. protocol-version				
		☐ field- type	e = Protocol Version			
		☐ field-leng	th = 4 bytes			
		☐ field- valu	ie=0x80 0x00 0x00 0x00			
		e. encoding rules				
		☐ field- type = EncodingRules				
		☐ field-length = 2 bytes				
		☐ field- value=				
		Bit 0 must be set (support for MDER)				
		Bits 1 (XER) and 2 (PER) may be set				
		• All c	other bits must be 0.			
		f. nomenclature	version			
		☐ field- type	e = NomenclatureVersion			
		☐ field-leng	th = 4 bytes			

	☐ field- value=0x80 0x00 0x00 0x00
	☐ This value indicates version1 is supported (nom-version1(0) is set).
	g. functional – units
	☐ field- type = FunctionalUnits
	☐ field-length = 4 bytes
	Bit 0 must be 0.
	Bits 1 and 2 may be set
	The rest of the bits must not be set
	h. System type
	☐ field- type = SystemType
	☐ field-length = 4 bytes
	☐ field- value = 0x00 0x80 0x00 0x00 (sys-type-agent)
	i. System-Id
	☐ field- type = OCTET STRING
	☐ field-length = 8 bytes
	☐ field- value = 0xXX 0xXX 0xXX 0xXX 0xXX 0xXX 0xXX 0x
	This value will be System Id attribute of MDS Object.
	j. dev-config-id
	☐ field- type = Configld(INT-U16)
	☐ field-length = 2 bytes
	☐ field- value =
	<ul> <li>0x06 0xA4 OR 0x06 A5 for standard configuration.</li> </ul>
	<ul> <li><between 0x00="" 0x40="" 0x7f="" 0xff="" and=""> for extended configuration.</between></li> </ul>
	k. data-req-mode-flags (DataReqModeCapab)
	☐ field- type = DataReqModeFlags
	☐ field-length = 2 bytes
	<ul> <li>If the agent supports agent-initiated measurement transfer → Bit 15 is set (data-req-supp-init-agent(15))</li> </ul>
	<ul> <li>If the agent supports requesting objects based on the object handle →Bit 6 will be set (data-req-supp-scope-handle(6)).</li> </ul>
	<ul> <li>If the agent supports single response →Bit 8 will be set (data-req-supp-mode-single-rsp(8)).</li> </ul>
	<ul> <li>If the agent supports time unlimited data request →Bit 10 will be set (data-req-supp-mode-time-no-limit(10)).</li> </ul>
	data-req-init-agent-count (DataReqModeCapab)
	☐ field- type = INT-U8
	☐ field-length = 2 bytes
	☐ field.value = 0x01
	m. data-req-init-manager-count (DataReqModeCapab)
	☐ field- type = INT-U8
	☐ field-length = 2 bytes
	☐ field.value = 0x00
Pass/Fail criteria	All checked values are as specified in the test procedure.

Notes
-------

TP ld		TP/PLT/AG/CLASS/GL/BV-019			
TP label		PM Segment Object for Extended Configuration			
Coverage	Spec	[IEEE 11073-10417]			
	Testable items	PMStrObjMeth 1; M			
Applicability	у	C_AG_OXP_000 AND C_AG_OXP_178 AND C_AG_OXP_041 AND C_AG_OXP_071			
Initial condi	tion	The simulated manager and the agent under test are in the operating state and the agent has at least one PM-Segment with data stored.			
Test proced	lure	Take measurements with the agent of a value that is stored on a PM-Segment.			
•		<ol> <li>The simulated manager shall send a Get request for the PM-Store object with an attribute-id-list set to 0 to indicate all PM-Store attributes.</li> </ol>			
		3. The agent issues a GET response with the PM-Store attributes, record the values of the PMStoreCapab attribute.			
		The simulated manager shall send a Get-Segment-Info object action with segmSelection set to all-segments to check that there are no Segments in use.			
		5. The simulated manager sends a Clear-Segment to all segments:			
		a. Data APDU			
		☐ Type = Invoke   Confirmed Action,			
		☐ HANDLE = obj-handle			
		☐ Action = MDC_ACT_SEG_CLEAR			
		☐ SegmSelection = all-segments			
		6. The agent under test operation response:			
		a. Data APDU			
		☐ Type = Response   Confirmed Action,			
		☐ HANDLE = obj-handle			
		☐ Action = MDC_ACT_SEG_CLEAR			
		7. Delay			
		8. The simulated manager sends a request for the PM-Segment Data with SegmSelection = all-segments to obtain all the segments:			
		a. Data APDU			
		☐ Type = Invoke   Confirmed Action,			
		☐ HANDLE = obj-handle			
		☐ Action = MDC_ACT_SEG_TRIG_XFER			
		SegmSelection = <instance action="" before="" clear-segment="" contained="" data="" number="" of="" pm-segment="" selected="" that="" the=""></instance>			
		9. The agent issues an action response with the Data			
		a. Data APDU			
		☐ Type = Response   Confirmed Action,			
		☐ HANDLE = obj-handle			
		☐ Action = MDC_ACT_SEG_TRIG_XFER			
		☐ TrigSegmXferRsp =			
		<ul> <li>IF pmsc-clear-segm-remove is NOT set THEN TrigSegmXferRsp = tsxr-</li> </ul>			

	fail-segm-empty
	<ul> <li>ELSE TrigSegmXferRsp = tsxr-fail-no-such-segment</li> </ul>
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP ld		TP/PLT/AG/CLASS/GL/BV-020		
TP label Config Changes Service. Contextual Attribute.		Config Changes Service. Contextual Attribute.		
Coverage	Spec	[ITU-T H.810]		
	Testable items	Communication 8; M		
Applicability	Ī	C_AG_OXP_000 AND C_AG_OXP_178 AND C_AG_GL_022		
Initial condit	tion	The simulated manager and the agent under test are in the operating state.		
Test procedure  1. Take some measurements with the agent under test. 2. Make a change to the contextual attribute Unit-Code for blood glucose. 3. The agent shall send an MDS event report indicating the new contextual attribute v. 4. Take some more measurements. 5. Wait for the manager to receive new event reports from the agent, which report the measurements from step 4.		<ol> <li>Make a change to the contextual attribute Unit-Code for blood glucose.</li> <li>The agent shall send an MDS event report indicating the new contextual attribute value.</li> <li>Take some more measurements.</li> <li>Wait for the manager to receive new event reports from the agent, which report the measurements from step 4.</li> <li>The agent sends an MDS event report to inform on the contextual attribute that has been</li> </ol>		
Notes				

TP Id		TP/PLT/AG/CLASS/GL/BV-021		
TP label		Operating State. Manager to Agent Maximum APDU Size		
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	CommonCharac 3; M		
	Spec	[IEEE 11073-10417]		
	Testable items	ComChar 2; M		
Applicability		C_AG_OXP_000 AND C_AG_OXP_178		
Initial condition		The simulated manager and the agent are in the operating state.		
Test procedure		<ol> <li>The simulated manager issues a "Remote Operation Invoke   Get" command with:         <ul> <li>a. Obj-handle set to 0 (to request for MDS object)</li> <li>b. attribute-id-list.count = 103</li> </ul> </li> <li>c. attribute-id-list: (MDC_ATTR_ID_MODEL, MDC_ATTR_SYS_ID, MDC_ATTR_DEV_CONFIG_ID) repeated 34 times followed by an additional MDC_ATTR_ID_MODEL</li> </ol>		

	2. Check the response of the agent.	
	<ol> <li>The simulated manager issues a "Remote Operation Invoke   Get" command with the handle set to 0 (to request for MDS object) and an empty attribute-id-list to indicate all attributes.</li> </ol>	
	4. Check the response of the agent.	
Pass/Fail criteria	<ul> <li>In step 2, the agent under test may respond with a rors-cmip-get listing all the requested attributes, or with a roer message. If PICS C_AG_OXP_100 =TRUE and the agent does not respond with a rors-cmip-get message, it responds with a roer message or rorj(resource-limitation) message, a WARNING will appear.</li> <li>If the response is a get response, the total size of the response cannot exceed the sum of the APDU sizes of the supported specializations</li> </ul>	
	(limited to an absolute limit of 64512 octets):	
	<ul> <li>Pulse oximeter -&gt; 9216 octets</li> </ul>	
	<ul> <li>Weighing scales -&gt; 896 octets</li> </ul>	
	<ul> <li>Glucose meter -&gt; 5120 octets or 64512 octets if the agent supports PM-Store</li> </ul>	
	<ul> <li>Blood pressure -&gt; 896 octets</li> </ul>	
	■ Thermometer -> 896 octets	
	<ul> <li>Independent activity hub -&gt; 5120 octets</li> </ul>	
	<ul> <li>Cardiovascular -&gt; 64512 octets or 6624 octets if the agent under test only supports a Step Counter Profile</li> </ul>	
	Strength -> 64512 octets:	
	<ul> <li>Adherence monitor -&gt; 1024 octets</li> </ul>	
	<ul><li>Peak flow -&gt; 2030 octets</li></ul>	
	<ul> <li>Body composition analyzer -&gt; 7730 octets</li> </ul>	
	<ul> <li>Basic ECG/Simple ECG -&gt; 7168 octets or 64512 octets if the agent supports a PM-Store</li> </ul>	
	<ul> <li>Basic ECG/Heart rate -&gt; 1280 octets or 64512 octets if the agent supports a PM-Store</li> </ul>	
	<ul> <li>International normalized ratio -&gt; 896 octets or 64512 if the agent supports a PM-Store</li> </ul>	
	o If it responds with a roer, the reason must not be a protocol-violation (23)	
	<ul> <li>In step 4, the agent must respond with a rors-cmip-get message.</li> </ul>	
Notes		

TP Id		TP/PLT/AG/CLASS/GL/BV-022	
TP label		Blood Glucose measurement above the capabilities of the device sensor	
Coverage	Spec	[IEEE 11073-10417]	
	Testable items	BloodGL 29; M	
Applicability		C_AG_OXP_000 AND C_AG_OXP_178 AND C_AG_GL_024 AND (NOT_C_AG_OXP_181)	
Initial condition		The simulated manager and the agent under test are in the operating state.	
Test procedure		Place a blood sample in the device sensor with a blood glucose level above the capabilities of the device sensor and acquire a measurement with the agent under test.	
		2. The test tool simulated manager waits to receive an event report from the agent under test. The event report shall contain the following value:	

	a. Data APDU
	event-type = MDC_NOTI_SCAN_REPORT_FIXED (0x0D 0x1D)
	□ obj-handle = 1 (Blood glucose)
	□ obs-val-data =
	<ul> <li>Basic-Nu-Observed-Value = 0x07FE</li> </ul>
	<ul> <li>Absolute-Time-Stamp = <not case="" for="" relevant="" test="" this=""></not></li> </ul>
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	The vendor must provide a blood sample (or a simulated blood solution) with a blood glucose level above the capabilities of device sensor.

TP ld		TP/PLT/AG/CLASS/GL/BV-023		
TP label		Blood Glucose measurement below the capabilities of the device sensor		
Coverage	Spec	[IEEE 11073-10417]		
	Testable items	BloodGL 30; M		
Applicability	,	C_AG_OXP_000 AND C_AG_OXP_178 AND C_AG_GL_024 AND (NOT_C_AG_OXP_181)		
Initial condit	ion	The simulated manager and the agent under test are in the operating state.		
Test procedure		Place a blood sample in the device sensor with a blood glucose level below the capabilities of the device sensor and acquire a measurement with the agent under test.		
		The test tool simulated manager waits to receive an event report from the agent under test. The event report shall contain the following value:		
		a. Data APDU		
		event-type = MDC_NOTI_SCAN_REPORT_FIXED (0x0D 0x1D)		
		□ obj-handle = 1 (Blood glucose)		
		☐ obs-val-data =		
		Basic-Nu-Observed-Value = 0x0802		
		<ul> <li>Absolute-Time-Stamp = <not case="" for="" relevant="" test="" this=""></not></li> </ul>		
Pass/Fail criteria All checked values are as specified in the test procedure.		All checked values are as specified in the test procedure.		
Notes  The vendor must provide a blood sample (or a simulated blood solution) with a blood glucose level below the capabilities of device sensor.		The vendor must provide a blood sample (or a simulated blood solution) with a blood glucose level below the capabilities of device sensor.		

TP ld		TP/PLT/AG/CLASS/GL/BV-024		
TP label		Control Solution measurement above the capabilities of the device sensor		
Coverage	Spec	[IEEE 11073-10417]		
	Testable items	CtrlSol 15; M		
Applicability		C_AG_OXP_000 AND C_AG_	OXP_178 AND C_AG_GL_024	AND (NOT_C_AG_OXP_181)
Initial condition		The simulated manager and the	ne agent under test are in the op	erating state.

Test procedure	Place a control solution sample in the device sensor with a blood glucose level above the capabilities of the device sensor and check it with the agent under test.	
	2. The test tool simulated manager waits to receive an event report from the agent under test. The event report shall contain the following value:	
	a. Data APDU	
	☐ event-type = MDC_NOTI_SCAN_REPORT_FIXED (0x0D 0x1D)	
	□ obj-handle = 2 (Control solution)	
	☐ obs-val-data =	
	Basic-Nu-Observed-Value = 0x07FE	
	<ul> <li>Absolute-Time-Stamp = <not case="" for="" relevant="" test="" this=""></not></li> </ul>	
Pass/Fail criteria	All checked values are as specified in the test procedure.	
Notes	The vendor must provide a control solution with a blood glucose level above the capabilities of the device sensor.	

TP ld		TP/PLT/AG/CLASS/GL/BV-025		
TP label		Control Solution measurement below the capabilities of the device sensor		
Coverage	Spec	[IEEE 11073-10417]		
	Testable items	CtrlSol 15; M		
Applicability	1	C_AG_OXP_000 AND C_AG_OXP_178 AND C_AG_GL_024 AND (NOT_C_AG_OXP_181)		
Initial condi	tion	The simulated manager and the agent under test are in the operating state.		
Test procedure		<ol> <li>Place a control solution sample in the device sensor with a blood glucose level above the capabilities of the device sensor and check it with the agent under test.</li> <li>The test tool simulated manager waits to receive an event report from the agent under</li> </ol>		
		test. The event report shall contain the following value:  a. Data APDU		
		event-type = MDC_NOTI_SCAN_REPORT_FIXED (0x0D 0x1D)		
		□ obj-handle = 2 (Control Solution)		
		□ obs-val-data =		
		<ul> <li>Basic-Nu-Observed-Value = 0x0802</li> </ul>		
		<ul> <li>Absolute-Time-Stamp = <not case="" for="" relevant="" test="" this=""></not></li> </ul>		
Pass/Fail cri	iteria	All checked values are as specified in the test procedure.		
Notes		The vendor must provide a control solution with a blood glucose level below the capabilities of the device sensor.		

## **Bibliography**

[b-CDG 1.0]	Continua Health Alliance, Continua Design Guidelines v1.0 (2008), <i>Continua Design Guidelines</i> .
[b-CDG 2010]	Continua Health Alliance, Continua Design Guidelines v1.5 (2010), <i>Continua Design Guidelines</i> .
[b-CDG 2011]	Continua Health Alliance, Continua Design Guidelines (2011) "Adrenaline", <i>Continua Design Guidelines</i> .
[b-CDG 2012]	Continua Health Alliance CDG, Continua Design Guidelines (2012), "Catalyst", <i>Continua Design Guidelines</i> .

[b-ETSI SR 001 262] ETSI SR 001 262 v1.8.1 (2003), *ETSI drafting rules*.

## SERIES OF ITU-T RECOMMENDATIONS

Series A	Organization of the work of ITU-T
Series D	General tariff principles
Series E	Overall network operation, telephone service, service operation and human factors
Series F	Non-telephone telecommunication services
Series G	Transmission systems and media, digital systems and networks
Series H	Audiovisual and multimedia systems
Series I	Integrated services digital network
Series J	Cable networks and transmission of television, sound programme and other multimedia signals
Series K	Protection against interference
Series L	Construction, installation and protection of cables and other elements of outside plant
Series M	Telecommunication management, including TMN and network maintenance
Series N	Maintenance: international sound programme and television transmission circuits
Series O	Specifications of measuring equipment
Series P	Terminals and subjective and objective assessment methods
Series Q	Switching and signalling
Series R	Telegraph transmission
Series S	Telegraph services terminal equipment
Series T	Terminals for telematic services
Series U	Telegraph switching
Series V	Data communication over the telephone network
Series X	Data networks, open system communications and security
Series Y	Global information infrastructure, Internet protocol aspects and next-generation networks
Series Z	Languages and general software aspects for telecommunication systems