

ITU-T

TELECOMMUNICATION
STANDARDIZATION SECTOR
OF ITU

H.845.7

(07/2016)

SERIES H: AUDIOVISUAL AND MULTIMEDIA SYSTEMS

E-health multimedia services and applications –
Interoperability compliance testing of personal health
systems (HRN, PAN, LAN, TAN and WAN)

**Conformance of ITU-T H.810 personal health
devices: PAN/LAN/TAN interface Part 5G:
Strength fitness equipment: Agent**

Recommendation ITU-T H.845.7

ITU-T H-SERIES RECOMMENDATIONS
AUDIOVISUAL AND MULTIMEDIA SYSTEMS

CHARACTERISTICS OF VISUAL TELEPHONE SYSTEMS	H.100–H.199
INFRASTRUCTURE OF AUDIOVISUAL SERVICES	
General	H.200–H.219
Transmission multiplexing and synchronization	H.220–H.229
Systems aspects	H.230–H.239
Communication procedures	H.240–H.259
Coding of moving video	H.260–H.279
Related systems aspects	H.280–H.299
Systems and terminal equipment for audiovisual services	H.300–H.349
Directory services architecture for audiovisual and multimedia services	H.350–H.359
Quality of service architecture for audiovisual and multimedia services	H.360–H.369
Telepresence	H.420–H.429
Supplementary services for multimedia	H.450–H.499
MOBILITY AND COLLABORATION PROCEDURES	
Overview of Mobility and Collaboration, definitions, protocols and procedures	H.500–H.509
Mobility for H-Series multimedia systems and services	H.510–H.519
Mobile multimedia collaboration applications and services	H.520–H.529
Security for mobile multimedia systems and services	H.530–H.539
Security for mobile multimedia collaboration applications and services	H.540–H.549
Mobility interworking procedures	H.550–H.559
Mobile multimedia collaboration inter-working procedures	H.560–H.569
BROADBAND, TRIPLE-PLAY AND ADVANCED MULTIMEDIA SERVICES	
Broadband multimedia services over VDSL	H.610–H.619
Advanced multimedia services and applications	H.620–H.629
Ubiquitous sensor network applications and Internet of Things	H.640–H.649
IPTV MULTIMEDIA SERVICES AND APPLICATIONS FOR IPTV	
General aspects	H.700–H.719
IPTV terminal devices	H.720–H.729
IPTV middleware	H.730–H.739
IPTV application event handling	H.740–H.749
IPTV metadata	H.750–H.759
IPTV multimedia application frameworks	H.760–H.769
IPTV service discovery up to consumption	H.770–H.779
Digital Signage	H.780–H.789
E-HEALTH MULTIMEDIA SERVICES AND APPLICATIONS	
Personal health systems	H.810–H.819
Interoperability compliance testing of personal health systems (HRN, PAN, LAN, TAN and WAN)	H.820–H.859
Multimedia e-health data exchange services	H.860–H.869

For further details, please refer to the list of ITU-T Recommendations.

Recommendation ITU-T H.845.7

Conformance of ITU-T H.810 personal health devices: PAN/LAN/TAN interface Part 5G: Strength fitness equipment: Agent

Summary

Recommendation ITU-T H.845.7 is a transposition of Continua Test Tool DG2013, Test Suite Structure & Test Purposes, PAN-LAN-TAN Interface; Part 5G: Device Specializations. Agent (Strength) (Version 1.4, 2014-01-24), that was developed by the Continua Health Alliance. A number of versions of this specification existed before transposition.

This Recommendation includes an electronic attachment with the protocol implementation conformance statements (PICS) and the protocol implementation extra information for testing (PIXIT) required for the implementation of Annex A

History

Edition	Recommendation	Approval	Study Group	Unique ID*
1.0	ITU-T H.845.7	2015-01-13	16	11.1002/1000/12268
2.0	ITU-T H.845.7	2016-07-14	16	11.1002/1000/12944

* To access the Recommendation, type the URL <http://handle.itu.int/> in the address field of your web browser, followed by the Recommendation's unique ID. For example, <http://handle.itu.int/11.1002/1000/11830-en>.

FOREWORD

The International Telecommunication Union (ITU) is the United Nations specialized agency in the field of telecommunications, information and communication technologies (ICTs). The ITU Telecommunication Standardization Sector (ITU-T) is a permanent organ of ITU. ITU-T is responsible for studying technical, operating and tariff questions and issuing Recommendations on them with a view to standardizing telecommunications on a worldwide basis.

The World Telecommunication Standardization Assembly (WTSA), which meets every four years, establishes the topics for study by the ITU-T study groups which, in turn, produce Recommendations on these topics.

The approval of ITU-T Recommendations is covered by the procedure laid down in WTSA Resolution 1.

In some areas of information technology which fall within ITU-T's purview, the necessary standards are prepared on a collaborative basis with ISO and IEC.

NOTE

In this Recommendation, the expression "Administration" is used for conciseness to indicate both a telecommunication administration and a recognized operating agency.

Compliance with this Recommendation is voluntary. However, the Recommendation may contain certain mandatory provisions (to ensure, e.g., interoperability or applicability) and compliance with the Recommendation is achieved when all of these mandatory provisions are met. The words "shall" or some other obligatory language such as "must" and the negative equivalents are used to express requirements. The use of such words does not suggest that compliance with the Recommendation is required of any party.

INTELLECTUAL PROPERTY RIGHTS

ITU draws attention to the possibility that the practice or implementation of this Recommendation may involve the use of a claimed Intellectual Property Right. ITU takes no position concerning the evidence, validity or applicability of claimed Intellectual Property Rights, whether asserted by ITU members or others outside of the Recommendation development process.

As of the date of approval of this Recommendation, ITU had not received notice of intellectual property, protected by patents, which may be required to implement this Recommendation. However, implementers are cautioned that this may not represent the latest information and are therefore strongly urged to consult the TSB patent database at <http://www.itu.int/ITU-T/ipr/>.

© ITU 2016

All rights reserved. No part of this publication may be reproduced, by any means whatsoever, without the prior written permission of ITU.

Table of Contents

	Page
1 Scope.....	1
2 References.....	2
3 Definitions	2
3.1 Terms defined elsewhere	2
3.2 Terms defined in this Recommendation.....	2
4 Abbreviations and acronyms	2
5 Conventions	3
6 Test suite structure (TSS)	4
7 Electronic attachment	6
Annex A – Test purposes (TPs)	7
A.1 TP definition conventions.....	7
A.2 Subgroup 1.3.7: Strength (ST).....	8
Bibliography.....	35

Electronic attachment: This Recommendation includes an electronic attachment with the protocol implementation conformance statements (PICS) and the protocol implementation extra information for testing (PIXIT) required for the implementation of Annex A.

Introduction

This Recommendation is a transposition of Continua Test Tool DG2013, Test Suite Structure & Test Purposes, PAN-LAN-TAN Interface; Part 5G: Device Specializations. Agent (Strength) (Version 1.4, 2014-01-24), that was developed by the Continua Health Alliance. A number of versions of this specification existed before transposition and these can be found in the table below

Version	Date	Revision history
1.2	2012-10-05	Initial release for Test Tool DG2011. This is the same version as "TSS&TP_1.5_PAN-LAN_PART_5G_v1.2.doc" because new features included in [b-CDG 2011] do not affect the test procedures specified in this document.
1.3	2013-05-24	Initial release for Test Tool DG2012. This uses "TSS&TP_DG2011_PAN-LAN_PART_5G_v1.2.doc" as a baseline and adds new features included in [b-CDG 2012]: Max APDU size for GM, BCA and ECG.
1.4	2014-01-24	Initial release for Test Tool DG2013. This uses "TSS&TP_DG2012_PAN-LAN_PART_5G_v1.3.doc" as a baseline and adds new features included in [ITU-T H.810 (2015)]: <ul style="list-style-type: none">• Adds glucose meter BLE• Adds BLE SSP support• Adds NFC new transport• Adds INR device specialization

Recommendation ITU-T H.845.7

Conformance of ITU-T H.810 personal health devices: PAN/LAN/TAN interface Part 5G: Strength fitness equipment: Agent

1 Scope

The scope of this Recommendation¹ is to provide a test suite structure and the test purposes (TSS & TP) for the PAN/LAN/TAN interface based on the requirements defined in the Continua Design Guidelines (CDG) [ITU-T H.810 (2015)]. The objective of this test specification is to provide a high probability of air interface interoperability between different devices.

The TSS and TP for the PAN/LAN/TAN interface document have been divided into ten parts. Each part is listed below:

- **Part 1:** Optimized exchange protocol [ISO/IEEE 11073-20601A] Agent
- **Part 2:** Optimized exchange protocol [ISO/IEEE 11073-20601A] Manager
- **Part 3:** Continua design guidelines. Agent
- **Part 4:** Continua design guidelines. Manager
- **Part 5:** Device specializations. Agent. This document is divided in 14 subparts:
 - **Part 5A:** Weighing scales
 - **Part 5B:** Glucose meter
 - **Part 5C:** Pulse oximeter
 - **Part 5D:** Blood pressure monitor
 - **Part 5E:** Thermometer
 - **Part 5F:** Cardiovascular fitness and activity monitor
 - **Part 5G:** Strength fitness equipment
 - **Part 5H:** Independent living activity hub
 - **Part 5I:** Adherence monitor
 - **Part 5J:** Insulin pump (Future development)
 - **Part 5K:** Peak flow
 - **Part 5L:** Body composition analyser
 - **Part 5M:** Basic electrocardiograph
 - **Part 5N:** International normalized ratio monitor
- **Part 6:** Device specializations. Manager
- **Part 7:** Continua design guidelines. Agent BLE
- **Part 8:** Continua design guidelines. Manager BLE
- **Part 9:** Personal health devices transcoding white paper. Agent
- **Part 10:** Personal health devices transcoding white paper. Manager

¹ This Recommendation includes an electronic attachment with the protocol implementation conformance statements (PICS) and the protocol implementation extra information for testing (PIXIT) required for the implementation of Annex A.

2 References

The following ITU-T Recommendations and other references contain provisions which, through reference in this text, constitute provisions of this Recommendation. At the time of publication, the editions indicated were valid. All Recommendations and other references are subject to revision; users of this Recommendation are therefore encouraged to investigate the possibility of applying the most recent edition of the Recommendations and other references listed below. A list of the currently valid ITU-T Recommendations is regularly published. The reference to a document within this Recommendation does not give it, as a stand-alone document, the status of a Recommendation.

- [ITU-T H.810 (2015)] Recommendation ITU-T H.810 (2015), *Interoperability design guidelines for personal health systems*.
- [ITU-T H.810 (2016)] Recommendation ITU-T H.810 (2016), *Interoperability design guidelines for personal health systems*.
- [IEEE 11073-10442] IEEE 11073-10442-2008, *Health informatics – Personal health device communication – Part 10442: Device specialization – Strength fitness equipment*.
- [ISO/IEEE 11073-104xx] ISO/IEEE 11073-104xx (in force), *Health informatics – Personal health device communication – Device specialization*.
- NOTE – Shorthand is used to refer to the collection of device specialization standards that utilize [ISO/IEEE 11073-20601A], where xx can be any number from 01 to 99, inclusive.
- [ISO/IEEE 11073-20601A] ISO/IEEE 11073-20601:2010, *Health informatics – Personal health device communication – Part 20601: Application profile – Optimized exchange protocol*, including ISO/IEEE 11073-20601:2010 Amd 1:2015.
- <http://www.iso.org/iso/home/store/catalogue_tc/catalogue_detail.htm?csnumber=54331>
with
<http://www.iso.org/iso/home/store/catalogue_tc/catalogue_detail.htm?csnumber=63972>

3 Definitions

3.1 Terms defined elsewhere

This Recommendation uses the following terms defined elsewhere:

3.1.1 agent [ISO/IEEE 11073-20601A]: A node that collects and transmits personal health data to an associated manager.

3.1.2 manager [ISO/IEEE 11073-20601A]: A node receiving data from one or more agent systems. Some examples of managers include a cellular phone, health appliance, set top box, or a computer system.

3.2 Terms defined in this Recommendation

None.

4 Abbreviations and acronyms

This Recommendation uses the following abbreviations and acronyms:

ATS Abstract Test Suite
CDG Continua Design Guidelines
DIMLESS Dimension-less

DUT	Device Under Test
GUI	Graphical User Interface
INR	International Normalized Ratio
IUT	Implementation Under Test
MDS	Medical Device System
NFC	Near Field Communication
PAN	Personal Area Network
PCT	Protocol Conformance Testing
PCO	Point of Control and Observation
PHD	Personal Healthcare Device
PHDC	Personal Healthcare Device Class
PHM	Personal Health Manager
PICS	Protocol Implementation Conformance Statement
PIXIT	Protocol Implementation extra Information for Testing
SABTE	Sleep Apnoea Breathing Therapy Equipment
SDP	Service Discovery Protocol
SOAP	Simple Object Access Protocol
TCWG	Test and Certification Working Group
TP	Test Purpose
TSS	Test Suite Structure
USB	Universal Serial Bus
WDM	Windows Driver Model

5 Conventions

The key words "SHALL", "SHALL NOT", "SHOULD", "SHOULD NOT", "MAY", "MAY NOT" in this Recommendation are to be interpreted as in [b-ETSI SR 001 262].

- SHALL is equivalent to 'must' or 'it is required to'.
- SHALL NOT is equivalent to 'must not' or 'it is not allowed'.
- SHOULD is equivalent to 'it is recommended to'.
- SHOULD NOT is equivalent to 'it is not recommended to'.
- MAY is equivalent to 'is permitted'.
- MAY NOT is equivalent to 'it is not required that'.

NOTE – The above-mentioned key words are capitalized for illustrative purposes only and they do not appear capitalized within this Recommendation.

Reference is made in the ITU-T H.800-series of Recommendations to different versions of the Continua Design Guidelines (CDG) by a specific designation. The list of terms that may be used in this Recommendation is provided in Table 1.

Table 1 – List of designations associated with the various versions of the CDG

CDG name	Transposed as	Version	Description	Designation
2016 plus errata	[ITU-T H.810 (2016)]	6.1	Release 2016 plus errata noting all ratified bugs [ITU-T H.810 (2016)].	–
2016	–	6.0	Release 2016 of the CDG including maintenance updates of the CDG 2015 and additional guidelines that cover new functionalities.	Iris
2015 plus errata	[ITU-T H.810 (2015)]	5.1	Release 2015 plus errata noting all ratified bugs [ITU-T H.810 (2015)].	–
2015	–	5.0	Release 2015 of the CDG including maintenance updates of the CDG 2013 and additional guidelines that cover new functionalities.	Genome
2013 plus errata	[b-ITU-T H.810 (2013)]	4.1	Release 2013 plus errata noting all ratified bugs [b-ITU-T H.810 (2013)].	–
2013	–	4.0	Release 2013 of the CDG including maintenance updates of the CDG 2012 and additional guidelines that cover new functionalities.	Endorphin
2012 plus errata	–	3.1	Release 2012 plus errata noting all ratified bugs [b-CDG 2012].	–
2012	–	3.0	Release 2012 of the CDG including maintenance updates of the CDG 2011 and additional guidelines that cover new functionalities.	Catalyst
2011 plus errata	–	2.1	CDG 2011 integrated with identified errata.	–
2011	–	2.0	Release 2011 of the CDG including maintenance updates of the CDG 2010 and additional guidelines that cover new functionalities [b-CDG 2011].	Adrenaline
2010 plus errata	–	1.6	CDG 2010 integrated with identified errata	–
2010	–	1.5	Release 2010 of the CDG with maintenance updates of the CDG Version 1 and additional guidelines that cover new functionalities [b-CDG 2010].	1.5
1.0	–	1.0	First released version of the CDG [b-CDG 1.0].	–

6 Test suite structure (TSS)

The test purposes (TPs) for the PAN/LAN/TAN interface have been divided into the main subgroups specified below. Annex A describes the TPs for subgroup 1.3.7 (shown in bold).

- Group 1: Agent (AG)
 - Group 1.1: Transport (TR)
 - Subgroup 1.1.1: Design guidelines: Common (DGC)
 - Subgroup 1.1.2: USB design guidelines (UDG)
 - Subgroup 1.1.3: Bluetooth design guidelines (BDG)
 - Subgroup 1.1.4: Pulse oximeter design guidelines (PODG)

- Subgroup 1.1.5: Cardiovascular design guidelines (CVDG)
- Subgroup 1.1.6: Activity hub design guidelines (HUBDG)
- Subgroup 1.1.7: ZigBee design guidelines (ZDG)
- Subgroup 1.1.8: Glucose meter design guidelines (GLDG)
- Subgroup 1.1.9: Bluetooth low energy design guidelines (BLEDG)
- Subgroup 1.1.10: Basic electrocardiograph design guidelines (ECGDG)
- Subgroup 1.1.11: NFC design guidelines (NDG)
- Group 1.2: 20601: Optimized exchange protocol (OXP)
 - Subgroup 1.2.1: PHD domain information model (DIM)
 - Subgroup 1.2.2: PHD service model (SER)
 - Subgroup 1.2.3: PHD communication model (COM)
- Group 1.3: Devices class specializations (CLASS)
 - Subgroup 1.3.1: Weighing scales (WEG)
 - Subgroup 1.3.2: Glucose meter (GL)
 - Subgroup 1.3.3: Pulse oximeter (PO)
 - Subgroup 1.3.4: Blood pressure monitor (BPM)
 - Subgroup 1.3.5: Thermometer (TH)
 - Subgroup 1.3.6: Cardiovascular (CV)
 - **Subgroup 1.3.7: Strength (ST)**
 - Subgroup 1.3.8: Activity hub (HUB)
 - Subgroup 1.3.9: Adherence monitor (AM)
 - Subgroup 1.3.10: Insulin pump (IP) (Future development)
 - Subgroup 1.3.11: Peak flow (PF)
 - Subgroup 1.3.12: Body composition analyzer (BCA)
 - Subgroup 1.3.13: Basic electrocardiograph (ECG)
 - Subgroup 1.3.14: International normalized ratio (INR)
 - Subgroup 1.3.15: Sleep apnoea breathing therapy equipment (SABTE)
- Group 1.4: Personal health device transcoding whitepaper (PHDTW)
 - Subgroup 1.4.1: Whitepaper general requirements (GEN)
 - Subgroup 1.4.2: Whitepaper thermometer requirements (TH)
 - Subgroup 1.4.3: Whitepaper blood pressure requirements (BPM)
 - Subgroup 1.4.4: Whitepaper heart rate requirements (HR)
 - Subgroup 1.4.5: Whitepaper glucose meter requirements (GL)
 - Subgroup 1.4.6: Whitepaper weight scale requirements (WS)

– Group 2: Manager (MAN)

- Group 2.1: Transport (TR)
 - Subgroup 2.1.1: Design guidelines: Common (DGC)
 - Subgroup 2.1.2: USB design guidelines (UDG)
 - Subgroup 2.1.3: Bluetooth design guidelines (BDG)
 - Subgroup 2.1.4: Cardiovascular design guidelines (CVDG)
 - Subgroup 2.1.5: Activity hub design guidelines (HUBDG)

- Subgroup 2.1.6: ZigBee design guidelines (ZDG)
- Subgroup 2.1.7: Bluetooth low energy design guidelines (BLEDG)
- Subgroup 2.1.8: NFC design guidelines (NDG)
- Group 2.2: 20601: Optimized exchange protocol (OXP)
 - Subgroup 2.2.1: General (GEN)
 - Subgroup 2.2.2: PHD domain information model (DIM)
 - Subgroup 2.2.3: PHD service model (SER)
 - Subgroup 2.2.4: PHD communication model (COM)
- Group 2.3: Devices class specializations (CLASS)
 - Subgroup 2.3.1: Weighing scales (WEG)
 - Subgroup 2.3.2: Glucose meter (GL)
 - Subgroup 2.3.3: Pulse oximeter (PO)
 - Subgroup 2.3.4: Blood pressure monitor (BPM)
 - Subgroup 2.3.5: Thermometer (TH)
 - Subgroup 2.3.6: Cardiovascular (CV)
 - Subgroup 2.3.7: Strength (ST)
 - Subgroup 2.3.8: Activity hub (HUB)
 - Subgroup 2.3.9: Adherence monitor (AM)
 - Subgroup 2.3.10: Insulin pump (IP) (Future development)
 - Subgroup 2.3.11: Peak flow (PF)
 - Subgroup 2.3.12: Body composition analyser (BCA)
 - Subgroup 2.3.13: Basic electrocardiograph (ECG)
 - Subgroup 2.3.14: International normalized ratio (INR)
 - Subgroup 2.3.15: Sleep apnoea breathing therapy equipment (SABTE)
- Group 2.4: Personal health device transcoding whitepaper (PHDTW)
 - Subgroup 2.4.1: Whitepaper general requirements (GEN)
 - Subgroup 2.4.2: Whitepaper thermometer requirements (TH)
 - Subgroup 2.4.3: Whitepaper blood pressure requirements (BPM)
 - Subgroup 2.4.4: Whitepaper heart rate requirements (HR)
 - Subgroup 2.4.5: Whitepaper glucose meter requirements (GL)
 - Subgroup 2.4.6: Whitepaper weight scale requirements (WS)

7 Electronic attachment

The protocol implementation conformance statements (PICS) and the protocol implementation extra information for testing (PIXIT) required for the implementation of Annex A can be downloaded from <http://handle.itu.int/11.1002/2000/12067>.

In the electronic attachment, letters "C" and "I" in the column labelled "Mandatory" are used to distinguish between "PICS" and "PIXIT" respectively during testing. If the cell is empty, the corresponding PICS is "independent". If the field contains a "C", the corresponding PICS is dependent on other PICS, and the logical expression is detailed in the "SCR_Expression" field. The static conformance review (SCR) is used in the test tool to assert whether the PICS selection is consistent.

Annex A

Test purposes (TPs)

(This annex forms an integral part of this Recommendation.)

A.1 TP definition conventions

The test purposes are defined according to the following rules:

- **TP Id:** This is a unique identifier (TP/<TT>/<DUT>/<GR>/<SGR>/<XX> – <NNN>). It is specified according to the naming convention defined below:
 - Each test purpose identifier is introduced by the prefix "TP".
 - <TT>: This is the test tool that will be used in the test case:
 - PAN: Personal area network (Bluetooth or USB)
 - LAN: Local area network (ZigBee)
 - PAN-LAN: Personal area network (Bluetooth or USB) – Local area network (ZigBee)
 - LP-PAN: Low power personal area network (Bluetooth low energy)
 - TAN: Touch area network (NFC)
 - PLT: Personal area network (Bluetooth or USB) – Local area network (ZigBee) – Touch area network (NFC)
 - <DUT>: This is the device under test:
 - AG: PAN/LAN Agent
 - MAN: PAN/LAN Manager
 - <GR>: This identifies a group of test cases.
 - <SGR>: This identifies a subgroup of test cases.
 - <XX>: This identifies the type of testing.
 - BV: Valid behaviour test
 - BI: Invalid behaviour test
 - <NNN>: This is a sequential number that identifies the test purpose.
- **TP label:** This is the TP's title.
- **Coverage:** This contains the specification reference and clause to be checked by the TP
 - Spec: This indicates the earliest version of the specification from which the testable items to be checked by the TP were included.
 - Testable item: This contains testable items to be checked by the TP.
- **Test purpose:** This is a description of the requirements to be tested.
- **Applicability:** This contains the PICS items that define if the test case is applicable or not for a specific device. When a TP contains an "ALL" in this field it means that it applies to the device under test within that scope of the test (specialization, transport used, etc.).
- **Other PICS:** It contains additional PICS items (apart from the PICS specified in the Applicability row) which are used within the test case implementation and can modify the final verdict. When this row is empty, it means that only the PICS specified in the Applicability row are used within the test case implementation.

- **Initial condition:** This indicates the state to which the DUT needs to be moved at the beginning of TC execution.
- **Test procedure:** This describes the steps to be followed in order to execute the test case.
- **Pass/Fail criteria:** This provides criteria to decide whether the DUT passes or fails the test case.

A.2 Subgroup 1.3.7: Strength (ST)

TP Id		TP/PLT/AG/CLASS/ST/BV-000		
TP label		Get MDS Object for Strength fitness equipment specialization: Mandatory, Conditional and Optional Attributes		
Coverage	Spec	[IEEE 11073-10442]		
	Testable items	StrenMDSCClassAttr 1; M	StrenMDSCClassAttr 2; M	StrenMDSCClassAttr 3; R
		StrenMDSCClassAttr 4; R	StrenMDSCClassAttr 5; R	StrenMDSCClassAttr 6; M
Test purpose		Check that: The MDS Object contains the attributes specified for a Strength Fitness Agent		
Applicability		C_AG_OXP_000 AND C_AG_OXP_175		
Other PICS				
Initial condition		The simulated manager and the agent under test are in the operating state.		
Test procedure		<div>1. The simulated manager issues a "roiv-cmip-get" command with the handle set to 0 (to request for an MDS object) and the attribute-id-list set to 0 to indicate all attributes.</div> <div>2. The agent responds with a "rors-cmip-get" service message in which the attribute-list contains a list of all implemented attributes of the MDS object:</div> <div>MDS attributes</div> <div>a. Mandatory attribute System-model</div> <div><div><input type="checkbox"/> attribute-id = MDC_ATTR_ID_MODEL</div><div><input type="checkbox"/> attribute-type = SystemModel</div><div><input type="checkbox"/> attribute-value.length = <variable></div><div><input type="checkbox"/> attribute-value = {Manufacturer, Model}</div></div> <div>b. Mandatory attribute Dev-Configuration-Id</div> <div><div><input type="checkbox"/> attribute-id = MDC_ATTR_DEV_CONFIG_ID</div><div><input type="checkbox"/> attribute-type = ConfigId(INT-U16)</div><div><input type="checkbox"/> attribute-value.length = 2 bytes</div><div><input type="checkbox"/> attribute-value = between < 0x4000 and 0x7FFF ></div></div> <div>c. Recommended attribute Power-Status</div> <div><div><input type="checkbox"/> attribute-id = MDC_ATTR_POWER_STAT</div><div><input type="checkbox"/> attribute-type = PowerStatus (BITS-16)</div><div><input type="checkbox"/> attribute-value.length = 2 bytes</div><div><input type="checkbox"/> attribute-value = ON_MAINS (0x8000) or ON_BATTERY(0x4000).</div></div> <div>d. Recommended attribute Battery-Level</div> <div><div><input type="checkbox"/> attribute-id = MDC_ATTR_VAL_BATT_CHARGE</div><div><input type="checkbox"/> attribute-type = INT-U16</div><div><input type="checkbox"/> attribute-value.length = 2 bytes</div></div>		

	<input type="checkbox"/> attribute-value = <undefined if value>100 > e. Recommended attribute Remaining-Battery-Time <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_BATT_REMAIN <input type="checkbox"/> attribute-type = BatMeasure <input type="checkbox"/> attribute-value.length = <variable> <input type="checkbox"/> attribute-value = <units shall be set to one of: MDC_DIM_MIN, MDC_DIM_HR, MDC_DIM_DAY > f. Mandatory attribute System-Type-Spec-List <input type="checkbox"/> attribute-id = MDC_ATTR_SYS_TYPE_SPEC_LIST <input type="checkbox"/> attribute-type = TypeVerList <input type="checkbox"/> attribute-value.length = <variable> <input type="checkbox"/> attribute-value =MDC_DEV_SPEC_PROFILE_HF_STRENGTH,1
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id		TP/PLT/AG/CLASS/ST/BV-004_A		
TP label		MDS-Configuration. Check Objects		
Coverage	Spec	[IEEE 11073-10442]		
	Testable items	RepCountAttr 1; O	ResisAttr 1; O	RepAttr 1; O
		SetAttr 1; O	ExeposAtt 1; O	ExLateAttr 1; O
		ExGripAttr 1; O	ExMovAttr 1; O	trenMDSObjEven 1; M
Test purpose		Check that: MDS-Configuration-Event shall be sent in Confirmed Mode. Configuration report contains objects declared by the vendor.		
Applicability		C_AG_OXP_000 AND C_AG_OXP_175		
Other PICS		C_AG_OXP_010, C_AG_ST_012, C_AG_ST_030, C_AG_ST_050, C_AG_ST_090, C_AG_ST_113, C_AG_ST_131 , C_AG_ST_153		
Initial condition		The simulated manager and the agent under test are in the unassociated state.		
Test procedure		1. The agent under test sends an AARQ message to the simulated manager. 2. The simulated manager issues an AARE message with result "accepted-unknown-config". 3. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager: a. APDU Type ❑ field- type = PrstApdu ❑ field-length =2 bytes ❑ field-value =0xE7 0x00 b. invoke-id ❑ field- type = InvokeIDType ❑ field-length =INT-U16 ❑ field- value=<Not relevant for this test>		

	<p>c. message</p> <ul style="list-style-type: none"> <input type="checkbox"/> field- type = roiv-cmip-confirmed-event-report <input type="checkbox"/> field-length =two bytes <input type="checkbox"/> field- value=0x01 0x01 (EventReportArgumentSimple) <p>d. obj-handle (EventReportArgumentSimple)</p> <ul style="list-style-type: none"> <input type="checkbox"/> field- type = HANDLE <input type="checkbox"/> field-length =INT-U16 <p>e. event-time (EventReportArgumentSimple)</p> <ul style="list-style-type: none"> <input type="checkbox"/> field- type = Relative Time <input type="checkbox"/> field-length =INT-U32 <input type="checkbox"/> field-value = <ul style="list-style-type: none"> • IF NOT C_AG_OXP_010 THEN value = 0xFF 0xFF 0xFF 0xFF <p>f. event-type (EventReportArgumentSimple)</p> <ul style="list-style-type: none"> <input type="checkbox"/> field- type = OID-Type <input type="checkbox"/> field-length =INT-U16 <input type="checkbox"/> field- value=0x 0D 0x 1C (MDC_NOTI_CONFIG) <p>g. config-report-id (ConfigReport)</p> <ul style="list-style-type: none"> <input type="checkbox"/> field- type = ConfigId <input type="checkbox"/> field-length = INT-U16 <input type="checkbox"/> field- value = <Between 0x40 0x00 and 0x7F 0xFF> <p>h. obj-class (ConfigReport → ConfigObjectList (ConfigObject))</p> <ul style="list-style-type: none"> <input type="checkbox"/> field- type = OID-Type <input type="checkbox"/> field-length = INT-U16 <input type="checkbox"/> field-value=Objects that will be cheked: <ul style="list-style-type: none"> <input type="checkbox"/> Set Object shall be present. <input type="checkbox"/> IF C_AG_ST_012 THEN Repetition numeric object is present, ELSE it is not present. <input type="checkbox"/> IF C_AG_ST_030 THEN Resistance numeric object is present, ELSE it is not present. <input type="checkbox"/> IF C_AG_ST_050 THEN Repetition numeric object is present, ELSE it is not present. <input type="checkbox"/> IF C_AG_ST_090 THEN Exercise Position enumeration object is present, ELSE it is not present. <input type="checkbox"/> IF C_AG_ST_113 THEN Exercise Laterality enumeration object is present, ELSE it is not present. <input type="checkbox"/> IF C_AG_ST_131 THEN Exercise Grip enumeration object is present, ELSE it is not present. Movement enumeration object is present, ELSE it is not present.
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id		TP/PLT/AG/CLASS/ST/BV-004_B
TP label		Repetition Count Numeric Object attributes
Coverage	Spec	[IEEE 11073-10442]

	Testable items	RepCountAttr 2; M	RepCountAttr 3; R	RepCountAttr 4; M
		RepCountAttr 5; R	RepCountAttr 6; M	RepCountAttr 7; R
		RepCountAttr 8; R	RepCountAttr 9; R	RepCountAttr 10; M
		RepCountAttr 11; M		
Test purpose		Check that: The Repetition Count Numeric object contains the attributes specified for Extended Configuration.		
Applicability		C_AG_OXP_000 AND C_AG_OXP_175 AND C_AG_ST_012		
Other PICS				
Initial condition		The simulated manager and the agent under test are in the unassociated state.		
Test procedure		<div>1. The agent under test sends an AARQ message to the simulated manager.</div> <div>2. The simulated manager issues an AARE message with result "accepted-unknown-config".</div> <div>3. The agent responds with a roiv-cmip-confirmed-event report message with an MDC_NOTI_CONFIG event to send its configuration to the manager.</div> <div>4. Record the handle of the Set object.</div> <div>5. The Repetition Count object shall be:<div>a. Mandatory attribute Type<div><input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE</div><div><input type="checkbox"/> attribute-type = TYPE</div><div><input type="checkbox"/> attribute-value = MDC_PART_PHD_HF (129) MDC_HF_REP_COUNT (202)</div></div><div>b. Mandatory attribute Metric-Spec-Small<div><input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL</div><div><input type="checkbox"/> attribute-type = MetricSpecSmall (BITS-16)</div><div><input type="checkbox"/> attribute-value ≠ 0x00 0x00<div><div>• bit 0 (mss-avail-intermittent(0)) shall be set.</div><div>• bit 1(mss-avail-stored-data(1)) shall be set.</div><div>• bit 2 (mss-upd-aperiodic(2)) shall be set.</div><div>• bit 3(mss-msmt-aperiodic(3)) shall be set</div><div>• bit 9 (mss-acc-agent-initiated(9)) shall be set.</div><div>• The other bits have to be 0.</div></div></div></div></div> <div>a. Mandatory attribute Source-Handle-Reference<div><input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF</div><div><input type="checkbox"/> attribute-type = HANDLE (INT-U16)</div><div><input type="checkbox"/> attribute-value.length = 2 bytes</div><div><input type="checkbox"/> attribute-value = Handle of the Set object to which this object is associated</div></div> <div>c. Not Recommended attribute Supplemental-Types<div><input type="checkbox"/> attribute-id = MDC_ATTR_SUPPLEMENTAL_TYPES</div><div><input type="checkbox"/> attribute-type = SupplementalTypeList</div><div><input type="checkbox"/> attribute-value.length =Sequence of TYPE (TYPE.length= 4 bytes)</div></div> <div>d. Not Recommended attribute Metric-Structure-Small<div><input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_STRUCTURE_SMALL</div></div>		

	<ul style="list-style-type: none"> <input type="checkbox"/> attribute-type = MetricStructureSmall <input type="checkbox"/> attribute-value.length = <variable> (Sequence of (ms-struct.length =1byte(INT-U8) + ms-comp-no =1byte(INT-U8))) <p>e. Not Recommended attribute Compound-Simple-Nu-Observed-Value</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_SIMP <input type="checkbox"/> attribute-type = SimpleNuObsValueCmp <input type="checkbox"/> attribute-value.length = <variable> ((SimpleNuObsValueCmp ::= SEQUENCE OF SimpleNuObsValue ; SimpleNuObsValue ::= FLOAT-Type) <p>f. Not Recommended attribute Compound-Basic-Nu-Observed-Value</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_BASIC <input type="checkbox"/> attribute-type = BasicNuObsValueCmp <input type="checkbox"/> attribute-value.length = <variable> (SimpleNuObsValueCmp ::= SEQUENCE OF BasicNuObsValue ; BasicNuObsValue ::= SFLOAT-Type) <p>g. Not Recommended attribute Compound-Nu-Observed-Value</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS <input type="checkbox"/> attribute-type = NuObsValueCmp <input type="checkbox"/> attribute-value.length = <variable> (NuObsValueCmp ::= SEQUENCE OF NuObsValue) <p>6. Wait for the agent under test and the simulated manager to reach the operating state.</p> <p>7. Take a measurement in the agent.</p> <p>8. Wait until the manager receives an event report.</p>
Pass/Fail criteria	<ul style="list-style-type: none"> • All checked values are as specified in the test procedure. • In step 8, check that only non-negative values are used (for observed values of the Repetition Count object).
Notes	

TP Id		TP/PLT/AG/CLASS/ST/BV-005		
TP label		Resistance Numeric Object attributes		
Coverage	Spec	[IEEE 11073-10442]		
	Testable items	ResisAttr 2; M	ResisAttr 3; R	ResisAttr 4; M
		ResisAttr 5; R	ResisAttr 6; M	ResisAttr 7; M
		ResisAttr 8; R	ResisAttr 9; R	ResisAttr 10; R
		ResisAttr 11; M	ResisAttr 12; M	ResisAttr 13; M
		ResisAttr 14; M		
Test purpose		Check that: The Resistance Numeric object contains the attributes specified for Extended Configuration.		
Applicability		C_AG_OXP_000 AND C_AG_OXP_175 AND C_AG_ST_030		
Other PICS				
Initial condition		The simulated manager and the agent under test are in the unassociated state.		
Test procedure		1. The agent under test sends an AARQ message to the simulated manager.		

2. The simulated manager issues an AARE message with result "accepted-unknown-config".
3. The agent responds with a roiv-cmip-confirmed-event report message with a MDC_NOTI_CONFIG event to send its configuration to the manager.
4. Record the handle of the Set object
5. The Resistance object shall be:
 - a. Mandatory attribute Type
 - ☐ attribute-id = MDC_ATTR_ID_TYPE
 - ☐ attribute-type = TYPE
 - ☐ attribute-value = MDC_PART_PHD_HF (129) | MDC_HF_RESISTANCE (203)
 - b. Mandatory attribute Metric-Spec-Small
 - ☐ attribute-id = MDC_ATTR_METRIC_SPEC_SMALL
 - ☐ attribute-type = MetricSpecSmall (BITS-16)
 - ☐ attribute-value ≠ 0x00 0x00
 - bit 0 (mss-avail-intermittent(0)) shall be set.
 - bit 1(mss-avail-stored-data(1)) shall be set.
 - bit 2 (mss-updt-aperiodic(2)) shall be set.
 - bit 3(mss-msmt-aperiodic(3)) shall be set
 - bit 9 (mss-acc-agent-initiated(9)) shall be set.
 - The other bits have to be 0.
 - c. Mandatory attribute Unit-Code
 - ☐ attribute-id = MDC_ATTR_UNIT_CODE
 - ☐ attribute-type = OID-Type(INT-U16)
 - ☐ attribute-value.length = 2 bytes
 - ☐ attribute-value=
 - If the resistance is measured in weight → attribute-value= MDC_DIM_X_G or MDC_DIM_LB
 - If the resistance is measured in an indexed value → attribute-value = MDC_DIM_DIMLESS
 - d. Mandatory attribute Source-Handle-Reference
 - ☐ attribute-id = MDC_ATTR_SOURCE_HANDLE_REF
 - ☐ attribute-type = HANDLE (INT-U16)
 - ☐ attribute-value.length = 2 bytes
 - ☐ attribute-value = Handle of the Set object to which this object is associated.
 - e. Optional attribute Unit-Label-String
 - ☐ attribute-id = MDC_ATTR_ID_LABEL_STRING
 - ☐ attribute-type = OCTET STRING
 - ☐ attribute-value.length = <variable>
 - ☐ attribute-value= If the resistance is measured in an indexed value → Label-string may provide additional information.
 - f. Not Recommended attribute Supplemental-Types
 - ☐ attribute-id = MDC_ATTR_SUPPLEMENTAL_TYPES
 - ☐ attribute-type = SupplementalTypeList
 - ☐ attribute-value.length= <variable> (Sequence of TYPE (TYPE.length= 4 bytes))
 - g. Not Recommended attribute Metric-Structure-Small

	<ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_STRUCTURE_SMALL <input type="checkbox"/> attribute-type = MetricStructureSmall <input type="checkbox"/> attribute-value.length = <variable> (Sequence of (ms-struct.length =1byte(INT-U8) + ms-comp-no =1byte(INT-U8))) <p>h. Not Recommended attribute Compound-Simple-Nu-Observed-Value</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_SIMP <input type="checkbox"/> attribute-type = SimpleNuObsValueCmp <input type="checkbox"/> attribute-value.length = <variable> (SimpleNuObsValueCmp ::= SEQUENCE OF SimpleNuObsValue ; SimpleNuObsValue ::= FLOAT-Type) <p>i. Not Recommended attribute Compound-Basic-Nu-Observed-Value</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_BASIC <input type="checkbox"/> attribute-type = BasicNuObsValueCmp <input type="checkbox"/> attribute-value.length = <variable> (SimpleNuObsValueCmp ::= SEQUENCE OF BasicNuObsValue ; BasicNuObsValue ::= SFLOAT-Type) <p>j. Not Recommended attribute Compound-Nu-Observed-Value</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_SIMP <input type="checkbox"/> attribute-type = NuObsValueCmp <input type="checkbox"/> attribute-value.length = <variable> (NuObsValueCmp ::= SEQUENCE OF NuObsValue) <p>6. Wait for the agent under test and the simulated manager to reach the operating state.</p> <p>7. Take a measurement in the agent.</p> <p>8. Wait until the manager receives an event report.</p>
Pass/Fail criteria	<ul style="list-style-type: none"> • All checked values are as specified in the test procedure. • In step 8, check that only non-negative values are used (for observed values of the Resistance object).
Notes	

TP Id		TP/PLT/AG/CLASS/ST/BV-006		
TP label		Repetition Numeric Object attributes		
Coverage	Spec	[IEEE 11073-10442]		
	Testable items	RepAttr 2; M	RepAttr 3; R	RepAttr 4; M
		RepAttr 5; R	RepAttr 6; M	RepAttr 7; M
		RepAttr 8; R	RepAttr 9; R	RepAttr 10; R
		RepAttr 11; M	RepAttr 13; M	
Test purpose		Check that: The Repetition Numeric object contains the attributes specified for Extended Configuration.		
Applicability		C_AG_OXP_000 AND C_AG_OXP_175 AND C_AG_ST_050		
Other PICS				
Initial condition		The simulated manager and the agent under test are in the unassociated state.		
Test procedure		1. The agent under test sends an AARQ message to the simulated manager.		

	<ol style="list-style-type: none"> 2. The simulated manager issues an AARE message with result "accepted-unknown-config". 3. The agent responds with a roiv-cmip-confirmed-event report message with a MDC_NOTI_CONFIG event to send its configuration to the manager. 4. Record the handle of the Set object. 5. The Repetition object shall be: <ol style="list-style-type: none"> a. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = MDC_PART_PHD_HF (129) MDC_HF_REPETITION (201) b. Mandatory attribute Metric-Spec-Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall (BITS-16) <input type="checkbox"/> attribute-value ≠ 0x00 0x00 <ul style="list-style-type: none"> • bit 0 (mss-avail-intermittent(0)) shall be set. • bit 1(mss-avail-stored-data(1)) shall be set. • bit 2 (mss-upd-aperiodic(2)) shall be set. • bit 3(mss-msmt-aperiodic(3)) shall be set • bit 9 (mss-acc-agent-initiated(9)) shall be set. • The other bits have to be 0. c. Mandatory attribute Unit-Code <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type(INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value= MDC_DIM_X_M or MDC_DIM_X_INCH d. Mandatory attribute Source-Handle-Reference <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF <input type="checkbox"/> attribute-type = HANDLE (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = Handle of the Set object to which this object is associated. Tested later e. Not Recommended attribute Supplemental-Types <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SUPPLEMENTAL_TYPES <input type="checkbox"/> attribute-type = SupplementalTypeList <input type="checkbox"/> attribute-value.length = <variable> (Sequence of TYPE (TYPE.length= 4 bytes)) f. Not Recommended attribute Metric-Structure-Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_STRUCTURE_SMALL <input type="checkbox"/> attribute-type = MetricStructureSmall <input type="checkbox"/> attribute-value.length = <variable> (Sequence of (ms-struct.length =1byte(INT-U8) + ms-comp-no =1byte(INT-U8))) g. Not Recommended attribute Compound-Simple-Nu-Observed-Value <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_SIMP <input type="checkbox"/> attribute-type = SimpleNuObsValueCmp <input type="checkbox"/> attribute-value.length = <variable> (SimpleNuObsValueCmp ::= SEQUENCE OF SimpleNuObsValue ; SimpleNuObsValue ::= FLOAT-Type)
--	---

	<p>h. Not Recommended attribute Compound-Basic-Nu-Observed-Value</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_BASIC <input type="checkbox"/> attribute-type = BasicNuObsValueCmp <input type="checkbox"/> attribute-value.length = <variable> SimpleNuObsValueCmp ::= SEQUENCE OF BasicNuObsValue ; BasicNuObsValue ::= SFLOAT-Type) <p>i. Not recommended attribute Compound-Nu-Observed-Value</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_CMPD_VAL_OBS_SIMP <input type="checkbox"/> attribute-type = NuObsValueCmp <input type="checkbox"/> attribute-value.length = <variable> (NuObsValueCmp ::= SEQUENCE OF NuObsValue) <p>6. Wait for the agent under test and the simulated manager to reach the operating state.</p> <p>7. Take a measurement in the agent.</p> <p>8. Wait until the manager receives an event report.</p>
Pass/Fail criteria	<ul style="list-style-type: none"> • All checked values are as specified in the test procedure. • In step 8, check that only non-negative values are used (for observed values of the Repetition object).
Notes	

TP Id		TP/PLT/AG/CLASS/ST/BV-007		
TP label		Repetition Count Numeric Object and Set object attributes		
Coverage	Spec	[IEEE 11073-10442]		
	Testable items	RepCountAttr 12; M		
Test purpose		Check that: This object shall use the same timestamp attribute and value as the associated Set object.		
Applicability		C_AG_OXP_000 AND C_AG_OXP_175 AND C_AG_ST_012		
Other PICS				
Initial condition		The simulated manager and the agent under test are in the unassociated state.		
Test procedure		<ol style="list-style-type: none"> 1. The agent under test sends an AARQ message to the simulated manager. 2. The simulated manager issues an AARE message with result "accepted-unknown-config". 3. The agent under test sends its configuration to the simulated manager. 4. Record the handle and Attribute-Value-Map of the Set object and the Repetition Count object that is associated to it. 5. Take a measurement. 6. Wait for the simulated manager to receive it. 		
Pass/Fail criteria		<input type="checkbox"/> In step 6 if the Repetition Count measurement contains a time stamp it shall be the same (attribute and value) as that for the Set object.		
Notes				

TP Id		TP/PLT/AG/CLASS/ST/BV-008		
TP label		Resistance Numeric Object and Set object attributes		
Coverage	Spec	[IEEE11073-10442]		
	Testable items	ResisAttr 15; M		
Test purpose		Check that: This object shall use the same timestamp attribute and value as the associated Set object.		
Applicability		C_AG_OXP_000 AND C_AG_OXP_175 AND C_AG_ST_030		
Other PICS				
Initial condition		The simulated manager and the agent under test are in the unassociated state.		
Test procedure		<ol style="list-style-type: none"> 1. The agent under test sends an AARQ message to the simulated manager. 2. The simulated manager issues an AARE message with result "accepted-unknown-config". 3. The agent under test sends its configuration to the simulated manager. 4. Record the handle and the Attribute-Value-Map of the Set object and the Resistance object that is associated to it. 5. Take a measurement. 6. Wait for the simulated manager to receive it. 		
Pass/Fail criteria		<input type="checkbox"/> In step 6 if the Resistance measurement contains a time stamp it shall be the same (attribute and value) as that for the Set object.		
Notes				

TP Id		TP/PLT/AG/CLASS/ST/BV-009		
TP label		Repetition Numeric Object and Set object attributes		
Coverage	Spec	[IEEE 11073-10442]		
	Testable items	RepAttr 14; M		
Test purpose		Check that: This object shall use the same timestamp attribute and value as the associated Set object.		
Applicability		C_AG_OXP_000 AND C_AG_OXP_175 AND C_AG_ST_050		
Other PICS				
Initial condition		The simulated manager and the agent under test are in the unassociated state.		
Test procedure		<ol style="list-style-type: none"> 1. The agent under test sends an AARQ message to the simulated manager. 2. The simulated manager issues an AARE message with result "accepted-unknown-config". 3. The agent under test sends its configuration to the simulated manager. 4. Record the handle and Attribute-Value-Map of the Set object and the Repetition Numeric Object that is associated to it. 5. Take a measurement. 		

	6. Wait for the simulated manager to receive it.
Pass/Fail criteria	<input type="checkbox"/> In step 6 if the Repetition measurement contains a time stamp it shall be the same (attribute) as that for the Set object.
Notes	

TP Id		TP/PLT/AG/CLASS/ST/BV-011		
TP label		Repetition Object observed values		
Coverage	Spec	[IEEE 11073-10442]		
	Testable items	RepAttr 15; M	RepAttr 16; M	
Test purpose		<p>Check that:</p> <p>If the related Set object has a Measure-Active-Period defined, the timestamp for this object shall fall within the time range defined by the associated Set object's timestamp attribute and lasts for the Measure-Active-Period defined by the related Set object's Measure-Active-Period attribute.</p> <p>[AND]</p> <p>If the Repetition object specifies a Measure-Active-Period, the time period defined by the Repetition object shall fall completely within the time range defined by the associated Set object.</p>		
Applicability		C_AG_OXP_000 AND C_AG_OXP_175 AND C_AG_ST_050		
Other PICS				
Initial condition		The simulated manager and the agent under test are in the operating state.		
Test procedure		<ol style="list-style-type: none"> 1. Take a measurement with the agent under test. 2. Wait for the simulated manager to receive it. Record the Time Stamp and the Measure-Active-Period of the Set object and of the Repetition object. 		
Pass/Fail criteria		<input type="checkbox"/> The Time-Stamp of the Repetition measurement shall fall between the time range defined by the Time-Stamp and the Measure-Active-Period of the Set measurement. <input type="checkbox"/> The Measure-Active-Period of the Repetition measurement shall be within the boundaries of the Set one.		
Notes				

TP Id		TP/PLT/AG/CLASS/ST/BV-012		
TP label		Set Enumeration Object attributes		
Coverage	Spec	[IEEE 11073-10442]		
	Testable items	SetAttr 2; M	SetAttr 3; R	SetAttr 4; M
		SetAttr 5; R	SetAttr 6; R	SetAttr 7; M
		SetAttr 8; R	SetAttr 9; R	SetAttr 10; R
		SetAttr 11; R	SetAttr 12; M	
Test purpose		Check that:		

	The Set Enumeration object contains the attributes specified for Extended Configuration.
Applicability	C_AG_OXP_000 AND C_AG_OXP_175
Other PICS	C_AG_ST_156
Initial condition	The simulated manager and the agent under test are in the unassociated state.
Test procedure	<ol style="list-style-type: none"> 1. The agent under test sends an AARQ message to the simulated manager. 2. The simulated manager issues an AARE message with result "accepted-unknown-config". 3. The agent under test sends its configuration to the simulated manager. 4. The Set object shall be: <ol style="list-style-type: none"> a. Mandatory Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value.length = <variable>(Sequence of partition (NomPartition (INT-U16)) and code (OID-Type))) <input type="checkbox"/> attribute-value = MDC_PART_PHD_HF MDC_HF_SET b. Not Recommended attribute Supplemental-Types <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SUPPLEMENTAL_TYPES <input type="checkbox"/> attribute-type = SupplementalTypeList <input type="checkbox"/> attribute-value.length = <variable> (Sequence of TYPE (TYPE.length= 4 bytes → partition NomPartition (INT-U16) and code (OID-Type))) c. Mandatory attribute Metric-Spec-Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall (BITS-16) <input type="checkbox"/> attribute-value ≠ 0x00 0x00 <ul style="list-style-type: none"> • bit 0 (mss-avail-intermittent(0)) shall be set • bit 1 is set(mss-avail-stored-data(1)) shall be set • bit 2 is set(mss-updt-aperiodic(2)) shall be set, • bit 3 is set(mss-msmt-aperiodic(3)) shall be set • bit 9 is set(mss-acc-agent-initiated(9)) shall be set • The other bits have to be 0. d. Not Recommended attribute Metric-Structure-Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_STRUCTURE_SMALL <input type="checkbox"/> attribute-type = MetricStructureSmall <input type="checkbox"/> attribute-value.length = <variable>(Sequence of (ms-struct.length =1byte(INT-U8) + ms-comp-no =1byte(INT-U8))) e. Not Recommended attribute Source-Handle-Reference <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF <input type="checkbox"/> attribute-type = HANDLE (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes f. Conditional attribute Measure-Active-Period <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_PD_MSMT_ACTIVE <input type="checkbox"/> attribute-type = FLOAT-Type (INT-U32) <input type="checkbox"/> attribute-value.length = 4 bytes

	<ul style="list-style-type: none"> g. Mandatory attribute Enum-Observed-Value-Simple-OID <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id= MDC_ATTR_ENUM_OBS_VAL_SIMP_OID <input type="checkbox"/> attribute-type = OID-Type (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_MUSC_* (See Annex C of [b-ISO/IEEE 11073-10408]) h. Not Recommended attribute Enum-Observed-Value-Simple-Bit-Str <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id= MDC_ATTR_ENUM_OBS_VAL_SIMP_BIT_STR <input type="checkbox"/> attribute-type = BITS-32 <input type="checkbox"/> attribute-value.length = 4 bytes i. Not Recommended attribute Enum-Observed-Value-Basic-Bit-Str <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id= MDC_ATTR_ENUM_OBS_VAL_BASIC_BIT_STR <input type="checkbox"/> attribute-type = BITS-16 <input type="checkbox"/> attribute-value.length = 2bytes j. Not Recommended attribute Enum-Observed-Value-Simple-Str <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id= MDC_ATTR_ENUM_OBS_VAL_SIM_STR <input type="checkbox"/> attribute-type = EnumPrintableString <input type="checkbox"/> attribute-value.length = <variable> k. Not Recommended attribute Enum-Observed-Value <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id= MDC_ATTR_VAL_ENUM_OBS <input type="checkbox"/> attribute-type = EnumObsValue <input type="checkbox"/> attribute-value.length = <variable> l. Mandatory attribute Enum-Observed-Value-Partition <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id= MDC_ATTR_ENUM_OBS_VAL_PART <input type="checkbox"/> attribute-type = NomPartition (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id		TP/PLT/AG/CLASS/ST/BV-013		
TP label		Exercise Position Enumeration Object attributes		
Coverage	Spec	[IEEE 11073-10442]		
	Testable items	ExeposAtt 2; M	ExeposAtt 3; R	ExeposAtt 4; M
		ExeposAtt 5; R	ExeposAtt 6; M	ExeposAtt 7; M
		ExeposAtt 8; R	ExeposAtt 9; R	ExeposAtt 10; R
		ExeposAtt 11; R	ExeposAtt 12; R	ExeposAtt 15; M
Test purpose		Check that: The Exercise Position Enumeration object contains the attributes specified for Extended Configuration.		
Applicability		C_AG_OXP_000 AND C_AG_OXP_175 AND C_AG_ST_090		

Other PICS	C_AG_ST_156
Initial condition	The simulated manager and the agent under test are in the unassociated state.
Test procedure	<ol style="list-style-type: none"> 1. The agent under test sends an AARQ message to the simulated manager. 2. The simulated manager issues an AARE message with result "accepted-unknown-config". 3. The agent under test sends its configuration to the simulated manager. 4. The Exercise Position object shall be: <ol style="list-style-type: none"> a. Mandatory Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value.length = <variable> (Sequence of partition (NomPartition (INT-U16)) and code (OID-Type))) <input type="checkbox"/> attribute-value= MDC_PART_PHD_HF MDC_HF_EXERCISE_POSITION b. Not recommended attribute Supplemental-Types <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SUPPLEMENTAL_TYPES <input type="checkbox"/> attribute-type = SupplementalTypeList <input type="checkbox"/> attribute-value.length = <variable> (Sequence of TYPE (TYPE.length= 4 bytes → partition NomPartition (INT-U16) and code (OID-Type))) c. Mandatory attribute Metric-Spec-Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall (BITS-16) <input type="checkbox"/> attribute-value ≠ 0x00 0x00 <ul style="list-style-type: none"> • bit 0 (mss-avail-intermittent(0)) shall be set • bit 1 is set(mss-avail-stored-data(1)) shall be set • bit 2 is set(mss-updt-aperiodic(2)) shall be set • bit 3 is set(mss-msmt-aperiodic(3)) shall be set • bit 9 is set(mss-acc-agent-initiated(9)) shall be set • The other bits have to be 0. d. Not Recommended attribute Metric-Structure-Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_STRUCTURE_SMALL <input type="checkbox"/> attribute-type = MetricStructureSmall <input type="checkbox"/> attribute-value.length = <variable> (Sequence of (ms-struct.length =1byte(INT-U8) + ms-comp-no =1byte(INT-U8))) e. Mandatory attribute Source-Handle-Reference <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF <input type="checkbox"/> attribute-type = HANDLE(INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = Handle of the Set object to which this object is associated f. Mandatory attribute Enum-Observed-Value-Simple-OID <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id= MDC_ATTR_ENUM_OBS_VAL_SIMP_OID <input type="checkbox"/> attribute-type = OID-Type (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_HF_POSITION_* (See Annex C of [b-ISO/IEEE 11073-10408]) g. Not Recommended attribute Enum-Observed-Value-Simple-Bit-Str

	<input type="checkbox"/> attribute-id= MDC_ATTR_ENUM_OBS_VAL_SIMP_BIT_STR <input type="checkbox"/> attribute-type = BITS-32 <input type="checkbox"/> attribute-value.length = 4 bytes h. Not Recommended attribute Enum-Observed-Value-Basic-Bit-Str <input type="checkbox"/> attribute-id= MDC_ATTR_ENUM_OBS_VAL_BASIC_BIT_STR <input type="checkbox"/> attribute-type = BITS-16 <input type="checkbox"/> attribute-value.length = 2 bytes i. Not Recommended attribute Enum-Observed-Value-Simple-Str <input type="checkbox"/> attribute-id= MDC_ATTR_ENUM_OBS_VAL_SIM_STR <input type="checkbox"/> attribute-type = EnumPrintableString <input type="checkbox"/> attribute-value.length = <variable> j. Not Recommended attribute Enum-Observed-Value <input type="checkbox"/> attribute-id= MDC_ATTR_VAL_ENUM_OBS <input type="checkbox"/> attribute-type = EnumObsValue <input type="checkbox"/> attribute-value.length = <variable> k. Not Recommended attribute Enum-Observed-Value-Partition <input type="checkbox"/> attribute-id= MDC_ATTR_ENUM_OBS_VAL_PART <input type="checkbox"/> attribute-type = NomPartition (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id		TP/PLT/AG/CLASS/ST/BV-014		
TP label		Exercise Position Object and Set object attributes		
Coverage	Spec	[IEEE 11073-10442]		
	Testable items	ExeposAtt 14; M		
Test purpose		Check that: The timestamp attribute used for an Exercise Position object instance shall be the same as that which is used for the Set object instance to which it is related		
Applicability		C_AG_OXP_000 AND C_AG_OXP_175 AND C_AG_ST_090		
Other PICS				
Initial condition		The simulated manager and the agent under test are in the unassociated state.		
Test procedure		<ol style="list-style-type: none"> 1. The agent under test sends an AARQ message to the simulated manager. 2. The simulated manager issues an AARE message with the result "accepted-unknown-config". 3. The agent under test sends its configuration to the simulated manager. 4. Record the handle and Attribute-Value-Mpa of the Set object and the Exercise Position object that is associated to it. 5. Take a measurement. 6. Wait for the simulated manager to receive it. 		

Pass/Fail criteria	<input type="checkbox"/> In step 6 if the Exercise Position measurement contains a time stamp it shall be the same (attribute and value) as that for the Set object.
Notes	

TP Id		TP/PLT/AG/CLASS/ST/BV-015		
TP label		Exercise Laterality enumeration Object attributes		
Coverage	Spec	[IEEE 11073-10442]		
	Testable items	ExLateAttr 2; M	ExLateAttr 3; R	ExLateAttr 4; M
		ExLateAttr 5; R	ExLateAttr 6; M	ExLateAttr 7; M
		ExLateAttr 8; R	ExLateAttr 9; R	ExLateAttr 10; R
		ExLateAttr 11; R	ExLateAttr 12; R	ExLateAttr 14; M
Test purpose		Check that: The Exercise Laterality Enumeration object contains the attributes specified for Extended Configuration.		
Applicability		C_AG_OXP_000 AND C_AG_OXP_175 AND C_AG_ST_113		
Other PICS		C_AG_ST_156		
Initial condition		The simulated manager and the agent under test are in the unassociated state.		
Test procedure		<div>1. The agent under test sends an AARQ message to the simulated manager.</div> <div>2. The simulated manager issues an AARE message with the result "accepted-unknown-config".</div> <div>3. The agent under test sends its configuration to the simulated manager.</div> <div>4. The Exercise Laterality object shall be:<div>a. Mandatory Type<div><div>❑ attribute-id = MDC_ATTR_ID_TYPE</div><div>❑ attribute-type = TYPE</div><div>❑ attribute-value.length = <variable> (Sequence of partition (NomPartition (INT-U16)) and code (OID-Type)))</div><div>❑ attribute-value= MDC_PART_PHD_HF MDC_HF_EXERCISE_LATERALITY</div></div></div><div>b. Not recommended attribute Supplemental-Types<div><div>❑ attribute-id = MDC_ATTR_SUPPLEMENTAL_TYPES</div><div>❑ attribute-type = SupplementalTypeList</div><div>❑ attribute-value.length = <variable> (Sequence of TYPE (TYPE.length= 4 bytes → partition NomPartition (INT-U16) and code (OID-Type)))</div></div></div><div>c. Mandatory attribute Metric-Spec-Small<div><div>❑ attribute-id = MDC_ATTR_METRIC_SPEC_SMALL</div><div>❑ attribute-type = MetricSpecSmall (BITS-16)</div><div>❑ attribute-value ≠ 0x00 0x00<div><div>▪ bit 0 (mss-avail-intermittent(0)) shall be set</div><div>▪ bit 1 is set (mss-avail-stored-data(1)) shall be set</div><div>▪ bit 2 is set (mss-updt-aperiodic(2)) shall be set</div><div>▪ bit 3 is set (mss-msmt-aperiodic(3)) shall be set</div></div></div></div></div></div>		

	<ul style="list-style-type: none"> ▪ bit 9 is set (mss-acc-agent-initiated(9)) shall be set ▪ The other bits have to be 0. <p>d. Not Recommended attribute Metric-Structure-Small</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_STRUCTURE_SMALL <input type="checkbox"/> attribute-type = MetricStructureSmall <input type="checkbox"/> attribute-value.length = <variable> (Sequence of (ms-struct.length =1byte(INT-U8) + ms-comp-no =1byte(INT-U8))) <p>e. Mandatory attribute Source-Handle-Reference</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF <input type="checkbox"/> attribute-type = HANDLE (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = Handle of the Set object to which this object is associated <p>f. Mandatory attribute Enum-Observed-Value-Simple-OID</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id= MDC_ATTR_ENUM_OBS_VAL_SIMP_OID <input type="checkbox"/> attribute-type = OID-Type (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_HF_LATERALITY_* (See Annex C of [b-ISO/IEEE 11073-10408]) <p>g. Not Recommended attribute Enum-Observed-Value-Simple-Bit-Str</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id= MDC_ATTR_ENUM_OBS_VAL_SIMP_BIT_STR <input type="checkbox"/> attribute-type = BITS-32 <input type="checkbox"/> attribute-value.length = 4 bytes <p>h. Not Recommended attribute Enum-Observed-Value-Basic-Bit-Str</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id= MDC_ATTR_ENUM_OBS_VAL_BASIC_BIT_STR <input type="checkbox"/> attribute-type = BITS-16 <input type="checkbox"/> attribute-value.length = 2 bytes <p>i. Not Recommended attribute Enum-Observed-Value-Simple-Str</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id= MDC_ATTR_ENUM_OBS_VAL_SIM_STR <input type="checkbox"/> attribute-type = EnumPrintableString <input type="checkbox"/> attribute-value.length = <variable> <p>j. Not Recommended attribute Enum-Observed-Value</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id= MDC_ATTR_VAL_ENUM_OBS <input type="checkbox"/> attribute-type = EnumObsValue <input type="checkbox"/> attribute-value.length = <variable> <p>k. Not Recommended attribute Enum-Observed-Value-Partition</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id= MDC_ATTR_ENUM_OBS_VAL_PART <input type="checkbox"/> attribute-type = NomPartition (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/AG/CLASS/ST/BV-016
TP label	Exercise Laterality Object and Set object attributes

Coverage	Spec	[IEEE 11073-10442]		
	Testable items	ExLateAttr 13; M		
Test purpose		Check that: This object's timestamp shall be equal to the timestamp of the associated Set Object		
Applicability		C_AG_OXP_000 AND C_AG_OXP_175 AND C_AG_ST_113		
Other PICS				
Initial condition		The simulated manager and the agent under test are in the unassociated state.		
Test procedure		<ol style="list-style-type: none"> 1. The agent under test sends an AARQ message to the simulated manager. 2. The simulated manager issues an AARE message with the result "accepted-unknown-config". 3. The agent under test sends its configuration to the simulated manager. 4. Record the handle and the Attribute-Value-Map of the Set object and the Exercise Laterality object that is associated to it. 5. Take a measurement 6. Wait for the simulated manager to receive it. 		
Pass/Fail criteria		<input type="checkbox"/> In step 6 if the Exercise Laterality measurement contains a time stamp it shall be the same (attribute and value) as that for the Set object.		
Notes				

TP Id		TP/PLT/AG/CLASS/ST/BV-017		
TP label		Exercise Grip enumeration Object attributes		
Coverage	Spec	[IEEE 11073-10442]		
	Testable items	ExGripAttr 2; M	ExGripAttr 3; R	ExGripAttr 4; M
		ExGripAttr 5; R	ExGripAttr 6; M	ExGripAttr 7; M
		ExGripAttr 8; R	ExGripAttr 9; R	ExGripAttr 10; R
		ExGripAttr 11; R	ExGripAttr 12; R	ExGripAttr 15; M
Test purpose		Check that: The Exercise Grip Enumeration object contains the attributes specified for Extended Configuration.		
Applicability		C_AG_OXP_000 AND C_AG_OXP_175 AND C_AG_ST_131		
Other PICS		C_AG_ST_156		
Initial condition		The simulated manager and the agent under test are in the unassociated state.		
Test procedure		1. The agent under test sends an AARQ message to the simulated manager. 2. The simulated manager issues an AARE message with the result "accepted-unknown-config". 3. The agent under test sends its configuration to the simulated manager. 4. The Exercise Grip object shall be: a. Mandatory Type		

	<ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value.length = <variable> (Sequence of partition (NomPartition (INT-U16)) and code (OID-Type))) <input type="checkbox"/> attribute-value= MDC_PART_PHD_HF MDC_HF_EXERCISE_GRIP <p>b. Not Recommended attribute Supplemental-Types</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SUPPLEMENTAL_TYPES <input type="checkbox"/> attribute-type = SupplementalTypeList <input type="checkbox"/> attribute-value.length = <variable> (Sequence of TYPE (TYPE.length= 4 bytes → partition NomPartition (INT-U16) and code (OID-Type))) <p>c. Mandatory attribute Metric-Spec-Small</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall (BITS-16) <input type="checkbox"/> attribute-value ≠ 0x00 0x00 <ul style="list-style-type: none"> • bit 0 (mss-avail-intermittent(0)) shall be set • bit 1 is set (mss-avail-stored-data(1)) shall be set • bit 2 is set (mss-updt-aperiodic(2)) shall be set • bit 3 is set (mss-msmt-aperiodic(3)) shall be set • bit 9 is set(mss-acc-agent-initiated(9)) shall be set • The other bits have to be 0. <p>d. Not Recommended attribute Metric-Structure-Small</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_STRUCTURE_SMALL <input type="checkbox"/> attribute-type = MetricStructureSmall <input type="checkbox"/> attribute-value.length = <variable> (Sequence of (ms-struct.length =1byte(INT-U8) + ms-comp-no =1byte(INT-U8))) <p>e. Mandatory attribute Source-Handle-Reference</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF <input type="checkbox"/> attribute-type = HANDLE (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = Handle of the Set object to which this object is associated. <p>f. Mandatory attribute Enum-Observed-Value-Simple-OID</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id= MDC_ATTR_ENUM_OBS_VAL_SIM_OID <input type="checkbox"/> attribute-type = OID-Type (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_HF_GRIP_* (See Annex C of [b-ISO/IEEE 11073-10408]) <p>g. Not Recommended attribute Enum-Observed-Value-Simple-Bit-Str</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id= MDC_ATTR_ENUM_OBS_VAL_SIMP_BIT_STR <input type="checkbox"/> attribute-type = BITS-32 <input type="checkbox"/> attribute-value.length = 4 bytes <p>h. Not Recommended attribute Enum-Observed-Value-Basic-Bit-Str</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id= MDC_ATTR_ENUM_OBS_VAL_BASIC_BIT_STR <input type="checkbox"/> attribute-type = BITS-16 <input type="checkbox"/> attribute-value.length = 2 bytes <p>i. Not Recommended attribute Enum-Observed-Value-Simple-Str</p>
--	---

	<input type="checkbox"/> attribute-id= MDC_ATTR_ENUM_OBS_VAL_SIMP_STR <input type="checkbox"/> attribute-type = EnumPrintableString <input type="checkbox"/> attribute-value.length = j. Not Recommended attribute Enum-Observed-Value <input type="checkbox"/> attribute-id= MDC_ATTR_VAL_ENUM_OBS <input type="checkbox"/> attribute-type = EnumObsValue <input type="checkbox"/> attribute-value.length = k. Not Recommended attribute Enum-Observed-Value-Partition <input type="checkbox"/> attribute-id= MDC_ATTR_ENUM_OBS_VAL_PART <input type="checkbox"/> attribute-type = NomPartition(INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id		TP/PLT/AG/CLASS/ST/BV-018		
TP label		Exercise Grip Object and Set object attributes		
Coverage	Spec	[IEEE 11073-10442]		
	Testable items	ExGripAttr 14; M		
Test purpose		Check that: The timestamp attribute used for an Exercise Grip object instance shall be the same as that which is used for the Set object instance to which it is related		
Applicability		C_AG_OXP_000 AND C_AG_OXP_175 AND C_AG_ST_131		
Other PICS				
Initial condition		The simulated manager and the agent under test are in the unassociated state.		
Test procedure		<ol style="list-style-type: none"> 1. The agent under test sends an AARQ message to the simulated manager. 2. The simulated manager issues an AARE message with result "accepted-unknown-config". 3. The agent under test sends its configuration to the simulated manager. 4. Record the handle and Attribute-Value-Map of the Set object and the Exercise Grip object that is associated to it. 5. Take a measurement. 6. Wait for the simulated manager to receive it. 		
Pass/Fail criteria		<ul style="list-style-type: none"> • In step 6 if the Exercise Grip measurement contains a time stamp it shall be the same (attribute and value) as that for the Set object. 		
Notes				

TP Id	TP/PLT/AG/CLASS/ST/BV-019
TP label	Exercise Movement enumeration Object attributes

Coverage	Spec	[IEEE 11073-10442]		
	Testable items	ExMovAttr 2; M	ExMovAttr 3; R	ExMovAttr 4; M
		ExMovAttr 5; R	ExMovAttr 6; M	ExMovAttr 7; M
		ExMovAttr 8; R	ExMovAttr 9; R	ExMovAttr 10; R
		ExMovAttr 11; R	ExMovAttr 12; R	ExMovAttr 15; M
Test purpose		Check that: The Exercise Movement Enumeration object contains the attributes specified for Extended Configuration.		
Applicability		C_AG_OXP_000 AND C_AG_OXP_175 AND C_AG_ST_153		
Other PICS		C_AG_ST_156		
Initial condition		The simulated manager and the agent under test are in the unassociated state.		
Test procedure		<div>1. The agent under test sends an AARQ message to the simulated manager.</div> <div>2. The simulated manager issues an AARE message with the result "accepted-unknown-config".</div> <div>3. The agent under test sends its configuration to the simulated manager.</div> <div>4. The Exercise Movement object shall be:<div>a. Mandatory Type<div><div>❑ attribute-id = MDC_ATTR_ID_TYPE</div><div>❑ attribute-type = TYPE</div><div>❑ attribute-value.length = <variable> (Sequence of partition (NomPartition (INT-U16)) and code (OID-Type)))</div><div>❑ attribute-value= MDC_PART_PHD_HF MDC_HF_EXERCISE_MOVEMENT</div></div></div><div>b. Not Recommended attribute Supplemental-Types<div><div>❑ attribute-id = MDC_ATTR_SPPLEMENTAL_TYPES</div><div>❑ attribute-type = SupplementalTypeList</div><div>❑ attribute-value.length =<variable>(Sequence of TYPE (TYPE.length= 4 bytes → partition NomPartition (INT-U16) and code (OID-Type)))</div></div></div><div>c. Mandatory attribute Metric-Spec-Small<div><div>❑ attribute-id = MDC_ATTR_METRIC_SPEC_SMALL</div><div>❑ attribute-type = MetricSpecSmall (BITS-16)</div><div>❑ attribute-value ≠ 0x00 0x00<div><div>• bit 0 (mss-avail-intermittent(0)) shall be set</div><div>• bit 1 (mss-avail-stored-data(1)) shall be set</div><div>• bit 2 (mss-updt-aperiodic(2)) shall be set</div><div>• bit 3 (mss-msmt-aperiodic(3)) shall be set</div><div>• bit 9 (mss-acc-agent-initiated(9)) shall be set</div><div>• The other bits have to be 0.</div></div></div></div></div><div>d. Not Recommended attribute Metric-Structure-Small<div><div>❑ attribute-id = MDC_ATTR_METRIC_STRUCTURE_SMALL</div><div>❑ attribute-type = MetricStructureSmall</div><div>❑ attribute-value.length = <variable> (Sequence of (ms-struct.length =1byte(INT-U8) + ms-comp-no = 1 byte(INT-U8)))</div></div></div></div>		

	<ul style="list-style-type: none"> e. Mandatory attribute Source-Handle-Reference <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF <input type="checkbox"/> attribute-type = HANDLE (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = Handle of the Set object to which this object is associated f. Mandatory attribute Enum-Observed-Value-Simple-OID <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id= MDC_ATTR_ENUM_OBS_VAL_SIMP_OID <input type="checkbox"/> attribute-type = OID-Type (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_HF_MOVEMENT_* (See Annex C of [b-ISO/IEEE 11073-10408]) g. Not Recommended attribute Enum-Observed-Value-Simple-Bit-Str <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id= MDC_ATTR_ENUM_OBS_VAL_SIMP_BIT_STR <input type="checkbox"/> attribute-type = BITS-32 <input type="checkbox"/> attribute-value.length = 4 bytes h. Not Recommended attribute Enum-Observed-Value-Basic-Bit-Str <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id= MDC_ATTR_ENUM_OBS_VAL_BASIC_BIT_STR <input type="checkbox"/> attribute-type = BITS-16 <input type="checkbox"/> attribute-value.length = 2 bytes i. Not Recommended attribute Enum-Observed-Value-Simple-Str <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id= MDC_ATTR_ENUM_OBS_VAL_SIM_STR <input type="checkbox"/> attribute-type = EnumPrintableString <input type="checkbox"/> attribute-value.length = j. Not Recommended attribute Enum-Observed-Value <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id= MDC_ATTR_VAL_ENUM_OBS <input type="checkbox"/> attribute-type = EnumObsValue <input type="checkbox"/> attribute-value.length = k. Not Recommended attribute Enum-Observed-Value-Partition <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id= MDC_ATTR_ENUM_OBS_VAL_PART <input type="checkbox"/> attribute-type = NomPartition(INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id		TP/PLT/AG/CLASS/ST/BV-020		
TP label		Exercise Movement Object and Set object attributes		
Coverage	Spec	[IEEE 11073-10442]		
	Testable items	ExMovAttr 14; M		
Test purpose		Check that: The timestamp attribute used for an Exercise Movement object instance shall be the same as that which is used for the Set object instance to which it is associated		

Applicability	C_AG_OXP_000 AND C_AG_OXP_175 AND C_AG_ST_153
Other PICS	
Initial condition	The simulated manager and the agent under test are in the unassociated state.
Test procedure	<ol style="list-style-type: none"> 1. The agent under test sends an AARQ message to the simulated manager. 2. The simulated manager issues an AARE message with the result "accepted-unknown-config". 3. The agent under test sends its configuration to the simulated manager. 4. Record the handle and the Attribute-Value-Map of the Set object and the Exercise Movement that is associated to it. 5. Take a measurement. 6. Wait for the simulated manager to receive it.
Pass/Fail criteria	<input type="checkbox"/> In step 6 if the Exercise Movement measurement contains a time stamp it shall be the same (attribute and value) as that for the Set object.
Notes	

TP Id		TP/PLT/AG/CLASS/ST/BV-021		
TP label		Association Request		
Coverage	Spec	[IEEE 11073-10442]		
	Testable items	StrenAssocReq 1; M	StrenAssocReq 2; M	StrenAssocReq 3; M
		StrenAssocReq 4; M	StrenAssocReq 5; M	StrenAssocReq 6; M
		StrenAssocReq 7; M	StrenAssocReq 8; M	StrenAssocReq 9; M
		StrenAssocReq 10; O	StrenAssocReq 11; M	StrenAssocReq 12; C
		StrenAssocReq 13; C		
Test purpose		Check that: The association procedure data exchange is correct.		
Applicability		C_AG_OXP_000 AND C_AG_OXP_175		
Other PICS		C_AG_OXP_017		
Initial condition		The simulated manager and the agent under test are in the unassociated state.		
Test procedure		1. The agent under test sends an AARQ message to the simulated manager. 2. The expected fields are: a. APDU Type ❑ field- type = AarqApdu ❑ field-length =2 bytes ❑ field-value =0xE2 0x00 b. Assoc-version ❑ field- type = AssociationVersion ❑ field-length =BITS-32 ❑ field- value=0x80 0x00 0x00 0x00		

	<ul style="list-style-type: none"> <input type="checkbox"/> assoc-version = 0x80 0x00 0x00 0x00 (asassoc-version1(0) set) indicates that version 1 of the association protocol is supported. <p>c. Data-proto-id</p> <ul style="list-style-type: none"> <input type="checkbox"/> field- type = DataProtold <input type="checkbox"/> field-length = INT-U16 <input type="checkbox"/> field- value = 0x50 0x79 (20601) <input type="checkbox"/> data-proto-id = 20601 indicates exchange protocol follows this standard, and data-proto-info shall contain PhdAssociationInformation. <p>d. Protocol-Version</p> <ul style="list-style-type: none"> <input type="checkbox"/> field- type = Protocol Version <input type="checkbox"/> field-length = BITS-32 <input type="checkbox"/> field- value = 0x80 0x00 0x00 0x00 <input type="checkbox"/> This value shows that version 1 of the data exchange protocol is supported (assoc-version1(0)=1) <p>e. Encoding-Rules</p> <ul style="list-style-type: none"> <input type="checkbox"/> field- type = EncodingRules <input type="checkbox"/> field-length = BITS-16 <input type="checkbox"/> field- value = depends on the encoding rules supported/selected. <input type="checkbox"/> mder(0) always is set (MDER always is supported) and xer(1) or/and per(2) may be set (optional). <p>f. Nomenclature-Version</p> <ul style="list-style-type: none"> <input type="checkbox"/> field- type = NomenclatureVersion <input type="checkbox"/> field-length = BITS-32 <input type="checkbox"/> field- value = 0x80 0x00 0x00 0x00 <input type="checkbox"/> This value indicates version1 is supported (nom-version1(0) is set). <p>g. Functional-Units</p> <ul style="list-style-type: none"> <input type="checkbox"/> field- type = FunctionalUnits <input type="checkbox"/> field-length = BITS-32 <input type="checkbox"/> If the agent has no Test Association capabilities: field- value= 0x00 0x00 0x00 0x00 <input type="checkbox"/> If the agent has tested capabilities that can be used within the Test Association: field- value= 0x40 0x00 0x00 0x00 <input type="checkbox"/> If the agent has tested capabilities that can be used within the Test Association and requires that the manager establish a Test Association: field- value= 0x60 0x00 0x00 0x00 <p>h. System-Type</p> <ul style="list-style-type: none"> <input type="checkbox"/> field- type = SystemType <input type="checkbox"/> field-length = BITS-32 <input type="checkbox"/> field- value = 0x00 0x80 0x00 0x00 (sys-type-agent) <p>i. System-id</p> <ul style="list-style-type: none"> <input type="checkbox"/> field- type = OCTET STRING <input type="checkbox"/> field-length = 0x00 0x0A <input type="checkbox"/> field- value = < Check with PIXITs > <p>j. Dev-config-id</p> <ul style="list-style-type: none"> <input type="checkbox"/> field- type = ConfigId <input type="checkbox"/> field-length = INT-U16 <input type="checkbox"/> field- value = <between 0x40 0x00 and 0x7F 0xFF>
--	--

	<ul style="list-style-type: none"> k. Data-req-mode-flags (DataReqModeCapab) <ul style="list-style-type: none"> <input type="checkbox"/> field- type = DataReqModeFlags <input type="checkbox"/> field-length = BITS-16 <input type="checkbox"/> If the agent implements only this Device Specialization: field- value = 0x00 0x01 – Agent initiated data request/flows l. Data-req-init-agent-count (DataReqModeCapab) <ul style="list-style-type: none"> <input type="checkbox"/> field- type = INT-U8 <input type="checkbox"/> field-length = INT-U8 <input type="checkbox"/> If the agent implements only this Device Specialization: field- value = 0x01 m. Data-req-init-manager-count (DataReqModeCapab) <ul style="list-style-type: none"> <input type="checkbox"/> field- type = INT-U8 <input type="checkbox"/> field-length = INT-U8 <input type="checkbox"/> If the agent implements only this Device Specialization: field- value = 0x00
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id		TP/PLT/AG/CLASS/ST/BV-022		
TP label		Config Changes Service. Resistance Contextual Attribute.		
Coverage	Spec	[IEEE 11073-10442]		
	Testable items	NumObj 1; M		
	Spec	[ITU-T H.810 (2015)]		
	Testable items	Communication 8; M		
Test purpose		Check that: Whenever a contextual attribute changes, the Agent shall report these changes to the Manager using an MDS object event prior to reporting any of the dependent values.		
Applicability		C_AG_OXP_000 AND C_AG_OXP_175 AND C_AG_ST_030 AND C_AG_ST_154		
Other PICS				
Initial condition		The simulated manager and the agent under test are in the operating state.		
Test procedure		<ol style="list-style-type: none"> 1. If the attribute that is going to be changed is reported in a Fixed format event report, take some measurements with the agent under test. 2. Make a change to the contextual attribute Unit-Code for Resistance object (grams to pounds, pounds to grams, grams or pounds to DIMLESS, or DIMLESS to grams or pounds). 3. The agent shall send an MDS event report indicating the new contextual attribute value. 4. Take some more measurements. 5. Wait for the manager to receive new event reports from the agent, which report the measurements from step 4. 		
Pass/Fail criteria		<input type="checkbox"/> The agent sends an MDS event report to inform about the contextual attribute that has been changed. <input type="checkbox"/> Data has changed accordingly to a new contextual attribute.		

Notes				
TP Id		TP/PLT/AG/CLASS/ST/BV-023		
TP label		Config Changes Service. Repetition Contextual Attribute.		
Coverage	Spec	[IEEE 11073-10442]		
	Testable items	NumObj 1; M		
	Spec	[b-CDG 2010]		
	Testable items	Communication 8; M		
Test purpose		Check that: Whenever a contextual attribute changes, the Agent shall report these changes to the Manager using an MDS object event prior to reporting any of the dependent values.		
Applicability		C_AG_OXP_000 AND C_AG_OXP_175 AND C_AG_ST_050 AND C_AG_ST_155		
Other PICS				
Initial condition		The simulated manager and the agent under test are in the operating state.		
Test procedure		<ol style="list-style-type: none"> 1. If the attribute that is going to be changed is reported in a Fixed format event report, take some measurements with the agent under test. 2. Make a change to the contextual attribute Unit-Code for Repetition object (meters to inches or inches to meters.) 3. The agent shall send an MDS event report indicating the new contextual attribute value. 4. Take some more measurements. 5. Wait for the manager to receive new event reports from the agent, which report the measurements from step 4. 		
Pass/Fail criteria		<input type="checkbox"/> The agent sends an MDS event report to inform about the contextual attribute that has been changed. <input type="checkbox"/> Data has changed accordingly to new contextual attribute.		
Notes				

TP Id		TP/PLT/AG/CLASS/ST/BV-024		
TP label		Operating State. Manager to Agent Maximum APDU Size		
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	CommonCharac 3; M		
Test purpose		Check that: The total size of the response do not exceed of the maximum APDU size established by the specialization		
Applicability		C_AG_OXP_000 AND C_AG_OXP_175		
Other PICS		C_AG_OXP_041, C_AG_OXP_100		

Initial condition	The simulated manager and the agent are in the operating state.
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager issues a "Remote Operation Invoke Get" command with: <ol style="list-style-type: none"> a. Obj-handle set to 0 (to request for MDS object) b. attribute-id-list.count = 4087 c. attribute-id-list: (MDC_ATTR_ID_MODEL, MDC_ATTR_SYS_ID, MDC_ATTR_DEV_CONFIG_ID) repeated 1362 times followed by an additional MDC_ATTR_ID_MODEL 2. Check the response of the agent. 3. The simulated manager issues a "Remote Operation Invoke Get" command with the handle set to 0 (to request for an MDS object) and an empty attribute-id-list to indicate all attributes. 4. Check the response of the agent.
Pass/Fail criteria	<ul style="list-style-type: none"> <input type="checkbox"/> In step 2, the agent under test may respond with a rors-cmip-get listing all the requested attributes, or with a roer message. If PICS C_AG_OXP_100 =TRUE and the agent does not respond with a rors-cmip-get message, it responds with a roer message or rorj (resource-limitation) message, a WARNING will appear. <ul style="list-style-type: none"> ○ If the response is a get response, the total size of the response cannot exceed the sum of the APDU sizes of the supported specializations (limited to an absolute limit of 64512 octets): <ul style="list-style-type: none"> ▪ Pulse oximeter -> 9216 octets ▪ Weighing scales -> 896 octets ▪ Glucose meter -> 5120 octets or 64512 octets if the agent supports PM-Store ▪ Blood pressure -> 896 octets ▪ Thermometer -> 896 octets ▪ Independent activity hub -> 5120 octets ▪ Cardiovascular -> 64512 octets or 6624 octets if the agent under test only supports Step Counter Profile ▪ Strength -> 64512 octets: ▪ Adherence monitor -> 1024 octets ▪ Peak flow -> 2030 octets ▪ Body composition analyser -> 7730 octets ▪ Basic ECG/Simple ECG -> 7168 octets or 64512 octets if agent supports PM-Store ▪ Basic ECG/Heart rate -> 1280 octets or 64512 octets if the agent supports PM-Store ▪ International normalized ratio -> 896 octets or 64512 if the agent supports PM-Store ○ In case it responds with a roer, the reason must not be protocol-violation (23) <input type="checkbox"/> In step 4, the agent must respond with a rors-cmip-get message.
Notes	

Bibliography

- [b-ITU-T H.810 (2013)] Recommendation ITU-T H.810 (2013), *Interoperability design guidelines for personal health systems*.
- [b-CDG 1.0] Continua Health Alliance, Continua Design Guidelines v1.0. (2008), *Continua Design Guidelines*.
- [b-CDG 2010] Continua Health Alliance, Continua Design Guidelines v1.5 (2010), *Continua Design Guidelines*.
- [b-CDG 2011] Continua Health Alliance, Continua Design Guidelines (2011), *Adrenaline, Continua Design Guidelines*.
- [b-CDG 2012] Continua Health Alliance, Continua Design Guidelines (2012), *Catalyst, Continua Design Guidelines*.
- [b-ETSI SR 001 262] ETSI SR 001 262 v1.8.1 (2003), *ETSI drafting rules*.
- [b-ETSI 300 406] ETSI ETS 300 406, *Methods for Testing and Specifications (MTS); Protocol and profile conformance testing specifications; Standardization methodology*.
- [b-ISO/IEC 9646-1] ISO/IEC 9646-1, *Information Technology – Open Systems Interconnection – Conformance testing methodology and framework – Part 1: General concepts*.
- [b-ISO/IEC 9646-7] ISO/IEC 9646-7, *Information technology – Open Systems Interconnection – Conformance testing methodology and framework – Part 7: Implementation Conformance Statements*.
- [b-ISO/IEEE 11073-10408] ISO/IEEE 11073-10408-2010, *Health informatics – Personal health device communication – Part 10408: Device specialization – Thermometer*.

SERIES OF ITU-T RECOMMENDATIONS

Series A	Organization of the work of ITU-T
Series D	General tariff principles
Series E	Overall network operation, telephone service, service operation and human factors
Series F	Non-telephone telecommunication services
Series G	Transmission systems and media, digital systems and networks
Series H	Audiovisual and multimedia systems
Series I	Integrated services digital network
Series J	Cable networks and transmission of television, sound programme and other multimedia signals
Series K	Protection against interference
Series L	Environment and ICTs, climate change, e-waste, energy efficiency; construction, installation and protection of cables and other elements of outside plant
Series M	Telecommunication management, including TMN and network maintenance
Series N	Maintenance: international sound programme and television transmission circuits
Series O	Specifications of measuring equipment
Series P	Terminals and subjective and objective assessment methods
Series Q	Switching and signalling
Series R	Telegraph transmission
Series S	Telegraph services terminal equipment
Series T	Terminals for telematic services
Series U	Telegraph switching
Series V	Data communication over the telephone network
Series X	Data networks, open system communications and security
Series Y	Global information infrastructure, Internet protocol aspects and next-generation networks, Internet of Things and smart cities
Series Z	Languages and general software aspects for telecommunication systems