

ITU-T

TELECOMMUNICATION
STANDARDIZATION SECTOR
OF ITU

H.845.9

(01/2015)

SERIES H: AUDIOVISUAL AND MULTIMEDIA SYSTEMS

E-health multimedia services and applications –
Interoperability compliance testing of personal health
systems (HRN, PAN, LAN, TAN and WAN)

**Conformance of ITU-T H.810 personal health
devices: PAN/LAN/TAN interface Part 5I:
Medication adherence monitor: Agent**

Recommendation ITU-T H.845.9

ITU-T H-SERIES RECOMMENDATIONS
AUDIOVISUAL AND MULTIMEDIA SYSTEMS

CHARACTERISTICS OF VISUAL TELEPHONE SYSTEMS	H.100–H.199
INFRASTRUCTURE OF AUDIOVISUAL SERVICES	
General	H.200–H.219
Transmission multiplexing and synchronization	H.220–H.229
Systems aspects	H.230–H.239
Communication procedures	H.240–H.259
Coding of moving video	H.260–H.279
Related systems aspects	H.280–H.299
Systems and terminal equipment for audiovisual services	H.300–H.349
Directory services architecture for audiovisual and multimedia services	H.350–H.359
Quality of service architecture for audiovisual and multimedia services	H.360–H.369
Telepresence	H.420–H.429
Supplementary services for multimedia	H.450–H.499
MOBILITY AND COLLABORATION PROCEDURES	
Overview of Mobility and Collaboration, definitions, protocols and procedures	H.500–H.509
Mobility for H-Series multimedia systems and services	H.510–H.519
Mobile multimedia collaboration applications and services	H.520–H.529
Security for mobile multimedia systems and services	H.530–H.539
Security for mobile multimedia collaboration applications and services	H.540–H.549
Mobility interworking procedures	H.550–H.559
Mobile multimedia collaboration inter-working procedures	H.560–H.569
BROADBAND, TRIPLE-PLAY AND ADVANCED MULTIMEDIA SERVICES	
Broadband multimedia services over VDSL	H.610–H.619
Advanced multimedia services and applications	H.620–H.629
Ubiquitous sensor network applications and Internet of Things	H.640–H.649
IPTV MULTIMEDIA SERVICES AND APPLICATIONS FOR IPTV	
General aspects	H.700–H.719
IPTV terminal devices	H.720–H.729
IPTV middleware	H.730–H.739
IPTV application event handling	H.740–H.749
IPTV metadata	H.750–H.759
IPTV multimedia application frameworks	H.760–H.769
IPTV service discovery up to consumption	H.770–H.779
Digital Signage	H.780–H.789
E-HEALTH MULTIMEDIA SERVICES AND APPLICATIONS	
Interoperability compliance testing of personal health systems (HRN, PAN, LAN, TAN and WAN)	H.820–H.859
Multimedia e-health data exchange services	H.860–H.869

For further details, please refer to the list of ITU-T Recommendations.

Recommendation ITU-T H.845.9

Conformance of ITU-T H.810 personal health devices: PAN/LAN/TAN interface Part 5I: Medication adherence monitor: Agent

Summary

Recommendation ITU-T H.845.9 is a transposition of Continua Test Tool DG2013, Test Suite Structure & Test Purposes, PAN-LAN-TAN Interface; Part 5I: Device Specializations. Agent (Adherence Monitor) (Version 1.4, 2014-01-24), that was developed by the Continua Health Alliance. A number of versions of this specification existed before transposition.

This Recommendation includes an electronic attachment with the protocol implementation conformance statements (PICS) and the protocol implementation extra information for testing (PIXIT) required for the implementation of Annex A

History

Edition	Recommendation	Approval	Study Group	Unique ID*
1.0	ITU-T H.845.9	2015-01-13	16	11.1002/1000/12270

* To access the Recommendation, type the URL <http://handle.itu.int/> in the address field of your web browser, followed by the Recommendation's unique ID. For example, <http://handle.itu.int/11.1002/1000/11830-en>.

FOREWORD

The International Telecommunication Union (ITU) is the United Nations specialized agency in the field of telecommunications, information and communication technologies (ICTs). The ITU Telecommunication Standardization Sector (ITU-T) is a permanent organ of ITU. ITU-T is responsible for studying technical, operating and tariff questions and issuing Recommendations on them with a view to standardizing telecommunications on a worldwide basis.

The World Telecommunication Standardization Assembly (WTSA), which meets every four years, establishes the topics for study by the ITU-T study groups which, in turn, produce Recommendations on these topics.

The approval of ITU-T Recommendations is covered by the procedure laid down in WTSA Resolution 1.

In some areas of information technology which fall within ITU-T's purview, the necessary standards are prepared on a collaborative basis with ISO and IEC.

NOTE

In this Recommendation, the expression "Administration" is used for conciseness to indicate both a telecommunication administration and a recognized operating agency.

Compliance with this Recommendation is voluntary. However, the Recommendation may contain certain mandatory provisions (to ensure, e.g., interoperability or applicability) and compliance with the Recommendation is achieved when all of these mandatory provisions are met. The words "shall" or some other obligatory language such as "must" and the negative equivalents are used to express requirements. The use of such words does not suggest that compliance with the Recommendation is required of any party.

INTELLECTUAL PROPERTY RIGHTS

ITU draws attention to the possibility that the practice or implementation of this Recommendation may involve the use of a claimed Intellectual Property Right. ITU takes no position concerning the evidence, validity or applicability of claimed Intellectual Property Rights, whether asserted by ITU members or others outside of the Recommendation development process.

As of the date of approval of this Recommendation, ITU had not received notice of intellectual property, protected by patents, which may be required to implement this Recommendation. However, implementers are cautioned that this may not represent the latest information and are therefore strongly urged to consult the TSB patent database at <http://www.itu.int/ITU-T/ipr/>.

© ITU 2015

All rights reserved. No part of this publication may be reproduced, by any means whatsoever, without the prior written permission of ITU.

Table of Contents

	Page
1 Scope.....	1
2 References.....	2
3 Definitions	2
3.1 Terms defined elsewhere	2
3.2 Terms defined in this Recommendation.....	2
4 Abbreviations and acronyms	2
5 Conventions	3
6 Test suite structure (TSS)	4
7 Electronic attachment	6
Annex A – Test purposes (TPs).....	7
A.1 TP definition conventions.....	7
A.1 Subgroup 1.3.9: Adherence monitor (AM)	8
Bibliography.....	34

Electronic attachment: Protocol implementation conformance statements (PICS) and protocol implementation extra information for testing (PIXIT) required for the implementation of Annex A.

Introduction

This Recommendation is a transposition of Continua Test Tool DG2013, Test Suite Structure & Test Purposes, PAN-LAN-TAN Interface; Part 5I: Device Specializations. Agent (Adherence Monitor) (Version 1.4, 2014-01-24), that was developed by the Continua Health Alliance. A number of versions of this specification existed before transposition and these can be found in the table below.

Version	Date	Revision history
1.2	2012-10-05	Initial release for Test Tool DG2011. This is the same version as "TSS&TP_1.5_PAN-LAN_PART_5I_v1.2.doc" because new features included in [b-CDG 2011] do not affect the test procedures specified in this document.
1.3	2013-05-24	Initial release for Test Tool DG2012. This uses "TSS&TP_DG2011_PAN-LAN_PART_5I_v1.2.doc" as a baseline and adds new features included in [b-CDG 2012]: Max APDU size for GM, BCA and ECG.
1.4	2014-01-24	Initial release for Test Tool DG2013. This uses "TSS&TP_DG2012_PAN-LAN_PART_5I_v1.4.doc" as a baseline and adds new features included in [ITU-T H.810]: <ul style="list-style-type: none">• Adds glucose meter BLE• Adds BLE SSP support• Adds NFC new transport• Adds INR device specialization

Recommendation ITU-T H.845.9

Conformance of ITU-T H.810 personal health devices: PAN/LAN/TAN interface Part 5I: Medication adherence monitor: Agent

1 Scope

The scope of this Recommendation¹ is to provide a test suite structure and the test purposes (TSS & TP) for the PAN/LAN/TAN interface based on the requirements defined in the Continua Design Guidelines (CDG) [ITU-T H.810]. The objective of this test specification is to provide a high probability of air interface interoperability between different devices.

The TSS and TP for the PAN/LAN/TAN interface document have been divided into ten parts. Each part is listed below:

- **Part 1:** Optimized exchange protocol [ISO/IEEE 11073-20601A] Agent
- **Part 2:** Optimized exchange protocol [ISO/IEEE 11073-20601A] Manager
- **Part 3:** Continua design guidelines. Agent
- **Part 4:** Continua design guidelines. Manager
- **Part 5:** Device specializations. Agent. This document is divided into 14 subparts:
 - **Part 5A:** Weighing scales
 - **Part 5B:** Glucose meter
 - **Part 5C:** Pulse oximeter
 - **Part 5D:** Blood pressure monitor
 - **Part 5E:** Thermometer
 - **Part 5F:** Cardiovascular fitness and activity monitor
 - **Part 5G:** Strength fitness equipment
 - **Part 5H:** Independent living activity hub
 - **Part 5I:** Adherence monitor
 - **Part 5J:** Insulin pump (Future development)
 - **Part 5K:** Peak flow
 - **Part 5L:** Body composition analyser
 - **Part 5M:** Basic electrocardiograph
 - **Part 5N:** International normalized ratio monitor
- **Part 6:** Device specializations. Manager
- **Part 7:** Continua design guidelines. Agent BLE
- **Part 8:** Continua design guidelines. Manager BLE
- **Part 9:** Personal health devices transcoding whitepaper. Agent
- **Part 10:** Personal health devices transcoding whitepaper. Manager

¹ This Recommendation includes an electronic attachment with the protocol implementation conformance statements (PICS) and the protocol implementation extra information for testing (PIXIT) required for the implementation of Annex A.

2 References

The following ITU-T Recommendations and other references contain provisions which, through reference in this text, constitute provisions of this Recommendation. At the time of publication, the editions indicated were valid. All Recommendations and other references are subject to revision; users of this Recommendation are therefore encouraged to investigate the possibility of applying the most recent edition of the Recommendations and other references listed below. A list of the currently valid ITU-T Recommendations is regularly published. The reference to a document within this Recommendation does not give it, as a stand-alone document, the status of a Recommendation.

- [ITU-T H.810] Recommendation ITU-T H.810 (2013), *Interoperability design guidelines for personal health systems*.
- [ISO/IEEE 11073-20601A] ISO/IEEE 11073-20601:2010, *Health informatics – Personal health device communication – Part 20601: Application profile – Optimized exchange protocol*, including ISO/IEEE 11073-20601:2010 Amd 1:2015.
<http://www.iso.org/iso/home/store/catalogue_tc/catalogue_detail.htm?csnumber=54331>
with
<http://www.iso.org/iso/home/store/catalogue_tc/catalogue_detail.htm?csnumber=63972>
- [ISO/IEEE 11073-104xx] ISO/IEEE 11073-104xx (in force), *Health informatics – Personal health device communication – Device specialization*.
NOTE – Shorthand is used to refer to the collection of device specialization standards that utilize [ISO/IEEE 11073-20601A], where xx can be any number from 01 to 99, inclusive.
- [ISO/IEEE 11073-10472] ISO/IEEE 11073-10472-2012, *Health informatics – Personal health device communication – Part 10472: Device specialization – Medication monitor*.

3 Definitions

3.1 Terms defined elsewhere

This Recommendation uses the following terms defined elsewhere:

3.1.1 agent [ISO/IEEE 11073-20601A]: A node that collects and transmits personal health data to an associated manager.

3.1.2 manager [ISO/IEEE 11073-20601A]: A node receiving data from one or more agent systems. Some examples of managers include a cellular phone, health appliance, set top box, or a computer system.

3.2 Terms defined in this Recommendation

None.

4 Abbreviations and acronyms

This Recommendation uses the following abbreviations and acronyms:

ATS	Abstract Test Suite
DUT	Device Under Test
CDG	Continua Design Guidelines
GUI	Graphical User Interface

INR	International Normalized Ratio
IUT	Implementation Under Test
MDS	Medical Device System
NFC	Near Field Communication
PAN	Personal Area Network
PCT	Protocol Conformance Testing
PCO	Point of Control and Observation
PHD	Personal Healthcare Device
PHDC	Personal Healthcare Device Class
PHM	Personal Health Manager
PICS	Protocol Implementation Conformance Statement
PIXIT	Protocol Implementation extra Information for Testing
SDP	Service Discovery Protocol
SOAP	Simple Object Access Protocol
TCRL	Test Case Reference List
TCWG	Test and Certification Working Group
TP	Test Purpose
TSS	Test Suite Structure
USB	Universal Serial Bus
WDM	Windows Driver Model

5 Conventions

The key words "SHALL", "SHALL NOT", "SHOULD", "SHOULD NOT", "MAY", "MAY NOT" in this document are to be interpreted as in [b-ETSI SR 001 262].

- SHALL is equivalent to 'must' or 'it is required to'.
- SHALL NOT is equivalent to 'must not' or 'it is not allowed'.
- SHOULD is equivalent to 'it is recommended to'.
- SHOULD NOT is equivalent to 'it is not recommended to'.
- MAY is equivalent to 'is permitted'.
- MAY NOT is equivalent to 'it is not required that'.

NOTE – The above-mentioned key words are capitalized for illustrative purposes only and they do not appear capitalized within this Recommendation.

Reference is made in the ITU-T H.800-series of Recommendations to different versions of the Continua Design Guidelines (CDG) by a specific designation. The list of terms that may be used in this Recommendation is provided in Table 1.

Table 1 – List of designations associated with the various versions of the CDG.

CDG name	Transposed as	Version	Description	Designation
2013 plus errata	ITU-T H.810	4.1	CDG 2013 plus errata noting all ratified bugs.	–
2013	–	4.0	Release 2013 of the CDG including maintenance updates of CDG 2012 and additional guidelines that cover new functionalities.	Endorphin
2012 plus errata	–	3.1	CDG 2012 plus errata noting all ratified bugs [b-CDG 2012].	–
2012	–	3.0	Release 2012 of the CDG including maintenance updates of CDG 2011 and additional guidelines that cover new functionalities.	Catalyst
2011 plus errata	–	2.1	CDG 2011 integrated with identified errata.	–
2011	–	2.0	Release 2011 of the CDG including maintenance updates of CDG 2010 and additional guidelines that cover new functionalities [b-CDG 2011].	Adrenaline
2010 plus errata	–	1.6	CDG 2010 integrated with identified errata	–
2010	–	1.5	Release 2010 of the CDG with maintenance updates of CDG Version 1 and additional guidelines that cover new functionalities [b-CDG 2010].	1.5
1.0	–	1.0	First released version of the CDG [b-CDG 1.0].	–

6 Test suite structure (TSS)

The test purposes (TPs) for the PAN/LAN/TAN interface have been divided into the main subgroups specified below. Annex A describes the TPs for subgroup 1.3.9 (shown in bold).

- Group 1: Agent (AG)
 - Group 1.1: Transport (TR)
 - Subgroup 1.1.1: Design guidelines: Common (DGC)
 - Subgroup 1.1.2: USB design guidelines (UDG)
 - Subgroup 1.1.3: Bluetooth design guidelines (BDG)
 - Subgroup 1.1.4: Pulse oximeter design guidelines (PODG)
 - Subgroup 1.1.5: Cardiovascular design guidelines (CVDG)
 - Subgroup 1.1.6: Activity hub design guidelines (HUBDG)
 - Subgroup 1.1.7: ZigBee design guidelines (ZDG)
 - Subgroup 1.1.8: Glucose meter design guidelines (GLDG)
 - Subgroup 1.1.9: Bluetooth low energy design guidelines (BLEDG)
 - Subgroup 1.1.10: Basic electrocardiograph design guidelines (ECGDG)
 - Subgroup 1.1.11: NFC design guidelines (NDG)
 - Group 1.2: Optimized exchange protocol (OXP)
 - Subgroup 1.2.1: PHD domain information model (DIM)

- Subgroup 1.2.2: PHD service model (SER)
- Subgroup 1.2.3: PHD communication model (COM)
- Group 1.3: Devices class specializations (CLASS)
 - Subgroup 1.3.1: Weighing scales (WEG)
 - Subgroup 1.3.2: Glucose meter (GL)
 - Subgroup 1.3.3: Pulse oximeter (PO)
 - Subgroup 1.3.4: Blood pressure monitor (BPM)
 - Subgroup 1.3.5: Thermometer (TH)
 - Subgroup 1.3.6: Cardiovascular (CV)
 - Subgroup 1.3.7: Strength (ST)
 - Subgroup 1.3.8: Activity hub (HUB)
 - **Subgroup 1.3.9: Adherence monitor (AM)**
 - Subgroup 1.3.10: Insulin pump (IP) (Future development)
 - Subgroup 1.3.11: Peak flow (PF)
 - Subgroup 1.3.12: Body composition analyser (BCA)
 - Subgroup 1.3.13: Basic electrocardiograph (ECG)
 - Subgroup 1.3.14: International normalized ratio (INR)
- Group 1.4: Personal health device transcoding whitepaper (PHDTW)
 - Subgroup 1.4.1: Whitepaper general requirements (GEN)
 - Subgroup 1.4.2: Whitepaper thermometer requirements (TH)
 - Subgroup 1.4.3: Whitepaper blood pressure requirements (BPM)
 - Subgroup 1.4.4: Whitepaper heart rate requirements (HR)
 - Subgroup 1.4.5: Whitepaper glucose meter requirements (GL)
- Group 2: Manager (MAN)
 - Group 2.1: Transport (TR)
 - Subgroup 2.1.1: Design guidelines: common (DGC)
 - Subgroup 2.1.2: USB design guidelines (UDG)
 - Subgroup 2.1.3: Bluetooth design guidelines (BDG)
 - Subgroup 2.1.4: Cardiovascular design guidelines (CVDG)
 - Subgroup 2.1.5: Activity hub design guidelines (HUBDG)
 - Subgroup 2.1.6: ZigBee design guidelines (ZDG)
 - Subgroup 2.1.7: Bluetooth low energy design guidelines (BLEDG)
 - Subgroup 2.1.8: NFC design guidelines (NDG)
 - Group 2.2: 20601: Optimized exchange protocol (OXP)
 - Subgroup 2.2.1: General (GEN)
 - Subgroup 2.2.2: PHD domain information model (DIM)
 - Subgroup 2.2.3: PHD service model (SER)
 - Subgroup 2.2.4: PHD communication model (COM)
 - Group 2.3: Devices class specializations (CLASS)
 - Subgroup 2.3.1: Weighing scales (WEG)
 - Subgroup 2.3.2: Glucose meter (GL)

- Subgroup 2.3.3: Pulse oximeter (PO)
- Subgroup 2.3.4: Blood pressure monitor (BPM)
- Subgroup 2.3.5: Thermometer (TH)
- Subgroup 2.3.6: Cardiovascular (CV)
- Subgroup 2.3.7: Strength (ST)
- Subgroup 2.3.8: Activity hub (HUB)
- Subgroup 2.3.9: Adherence monitor (AM)
- Subgroup 2.3.10: Insulin pump (IP) (Future development)
- Subgroup 2.3.11: Peak flow (PF)
- Subgroup 2.3.12: Body composition analyser (BCA)
- Subgroup 2.3.13: Basic electrocardiograph (ECG)
- Subgroup 2.3.14: International normalized ratio (INR)
- Group 2.4: Personal health device transcoding whitepaper (PHDTW)
 - Subgroup 2.4.1: Whitepaper general requirements (GEN)
 - Subgroup 2.4.2: Whitepaper thermometer requirements (TH)
 - Subgroup 2.4.3: Whitepaper blood pressure measurement requirements (BPM)
 - Subgroup 2.4.4: Whitepaper heart rate requirements (HR)
 - Subgroup 2.4.5: Whitepaper glucose meter requirements (GL)

7 Electronic attachment

The protocol implementation conformance statements (PICS) and the protocol implementation extra information for testing (PIXIT) required for the implementation of this Annex can be downloaded from <http://handle.itu.int/11.1002/2000/12067>.

In the electronic attachment, letters "C" and "I" in the column labelled "Mandatory" are used to distinguish between "PICS" and "PIXIT" respectively during testing. If the cell is empty, the corresponding PICS is "independent". If the field contains a "C", the corresponding PICS is dependent on other PICS, and the logical expression is detailed in the "SCR_Expression" field. The static conformance review (SCR) is used in the test tool to assert whether the PICS selection is consistent.

Annex A

Test purposes (TPs)

(This annex forms an integral part of this Recommendation.)

A.1 TP definition conventions

The test purposes are defined according to the following rules:

- **TP Id:** This is a unique identifier (TP/<TT>/<DUT>/<GR>/<SGR>/<XX> – <NNN>). It is specified according to the naming convention defined below:
 - Each test purpose identifier is introduced by the prefix "TP".
 - <TT>: This is the test tool that will be used in the test case:
 - PAN: Personal area network (Bluetooth or USB)
 - LAN: Local area network (ZigBee)
 - PAN-LAN: Personal area network (Bluetooth or USB) - Local area network (ZigBee)
 - LP-PAN: Low power personal area network (Bluetooth low energy)
 - TAN: Touch area network (NFC)
 - PLT: Personal area network (Bluetooth or USB) – Local area network (ZigBee) – Touch area network (NFC)
 - <DUT>: This is the device under test:
 - AG: PAN/LAN Agent
 - MAN: PAN/LAN Manager
 - <GR>: This identifies a group of test cases.
 - <SGR>: This identifies a subgroup of test cases.
 - <XX>: This identifies the type of testing:
 - BV: Valid behaviour test
 - BI: Invalid behaviour test
 - <NNN>: This is a sequential number that identifies the test purpose.
- **TP label:** This is the TP's title.
- **Coverage:** This contains the specification reference and clause to be checked by the TP.
 - Spec: This indicates the earliest version of the specification from which the testable items to be checked by the TP were included.
 - Testable item: This contains testable items to be checked by the TP.
- **Test purpose:** This is a description of the requirements to be tested.
- **Applicability:** This contains the PICS items that define if the test case is applicable or not for a specific device. When a TP contains an "ALL" in this field it means that it applies to the device under test within that scope of the test (specialization, transport used, etc.).
- **Initial condition:** This indicates the state to which the DUT needs to be moved at the beginning of TC execution.
- **Test procedure:** This describes the steps to be followed in order to execute the test case.
- **Pass/Fail criteria:** This provides criteria to decide whether the DUT passes or fails the test case.

A.1 Subgroup 1.3.9: Adherence monitor (AM)

TP Id		TP/PLT/AG/CLASS/AM/BV-000		
TP label		Get MDS Object for Adherence Monitor specialization: Mandatory, Conditional and Optional Attributes		
Coverage	Spec	[ISO/IEEE 11073-10472]		
	Testable items	MM_MDSAttr1; M	MM_MDSAttr2; M	MM_MDSAttr3; M
		MM_MDSAttr4; R	MM_MDSAttr5; R	MM_MDSAttr6; R
		MM_MDSAttr7; M	MM_MDSAttr8; M	MM_GETServ1; M
		MM_GETServ4; M	MM_OperProc2; M	
Applicability		C_AG_OXP_168 AND C_AG_OXP_000		
Initial condition		The simulated manager and the agent under test are in the operating state.		
Test procedure		<p>1. The simulated manager issues a "roiv-cmip-get" command with the handle set to 0 (to request for an MDS object) and the attribute-id-list set to 0 to indicate all attributes.</p> <p>2. The agent responds with a "rors-cmip-get" service message in which the attribute-list contains a list of all implemented attributes of the MDS object:</p> <p>MDS Attributes:</p> <ol style="list-style-type: none"> a. Mandatory attribute System-model <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_MODEL <input type="checkbox"/> attribute-type = SystemModel <input type="checkbox"/> attribute-value.length = <variable> <input type="checkbox"/> attribute-value = {Manufacturer, Model} b. Mandatory attribute Dev-Configuration-Id <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_DEV_CONFIG_ID <input type="checkbox"/> attribute-type = ConfigId <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <ul style="list-style-type: none"> - IF C_AG_AM_001 then attribute-value = 0x1C 0x20 - ELSE IF C_AG_AM_002 then attribute-value = 0x1C 0x21 - ELSE IF C_AG_AM_003 then attribute-value = 0x1C 0x22 - ELSE IF C_AG_AM_004 then attribute-value = 0x1C 0x23 - ELSE attribute-value = < between 0x4000 and 0x7FFF c. Recommended attribute Power-Status <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_POWER_STAT <input type="checkbox"/> attribute-type = PowerStatus (BITS-16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <ul style="list-style-type: none"> ON_MAINS (0x8000) or ON_BATTERY(0x4000) <p>Only one of the following may be active:</p> <ul style="list-style-type: none"> ▪ chargingFull(8), ▪ chargingTrickle(9), 		

	<ul style="list-style-type: none"> ▪ chargingOff(10). ▪ The rest of the bits must not be set <p>d. Recommended attribute Battery-Level</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_VAL_BATT_CHARGE <input type="checkbox"/> attribute-type = INT-U16 <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <undefined if value>100 > <p>e. Recommended attribute Remaining-Battery-Time</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_BATT_REMAIN <input type="checkbox"/> attribute-type = BatMeasure <input type="checkbox"/> attribute-value.length = <variable> <input type="checkbox"/> attribute-value = <units shall be set to one of: MDC_DIM_MIN, MDC_DIM_HR, MDC_DIM_DAY > <p>f. Mandatory attribute System-Type-Spec_List</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SYS_TYPE_SPEC_LIST <input type="checkbox"/> attribute-type = TypeVerList <input type="checkbox"/> attribute-value.length = 4 bytes attribute-value = MDC_DEV_SPEC_PROFILE_AI_MED_MINDER, 1 <input type="checkbox"/> Attribute System-Type must not be present.
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/AG/CLASS/AM/BV-001			
TP label	MDS Configuration objects events for Adherence Monitor			
Coverage	Spec	[ISO/IEEE 11073-10472]		
	Testable items	MedDispensed1; M	StatusRep1; O	UserFeedback1; O
		MM_StandConfig1; C	MM_StandConfig2;C	MM_StandConfig3;C
		MM_StandConfig4;C	MM_MDSEvent1; M	MM_GenNumObj1;M
		MM_GenNumObj2: O	FixedDosage1; M	VarDosage1; M
		StatReporter1; O	MM_EventRepServ1; M	MM_ConfProc1; M
Applicability	C_AG_OXP_168 AND C_AG_OXP_000			
Initial condition	The simulated manager and the agent under test are in the unassociated state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager: <ol style="list-style-type: none"> a. APDU Type <ul style="list-style-type: none"> <input type="checkbox"/> field- type = PrstApdu <input type="checkbox"/> field-length =2 bytes 			

- field-value =0xE7 0x00
- b. invoke-id
 - field- type = InvokeIDType
 - field-length =INT-U16
 - field- value=<Not relevant for this test>
- c. message
 - field- type = roiv-cmip-confirmed-event-report
 - field-length =two bytes
 - field- value=0x01 0x01 (EventReportArgumentSimple)
- d. obj-handle (EventReportArgumentSimple)
 - field- type = HANDLE
 - field-length =INT-U16
- e. event-time (EventReportArgumentSimple)
 - field- type = Relative Time
 - field-length =INT-U32
 - field-value =
 - IF NOT C_AG_OXP_010 THEN value = 0xFF 0xFF 0xFF 0xFF
- f. event-type (EventReportArgumentSimple)
 - field- type = OID-Type
 - field-length =INT-U16
 - field- value=0x 0D 0x 1C (MDC_NOTI_CONFIG)
- g. config-report-id (ConfigReport)
 - field- type = ConfigId
 - field-length = INT-U16
 - field- value =
 - IF C_AG_AM_001 then attribute-value = 0x1C 0x20
 - ELSE IF C_AG_AM_002 then attribute-value = 0x1C 0x21
 - ELSE IF C_AG_AM_003 then attribute-value = 0x1C 0x22
 - ELSE IF C_AG_AM_004 then attribute-value = 0x1C 0x23
 - ELSE IF C_AG_OXP_181=TRUE then attribute-value = < between 0x4000 and 0x7FFF >
- h. obj-class (ConfigReport → ConfigObjectList (ConfigObject)). To check the objects that are supported by the agent, Type Attribute will be checked in AttributeList.
 - field- type = OID-Type
 - field-length = INT-U16
 - field- value =
 - IF C_AG_AM_001 then 1 Fixed Dosage Medication object is present.
 - ELSE IF C_AG_AM_002 then 1 Fixed Dosage Medication, 1 Status Reporter and 1 User Feedback object are present.
 - ELSE IF C_AG_AM_003 then 1 Variable Dosage Medication object is present.
 - ELSE IF C_AG_AM_004 then 1 Variable Dosage Medication, 1 Status Reporter and 1 User Feedback object are present.

	<ul style="list-style-type: none"> • ELSE : <ul style="list-style-type: none"> ▪ IF C_AG_AM_005 then 1 Fixed Dosage Medication is present, ELSE this object is not present. ▪ IF C_AG_AM_006 then 1 Variable Dosage Medication is present, ELSE this object is not present. ▪ Exactly one of the fixed dosage medication dispensed numeric object or the variable dosage medication dispensed numeric object shall be supported. ▪ IF C_AG_AM_007 then User Feedback is present, ELSE this object is not present. ▪ IF C_AG_AM_008 then Status Reporter is present, ELSE this object is not present.
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/AG/CLASS/AM/BV-002			
TP label	MDS objects events Adherence Monitor & PM-Store Object			
Coverage	Spec	[ISO/IEEE 11073-10472]		
	Testable items	MM_MDSEvent2; M	MM_MDSEvent3; M	MM_MDSEvent4; M
		MM_MDSEvent5; M	MM_MDSEvent6; M	MM_MDSEvent7; M
		MM_MDSEvent8; M	MM_MDSEvent9; M	MM_MDSEvent10; M
		MM_EventRepServ1; M	MM_OperProc11; M	MM_OperProc12; M
		MM_PMStoreGen1; M	MM_PMStoreGen2; M	MM_EventRepServ2; M
		MM_OperProc5; M	MM_OperProc6; M	
Applicability	C_AG_OXP_168 AND C_AG_OXP_000			
Initial condition	The simulated manager and the agent under test are in the anassociated state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. Check that the field Dev-Config-Id is set to the tested configuration. If it is not, the manager responds with an "unsupported-config" and waits for a new configuration. Repeat this step until a Dev-config-Id equal to the tested configuration is received. 5. Record the agent configuration. 6. Take Measurements for every supported object in the agent under test. 7. Wait to receive every event report and check: <ul style="list-style-type: none"> • IF the agent does not support PM-Store, THEN MDS-Event Report is sent by the agent to report the measurements. • IF the agent supports PM-Store, THEN the agent shall not send MDS event reports. • For MDS Event Reports: <ul style="list-style-type: none"> <input type="checkbox"/> field- type = Event Report <input type="checkbox"/> field-length = 2 bytes 			

	<ul style="list-style-type: none"> <input type="checkbox"/> field- value=0x01 0x01 (EventReportArgumentSimple, confirmed) This field identifies the type of message sent by the agent, for the confirmed event configuration, roiv-cmip-confirmed-event-report.
Pass/Fail criteria	<ul style="list-style-type: none"> • Check that every received MDS Event report is a one of the following Data APDU and that it is confirmed. • For a Standard Configuration (C_AG_AM_001 or C_AG_AM_002 or C_AG_AM_003 or C_AG_AM_004): the MDS Event Report is sent by the agent to report measurements for every object. <ul style="list-style-type: none"> <input type="checkbox"/> MDC_NOTI_SCAN_REPORT_FIXED <input type="checkbox"/> MDC_NOTI_SCAN_REPORT_MP_FIXED • For an Extended Configuration that does not support the PM-Store object, an MDS Event Report is sent by the agent to report measurements for every object: <ul style="list-style-type: none"> <input type="checkbox"/> MDC_NOTI_SCAN_REPORT_FIXED <input type="checkbox"/> MDC_NOTI_SCAN_REPORT_MP_FIXED <input type="checkbox"/> MDC_NOTI_SCAN_REPORT_VAR <input type="checkbox"/> MDC_NOTI_SCAN_REPORT_MP_VAR • For an Extended Configuration that supports the PM-Store object, an MDS Event Report is not sent by the agent to report measurements for objects.
Notes	

TP Id	TP/PLT/AG/CLASS/AM/BV-003			
TP label	Fixed Dosage Medication Dispensed Object for Standard Configuration (0x1C20 or 0x1C21)			
Coverage	Spec	[ISO/IEEE 11073-10472]		
	Testable items	FixedDosage2; M	FixedDosage3; M	FixedDosage4; R
		FixedDosage5; M	FixedDosage6; R	FixedDosage7; O
		FixedDosage8; R	FixedDosage9; R	FixedDosage10; R
		FixedDosage11; R	FixedDosage12; M	FixedDosage13; R
		FixedDosage14; O	FixedDosage15; O	FixedDosage16; C
		FixedDosage17; R	FixedDosage18; C	FixedDosage19; R
		FixedDosage20; C	FixedDosage21; C	FixedDosage22; C
		FixedDosage23; C	FixedDosage24; C	FixedDosage25; C
		FixedDosage26; R	FixedDosage39; M	MM_ConfProc2; M
Applicability	C_AG_OXP_168 AND (C_AG_AM_001 OR C_AG_AM_002) AND C_AG_OXP_000			
Initial condition	The simulated manager and the agent under test are in the unassociated state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. Check that the field Dev-Config-Id is set to 0x1C20 OR 0x1C21. If it is not, the manager responds with an "unsupported-config" and waits for a new configuration. Repeat this step until a Dev-config-Id equal to 0x1C20 or 0x1C21 is received. 			

	<p>5. Once the agent under test sends a standard configuration, check the Fixed Dosage Medication object.</p> <p>6. The Fixed Dosage Medication object contents shall be:</p> <p>a. Mandatory attribute Handle</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_HANDLE <input type="checkbox"/> attribute-type = HANDLE <input type="checkbox"/> attribute-value = 0x00 0x01 <p>b. Mandatory attribute Type</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = MDC_PART_PHD_AI, MDC_AI_MED_DISPENSED_FIXED <p>c. Mandatory attribute Metric-Spec-Small</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value ≠ 0x00 0x00 <ul style="list-style-type: none"> • Bit 0 (mss-avail-intermittent(0)) must be set. • Bit 1 (mss-avail-stored-data(1)) must be set. • Bit 2 (mss-upd-aperiodic(2)) must be set. • Bit 3 (mss-msmt-aperiodic(3)) is set. • Bit 9 (mss-acc-agent-initiated(9)) is set. <p>d. Mandatory attribute Attribute-Value-Map</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ATTRIBUTE_VAL_MAP <input type="checkbox"/> attribute-type = AttrValMap <input type="checkbox"/> attribute-count = 2 <input type="checkbox"/> attribute-value = (MDC_ATTR_TIME_STAMP_ABS ,8 MDC_ATTR_NU_VAL_OBS_BASIC,2) <p>7. Check that no other attributes are present in the initial configuration.</p>
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/AG/CLASS/AM/BV-004			
TP label	Fixed Dosage Medication Dispensed Object for Extended Configuration			
Coverage	Spec	[ISO/IEEE 11073-10472]		
	Testable items	FixedDosage27; M	FixedDosage28; R	FixedDosage29; R
		FixedDosage30; O	FixedDosage31; R	FixedDosage32; R
		FixedDosage33; R	FixedDosage34; R	FixedDosage35; R
		FixedDosage36; R	FixedDosage37; R	FixedDosage38; R
Applicability	C_AG_OXP_168 AND C_AG_OXP_181 AND C_AG_AM_005 AND C_AG_OXP_000			
Initial condition	The simulated manager and the agent under test are in the unassociated state.			

<p>Test procedure</p>	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. Check that the field Dev-Config-Id is set to the tested extended configuration. If it is not, the manager responds with an "unsupported-config" and waits for a new configuration. Repeat this step until a Dev-config-Id equal to the extended configuration is received. 5. Once the agent under test sends the tested configuration, check the Fixed Dosage Medication object. 6. The Fixed Dosage Medication object contents shall be: <ol style="list-style-type: none"> a. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE attribute-value = = MDC_PART_PHD_AI, MDC_AI_MED_DISPENSED_FIXED b. IF Not Recommended attribute Supplemental-Types <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SPPLEMENTAL_TYPES <input type="checkbox"/> attribute-type = SupplementalTypeList <input type="checkbox"/> attribute-value.length = <variable>Sequence of TYPE (TYPE.length= 4 bytes) <input type="checkbox"/> attribute-value = <Not relevant for this test> c. IF Not recommended attribute Metric-Structure-Small is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_STRUCTURE_SMALL <input type="checkbox"/> attribute-type = MetricStructureSmall <input type="checkbox"/> attribute-length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> d. IF Optional attribute Measurement-Status is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_MSMT_STAT <input type="checkbox"/> attribute-type = MeasurementStatus <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> e. IF Not recommended attribute Metric-Id is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_PHYSIO <input type="checkbox"/> attribute-type = OID-Type(INT-U16) <input type="checkbox"/> attribute-value.length =2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> f. IF Not Recommended attribute Metric-Id-List is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_PHYSIO_LIS <input type="checkbox"/> attribute-type = MetricIdList <input type="checkbox"/> attribute-value = <Not relevant for this test> g. IF Not recommended attribute Metric-Id-Partition is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_ID_PART <input type="checkbox"/> attribute-type = NomPartition(INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> h. IF Not recommended attribute Unit-Code <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE
------------------------------	---

	<ul style="list-style-type: none"> <input type="checkbox"/> attribute-type = OID-Type(INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> i. IF Not recommended attribute Source-Handle-Reference is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF <input type="checkbox"/> attribute-type = HANDLE(INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> j. IF Not recommended attribute Relative-Time-Stamp <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_STAMP_REL <input type="checkbox"/> attribute-type = RelativeTime (INT-U32) <input type="checkbox"/> attribute-value.length =4 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> k. IF Not recommended attribute Measure-Active-Period <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_PD_MSMT_ACTIVE <input type="checkbox"/> attribute-type = FLOAT-Type (INT-U32) <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> l. IF Not Recommended attribute Accuracy is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_ACCUR_MSMT <input type="checkbox"/> attribute-type = FLOAT-Type (INT-U32) <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test>
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/AG/CLASS/AM/BV-005		
TP label	Variable Dosage Medication Dispensed Object for Standard Configuration (0x1C22 or 0x1C23)		
Coverage	Spec	[ISO/IEEE 11073-10472]	
	Testable items	VarDosage2; M	VarDosage3; M
		VarDosage4; R	
		VarDosage5; M	VarDosage6; R
		VarDosage7; O	
		VarDosage8; R	VarDosage9; R
		VarDosage10; R	
		VarDosage11; M	VarDosage12; M
		VarDosage13; R	
		VarDosage14; O	VarDosage15; O
		VarDosage16; C	
		VarDosage17; R	VarDosage18; C
		VarDosage19; R	
		VarDosage20; C	VarDosage21; C
		VarDosage22; C	
		VarDosage23; C	VarDosage24; C
		VarDosage25; C	
		VarDosage26; R	VarDosage39; M
		MM_ConfProc2; M	

Applicability	C_AG_OXP_168 AND (C_AG_AM_003 OR C_AG_AM_004) AND C_AG_OXP_000
Initial condition	The simulated manager and the agent under test are in the unassociated state.
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. Check that the field Dev-Config-Id is set to 0x1C22 OR 0x1C23. If it is not, the manager responds with an "unsupported-config" and waits for a new configuration. Repeat this step until a Dev-config-Id equal to 0x1C22 or 0x1C23 is received. 5. Once the agent under test sends a standard configuration, check the Variable Dosage Medication object. 6. The Variable Dosage Medication object contents shall be: <ol style="list-style-type: none"> a. Mandatory attribute Handle <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_HANDLE <input type="checkbox"/> attribute-type = HANDLE <input type="checkbox"/> attribute-value = 0x00 0x02 b. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = MDC_PART_PHD_AI, MDC_AI_MED_DISPENSED_VARIABLE c. Mandatory attribute Metric-Spec-Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value ≠ 0x00 0x00 <ul style="list-style-type: none"> • Bit 0 (mss-avail-intermittent(0)) must be set. • Bit 1 (mss-avail-stored-data(1)) must be set. • Bit 2 (mss-upd-aperiodic(2)) must be set. • Bit 3 (mss-msmt-aperiodic(3)) is set. • Bit 9 (mss-acc-agent-initiated(9)) is set. d. Mandatory attribute Unit-Code <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = MDC_DIM_MILLI_L e. Mandatory attribute Attribute-Value-Map <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ATTRIBUTE_VAL_MAP <input type="checkbox"/> attribute-type = AttrValMap <input type="checkbox"/> attribute-count = 2 <input type="checkbox"/> attribute-value = (MDC_ATTR_TIME_STAMP_ABS ,8 MDC_ATTR_NU_VAL_OBS_SIMP,4) 7. Check that no other attributes are present in the initial configuration.
Pass/Fail criteria	All checked values are as specified in the test procedure.

Notes	
--------------	--

TP Id	TP/PLT/AG/CLASS/AM/BV-006			
TP label	Variable Dosage Medication Dispensed Object for Extended Configuration			
Coverage	Spec	[ISO/IEEE 11073-10472]		
	Testable items	VarDosage27; M	VarDosage28; R	VarDosage29; R
		VarDosage30; O	VarDosage31; R	VarDosage32; R
		VarDosage33; R	VarDosage34; M	VarDosage35; R
		VarDosage36; R	VarDosage37; R	VarDosage38; R
Applicability	C_AG_OXP_168 AND C_AG_OXP_181 AND C_AG_AM_006 AND C_AG_OXP_000			
Initial condition	The simulated manager and the agent under test are in the unassociated state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. Check that the field Dev-Config-Id is set to the tested extended configuration. If it is not, the manager responds with an "unsupported-config" and waits for a new configuration. Repeat this step until a Dev-config-Id equal to the extended configuration is received. 5. Once the agent under test sends the tested configuration, check the Variable Dosage Medication object. 6. The Variable Dosage Medication object contents shall be: <ol style="list-style-type: none"> a. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE attribute-value = MDC_PART_PHD_AI, MDC_AI_MED_DISPENSED_VARIABLE b. IF Not Recommended attribute Supplemental-Types <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SPPLEMENTAL_TYPES <input type="checkbox"/> attribute-type = SupplementalTypeList <input type="checkbox"/> attribute-value.length = <variable>Sequence of TYPE (TYPE.length= 4 bytes) <input type="checkbox"/> attribute-value = <Not relevant for this test> c. IF Not recommended attribute Metric-Structure-Small is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_STRUCTURE_SMALL <input type="checkbox"/> attribute-type = MetricStructureSmall <input type="checkbox"/> attribute-length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> d. IF Optional attribute Measurement-Status is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_MSMT_STAT <input type="checkbox"/> attribute-type = MeasurementStatus <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> e. IF Not recommended attribute Metric-Id is present 			

	<ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_PHYSIO <input type="checkbox"/> attribute-type = OID-Type(INT-U16) <input type="checkbox"/> attribute-value.length =2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> f. IF Not Recommended attribute Metric-Id-List is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_PHYSIO_LIS <input type="checkbox"/> attribute-type = MetricIdList <input type="checkbox"/> attribute-value = <Not relevant for this test> g. IF Not recommended attribute Metric-Id-Partition is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_ID_PART <input type="checkbox"/> attribute-type = NomPartition(INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> h. Mandatory recommended attribute Unit-Code <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type(INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> i. IF Not recommended attribute Source-Handle-Reference is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF <input type="checkbox"/> attribute-type = HANDLE(INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> j. IF Not recommended attribute Relative-Time-Stamp <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_STAMP_REL <input type="checkbox"/> attribute-type = RelativeTime (INT-U32) <input type="checkbox"/> attribute-value.length =4 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> k. IF Not recommended attribute Measure-Active-Period <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_PD_MSMT_ACTIVE <input type="checkbox"/> attribute-type = FLOAT-Type (INT-U32) <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> l. IF Not Recommended attribute Accuracy is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_ACCUR_MSMT <input type="checkbox"/> attribute-type = FLOAT-Type (INT-U32) <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test>
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id		TP/PLT/AG/CLASS/AM/BV-007		
TP label		User Feedback Object for Standard Configuration (0x1C21 or 0x1C23)		
Coverage	Spec	[ISO/IEEE 11073-10472]		
	Testable items	UserFeedback2; M	UserFeedback3; M	UserFeedback4; R
		UserFeedback5; M	UserFeedback6; R	UserFeedback7; O
		UserFeedback8; R	UserFeedback9; M	UserFeedback10 ;R
		UserFeedback11 ;R	UserFeedback12 ;M	UserFeedback13 ;R
		UserFeedback14 ;O	UserFeedback15 ;O	UserFeedback16 ;C
		UserFeedback17 ;R	UserFeedback18 ;C	UserFeedback19 ;R
		UserFeedback20 ;C	UserFeedback21 ;C	UserFeedback22 ;C
		UserFeedback23 ;C	UserFeedback24 ;C	UserFeedback25 ;C
		UserFeedback26 ;R	UserFeedback38; M	MM_ConfProc2; M
Applicability		C_AG_OXP_168 AND (C_AG_AM_002 OR C_AG_AM_004) AND C_AG_OXP_000		
Initial condition		The simulated manager and the agent under test are in the unassociated state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. Check that the field Dev-Config-Id is set to 0x1C21 OR 0x1C23. If it is not, the manager responds with an "unsupported-config" and waits for a new configuration. Repeat this step until a Dev-config-Id equal to 0x1C21 or 0x1C23 is received. 5. Once the agent under test sends a standard configuration, check the User Feedback object: 6. The User Feedback object contents shall be: <ol style="list-style-type: none"> a. Mandatory attribute Handle <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_HANDLE <input type="checkbox"/> attribute-type = HANDLE <input type="checkbox"/> attribute-value = 0x00 0x04 b. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = MDC_PART_PHD_AI, MDC_AI_MED_FEEDBACK c. Mandatory attribute Metric-Spec-Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value ≠ 0x00 0x00 <ul style="list-style-type: none"> • Bit 0 (mss-avail-intermittent(0)) must be set. • Bit 1 (mss-avail-stored-data(1)) must be set. • Bit 2 (mss-upd-aperiodic(2)) must be set. 		

	<ul style="list-style-type: none"> • Bit 3 (mss-msmt-aperiodic(3)) is set. • Bit 9 (mss-acc-agent-initiated(9)) is set. • Bit 12 (mss-cat-manual(12)) is set. <p>d. Mandatory attribute Metric-Id-List is present</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_PHYSIO_LIS <input type="checkbox"/> attribute-type = MetricIdList <input type="checkbox"/> attribute-value = MDC_AI_MED_UF_LOCATION, MDC_AI_MED_UF_RESPONSE <p>e. Mandatory attribute Attribute-Value-Map</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ATTRIBUTE_VAL_MAP <input type="checkbox"/> attribute-type = AttrValMap <input type="checkbox"/> attribute-count = 2 <input type="checkbox"/> attribute-value = (MDC_ATTR_TIME_STAMP_ABS ,8 MDC_ATTR_NU_CMPD_VAL_OBS_BASIC,8) <p>7. Check that no other attributes are present in the initial configuration.</p>
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/AG/CLASS/AM/BV-007_A		
TP label	User Feedback Object format for Standard Configuration		
Coverage	Spec	[ISO/IEEE 11073-10472]	
	Testable items	UserFeedback39; M	UserFeedback23; C
Applicability	C_AG_OXP_168 AND (C_AG_AM_002 OR C_AG_AM_004) AND C_AG_OXP_000		
Initial condition	The simulated manager and the agent under test are in the unassociated state.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The Agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. Check that the field Dev-Config-Id is set to 0x1C21 OR 0x1C23. If it is not, the manager responds with an "unsupported-config" and waits for a new configuration. Repeat this step until a Dev-config-Id equal to 0x1C21 or 0x1C23 is received. 5. Once the agent under test sends the tested configuration, the simulated manager sends a "roiv-cmip-get" to get all the attributes of the MDS, record the value of Date-and-Time. 6. Once the agent under test is in the operating state, take a measurement and record the value of the measurement. 7. Wait until the agent under test sends an Event Report to the simulated manager, the relevant fields are: <ol style="list-style-type: none"> a. event-type = MDC_NOTI_SCAN_REPORT_FIXED b. ScanReportInfoFixed <ul style="list-style-type: none"> <input type="checkbox"/> obj-handle = 4 <input type="checkbox"/> Compound Object Count =2 <input type="checkbox"/> obs-val-data.value = 		

	<ul style="list-style-type: none"> • Time Stamp (8 bytes) • Location (2 bytes) • Response (2 bytes)
Pass/Fail criteria	<ul style="list-style-type: none"> • The received data must be coherent with that previously recorded. • The Time Stamp must be coherent with that received in the MDS attribute. • The data must be received in this exact same order and the Compound value contains two fields, the first one that represents the "relative location" and the second one that is the user response (numeric form).
Notes	

TP Id	TP/PLT/AG/CLASS/AM/BV-008			
TP label	User Feedback Object for Extended Configuration			
Coverage	Spec	[ISO/IEEE 11073-10472]		
	Testable items	UserFeedback27; M	UserFeedback28; R	UserFeedback29; R
		UserFeedback30; O	UserFeedback31; R	UserFeedback32; R
		UserFeedback33; R	UserFeedback34; R	UserFeedback35; R
		UserFeedback36; R	UserFeedback37; R	
Applicability	C_AG_OXP_168 AND C_AG_OXP_181 AND C_AG_AM_007 AND C_AG_OXP_000			
Initial condition	The simulated manager and the agent under test are in the unassociated state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. Check that the field Dev-Config-Id is set to the tested extended configuration. If it is not, the manager responds with an "unsupported-config" and waits for a new configuration. Repeat this step until a Dev-config-Id equal to the extended configuration is received. 5. Once the agent under test sends the tested configuration, check User Feedback object: 6. User Feedback Object contents shall be: <ol style="list-style-type: none"> a. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = MDC_PART_PHD_AI, MDC_AI_MED_FEEDBACK b. IF Not Recommended attribute Supplemental-Types <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SPPLEMENTAL_TYPES <input type="checkbox"/> attribute-type = SupplementalTypeList <input type="checkbox"/> attribute-value.length = <variable>Sequence of TYPE (TYPE.length= 4 bytes) <input type="checkbox"/> attribute-value = <Not relevant for this test> c. IF Not recommended attribute Metric-Structure-Small is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_STRUCTURE_SMALL <input type="checkbox"/> attribute-type = MetricStructureSmall <input type="checkbox"/> attribute-length = 2 bytes 			

	<ul style="list-style-type: none"> <input type="checkbox"/> attribute-value = <Not relevant for this test> d. IF Optional attribute Measurement-Status is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_MSMT_STAT <input type="checkbox"/> attribute-type = MeasurementStatus <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> e. IF Not recommended attribute Metric-Id is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_PHYSIO <input type="checkbox"/> attribute-type = OID-Type(INT-U16) <input type="checkbox"/> attribute-value.length =2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> f. IF Not recommended attribute Metric-Id-Partition is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_ID_PART <input type="checkbox"/> attribute-type = NomPartition(INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> g. IF Not recommended attribute Unit-Code <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_UNIT_CODE <input type="checkbox"/> attribute-type = OID-Type(INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> h. IF Not recommended attribute Source-Handle-Reference is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SOURCE_HANDLE_REF <input type="checkbox"/> attribute-type = HANDLE(INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> i. IF Not recommended attribute Relative-Time-Stamp <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_STAMP_REL <input type="checkbox"/> attribute-type = RelativeTime (INT-U32) <input type="checkbox"/> attribute-value.length =4 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> j. IF Not recommended attribute Measure-Active-Period <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_PD_MSMT_ACTIVE <input type="checkbox"/> attribute-type = FLOAT-Type (INT-U32) <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> k. IF Not Recommended attribute Accuracy is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_NU_ACCUR_MSMT <input type="checkbox"/> attribute-type = FLOAT-Type (INT-U32) <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test>
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id		TP/PLT/AG/CLASS/AM/BV-009		
TP label		Status Reporter Object for Standard Configuration (0x1C21 or 0x1C23)		
Coverage	Spec	[ISO/IEEE 11073-10472]		
	Testable items	StatReporter2; M	StatReporter3; M	StatReporter4; R
		StatReporter5; M	StatReporter6; R	StatReporter7; O
		StatReporter8; R	StatReporter9; R	StatReporter10; R
		StatReporter11; R	StatReporter12; M	StatReporter13; R
		StatReporter14; O	StatReporter15; O	StatReporter16; C
		StatReporter17; R	StatReporter18; C	StatReporter19; R
		StatReporter20; R	StatReporter21; C	StatReporter22; C
		StatReporter23; C	StatReporter24; C	StatReporter25; C
		StatReporter26; R	StatReporter27; O	StatReporter44; M
	MM_ConfProc2; M			
Applicability		C_AG_OXP_168 AND (C_AG_AM_002 OR C_AG_AM_004) AND C_AG_OXP_000		
Initial condition		The simulated manager and the agent under test are in the unassociated state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. Check that the field Dev-Config-Id is set to 0x1C21 OR 0x1C23. If it is not, the manager responds with an "unsupported-config" and waits for a new configuration. Repeat this step until a Dev-config-Id equal to 0x1C21 or 0x1C23 is received. 5. Once the agent under test sends a standard configuration, check the Status Reporter object 6. The Status Reporter object contents shall be: <ol style="list-style-type: none"> a. Mandatory attribute Handle <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_HANDLE <input type="checkbox"/> attribute-type = HANDLE <input type="checkbox"/> attribute-value = 0x00 0x03 b. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = MDC_PART_PHD_AI, MDC_AI_MED_STATUS c. Mandatory attribute Metric-Spec-Small <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_SPEC_SMALL <input type="checkbox"/> attribute-type = MetricSpecSmall (BITS-16) <input type="checkbox"/> attribute-value ≠ 0x00 0x00 <ul style="list-style-type: none"> • Bit 0 (mss-avail-intermittent(0)) must be set. • Bit 1 (mss-avail-stored-data(1)) must be set. 		

	<ul style="list-style-type: none"> • Bit 2 (mss-upd-aperiodic(2)) must be set. • Bit 3 (mss-msmt-aperiodic(3)) is set. • Bit 9 (mss-acc-agent-initiated(9)) is set. <p>d. Mandatory attribute Attribute-Value-Map</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ATTRIBUTE_VAL_MAP <input type="checkbox"/> attribute-type = AttrValMap <input type="checkbox"/> attribute-count = 2 <input type="checkbox"/> attribute-value = (MDC_ATTR_TIME_STAMP_ABS ,8 MDC_ATTR_ENUM_OBS_VAL_BASIC_BIT_STR,2) <p>7. Check that no other attributes are present in the initial configuration.</p>
Pass/Fail criteria	All checked values are as specified in the test procedure.

TP Id		TP/PLT/AG/CLASS/AM/BV-010		
TP label		Status Reporter Object for Extended Configuration		
Coverage	Spec	[ISO/IEEE 11073-10472]		
	Testable items	StatReporter28; M	StatReporter29; R	StatReporter30; R
		StatReporter31; O	StatReporter32; R	StatReporter33; R
		StatReporter34; R	StatReporter35; R	StatReporter36; R
		StatReporter37; R	StatReporter38; R	StatReporter39; R
		StatReporter40; R	StatReporter41; O	StatReporter42; O
		StatReporter43; M		
Applicability		C_AG_OXP_168 AND C_AG_OXP_181 AND C_AG_AM_008 AND C_AG_OXP_000		
Initial condition		The simulated manager and the agent under test are in the unassociated state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated manager receives an association request from the agent under test. 2. The simulated manager responds with a result = accepted-unknown-config. 3. The agent responds with a "Remote Operation Invoke Confirmed Event Report" message with an MDC_NOTI_CONFIG event to send its configuration to the manager. 4. Check that the field Dev-Config-Id is set to extended configuration. If it is not, the manager responds with an "unsupported-config" and waits for a new configuration. Repeat this step until a Dev-config-Id equal to tested extended configuration is received. 5. Once the agent under test sends the tested configuration, check the Status Reporter object. 6. The Status Reporter object contents shall be: <ol style="list-style-type: none"> a. Mandatory attribute Type <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_ID_TYPE <input type="checkbox"/> attribute-type = TYPE <input type="checkbox"/> attribute-value = MDC_PART_PHD_AI, MDC_AI_MED_STATUS b. IF Not Recommended attribute Supplemental-Types <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SPPLEMENTAL_TYPES <input type="checkbox"/> attribute-type = SupplementalTypeList <input type="checkbox"/> attribute-value.length = <variable>Sequence of TYPE (TYPE.length= 4 bytes) 		

- attribute-value = <Not relevant for this test>
- c. IF Not recommended attribute Metric-Structure-Small is present
 - attribute-id = MDC_ATTR_METRIC_STRUCTURE_SMALL
 - attribute-type = MetricStructureSmall
 - attribute-length = 2 bytes
 - attribute-value = <Not relevant for this test>
- d. IF Optional attribute Measurement-Status is present
 - attribute-id = MDC_ATTR_MSMT_STAT
 - attribute-type = MeasurementStatus
 - attribute-value.length = 2 bytes
 - attribute-value = <Not relevant for this test>
- e. IF Not recommended attribute Metric-Id is present
 - attribute-id = MDC_ATTR_ID_PHYSIO
 - attribute-type = OID-Type(INT-U16)
 - attribute-value.length = 2 bytes
 - attribute-value = <Not relevant for this test>
- f. IF Not Recommended attribute Metric-Id-List is present
 - attribute-id = MDC_ATTR_ID_PHYSIO_LIS
 - attribute-type = MetricIdList
 - attribute-value = <Not relevant for this test>
- g. IF Not recommended attribute Metric-Id-Partition is present
 - attribute-id = MDC_ATTR_METRIC_ID_PART
 - attribute-type = NomPartition(INT-U16)
 - attribute-value.length = 2 bytes
 - attribute-value = <Not relevant for this test>
- h. IF Not recommended attribute Unit-Code is present
 - attribute-id = MDC_ATTR_UNIT_CODE
 - attribute-type = OID-Type(INT-U16)
 - attribute-value.length = 2 bytes
 - attribute-value = <Not relevant for this test>
- i. IF Not recommended attribute Source-Handle-Reference is present
 - attribute-id = MDC_ATTR_SOURCE_HANDLE_REF
 - attribute-type = HANDLE(INT-U16)
 - attribute-value.length = 2 bytes
 - attribute-value = <Not relevant for this test>
- j. IF Not recommended attribute Relative-Time-Stamp is present
 - attribute-id = MDC_ATTR_TIME_STAMP_REL
 - attribute-type = RelativeTime (INT-U32)
 - attribute-value.length = 4 bytes
 - attribute-value = <Not relevant for this test>
- k. IF Not recommended attribute Measure-Active-Period is present
 - attribute-id = MDC_ATTR_TIME_PD_MSMT_ACTIVE
 - attribute-type = FLOAT-Type (INT-U32)
 - attribute-value.length = 4 bytes

	<ul style="list-style-type: none"> <input type="checkbox"/> attribute-value = <Not relevant for this test> i. IF Not recommended attribute Enum-Observed-Value-Simple-OID is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id= MDC_ATTR_ENUM_OBS_VAL_SIM_OID <input type="checkbox"/> attribute-type = OID-Type (INT-U16) <input type="checkbox"/> attribute-value.length = 2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> m. IF Not recommended attribute Enum-Observed-Value-Partition is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id= MDC_ATTR_ENUM_OBS_VAL_PART <input type="checkbox"/> attribute-type = NomPartition (INT-U16) <input type="checkbox"/> attribute-value-length=2 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> n. IF Optional attribute Context-Key is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_CONTEXT_KEY <input type="checkbox"/> attribute-type = OCTET STRING(Size(8)) <input type="checkbox"/> attribute-value.length =10 bytes <input type="checkbox"/> attribute-value = Check against PIXIT(I_AG_OXP_009)
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/AG/CLASS/AM/BV-011			
TP label	Adherence Monitor PM-Store object			
Coverage	Spec	[ISO/IEEE 11073-10472]		
	Testable items	MM_PMStoreAttr1; M	MM_PMStoreAttr2; M	MM_PMStoreAttr3; M
		MM_PMStoreAttr4; M		
Applicability	C_AG_OXP_168 AND C_AG_OXP_041 AND C_AG_OXP_000			
Initial condition	The simulated manager and the agent under test are in the operating state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager issues a "Remote Operation Invoke Get" command with the handle set to the PM-Store and the attribute-id-list set to 0 to indicate all attributes. 2. The agent response must contain: <ol style="list-style-type: none"> a. Mandatory Storage-Capacity-Count is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_STORE_CAPAC_CNT <input type="checkbox"/> attribute-type = INT-U32 <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant in this test> b. Mandatory Storage-Usage-Count is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_METRIC_STORE_USAGE_CNT <input type="checkbox"/> attribute-type = INT-U32 <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant in this test> c. Mandatory attribute PM-Store-Label 			

	<ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_PM_STORE_LABEL_STRING <input type="checkbox"/> attribute-type = OCTET STRING <input type="checkbox"/> attribute-value.length <= 255 octets <input type="checkbox"/> attribute-value = <Not relevant for this Test> <p>d. IF Not recommended attribute Sample-Period is present</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_PD_SAMP <input type="checkbox"/> attribute-type = RelativeTime <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant in this test>
Pass/Fail criteria	All checked values are as specified in the test procedure.
Notes	

TP Id	TP/PLT/AG/CLASS/AM/BV-012		
TP label	Adherence Monitor Segment-Data-Event size		
Coverage	Spec	[ISO/IEEE 11073-10472]	
	Testable items	MM_PMStoreEvent3; M	MM_PMStoreEvent4; M
Applicability	C_AG_OXP_168 AND C_AG_OXP_041 AND C_AG_OXP_000		
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager issues a Get-Segment-Info with SegmSelection set to all-segments. 2. The simulated manager sends a request for the PM-Segment Data to one of the PM-Segments that contains data: <ol style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_ACT_SEG_TRIG_XFER <input type="checkbox"/> TrigSegmDataXferReq = <Instance number of the selected PM-Segment that contains the data> 3. The agent issues an action response: <ol style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action, <input type="checkbox"/> HANDLE = obj-handle b. Action = MDC_ACT_SEG_TRIG_XFER <ul style="list-style-type: none"> <input type="checkbox"/> TrigSegmDataXferRsp = <Same Instance number> tsxr-succesful (0x00 0x00) 4. The agent under test starts Data transfer: <ol style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Invoke CfmEventReport <input type="checkbox"/> Action = MDC_NOTI_SEGMENT_DATA <input type="checkbox"/> SegmentDataEvent 5. The simulated manager response to transferred data APDU's: 		

	<ul style="list-style-type: none"> a. Data APDU <ul style="list-style-type: none"> <input type="checkbox"/> Type = Invoke Confirmed Action,, <input type="checkbox"/> HANDLE = obj-handle <input type="checkbox"/> Action = MDC_NOTI_SEGMENT_DATA <input type="checkbox"/> SegmentDataResult <input type="checkbox"/> Agent under test repeats steps 3 and 4 until all the data is transferred
Pass/Fail criteria	The size of the Segment-Data-Event (MDC_NOTI_SEGMENT_DATA) cannot exceed 1024 octets.
Notes	

TP Id	TP/PLT/AG/CLASS/AM/BV-013			
TP label	Adherence Monitor PM-Segment			
Coverage	Spec	[ISO/IEEE 11073-10472]		
	Testable items	MM_PStoreModel1; M	MM_PStoreModel2; M	MM_PMSegmAttr1; M
		MM_PMSegmAttr2; M	MM_PMSegmAttr3; M	MM_PMSegmAttr4; M
Applicability	C_AG_OXP_168 AND C_AG_OXP_041 AND C_AG_OXP_000			
Initial condition	The simulated manager and the agent under test are in the operating state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager shall send a Get-Segment-Info action for a PM-Store object with SegmSelection = all-segments to indicate the PM-Segments attributes of all available PM-Segments. 2. The agent issues a "rors-cmip-confirmed-action" response with the PM-Segment attributes it supports: <ol style="list-style-type: none"> a. Mandatory attribute PM-Segment-Entry-Map <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_PM_SEG_MAP <input type="checkbox"/> attribute-type = PmSegmentEntryMap <input type="checkbox"/> attribute-value = SEQUENCE of <ul style="list-style-type: none"> ▪ segm-entry-header shall include one of the time formats. ▪ segm-entry-elem-list: Record this value b. Mandatory attribute Segment-Label is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_PM_SEG_LABEL_STRING <input type="checkbox"/> attribute-type = OCTET STRING <input type="checkbox"/> attribute-value.length = consistent with value <input type="checkbox"/> attribute-value = <Not relevant for this test> c. Mandatory attribute Segment-Start-Abs-Time is present <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_START_SEG <input type="checkbox"/> attribute-type = AbsoluteTime <input type="checkbox"/> attribute-value.length = 8 bytes <input type="checkbox"/> attribute-value = <ul style="list-style-type: none"> ▪ century = ▪ year ≤ 99 ▪ month ≤ 12 			

	<ul style="list-style-type: none"> ▪ day ≤ 31 ▪ hour ≤ 24 ▪ minute ≤ 60 ▪ second ≤ 60 ▪ sec-fractions ≤ 100 <p>d. Mandatory attribute Segment-End-Abs-Time is present</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_TIME_END_SEG <input type="checkbox"/> attribute-type = AbsoluteTime <input type="checkbox"/> attribute-value.length = 8 bytes <input type="checkbox"/> attribute-value = <ul style="list-style-type: none"> ▪ century = ▪ year ≤ 99 ▪ month ≤ 12 ▪ day ≤ 31 ▪ hour ≤ 24 ▪ minute ≤ 60 ▪ second ≤ 60 ▪ sec-fractions ≤ 100 <p>e. Mandatory attribute Segment-Usage-Count is present</p> <ul style="list-style-type: none"> <input type="checkbox"/> attribute-id = MDC_ATTR_SEG_USAGE_CNT <input type="checkbox"/> attribute-type = INT-U32 <input type="checkbox"/> attribute-value.length = 4 bytes <input type="checkbox"/> attribute-value = <Not relevant for this test> <p>3. Repeat step 1 and 2 for every PM-Store object</p>
Pass/Fail criteria	All checked attributes are as specified in the test procedure. There is at least one segment for every implemented object.
Notes	

TP Id	TP/PLT/AG/CLASS/AM/BV-014		
TP label	Association Adherence Monitor Agent		
Coverage	Spec	[ISO/IEEE 11073-10472]	
	Testable items	MM_AssocReq1; M	MM_AssocReq2 ; M
		MM_AssocReq4 ; M	MM_AssocReq5 ; M
		MM_AssocReq7 ; M	MM_AssocReq8 ; M
		MM_AssocReq10 ; M	MM_AssocReq11 ; M
		MM_MDSMethod4 ; M	
Applicability	C_AG_OXP_168 AND C_AG_OXP_000		
Initial condition	The simulated manager and the agent under test are in the unassociated state.		

Test procedure

1. The agent sends a message to associate to the simulated manager, the expected fields sent by the agent are:
 - a. APDU Type
 - field- type = AarqApdu
 - field-length =2 bytes
 - field-value =0xE2 0x00.
 - b. assoc-version
 - field- type = AssociationVersion
 - field-length =BITS-32
 - field- value=0x80 0x00 0x00 0x00
 - c. data-proto-id
 - field- type = DataProtold(INT-U16)
 - field-length =2 bytes
 - field- value=0x50 0x79 (20601)
 - d. protocol-version
 - field- type = Protocol Version
 - field-length = 4 bytes
 - field- value=0x80 0x00 0x00 0x00
 - e. encoding rules
 - field- type = EncodingRules
 - field-length = 2 bytes
 - field- value=
 - Bit 0 must be set (support MDER)
 - Bits 1 and 2 may be set
 - The rest of the bits must be 0
 - f. nomenclature version
 - field- type = NomenclatureVersion
 - field-length = 4 bytes
 - field- value=0x80 0x00 0x00 0x00
 - This value indicates version1 is supported (nom-version1(0) is set).
 - g. functional–units
 - field- type = FunctionalUnits
 - field-length = 4 bytes
 - field-value =
 - Bit 0 must not be set , only bit 1 or 2 may be set to 1.
 - h. System type
 - field- type = SystemType
 - field-length = 4 bytes
 - field- value = 0x00 0x80 0x00 0x00 (sys-type-agent)
 - i. System-Id
 - field- type = OCTET STRING
 - field-length = 8 bytes
 - field- value = 0xXX 0xXX 0xXX 0xXX 0xXX 0xXX 0xXX 0xXX (octet string length = 8 | UI-64 manufacturer and device)

	<ul style="list-style-type: none"> <input type="checkbox"/> This value will be the System Id attribute of the MDS object. j. dev-config-id <ul style="list-style-type: none"> <input type="checkbox"/> field- type = ConfigId(INT-U16) <input type="checkbox"/> field-length = 2 bytes <input type="checkbox"/> field- value = <ul style="list-style-type: none"> ▪ <0x1C20 or 0x1C21 or 0x1C22 or 0x1C23> for standard configuration ▪ <between 0x40 0x00 and 0x7F 0xFF > for extended configuration. k. data-req-mode-flags (DataReqModeCapab) <ul style="list-style-type: none"> <input type="checkbox"/> field- type = DataReqModeFlags <input type="checkbox"/> field-length = 2 bytes <input type="checkbox"/> If the agent supports only Medication Monitor specialization →Only bit 15 is set (data-req-supp-init-agent(15)) l. data-req-init-agent-count (DataReqModeCapab) <ul style="list-style-type: none"> <input type="checkbox"/> field- type = INT-U8 <input type="checkbox"/> field-length = 2 bytes <input type="checkbox"/> field.value = 0x01 m. data-req-init-manager-count (DataReqModeCapab) <ul style="list-style-type: none"> <input type="checkbox"/> field- type = INT-U8 <input type="checkbox"/> field-length = 2 bytes <input type="checkbox"/> field.value = 0x00
Pass/Fail criteria	All checked attributes have proper values.
Notes	

TP Id	TP/PLT/AG/CLASS/AM/BV-015		
TP label	Get Request Adherence Monitor Agent		
Coverage	Spec	[ISO/IEEE 11073-10472]	
	Testable items	MM_OperProc4; M	
Applicability	C_AG_OXP_168 AND C_AG_OXP_000		
Initial condition	The simulated manager and the agent under test are in the operating state.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated manager issues a "Remote Operation Invoke Get" command with <ol style="list-style-type: none"> a. Obj-handle set to 0 (to request for MDS object) b. attribute-id-list.count=1 and a single AVA_Type MDC_ATTR_DEV_CONFIG_ID (0X0A 0X44) to retrieve the mandatory "Dev-Configuration-Id" attribute 2. The agent under test responds with: <ul style="list-style-type: none"> • IF C_AG_OXP_100 THEN: with a "rors-cmip-get" service message which contains the "Dev-Configuration-Id" • ELSE: with a "roer" service message with error-value set to no-such-an-action (9) 		
Pass/Fail criteria	In step 2 the agent properly sends the requested attribute or the error (no-such-action) message.		
Notes			

TP Id		TP/PLT/AG/CLASS/AM/BV-016		
TP label		Operating State. Manager to Agent Maximum APDU Size		
Coverage	Spec	[ISO/IEEE 11073-20601A]		
	Testable items	CommonCharac 3; M		
	Spec	[ISO/IEEE 11073-10472]		
	Testable items	MM_ComModel1; M	MM_ComModel2; M	
Applicability		C_AG_OXP_000 AND C_AG_OXP_168		
Initial condition		The simulated manager and the agent are in the operating state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated manager issues a "Remote Operation Invoke Get" command with: <ol style="list-style-type: none"> a. Obj-handle set to 0 (to request for an MDS object) b. attribute-id-list.count = 23 c. attribute-id-list: (MDC_ATTR_ID_MODEL, MDC_ATTR_SYS_ID, MDC_ATTR_DEV_CONFIG_ID) repeated 7 times followed by an additional MDC_ATTR_ID_MODEL and MDC_ATTR_SYS_ID 2. Check the response of the agent. 3. The simulated manager issues a "Remote Operation Invoke Get" command with the handle set to 0 (to request for an MDS object) and an empty attribute-id-list to indicate all attributes. 4. Check the response of the agent. 		
Pass/Fail criteria		<ul style="list-style-type: none"> • In step 2, the agent under test may respond with a rors-cmip-get listing all the requested attributes, or with a roer message. If PICS C_AG_OXP_100 =TRUE and the agent does not respond with a rors-cmip-get message, it responds with a roer message or rorj(resource-limitation) message, a WARNING will appear. <ul style="list-style-type: none"> ○ If the response is a get response, the total size of the response cannot exceed the sum of the APDU sizes of the supported specializations (limited to an absolute limit of 64512 octets): <ul style="list-style-type: none"> ▪ Pulse oximeter -> 9216 octets ▪ Weighing scales -> 896 octets ▪ Glucose meter -> 5120 octets or 64512 octets if the agent supports PM-Store ▪ Blood pressure -> 896 octets ▪ Thermometer -> 896 octets ▪ Independent activity hub -> 5120 octets ▪ Cardiovascular -> 64512 octets or 6624 octets if agent under test only supports Step Counter Profile ▪ Strength -> 64512 octets: ▪ Adherence monitor -> 1024 octets ▪ Peak flow -> 2030 octets ▪ Body composition analyser -> 7730 octets ▪ Basic ECG/Simple ECG -> 7168 octets or 64512 octets if the agent supports PM-Store ▪ Basic ECG/Heart rate -> 1280 octets or 64512 octets if the agent supports PM-Store 		

	<ul style="list-style-type: none">▪ International normalized ratio -> 896 octets or 64512 if the agent supports PM-Store○ In the case where it responds with a roer, the reason must not be protocol-violation (23)• In step 4, the agent must respond with a rors-cmip-get message.
Notes	

Bibliography

- [b-CDG 1.0] Continua Health Alliance, Continua Design Guidelines v1.0 (2008), *Continua Design Guidelines*.
- [b-CDG 2010] Continua Health Alliance, Continua Design Guidelines v1.5 (2010), *Continua Design Guidelines*.
- [b-CDG 2011] Continua Health Alliance, Continua Design Guidelines (2011), "Adrenalin", *Continua Design Guidelines*.
- [b-CDG 2012] Continua Health Alliance, Continua Design Guidelines (2012), "Catalyst", *Continua Design Guidelines*.
- [b-ETSI SR 001 262] ETSI SR 001 262 v1.8.1 (2003), *ETSI drafting rules*.

SERIES OF ITU-T RECOMMENDATIONS

Series A	Organization of the work of ITU-T
Series D	General tariff principles
Series E	Overall network operation, telephone service, service operation and human factors
Series F	Non-telephone telecommunication services
Series G	Transmission systems and media, digital systems and networks
Series H	Audiovisual and multimedia systems
Series I	Integrated services digital network
Series J	Cable networks and transmission of television, sound programme and other multimedia signals
Series K	Protection against interference
Series L	Construction, installation and protection of cables and other elements of outside plant
Series M	Telecommunication management, including TMN and network maintenance
Series N	Maintenance: international sound programme and television transmission circuits
Series O	Specifications of measuring equipment
Series P	Terminals and subjective and objective assessment methods
Series Q	Switching and signalling
Series R	Telegraph transmission
Series S	Telegraph services terminal equipment
Series T	Terminals for telematic services
Series U	Telegraph switching
Series V	Data communication over the telephone network
Series X	Data networks, open system communications and security
Series Y	Global information infrastructure, Internet protocol aspects and next-generation networks
Series Z	Languages and general software aspects for telecommunication systems