

International Telecommunication Union

ITU-T

TELECOMMUNICATION
STANDARDIZATION SECTOR
OF ITU

H.848

(07/2016)

SERIES H: AUDIOVISUAL AND MULTIMEDIA SYSTEMS

E-health multimedia services and applications –
Interoperability compliance testing of personal health
systems (HRN, PAN, LAN, TAN and WAN)

**Conformance of ITU-T H.810 personal health
devices: PAN/LAN/TAN interface Part 8:
Bluetooth low energy (BLE): Manager**

Recommendation ITU-T H.848



ITU-T H-SERIES RECOMMENDATIONS
AUDIOVISUAL AND MULTIMEDIA SYSTEMS

CHARACTERISTICS OF VISUAL TELEPHONE SYSTEMS	H.100–H.199
INFRASTRUCTURE OF AUDIOVISUAL SERVICES	
General	H.200–H.219
Transmission multiplexing and synchronization	H.220–H.229
Systems aspects	H.230–H.239
Communication procedures	H.240–H.259
Coding of moving video	H.260–H.279
Related systems aspects	H.280–H.299
Systems and terminal equipment for audiovisual services	H.300–H.349
Directory services architecture for audiovisual and multimedia services	H.350–H.359
Quality of service architecture for audiovisual and multimedia services	H.360–H.369
Telepresence	H.420–H.429
Supplementary services for multimedia	H.450–H.499
MOBILITY AND COLLABORATION PROCEDURES	
Overview of Mobility and Collaboration, definitions, protocols and procedures	H.500–H.509
Mobility for H-Series multimedia systems and services	H.510–H.519
Mobile multimedia collaboration applications and services	H.520–H.529
Security for mobile multimedia systems and services	H.530–H.539
Security for mobile multimedia collaboration applications and services	H.540–H.549
Mobility interworking procedures	H.550–H.559
Mobile multimedia collaboration inter-working procedures	H.560–H.569
BROADBAND, TRIPLE-PLAY AND ADVANCED MULTIMEDIA SERVICES	
Broadband multimedia services over VDSL	H.610–H.619
Advanced multimedia services and applications	H.620–H.629
Ubiquitous sensor network applications and Internet of Things	H.640–H.649
IPTV MULTIMEDIA SERVICES AND APPLICATIONS FOR IPTV	
General aspects	H.700–H.719
IPTV terminal devices	H.720–H.729
IPTV middleware	H.730–H.739
IPTV application event handling	H.740–H.749
IPTV metadata	H.750–H.759
IPTV multimedia application frameworks	H.760–H.769
IPTV service discovery up to consumption	H.770–H.779
Digital Signage	H.780–H.789
E-HEALTH MULTIMEDIA SERVICES AND APPLICATIONS	
Personal health systems	H.810–H.819
Interoperability compliance testing of personal health systems (HRN, PAN, LAN, TAN and WAN)	H.820–H.859
Multimedia e-health data exchange services	H.860–H.869

For further details, please refer to the list of ITU-T Recommendations.

Recommendation ITU-T H.848

Conformance of ITU-T H.810 personal health devices: PAN/LAN/TAN interface Part 8: Bluetooth low energy (BLE): Manager

Summary

Recommendation ITU-T H.848 is a transposition of Continua Health Alliance Test Tool DG2013, Test Suite Structure & Test Purposes, PAN-LAN-TAN Interface; Part 8: Continua Design Guidelines. Manager BLE (Version 1.2, 2014-01-24), that was developed by the Continua Health Alliance. A number of versions of this specification existed before transposition.

This Recommendation includes an electronic attachment with the protocol implementation conformance statements (PICS) and the protocol implementation extra information for testing (PIXIT) required for the implementation of Annex A.

History

Edition	Recommendation	Approval	Study Group	Unique ID*
1.0	ITU-T H.848	2015-01-13	16	11.1002/1000/12277
2.0	ITU-T H.848	2016-07-14	16	11.1002/1000/12955

Keywords

Conformance testing, continua design guidelines, e-health, H.810, PAN/LAN/TAN interface, personal area network, personal connected health devices, touch area network.

* To access the Recommendation, type the URL <http://handle.itu.int/> in the address field of your web browser, followed by the Recommendation's unique ID. For example, <http://handle.itu.int/11.1002/1000/11830-en>.

FOREWORD

The International Telecommunication Union (ITU) is the United Nations specialized agency in the field of telecommunications, information and communication technologies (ICTs). The ITU Telecommunication Standardization Sector (ITU-T) is a permanent organ of ITU. ITU-T is responsible for studying technical, operating and tariff questions and issuing Recommendations on them with a view to standardizing telecommunications on a worldwide basis.

The World Telecommunication Standardization Assembly (WTSA), which meets every four years, establishes the topics for study by the ITU-T study groups which, in turn, produce Recommendations on these topics.

The approval of ITU-T Recommendations is covered by the procedure laid down in WTSA Resolution 1.

In some areas of information technology which fall within ITU-T's purview, the necessary standards are prepared on a collaborative basis with ISO and IEC.

NOTE

In this Recommendation, the expression "Administration" is used for conciseness to indicate both a telecommunication administration and a recognized operating agency.

Compliance with this Recommendation is voluntary. However, the Recommendation may contain certain mandatory provisions (to ensure, e.g., interoperability or applicability) and compliance with the Recommendation is achieved when all of these mandatory provisions are met. The words "shall" or some other obligatory language such as "must" and the negative equivalents are used to express requirements. The use of such words does not suggest that compliance with the Recommendation is required of any party.

INTELLECTUAL PROPERTY RIGHTS

ITU draws attention to the possibility that the practice or implementation of this Recommendation may involve the use of a claimed Intellectual Property Right. ITU takes no position concerning the evidence, validity or applicability of claimed Intellectual Property Rights, whether asserted by ITU members or others outside of the Recommendation development process.

As of the date of approval of this Recommendation, ITU had not received notice of intellectual property, protected by patents, which may be required to implement this Recommendation. However, implementers are cautioned that this may not represent the latest information and are therefore strongly urged to consult the TSB patent database at <http://www.itu.int/ITU-T/ipr/>.

© ITU 2016

All rights reserved. No part of this publication may be reproduced, by any means whatsoever, without the prior written permission of ITU.

Table of Contents

	Page
1 Scope.....	1
2 References.....	2
3 Definitions	2
3.1 Terms defined elsewhere	2
3-2 Terms defined in this Recommendation.....	2
4 Abbreviations and acronyms	2
5 Conventions	3
6 Test suite structure (TSS)	4
7 Electronic attachment	7
Annex A – Test purposes	8
A.1 TP definition conventions.....	8
A.2 Subgroup 2.1.7: Bluetooth low energy design guidelines (BLEDG).....	9
Bibliography.....	17

Electronic attachment: This Recommendation includes an electronic attachment with the protocol implementation conformance statements (PICS) and the protocol implementation extra information for testing (PIXIT) required for the implementation of Annex A.

Introduction

This Recommendation is a transposition of Continua Health Alliance Test Tool DG2013, Test Suite Structure & Test Purposes, PAN-LAN-TAN Interface; Part 8: Continua Design Guidelines. Manager BLE (Version 1.2, 2014-01-24), that was developed by the Continua Health Alliance. A number of versions of this specification existed before transposition and these can be found in the table below.

Version	Date	Revision history
1.0	2012-10-05	Initial release for Test Tool DG2011.
1.1	2013-05-24	Initial release for Test Tool DG2012. This uses "TSS&TP_DG2011_PAN-LAN-TAN_PART_8_v1.0.doc" as a baseline and it adds new features included in [b-CDG 2012] (BPM and HR profiles)
1.2	2014-01-24	Initial release for Test Tool DG2013. This uses "TSS&TP_DG2012_LP-PAN PART_8_v1.1.doc" as a baseline and it adds new features included in [ITU-T H.810]: <ul style="list-style-type: none">• Adds glucose meter BLE• Adds BLE SSP support• Adds NFC new transport• Adds INR device specialization

Recommendation ITU-T H.848

Conformance of ITU-T H.810 personal health devices conformance: PAN/LAN/TAN interface Part 8: Bluetooth low energy (BLE): Manager

1 Scope

The scope of this Recommendation¹ is to provide test suite structure and the test purposes (TSS & TP) for the PAN/LAN/TAN interface based on the requirements defined in the Continua design guidelines (CDG) [ITU-T H.810 (2015)]. The objective of this test specification is to provide a high probability of air interface interoperability between different devices.

TSS & TP for the PAN/LAN/TAN interface have been divided into the 10 parts specified below. This Recommendation covers Part 8.

- **Part 1:** Optimized exchange protocol [IEEE 11073-20601A] Agent
- **Part 2:** Optimized exchange protocol [IEEE 11073-20601A] Manager
- **Part 3:** Continua design guidelines [ITU-T H.810 (2015)] Agent
- **Part 4:** Continua design guidelines [ITU-T H.810 (2015)] Manager
- **Part 5:** Device specializations Agent [ISO/IEEE 11073-104xx]. This document is divided into 14 subparts:
 - **Part 5A:** Weighing scales
 - **Part 5B:** Glucose meter
 - **Part 5C:** Pulse oximeter
 - **Part 5D:** Blood pressure monitor
 - **Part 5E:** Thermometer
 - **Part 5F:** Cardiovascular fitness and activity monitor
 - **Part 5G:** Strength fitness equipment
 - **Part 5H:** Independent living activity hub
 - **Part 5I:** Adherence monitor
 - **Part 5J:** Insulin pump (Future development)
 - **Part 5K:** Peak flow
 - **Part 5L:** Body composition analyzer
 - **Part 5M:** Basic electrocardiograph
 - **Part 5N:** International normalized ratio monitor
- **Part 6:** Device specializations [ISO/IEEE 11073-104xx] Manager
- **Part 7:** Continua design guidelines [ITU-T H.810 (2015)] Agent BLE
- **Part 8: Continua design guidelines [ITU-T H.810 (2015)] Manager BLE**
- **Part 9:** Personal health devices transcoding Whitepaper [Bluetooth PHDT v1.4] Agent
- **Part 10:** Personal health devices transcoding Whitepaper [Bluetooth PHDT v1.4] Manager

¹ This Recommendation includes an electronic attachment with the protocol implementation conformance statements (PICS) and the protocol implementation extra information for testing (PIXIT) required for the implementation of Annex A.

2 References

The following ITU-T Recommendations and other references contain provisions which, through reference in this text, constitute provisions of this Recommendation. At the time of publication, the editions indicated were valid. All Recommendations and other references are subject to revision; users of this Recommendation are therefore encouraged to investigate the possibility of applying the most recent edition of the Recommendations and other references listed below. A list of the currently valid ITU-T Recommendations is regularly published. The reference to a document within this Recommendation does not give it, as a stand-alone document, the status of a Recommendation.

- [ITU-T H.810 (2015)] Recommendation ITU-T H.810 (2015), *Interoperability design guidelines for personal health systems*.
- [ITU-T H.810 (2016)] Recommendation ITU-T H.810 (2016), *Interoperability design guidelines for personal health systems*.
- [Bluetooth PHDT v1.4] Bluetooth SIG (2013), *Personal Health Devices Transcoding White Paper, v1.4*. https://www.bluetooth.org/DocMan/handlers/DownloadDoc.ashx?doc_id=294539
- [IEEE 11073-20601A] IEEE 11073-20601A-2010, *IEEE Health informatics – Personal health device communication Part 20601: Application profile – Optimized Exchange Protocol Amendment 1*.
<<http://standards.ieee.org/findstds/standard/11073-20601a-2010.html>>
- [ISO/IEEE 11073-104xx] ISO/IEEE 11073-104xx (in force), *Health informatics – Personal health device communication – Device specialization*.

NOTE – Shorthand is used to refer to the collection of device specialization standards that utilize [b-ISO/IEEE 11073-20601], where xx can be any number from 01 to 99, inclusive.

3 Definitions

3.1 Terms defined elsewhere

This Recommendation uses the following terms defined elsewhere:

3.1.1 agent [IEEE 11073-20601A]: A node that collects and transmits personal health data to an associated manager.

3.1.2 manager [IEEE 11073-20601A]: A node receiving data from one or more agent systems. Some examples of managers include a cellular phone, health appliance, set top box, or a computer system.

3.2 Terms defined in this Recommendation

None.

4 Abbreviations and acronyms

This Recommendation uses the following abbreviations and acronyms:

ATS	Abstract Test Suite
CDG	Continua Design Guidelines
DUT	Device Under Test
GUI	Graphical User Interface
INR	International Normalized Ratio
IUT	Implementation Under Test

MDS	Medical Device System
NFC	Near Field Communication
PAN	Personal Area Network
PCO	Point of Control and Observation
PCT	Protocol Conformance Testing
PHD	Personal Healthcare Device
PHDC	Personal Healthcare Device Class
PHM	Personal Health Manager
PICS	Protocol Implementation Conformance Statement
PIXIT	Protocol Implementation extra Information for Testing
SABTE	Sleep Apnoea Breathing Therapy Equipment
SDP	Service Discovery Protocol
SOAP	Simple Object Access Protocol
TCRL	Test Case Reference List
TCWG	Test and Certification Working Group
TP	Test Purpose
TSS	Test Suite Structure
USB	Universal Serial Bus
WDM	Windows Driver Model

5 Conventions

The key words "SHALL", "SHALL NOT", "SHOULD", "SHOULD NOT", "MAY", "MAY NOT" in this Recommendation are to be interpreted as in [b-ETSI SR 001 262].

- SHALL is equivalent to 'must' or 'it is required to'.
- SHALL NOT is equivalent to 'must not' or 'it is not allowed'.
- SHOULD is equivalent to 'it is recommended to'.
- SHOULD NOT is equivalent to 'it is not recommended to'.
- MAY is equivalent to 'is permitted'.
- MAY NOT is equivalent to 'it is not required that'.

NOTE – The above-mentioned key words are capitalized for illustrative purposes only and they do not appear capitalized within this Recommendation.

Reference is made in the ITU-T H.800-series of Recommendations to different versions of the Continua Design Guidelines (CDG) by a specific designation. The list of terms that may be used in this Recommendation is provided in Table 1.

Table 1 – List of designations associated with the various versions of the CDG

CDG release	Transposed as	Version	Description	Designation
2016 plus errata	[ITU-T H.810 (2016)]	6.1	Release 2016 plus errata noting all ratified bugs [ITU-T H.810 (2016)].	–
2016	–	6.0	Release 2016 of the CDG including maintenance updates of the CDG 2015 and additional guidelines that cover new functionalities.	Iris
2015 plus errata	[ITU-T H.810 (2015)]	5.1	Release 2015 plus errata noting all ratified bugs [ITU-T H.810 (2015)].	–
2015	–	5.0	Release 2015 of the CDG including maintenance updates of the CDG 2013 and additional guidelines that cover new functionalities.	Genome
2013 plus errata	[ITU-T H.810 (2013)]	4.1	Release 2013 plus errata noting all ratified bugs [b-ITU-T H.810 (2013)].	–
2013	–	4.0	Release 2013 of the CDG including maintenance updates of the CDG 2012 and additional guidelines that cover new functionalities.	Endorphin
2012 plus errata	–	3.1	Release 2012 plus errata noting all ratified bugs [b-CDG 2012].	–
2012	–	3.0	Release 2012 of the CDG including maintenance updates of the CDG 2011 and additional guidelines that cover new functionalities.	Catalyst
2011 plus errata	–	2.1	CDG 2011 integrated with identified errata.	–
2011	–	2.0	Release 2011 of the CDG including maintenance updates of the CDG 2010 and additional guidelines that cover new functionalities [b-CDG 2011].	Adrenaline
2010 plus errata	–	1.6	CDG 2010 integrated with identified errata	–
2010	–	1.5	Release 2010 of the CDG with maintenance updates of the CDG Version 1 and additional guidelines that cover new functionalities [b-CDG 2010].	1.5
1.0	–	1.0	First released version of the CDG [b-CDG 1.0].	–

6 Test suite structure (TSS)

The test purposes (TPs) for the PAN/LAN/TAN interface have been divided into the main subgroups specified below. Annex A describes the TPs for subgroup 2.1.7 (shown in bold).

- Group 1: Agent (AG)
 - Group 1.1: Transport (TR)
 - Subgroup 1.1.1: Design guidelines: Common (DGC)
 - Subgroup 1.1.2: USB design guidelines (UDG)
 - Subgroup 1.1.3: Bluetooth design guidelines (BDG)

- Subgroup 1.1.4: Pulse oximeter design guidelines (PODG)
- Subgroup 1.1.5: Cardiovascular design guidelines (CVDG)
- Subgroup 1.1.6: Activity hub design guidelines (HUBDG)
- Subgroup 1.1.7: ZigBee design guidelines (ZDG)
- Subgroup 1.1.8: Glucose meter design guidelines (GLDG)
- Subgroup 1.1.9: Bluetooth low energy design guidelines (BLEDG)
- Subgroup 1.1.10: Basic electrocardiograph design guidelines (ECGDG)
- Subgroup 1.1.11: NFC design guidelines (NDG)
- Group 1.2: 20601: Optimized exchange protocol (OXP)
 - Subgroup 1.2.1: PHD domain information model (DIM)
 - Subgroup 1.2.2: PHD service model (SER)
 - Subgroup 1.2.3: PHD communication model (COM)
- Group 1.3: Devices class specializations (CLASS)
 - Subgroup 1.3.1: Weighing scales (WEG)
 - Subgroup 1.3.2: Glucose meter (GL)
 - Subgroup 1.3.3: Pulse oximeter (PO)
 - Subgroup 1.3.4: Blood pressure monitor (BPM)
 - Subgroup 1.3.5: Thermometer (TH)
 - Subgroup 1.3.6: Cardiovascular (CV)
 - Subgroup 1.3.7: Strength (ST)
 - Subgroup 1.3.8: Activity hub (HUB)
 - Subgroup 1.3.9: Adherence monitor (AM)
 - Subgroup 1.3.10: Insulin pump (IP) (Future development)
 - Subgroup 1.3.11: Peak flow (PF)
 - Subgroup 1.3.12: Body composition analyzer (BCA)
 - Subgroup 1.3.13: Basic electrocardiograph (ECG)
 - Subgroup 1.3.14: International normalized ratio (INR)
 - Subgroup 1.3.15: Sleep apnoea breathing therapy equipment (SABTE)
- Group 1.4: Personal health device transcoding whitepaper (PHDTW)
 - Subgroup 1.4.1: Whitepaper general requirements (GEN)
 - Subgroup 1.4.2: Whitepaper thermometer requirements (TH)
 - Subgroup 1.4.3: Whitepaper blood pressure requirements (BPM)
 - Subgroup 1.4.4: Whitepaper heart rate requirements (HR)
 - Subgroup 1.4.5: Whitepaper glucose meter requirements (GL)
 - Subgroup 1.4.6: Whitepaper weight scale requirements (WS)
- Group 2: Manager (MAN)
 - Group 2.1: Transport (TR)
 - Subgroup 2.1.1: Design guidelines: Common (DGC)

- Subgroup 2.1.2: USB design guidelines (UDG)
- Subgroup 2.1.3: Bluetooth design guidelines (BDG)
- Subgroup 2.1.4: Cardiovascular design guidelines (CVDG)
- Subgroup 2.1.5: Activity hub design guidelines (HUBDG)
- Subgroup 2.1.6: ZigBee design guidelines (ZDG)
- **Subgroup 2.1.7: Bluetooth low energy design guidelines (BLEDG)**
- Subgroup 2.1.8: NFC design guidelines (NDG)
- Group 2.2: 20601: Optimized exchange protocol (OXP)
 - Subgroup 2.2.1: General (GEN)
 - Subgroup 2.2.2: PHD domain information model (DIM)
 - Subgroup 2.2.3: PHD service model (SER)
 - Subgroup 2.2.4: PHD communication model (COM)
- Group 2.3: Devices class specializations (CLASS)
 - Subgroup 2.3.1: Weighing scales (WEG)
 - Subgroup 2.3.2: Glucose meter (GL)
 - Subgroup 2.3.3: Pulse oximeter (PO)
 - Subgroup 2.3.4: Blood pressure monitor (BPM)
 - Subgroup 2.3.5: Thermometer (TH)
 - Subgroup 2.3.6: Cardiovascular (CV)
 - Subgroup 2.3.7: Strength (ST)
 - Subgroup 2.3.8: Activity hub (HUB)
 - Subgroup 2.3.9: Adherence monitor (AM)
 - Subgroup 2.3.10: Insulin pump (IP) (Future development)
 - Subgroup 2.3.11: Peak flow (PF)
 - Subgroup 2.3.12: Body composition analyzer (BCA)
 - Subgroup 2.3.13: Basic electrocardiograph (ECG)
 - Subgroup 2.3.14: International normalized ratio (INR)
 - Subgroup 2.3.15: Sleep apnoea breathing therapy equipment (SABTE)
- Group 2.4: Personal health device transcoding whitepaper (PHDTW)
 - Subgroup 2.4.1: Whitepaper general requirements (GEN)
 - Subgroup 2.4.2: Whitepaper thermometer requirements (TH)
 - Subgroup 2.4.3: Whitepaper blood pressure requirements (BPM)
 - Subgroup 2.4.4: Whitepaper heart rate requirements (HR)
 - Subgroup 2.4.5: Whitepaper glucose meter requirements (GL)
 - Subgroup 2.4.6: Whitepaper weight scale requirements (WS)

7 Electronic attachment

The protocol implementation conformance statements (PICS) and the protocol implementation extra information for testing (PIXIT) required for the implementation of Annex A can be downloaded from <http://handle.itu.int/11.1002/2000/12067>.

In the electronic attachment, letters "C" and "I" in the column labelled "Mandatory" are used to distinguish between "PICS" and "PIXIT" respectively during testing. If the cell is empty, the corresponding PICS is "independent". If the field contains a "C", the corresponding PICS is dependent on other PICS, and the logical expression is detailed in the "SCR_Expression" field. The static conformance review (SCR) is used in the test tool to assert whether the PICS selection is consistent.

Annex A

Test purposes

(This annex forms an integral part of this Recommendation.)

A.1 TP definition conventions

The test purposes (TPs) are defined according to the following rules:

- **TP Id:** This is a unique identifier (TP/<TT>/<DUT>/<GR>/<SGR>/<XX> – <NNN>). It is specified according to the naming convention defined below:
 - Each test purpose identifier is introduced by the prefix "TP".
 - <TT>: This is the test tool that will be used in the test case.
 - PAN: Personal area network (Bluetooth or USB)
 - LAN: Local area network (ZigBee)
 - PAN-LAN: Personal area network (Bluetooth or USB) – Local area network (ZigBee)
 - LP-PAN: Low power personal area network (Bluetooth Low Energy)
 - TAN: Touch area network (NFC)
 - PAN-LAN-TAN: Personal area network (Bluetooth or USB) – Local area network (ZigBee) – Touch area network (NFC)
 - <DUT>: This is the device under test.
 - AG: PAN/LAN Agent
 - MAN: PAN/LAN Manager
 - <GR>: This identifies a group of test cases.
 - <SGR>: This identifies a subgroup of test cases.
 - <XX>: This identifies the type of testing.
 - BV: valid behaviour test
 - BI: invalid behaviour test
 - <NNN>: This is a sequential number that identifies the test purpose (TP).
- **TP label:** This is the title of the TP.
- **Coverage:** This contains the specification reference and clause to be checked by the TP.
 - Spec: This indicates the earliest version of the specification from which the testable items to be checked by the TP are included.
 - Testable item: This contains testable items to be checked by the TP.
- **Test purpose:** This is a description of the requirements to be tested.
- **Applicability:** This contains the PICS items that define if the test case is applicable or not for a specific device. When a TP contains an "ALL" in this field it means that it applies to the device under test within that scope of the test (specialization, transport used, etc.).
- **Other PICS:** This contains additional PICS items (apart from the PICS specified in the Applicability row) which are used within the test case implementation and can modify the final verdict. When this row is empty, it means that only the PICS specified in the Applicability row are used within the test case implementation.
- **Initial condition:** This indicates the state to which the DUT needs to be moved at the beginning of TC execution.

- **Test procedure:** This describes the steps to be followed in order to execute the test case.
- **Pass/Fail criteria:** This provides criteria to decide whether the DUT passes or fails the test case.

A.2 Subgroup 2.1.7: Bluetooth low energy design guidelines (BLEDG)

TP Id		TP/LP-PAN/MAN/TR/BLEDG/BV-000		
TP label		Discovery and pairing process and Pairing/Auth notification		
Coverage	Spec	[ITU-T H.810 (2015)]		
	Testable items	Discovery_Pairing BT LE 1; M	Notify BT LE 2; M	
Test purpose		<p>Check that:</p> <p>Once a BLE Manager has discovered a BLE Agent that supports a compatible service, it shall support pairing with that BLE Agent</p> <p>[AND]</p> <p>If supported by the UI, BLE Manager shall inform the user that a pairing and authentication was successful</p>		
Applicability		C_MAN_BLE_000		
Other PICS		C_MAN_BLEDG_001		
Initial condition		The manager under test and the simulated agent are in a standby state and they have not been paired before.		
Test procedure		<ol style="list-style-type: none"> 1. Reset the manager under test to default configuration and turn it on. 2. Set the simulated agent in discoverable mode (advertising state). 3. The manager under test initiates a search for discoverable agents (scanning state). 4. Once the simulated agent has been discovered, make the manager under test pair with it as stated in the documentation (initiating state). 5. When the pairing has been completed (connection state), IF PICS C_MAN_BLEDG_001 = TRUE then the manager under test notifies the successful pairing and authentication process. 		
Pass/Fail criteria		<p>In step 4, the manager under test completes the pairing process successfully.</p> <p>In step 5, if the manager supports an UI that provides information about the Bluetooth connection (C_MAN_BLEDG_001 = TRUE) then the manager has notified the successful pairing and authentication process.</p>		
Notes				

TP Id		TP/LP-PAN/MAN/TR/BLEDG/BV-001		
TP label		Store pairing data		
Coverage	Spec	[ITU-T H.810 (2015)]		
	Testable items	Discovery_Pairing BT LE 12; M	Discovery_Pairing BT LE 13; R	
Test purpose		<p>Check that:</p> <p>BLE Manager should store pairing data from at least the most recently paired device such that the data is persistent (e.g. With loss of power including removal of a battery)</p> <p>[AND]</p>		

	BLE Manager should store pairing data for at least the number of devices they are intended to simultaneously support
Applicability	C_MAN_BLE_000
Other PICS	
Initial condition	The manager under test and the simulated agent are in a standby state and they have not been paired before.
Test procedure	<ol style="list-style-type: none"> 1. Reset the manager under test to the default configuration and turn it on. 2. Set the simulated agent to discoverable and pairable mode (advertising state). 3. The manager initiates a discovery process (scanning state), it finds the simulated, agent, establishes a pairing with it (initiating state) and starts a Bluetooth connection (connection state). 4. Turn off the manager under test and remove the batteries or unplug the power supply. 5. Turn on the manager under test again (standby state). 6. Set the simulated agent in discoverable and pairable mode (advertising state). 7. The manager under test initiates a discovery process (scanning state), it finds the simulated agent and starts a Bluetooth connection with it (connection state).
Pass/Fail criteria	In step 7, the pairing process should not be dispatched again because both devices should have stored the pairing data from the previous pairing process. If the pairing process is dispatched again then the test tool gives a WARNING message.
Notes	

TP Id	TP/LP-PAN/MAN/TR/BLEDG/BV-002		
TP label	Discovered devices filter		
Coverage	Spec	[ITU-T H.810 (2015)]	
	Testable items	Notify BT LE 3; R	
Test purpose	<p>Check that:</p> <p>Manager shall support all pairing methods for Bluetooth 4.0, including Just Works and Passkey Entry, if the manager has the appropriate I/O capabilities</p>		
Applicability	C_MAN_BLE_000		
Other PICS			
Initial condition	The manager under test and the simulated agent are in a standby state and they have not been paired before.		
Test procedure	<ol style="list-style-type: none"> 1. Reset the manager under test to the default configuration and turn it on. 2. Set the simulated agent to discoverable and pairable mode. 3. The manager initiates a discovery process, it finds the simulated agent and establishes a pairing with it. 		
Pass/Fail criteria	<p>In step 3, if the manager under test does not filter the discovered devices to include only the test tool simulated agent that supports the supported service/profile, the test tool displays a warning message.</p> <p>In step 3, if the manager supports an UI that provides information about the Bluetooth connection (C_MAN_BLE_DG_001= TRUE) check that the manager has notified the successful pairing and authentication process.</p>		

Notes	
--------------	--

TP Id	TP/LP-PAN/MAN/TR/BLEDG/BV-003		
TP label	Failures notification		
Coverage	Spec	[ITU-T H.810 (2015)]	
	Testable items	Notify BT LE 4; M	
Test purpose	<p>Check that:</p> <p>If there is a failure during the discovery, pairing and authentication process, and if supported by the UI, the BLE Manager shall inform the user whether the failure is because 1) no compatible BLE Agent was found (compatible device not found) or 2) the pairing failed (pairing failure) or 3) the authentication process timed out (authentication timeout) or 4) the user entered the incorrect passkey (incorrect PIN)</p>		
Applicability	C_MAN_BLE_000 AND C_MAN_BLEDG_001		
Other PICS			
Initial condition	The manager under test and the simulated agent are in a standby state and they have not been paired before.		
Test procedure	<ol style="list-style-type: none"> 1. Disable the simulated agent (it is not discoverable). 2. The manager under test initiates discovery as stated in the product documentation. 3. The simulated agent is not discovered because it has not been initialized yet. Check the manager under test for error messages. 4. Configure the simulated agent with a device specialization not supported by the manager under test. 5. Set the simulated agent to discoverable mode. 6. The manager under test starts a discovery and pairing process with the simulated agent. 7. Check the manager under test for error messages. 8. Restart the simulated agent. 9. Restart the manager under test. 10. Configure the simulated agent with a device specialization that is supported by the manager under test. 11. Disable the pairable mode in the simulated agent. 12. The manager under test starts a discovery and pairing process with the simulated agent. 13. Check the manager under test for error messages. 14. Restart the simulated agent. 15. Restart the manager under test. 16. Configure simulated agent with a device specialization that is supported by the manager under test. 17. Set the simulated agent to discoverable and pairable mode. 18. The manager under test starts a discovery and pairing process with the simulated agent. 19. Force the test tool simulated agent not to complete the authentication process. 20. Check the manager under test for error messages. 		

Pass/Fail criteria	<ul style="list-style-type: none"> In step 3, the manager under test shall inform the user that the pairing process cannot be completed (the simulated agent has not been found). In step 7, the manager under test shall inform the user that the pairing process cannot be completed (the simulated agent implements an unsupported specialization). In step 13, the manager under test shall inform the user that the pairing process cannot be completed (the simulated agent is not in pairable mode). In step 20, the manager under test shall inform to the user that authentication process cannot be completed (authentication timeout).
Notes	

TP Id	TP/LP-PAN/MAN/TR/BLEDG/BV-004		
TP label	Secure Simple Pairing with agent with NoInputNoOutput capabilities		
Coverage	Spec	[ITU-T H.810 (2015)]	
	Testable items	Authentication BT LE 2; M	
Test purpose	<p>Check that:</p> <p>Manager shall support all pairing methods for Bluetooth 4.0, including Just Works and Passkey Entry, if the manager has the appropriate I/O capabilities</p>		
Applicability	C_MAN_BLE_000		
Other PICS			
Initial condition	The manager under test and the simulated agent support the same device specialization, they are in a disconnected state and they have not been paired before.		
Test procedure	<ol style="list-style-type: none"> Check that the manager under test IO capabilities are declared in PIXIT I_MAN_BLEDG_002 and the man in the middle (MITM) protection is declared in PIXIT I_MAN_BLEDG_003 <ol style="list-style-type: none"> IF the manager under test does not support MITM protection (PIXIT I_MAN_BDG_003 = FALSE) THEN the test tool simulated agent is configured with NoInputNoOutput capabilities and without MITM protection and the just works association model shall be used during the pairing process and the generated link key will be unauthenticated (without MITM protection). IF the manager under test supports MITM protection (PIXIT I_MAN_BDG_003 = TRUE) THEN <ul style="list-style-type: none"> IF the Manager under test supports NoInputNoOutput capabilities (PIXIT I_MAN_BDG_002 = 4) THEN the combination of IO capabilities and MITM support declared by the manager under test in PIXITs is not feasible and the test case ends by giving a FAIL verdict due to inconsistency among the manager under test SSP features declared in PIXITs IF the manager under test supports other IO capabilities (PIXIT I_MAN_BDG_002 = 0 or 1 or 2 or 3) it will not pair with agents with NoInputNoOutput capabilities because they do not fulfil the security level required (i.e., MITM protection) and the test case execution ends by giving a PASS verdict Set the test tool simulated agent in discoverable and pairable mode. The manager under test initiates a discovery process as stated in the product documentation. Once the simulated agent has been discovered, make the manager under test pair with it as stated in the documentation. 		

Pass/Fail criteria	In step 4, the manager under test completes the pairing process successfully.
Notes	

TP Id	TP/LP-PAN/MAN/TR/BLEDG/BV-005		
TP label	Secure Simple Pairing with agent with DisplayOnly capabilities		
Coverage	Spec	[ITU-T H.810 (2015)]	
	Testable items	Authentication BT LE 2; M	
Test purpose	<p>Check that:</p> <p>Manager shall support all pairing methods for Bluetooth 4.0, including Just Works and Passkey Entry, if the manager has the appropriate I/O capabilities</p>		
Applicability	C_MAN_BLE_000		
Other PICS			
Initial condition	The manager under test and the simulated agent support the same device specialization, they are in a disconnected state and they have not been paired before.		
Test procedure	<ol style="list-style-type: none"> 1. Check that manager under test IO capabilities are declared in PIXIT I_MAN_BDG_002 and that the MITM protection is declared in PIXIT I_MAN_BDG_003 <ol style="list-style-type: none"> a. IF the manager under test does not support MITM protection (PIXIT I_MAN_BDG_003 = FALSE) THEN the test tool simulated agent is configured with secure simple pairing, NoInputNoOutput capabilities and without MITM protection in addition the just works association model shall be used during the pairing process and the generated link key will be unauthenticated (without MITM protection) b. IF the manager under test supports MITM protection (PIXIT I_MAN_BDG_003 = TRUE) THEN <ul style="list-style-type: none"> • IF the manager under test supports KeyboardOnly or KeyboardDisplaycapabilities (PIXIT I_MAN_BDG_002 = 2 or 3) THEN the test tool simulated agent is configured with secure simple pairing, DisplayOnly capabilities and MITM protection in addition the passkey entry association model shall be used during the pairing process and the generated link key will be authenticated (with MITM protection) • IF the manager under test supports DisplayOnly or DisplayYesNo capabilities (PIXIT I_MAN_BDG_002 = 0 or 1) THEN it will not pair with agents with DisplayOnly capabilities because they do not fulfil the security level required (i.e., MITM protection) and the test case execution ends by giving a PASS verdict • IF manager under test supports NoInputNoOutput capabilities (PIXIT I_MAN_BDG_002 = 4) THEN the combination of IO capabilities and MITM support declared by the manager under test in PIXITs is not feasible and the test case ends by giving a FAIL verdict due to inconsistency among the agent under test SSP features declared in PIXITs 2. Set the test tool simulated agent in discoverable and pairable mode. 3. The manager under test initiates a discovery process as stated in the product documentation. 4. Once the simulated agent has been discovered, make the manager under test pair with it as stated in the documentation. 		
Pass/Fail criteria	In step 4, the manager under test completes the pairing process successfully.		
Notes			

TP Id		TP/LP-PAN/MAN/TR/BLEDG/BV-006	
TP label		Secure Simple Pairing with agent with DisplayYesNo capabilities	
Coverage	Spec	[ITU-T H.810 (2015)]	
	Testable items	Authentication BT LE 2; M	
Test purpose		Check that: Manager shall support all pairing methods for Bluetooth 4.0, including Just Works and Passkey Entry, if the manager has the appropriate I/O capabilities	
Applicability		C_MAN_BLE_000	
Other PICS			
Initial condition		The manager under test and the simulated agent support the same device specialization, they are in a disconnected state and they have not been paired before.	
Test procedure		<ol style="list-style-type: none"> 1. Check that the manager under test IO capabilities are declared in PIXIT I_MAN_BDG_002 and that the MITM protection is declared in PIXIT I_MAN_BDG_003 <ol style="list-style-type: none"> a. IF the manager under test does not support MITM protection (PIXIT I_MAN_BDG_003 = FALSE) THEN the test tool simulated agent is configured with secure simple pairing, NoInputNoOutput capabilities and without MITM protection in addition the just works association model shall be used during the pairing process and the generated link key will be unauthenticated (without MITM protection) b. IF the manager under test supports MITM protection (PIXIT I_MAN_BDG_003 = TRUE) THEN <ul style="list-style-type: none"> • IF the manager under test supports KeyboardOnly or KeyboardDisplay capabilities (PIXIT I_MAN_BDG_002 = 2 or 3) THEN the test tool simulated agent is configured with secure simple pairing, DisplayYesNo capabilities and with MITM protection in addition the passkey entry association model shall be used during the pairing process and the generated link key will be authenticated (with MITM protection) • IF the manager under test supports DisplayOnly or DisplayYesNocapabilities (PIXIT I_MAN_BDG_002 = 0 or 1) THEN it will not pair with agents with DisplayYesNo capabilities because they do not fulfil the security level required (i.e., MITM protection) and the test case execution ends by giving a PASS verdict • IF the manager under test supports NoInputNoOutput capabilities (PIXIT I_MAN_BDG_002 = 4) THEN the combination of IO capabilities and MITM support declared by the manager under test in PIXITs is not feasible and the test case ends by giving a FAIL verdict due to inconsistency among the agent under test SSP features declared in PIXITs 2. Set the test tool simulated agent in discoverable and pairable mode. 3. The manager under test initiates a discovery process as stated in the product documentation. 4. Once the simulated agent has been discovered, make the manager under test pair with it as stated in the documentation. 	
Pass/Fail criteria		In step 4, the manager under test completes the pairing process successfully.	
Notes			

TP Id		TP/LP-PAN/MAN/TR/BLEDG/BV-007	
TP label		Secure Simple Pairing with agent with KeyboardOnly capabilities	
Coverage	Spec	[ITU-T H.810 (2015)]	
	Testable items		

	Testable items	Authentication BT LE 2; M		
Test purpose	Check that: Manager shall support all pairing methods for Bluetooth 4.0, including Just Works and Passkey Entry, if the manager has the appropriate I/O capabilities			
Applicability	C_MAN_BLE_000			
Other PICS				
Applicability	C_MAN_BLE_000			
Initial condition	The manager under test and the simulated agent support the same device specialization, they are in a disconnected state and they have not been paired before.			
Test procedure	<ol style="list-style-type: none"> 1. Check that the manager under test IO capabilities are declared in PIXIT I_MAN_BDG_002 and that the MITM protection is declared in PIXIT I_MAN_BDG_003 <ol style="list-style-type: none"> a. IF the manager under test does not support MITM protection (PIXIT I_MAN_BDG_003 = FALSE) THEN the test tool simulated agent is configured with secure simple pairing, NoInputNoOutput capabilities and without MITM protection and the just works association model shall be used during the pairing process and the generated link key will be unauthenticated (without MITM protection) b. IF the manager under test supports MITM protection (PIXIT I_MAN_BDG_003 = TRUE) THEN <ul style="list-style-type: none"> • IF the manager under test supports DisplayOnly or DisplayYesNo or KeyboardOnly or KeyboardDisplay capabilities (PIXIT I_MAN_BDG_002 = 0 or 1 or 2 or 3) THEN the test tool simulated agent is configured with secure simple pairing, KeyboardOnly capabilities and with MITM protection in addition the passkey entry association model shall be used during the pairing process and the generated link key will be authenticated (with MITM protection) • IF the manager under test supports NoInputNoOutput capabilities (PIXIT I_MAN_BDG_002 = 4) THEN the combination of IO capabilities and MITM support declared by the manager under test in PIXITs is not feasible and the test case ends by giving a FAIL verdict due to inconsistency among the agent under test SSP features declared in PIXITs 2. Set the test tool simulated agent in discoverable and pairable mode. 3. The manager under test initiates a discovery process as stated in the product documentation. 4. Once the simulated agent has been discovered, make the manager under test pair with it as stated in the documentation. 			
Pass/Fail criteria	In step 4, the manager under test completes the pairing process successfully.			
Notes				

TP Id	TP/LP-PAN/MAN/TR/BLEDG/BV-008			
TP label	Secure Simple Pairing with agent with KeyboardDisplay capabilities			
Coverage	Spec	[ITU-T H.810 (2015)]		
	Testable items	Authentication BT LE 2; M		
Test purpose	Check that: Manager shall support all pairing methods for Bluetooth 4.0, including Just Works and Passkey Entry, if the manager has the appropriate I/O capabilities			

Applicability	C_MAN_BLE_000
Other PICS	
Applicability	C_MAN_BLE_000
Initial condition	The manager under test and the simulated agent support the same device specialization, they are in a disconnected state and they have not been paired before.
Test procedure	<ol style="list-style-type: none"> 1. Check the manager under test IO capabilities declared in PIXIT I_MAN_BDG_002 and the MITM protection declared in PIXIT I_MAN_BDG_003 <ol style="list-style-type: none"> a. IF the manager under test does not support MITM protection (PIXIT I_MAN_BDG_003 = FALSE) THEN the test tool simulated agent is configured with secure simple pairing, NoInputNoOutput capabilities and without MITM protection in addition the just works association model shall be used during the pairing process and the generated link key will be unauthenticated (without MITM protection) b. IF the manager under test supports MITM protection (PIXIT I_MAN_BDG_003 = TRUE) THEN <ul style="list-style-type: none"> • IF the manager under test supports DisplayOnly or DisplayYesNo or KeyboardOnly or KeyboardDisplay capabilities (PIXIT I_MAN_BDG_002 = 0 or 1 or 2 or 3) THEN the test tool simulated agent is configured with secure simple pairing, KeyboardDisplay capabilities and with MITM protection in addition the passkey entry association model shall be used during the pairing process and the generated link key will be authenticated (with MITM protection) • IF the manager under test supports NoInputNoOutput capabilities (PIXIT I_MAN_BDG_002 = 4) THEN the combination of IO capabilities and MITM support declared by the Manager under test in PIXITs is not feasible and the test case ends giving a FAIL verdict due to inconsistency among the Agent under test SSP features declared in PIXITs 2. Set the test tool simulated agent in discoverable and pairable mode. 3. The manager under test initiates a discovery process as stated in the product documentation. 4. Once the simulated agent has been discovered, make the manager under test pair with it as stated in the documentation.
Pass/Fail criteria	In step 4, the manager under test completes the pairing process successfully.
Notes	

Bibliography

- [b-ITU-T H.810 (2013)] Recommendation ITU-T H.810 (2013), *Interoperability design guidelines for personal health systems*.
- [b-BT HDP] Bluetooth Health Device Profile and Multi-Channel Adaptation Protocol. Version 1.0.
https://www.bluetooth.org/docman/handlers/DownloadDoc.ashx?doc_id=119999
- [b-CDG 1.0] Continua Health Alliance, Continua Design Guidelines v1.0. (2008), *Continua Design Guidelines*.
- [b-CDG 2010] Continua Health Alliance, Continua Design Guidelines v1.5 (2010), *Continua Design Guidelines*.
- [b-CDG 2011] Continua Health Alliance, Continua Design Guidelines (2011), "Adrenaline", *Continua Design Guidelines*.
- [b-CDG 2012] Continua Health Alliance, Continua Design Guidelines (2012), "Catalyst", *Continua Design Guidelines*.
- [b-ETSI SR 001 262] ETSI SR 001 262 v1.8.1 (2003): *ETSI drafting rules*.
<https://docbox.etsi.org/MTS/MTS/10-PromotionalMaterial/MBS-20111118/Referenced%20Documents/Drafting%20Rules.pdf>
- [b-ISO/IEEE 11073-20601] ISO/IEEE 11073-20601-2010, *Health informatics – Personal health device communication – Part 20601: Application profile – Optimized exchange protocol*.
- [b-Agent PICS & PIXIT] PLT Agent DG2013 *PICS and PIXIT excel sheet v1.6*
<http://handle.itu.int/11.1002/2000/12067>
- [b-Bluetooth PHDT] Bluetooth SIG (in force), *Personal Health Devices Transcoding White Paper*.
- [b-Manager PICS & PIXIT] PLT Manager DG2013 *PICS and PIXIT excel sheet v1.6*
<http://handle.itu.int/11.1002/2000/12067>
- [b-USB PHDC] USB Personal Healthcare Device Class v1.0 plus (2008), errata.
http://www.usb.org/developers/docs/devclass_docs/Personal_Healthcare_1.zip

SERIES OF ITU-T RECOMMENDATIONS

Series A	Organization of the work of ITU-T
Series D	General tariff principles
Series E	Overall network operation, telephone service, service operation and human factors
Series F	Non-telephone telecommunication services
Series G	Transmission systems and media, digital systems and networks
Series H	Audiovisual and multimedia systems
Series I	Integrated services digital network
Series J	Cable networks and transmission of television, sound programme and other multimedia signals
Series K	Protection against interference
Series L	Environment and ICTs, climate change, e-waste, energy efficiency; construction, installation and protection of cables and other elements of outside plant
Series M	Telecommunication management, including TMN and network maintenance
Series N	Maintenance: international sound programme and television transmission circuits
Series O	Specifications of measuring equipment
Series P	Terminals and subjective and objective assessment methods
Series Q	Switching and signalling
Series R	Telegraph transmission
Series S	Telegraph services terminal equipment
Series T	Terminals for telematic services
Series U	Telegraph switching
Series V	Data communication over the telephone network
Series X	Data networks, open system communications and security
Series Y	Global information infrastructure, Internet protocol aspects and next-generation networks, Internet of Things and smart cities
Series Z	Languages and general software aspects for telecommunication systems