

International Telecommunication Union

ITU-T

TELECOMMUNICATION
STANDARDIZATION SECTOR
OF ITU

H.850.2

(08/2020)

SERIES H: AUDIOVISUAL AND MULTIMEDIA SYSTEMS

E-health multimedia systems, services and applications –
Interoperability compliance testing of personal health
systems (HRN, PAN, LAN, TAN and WAN)

**Conformance of ITU-T H.810 personal health
system: Personal Health Devices interface
Part 10B: Transcoding for Bluetooth Low
Energy: Personal Health Gateway – Blood
pressure**

Recommendation ITU-T H.850.2



ITU-T H-SERIES RECOMMENDATIONS
AUDIOVISUAL AND MULTIMEDIA SYSTEMS

CHARACTERISTICS OF VISUAL TELEPHONE SYSTEMS	H.100–H.199
INFRASTRUCTURE OF AUDIOVISUAL SERVICES	
General	H.200–H.219
Transmission multiplexing and synchronization	H.220–H.229
Systems aspects	H.230–H.239
Communication procedures	H.240–H.259
Coding of moving video	H.260–H.279
Related systems aspects	H.280–H.299
Systems and terminal equipment for audiovisual services	H.300–H.349
Directory services architecture for audiovisual and multimedia services	H.350–H.359
Quality of service architecture for audiovisual and multimedia services	H.360–H.369
Telepresence, immersive environments, virtual and extended reality	H.420–H.439
Supplementary services for multimedia	H.450–H.499
MOBILITY AND COLLABORATION PROCEDURES	
Overview of Mobility and Collaboration, definitions, protocols and procedures	H.500–H.509
Mobility for H-Series multimedia systems and services	H.510–H.519
Mobile multimedia collaboration applications and services	H.520–H.529
Security for mobile multimedia systems and services	H.530–H.539
Security for mobile multimedia collaboration applications and services	H.540–H.549
VEHICULAR GATEWAYS AND INTELLIGENT TRANSPORTATION SYSTEMS (ITS)	
Architecture for vehicular gateways	H.550–H.559
Vehicular gateway interfaces	H.560–H.569
BROADBAND, TRIPLE-PLAY AND ADVANCED MULTIMEDIA SERVICES	
Broadband multimedia services over VDSL	H.610–H.619
Advanced multimedia services and applications	H.620–H.629
Ubiquitous sensor network applications and Internet of Things	H.640–H.649
IPTV MULTIMEDIA SERVICES AND APPLICATIONS FOR IPTV	
General aspects	H.700–H.719
IPTV terminal devices	H.720–H.729
IPTV middleware	H.730–H.739
IPTV application event handling	H.740–H.749
IPTV metadata	H.750–H.759
IPTV multimedia application frameworks	H.760–H.769
IPTV service discovery up to consumption	H.770–H.779
Digital Signage	H.780–H.789
E-HEALTH MULTIMEDIA SYSTEMS, SERVICES AND APPLICATIONS	
Personal health systems	H.810–H.819
Interoperability compliance testing of personal health systems (HRN, PAN, LAN, TAN and WAN)	H.820–H.859
Multimedia e-health data exchange services	H.860–H.869
Safe listening	H.870–H.879

For further details, please refer to the list of ITU-T Recommendations.

Recommendation ITU-T H.850.2

Conformance of ITU-T H.810 personal health system: Personal Health Devices interface Part 10B: Transcoding for Bluetooth Low Energy: Personal Health Gateway – Blood pressure

Summary

Recommendation ITU-T H.850.2 provides a test suite structure (TSS) and the test purposes (TP) for the transcoding of blood pressure data by personal health gateways in the Personal Health Devices (PHD) interface of application-level data between the Bluetooth Low Energy Bluetooth Generic Attribute Profile (GATT) format and the IEEE 11073-20601 data format, of which Recommendation ITU-T H.810 (2016) is the base Recommendation. The objective of this test specification is to provide a high probability of interoperability at this interface.

Recommendation ITU-T H.850.2 is a transposition of clause 3.4 of Continua Test Tool DG2016, Test Suite Structure & Test Purposes, Personal Health Devices Interface; Part 10: PHD Transcoding Whitepaper. Personal Health Gateway (Version 1.7, 2017-07-18), that was developed by the Personal Connected Health Alliance. A number of versions of this specification existed before transposition.

This Recommendation includes an electronic attachment with the protocol implementation conformance statements (PICS) and the protocol implementation extra information for testing (PIXIT) required for the implementation of Annex A.

History

Edition	Recommendation	Approval	Study Group	Unique ID*
1.0	ITU-T H.850.2	2017-04-29	16	11.1002/1000/13355
2.0	ITU-T H.850.2	2020-08-13	16	11.1002/1000/14346

Keywords

Bluetooth Low Energy (BLE), Bluetooth Generic Attribute Profile, blood pressure, Conformance testing, Continua Design Guidelines, data format transcoding, e-health, IEEE 11073-20601, ITU-T H.810, Personal Health Devices interface, personal area network, personal connected health devices, Personal Health Gateway, touch area network.

* To access the Recommendation, type the URL <http://handle.itu.int/> in the address field of your web browser, followed by the Recommendation's unique ID. For example, <http://handle.itu.int/11.1002/1000/11830-en>.

FOREWORD

The International Telecommunication Union (ITU) is the United Nations specialized agency in the field of telecommunications, information and communication technologies (ICTs). The ITU Telecommunication Standardization Sector (ITU-T) is a permanent organ of ITU. ITU-T is responsible for studying technical, operating and tariff questions and issuing Recommendations on them with a view to standardizing telecommunications on a worldwide basis.

The World Telecommunication Standardization Assembly (WTSA), which meets every four years, establishes the topics for study by the ITU-T study groups which, in turn, produce Recommendations on these topics.

The approval of ITU-T Recommendations is covered by the procedure laid down in WTSA Resolution 1.

In some areas of information technology which fall within ITU-T's purview, the necessary standards are prepared on a collaborative basis with ISO and IEC.

NOTE

In this Recommendation, the expression "Administration" is used for conciseness to indicate both a telecommunication administration and a recognized operating agency.

Compliance with this Recommendation is voluntary. However, the Recommendation may contain certain mandatory provisions (to ensure, e.g., interoperability or applicability) and compliance with the Recommendation is achieved when all of these mandatory provisions are met. The words "shall" or some other obligatory language such as "must" and the negative equivalents are used to express requirements. The use of such words does not suggest that compliance with the Recommendation is required of any party.

INTELLECTUAL PROPERTY RIGHTS

ITU draws attention to the possibility that the practice or implementation of this Recommendation may involve the use of a claimed Intellectual Property Right. ITU takes no position concerning the evidence, validity or applicability of claimed Intellectual Property Rights, whether asserted by ITU members or others outside of the Recommendation development process.

As of the date of approval of this Recommendation, ITU had not received notice of intellectual property, protected by patents, which may be required to implement this Recommendation. However, implementers are cautioned that this may not represent the latest information and are therefore strongly urged to consult the TSB patent database at <http://www.itu.int/ITU-T/ipr/>.

© ITU 2020

All rights reserved. No part of this publication may be reproduced, by any means whatsoever, without the prior written permission of ITU.

Table of Contents

	Page
1 Scope	1
2 References.....	2
3 Definitions	3
3.1 Terms defined elsewhere.....	3
3.2 Terms defined in this Recommendation.....	3
4 Abbreviations and acronyms	3
5 Conventions	4
6 Test suite structure.....	6
7 Electronic attachment	8
Annex A Test purposes	9
A.1 TP definition conventions.....	9
A.2 Subgroup 2.4.3 – Whitepaper Blood pressure requirements (BP)	11
Bibliography.....	61

Electronic attachment: This Recommendation includes an electronic attachment with the protocol implementation conformance statements (PICS) and the protocol implementation extra information for testing (PIXIT) required for the implementation of Annex A.

Introduction

This Recommendation is a transposition of clause 3.4 of Continua Test Tool DG2016, Test Suite Structure & Test Purposes, Personal Health Devices Interface; Part 10: PHD Transcoding Whitepaper. Personal Health Gateway (Version 1.7, 2017-07-18), that was developed by the Personal Connected Health Alliance. The table below shows the revision history of this test specification; it may contain versions that existed before transposition.

Version	Date	Revision history
1.0	2012-10-05	Initial release for Test Tool DG2011 based on the requirements in [b-CDG 2011].
1.1	2013-05-24	Initial release for Test Tool DG2012. It uses "TSS&TP_DG2011_LP-PAN_PART_10_v1.0.doc" as a baseline and adds new features included in [b-CDG 2012] (BPM and HR profiles).
1.2	2014-01-24	Initial release for Test Tool DG2013. It uses "TSS&TP_DG2012_LP-PAN_PART_10_v1.1.doc" as a baseline and adds new features included in [b-ITU-T H.810 (2013)]/[b-CDG 2013]: <ul style="list-style-type: none"> • Adds glucose meter BLE • Adds BLE SSP support • Adds NFC new transport • Adds INR device specialization
1.3	2014-04-24	TM Lite & Doc Enhancements (Test Tool v4.0 Maintenance Release 1). It uses "TSS&TP_DG2013_LP-PAN_PART_10_v1.2.doc" as a baseline and adds new features included in Documentation Enhancements: <ul style="list-style-type: none"> • "Other PICS" row has been added
1.4	2015-07-01	Initial release for Test Tool DG2015. It uses "TSS&TP_DG2013_LP-PAN_PART_10_v1.3.doc" as a baseline and adds new features included in [b-ITU-T H.810 (2015)]/[b-CDG 2015]: <ul style="list-style-type: none"> • Adds WS/BCA BLE device specialization • Adds SABTE IEEE device specialization
1.5	2016-01-26	First maintenance release for Test Tool DG2015. It uses "TSS&TP_DG2015_LP-PAN_PART_10_v1.4.doc" as a baseline and adds some updates according to the Maintenance 2015 activity.
1.6	2016-09-20	Initial release for Test Tool DG2016. It uses "TSS&TP_DG2016_LP-PAN_PART_10_v1.5.doc" as a baseline and adds new features included in [ITU-T H.810 (2016)]/[b-CDG 2016]: <ul style="list-style-type: none"> • Adds PLX BLE device specialization • Adds PLX CGM device specialization
1.7	2017-07-18	Second Maintenance Release for Test Tool DG2016. It uses "TSS&TP_DG2016_LP-PAN_PART_10_v1.6.doc" as a baseline and corrects minor typos.
1.8	2020-06-02	Updates related to the value of the Reg-Cert-Data-List according to [b-CDG 2017].

Recommendation ITU-T H.850.2

Conformance of ITU-T H.810 personal health devices: Personal Health Devices interface Part 10B: Transcoding for Bluetooth Low Energy: Personal Health Gateway – Blood pressure

1 Scope

The scope of this Recommendation¹ is to provide a test suite structure (TSS) and the test purposes (TP) for the Personal Health Devices interface based on the requirements defined in the Continua Design Guidelines (CDG) [ITU-T H.810 (2016)]. The objective of this test specification is to provide a high probability of interoperability at this interface.

The TSS and TP for the Personal Health Devices interface have been divided into the parts specified below. This Recommendation covers Part 10B.

- Part 1: Optimized exchange protocol. Personal Health Device
- Part 2: Optimized exchange protocol. Personal Health Gateway
- Part 3: Continua design guidelines. Personal Health Device
- Part 4: Continua design guidelines. Personal Health Gateway
- Part 5: Device specializations. Personal Health Devices interface. This document is divided into the following subparts:
 - Part 5A: Weighing scales
 - Part 5B: Glucose meter
 - Part 5C: Pulse oximeter
 - Part 5D: Blood pressure monitor
 - Part 5E: Thermometer
 - Part 5F: Cardiovascular fitness and activity monitor
 - Part 5G: Strength fitness equipment
 - Part 5H: Independent living activity hub
 - Part 5I: Adherence monitor
 - Part 5J: Insulin pump
 - Part 5K: Peak expiratory flow monitor
 - Part 5L: Body composition analyser
 - Part 5M: Basic electrocardiograph
 - Part 5N: International normalized ratio monitor
 - Part 5O: Sleep apnoea breathing therapy equipment (SABTE)
 - Part 5P: Continuous glucose monitor (CGM)
- Part 6: Device specializations. Personal Health Gateway

¹ This Recommendation includes an electronic attachment with the protocol implementation conformance statements (PICS) and the protocol implementation extra information for testing (PIXIT) required for the implementation of Annex A.

- Part 7: Continua Design Guidelines. BLE Personal Health Device
- Part 8: Continua Design Guidelines. BLE Personal Health Gateway
- Part 9: Personal Health Devices Transcoding Whitepaper. Personal Health Devices
- Part 10: Personal Health Devices Transcoding Whitepaper. Personal Health Gateway. In addition to the main part, the document is subdivided into the following subparts:
 - Part 10A: Whitepaper Thermometer requirements
 - **Part 10B: Whitepaper Blood pressure requirements**
 - Part 10C: Whitepaper Heart rate requirements
 - Part 10D: Whitepaper Glucose meter requirements
 - Part 10E: Whitepaper Weighing scales requirements
 - Part 10F: Whitepaper Pulse oximeter requirements
 - Part 10G: Whitepaper Continuous glucose monitoring requirements

2 References

The following ITU-T Recommendations and other references contain provisions which, through reference in this text, constitute provisions of this Recommendation. At the time of publication, the editions indicated were valid. All Recommendations and other references are subject to revision; users of this Recommendation are therefore encouraged to investigate the possibility of applying the most recent edition of the Recommendations and other references listed below. A list of the currently valid ITU-T Recommendations is regularly published. The reference to a document within this Recommendation does not give it, as a stand-alone document, the status of a Recommendation.

- | | |
|------------------------------|--|
| [ITU-T H.810 (2016)] | Recommendation ITU-T H.810 (2016), <i>Interoperability design guidelines for personal health systems</i> . |
| [Bluetooth PHDT v1.4] | Bluetooth SIG (2013), <i>Personal Health Devices Transcoding White Paper, v1.4</i> .
https://www.bluetooth.org/DocMan/handlers/DownloadDoc.ashx?doc_id=294539 |
| [Bluetooth PHDT v1.5] | Bluetooth SIG (2014), <i>Personal Health Devices Transcoding White Paper, v1.5</i> .
https://www.bluetooth.org/DocMan/handlers/DownloadDoc.ashx?doc_id=272346 |
| [Bluetooth PHDT v1.6] | Bluetooth SIG (2015), <i>Personal Health Devices Transcoding White Paper, v1.6</i> .
https://www.bluetooth.org/DocMan/handlers/DownloadDoc.ashx?doc_id=310657 |
| [ISO/IEEE 11073-104xx] | ISO/IEEE 11073-104xx (in force), <i>Health informatics – Personal health device communication – Device specialization</i> .
NOTE – Shorthand to refer to the collection of device specialization standards that utilize [ISO/IEEE 11073-20601-2015A], where xx can be any number from 01 to 99, inclusive. |
| [ISO/IEEE 11073-20601-2015A] | ISO/IEEE 11073-20601:2010, <i>Health informatics – Personal health device communication – Part 20601: Application profile – Optimized exchange protocol</i> , including ISO/IEEE 11073-20601:2010 Amd 1:2015.
https://www.iso.org/standard/54331.html with
https://www.iso.org/standard/63972.html |

- [ISO/IEEE 11073-20601-2016C] ISO/IEEE 11073-20601:2016, *Health informatics – Personal health device communication – Part 20601: Application profile – Optimized exchange protocol*, including ISO/IEEE 11073-20601:2016/Cor.1:2016.
<https://www.iso.org/standard/66717.html> with
<https://www.iso.org/standard/71886.html>
- [IHE PCD TF 1] IHE PCD TF 1 (2012), *IHE Patient Care Device Technical Framework – Revision 2.0. Volume 1: Integration Profiles*.
http://www.ihe.net/Technical_Framework/upload/IHE_PCD_TF_Rev2-0_Vol1_FT_2012-08-16.pdf
- [IHE PCD TF 2] IHE PCD TF 2 (2012), *IHE Patient Care Device Technical Framework – Revision 2.0. Volume 2: Transactions*.
http://www.ihe.net/Technical_Framework/upload/IHE_PCD_TF_Rev2-0_Vol2_FT_2012-08-16.pdf
- [IHE PCD TF 3] IHE PCD TF 3 (2012), *IHE Patient Care Device Technical Framework – Revision 2.0. Volume 3: Semantic Content*.
http://www.ihe.net/Technical_Framework/upload/IHE_PCD_TF_Rev2-0_Vol3_FT_2012-08-16.pdf

3 Definitions

3.1 Terms defined elsewhere

This Recommendation uses the following terms defined elsewhere:

3.1.1 agent [ISO/IEEE 11073-20601-2016C]: A node that collects and transmits personal health data to an associated manager.

3.1.2 manager [ISO/IEEE 11073-20601-2016C]: A node receiving data from one or more agent systems. Some examples of managers include a cellular phone, health appliance, set top box, or a computer system.

3.2 Terms defined in this Recommendation

None.

4 Abbreviations and acronyms

This Recommendation uses the following abbreviations and acronyms:

ATS	Abstract Test Suite
CDG	Continua Design Guidelines
CGM	Continuous Glucose Monitor
DUT	Device Under Test
GUI	Graphical User Interface
INR	International Normalized Ratio
IP	Insulin Pump
IUT	Implementation Under Test
LSB	Least Significant Bit
MDS	Medical Device System
MSB	Most Significant Bit

NFC	Near Field Communication
PAN	Personal Area Network
PCD	Patient Care Device
PCO	Point of Control and Observation
PCT	Protocol Conformance Testing
PHD	Personal Health Device
PHDC	Personal Healthcare Device Class
PHG	Personal Health Gateway
PICS	Protocol Implementation Conformance Statement
PIXIT	Protocol Implementation extra Information for Testing
RACP	Record Access Control Point
SABTE	Sleep Apnoea Breathing Therapy Equipment
SCR	Static Conformance Review
SDP	Service Discovery Protocol
SOAP	Simple Object Access Protocol
TCRL	Test Case Reference List
TCWG	Test and Certification Working Group
TP	Test Purposes
TSS	Test Suite Structure
USB	Universal Serial Bus
WDM	Windows Driver Model

5 Conventions

In this text, the uppercase letter L is used as the symbol for litre.

Several of the test purposes in Annex A refer to "WAN PCD-01 messages"; these messages are specified in the patient care device (PCD) technical framework defined in [IHE PCD TF 1], [IHE PCD TF 2] and [IHE PCD TF 3]. Similarly, the "IEEE 11073 Objects and Attributes" are defined in [ISO/IEEE 11073-104xx].

The key words "SHALL", "SHALL NOT", "SHOULD", "SHOULD NOT", "MAY", "MAY NOT" in this Recommendation are to be interpreted as in [b-ETSI SR 001 262].

- SHALL is equivalent to 'must' or 'it is required to'.
- SHALL NOT is equivalent to 'must not' or 'it is not allowed'.
- SHOULD is equivalent to 'it is recommended to'.
- SHOULD NOT is equivalent to 'it is not recommended to'.
- MAY is equivalent to 'is permitted'.
- MAY NOT is equivalent to 'it is not required that'.

NOTE – The above-mentioned key words are capitalized for illustrative purposes only and they do not appear capitalized within this Recommendation.

In this document, hexadecimal numbers are denoted either with the prefix "0x" or by "(hex)" after the number; "(dec)" after a number indicates it is expressed in decimal format.

Reference is made in the ITU-T H.800-series of Recommendations to different versions of the Continua Design Guidelines (CDG) by a specific designation. The list of terms that may be used in this Recommendation is provided in Table 1.

Table 1 – List of designations associated with the various versions of the CDG

CDG release	Transposed as	Version	Description	Designation
2016 plus errata	[ITU-T H.810 (2016)]	6.1	Release 2016 plus errata noting all ratified bugs [b-CDG 2016].	–
2016	–	6.0	Release 2016 of the CDG including maintenance updates of the CDG 2015 and additional guidelines that cover new functionalities.	Iris
2015 plus errata	[b-ITU-T H.810 (2015)]	5.1	Release 2015 plus errata noting all ratified bugs [b-CDG 2015]. The 2013 edition of ITU-T H.810 is split into eight parts in the ITU-T H.810-series.	–
2015	–	5.0	Release 2015 of the CDG including maintenance updates of the CDG 2013 and additional guidelines that cover new functionalities.	Genome
2013 plus errata	[b-ITU-T H.810 (2013)]	4.1	Release 2013 plus errata noting all ratified bugs [b-CDG 2013].	–
2013	–	4.0	Release 2013 of the CDG including maintenance updates of the CDG 2012 and additional guidelines that cover new functionalities.	Endorphin
2012 plus errata	–	3.1	Release 2012 plus errata noting all ratified bugs [b-CDG 2012].	–
2012	–	3.0	Release 2012 of the CDG including maintenance updates of the CDG 2011 and additional guidelines that cover new functionalities.	Catalyst
2011 plus errata	–	2.1	CDG 2011 integrated with identified errata.	–
2011	–	2.0	Release 2011 of the CDG including maintenance updates of the CDG 2010 and additional guidelines that cover new functionalities [b-CDG 2011].	Adrenaline
2010 plus errata	–	1.6	CDG 2010 integrated with identified errata	–
2010	–	1.5	Release 2010 of the CDG with maintenance updates of the CDG Version 1 and additional guidelines that cover new functionalities [b-CDG 2010].	1.5
1.0	–	1.0	First released version of the CDG [b-CDG 1.0].	–

6 Test suite structure

The test purposes (TP) for the Personal Health Devices interface have been divided into the groups and subgroups specified below. Annex A describes the TPs for subgroup 2.4.3 (shown in bold).

- Group 1: Personal Health Device (PHD)
 - Group 1.1: Transport (TR)
 - Subgroup 1.1.1: Design guidelines: Common (DGC)
 - Subgroup 1.1.2: USB design guidelines (UDG)
 - Subgroup 1.1.3: Bluetooth design guidelines (BDG)
 - Subgroup 1.1.4: Pulse oximeter design guidelines (PODG)
 - Subgroup 1.1.5: Cardiovascular design guidelines (CVDG)
 - Subgroup 1.1.6: Activity hub design guidelines (HUBDG)
 - Subgroup 1.1.7: ZigBee design guidelines (ZDG)
 - Subgroup 1.1.8: Glucose meter design guidelines (GLDG)
 - Subgroup 1.1.9: Bluetooth low energy design guidelines (BLEDDG)
 - Subgroup 1.1.10: Basic electrocardiograph design guidelines (ECGDG)
 - Subgroup 1.1.11: NFC design guidelines (NDG)
 - Group 1.2: IEEE 20601 Optimized exchange protocol (OXP)
 - Subgroup 1.2.1: PHD domain information model (DIM)
 - Subgroup 1.2.2: PHD service model (SER)
 - Subgroup 1.2.3: PHD communication model (COM)
 - Group 1.3: Devices class specializations (CLASS)
 - Subgroup 1.3.1: Weighing scales (WEG)
 - Subgroup 1.3.2: Glucose meter (GL)
 - Subgroup 1.3.3: Pulse oximeter (PO)
 - Subgroup 1.3.4: Blood pressure monitor (BPM)
 - Subgroup 1.3.5: Thermometer (TH)
 - Subgroup 1.3.6: Cardiovascular (CV)
 - Subgroup 1.3.7: Strength (ST)
 - Subgroup 1.3.8: Activity hub (HUB)
 - Subgroup 1.3.9: Adherence monitor (AM)
 - Subgroup 1.3.10: Insulin pump (IP)
 - Subgroup 1.3.11: Peak flow (PF)
 - Subgroup 1.3.12: Body composition analyser (BCA)
 - Subgroup 1.3.13: Basic electrocardiograph (ECG)
 - Subgroup 1.3.14: International normalized ratio (INR)
 - Subgroup 1.3.15: Sleep apnoea breathing therapy equipment (SABTE)
 - Subgroup 1.3.16: Continuous glucose monitor (CGM)
 - Group 1.4: Personal health device transcoding whitepaper (PHDTW)
 - Subgroup 1.4.1: Whitepaper general requirements (GEN)
 - Subgroup 1.4.2: Whitepaper thermometer requirements (TH)
 - Subgroup 1.4.3: Whitepaper blood pressure requirements (BPM)

- Subgroup 1.4.4: Whitepaper heart rate requirements (HR)
 - Subgroup 1.4.5: Whitepaper glucose meter requirements (GL)
 - Subgroup 1.4.6: Whitepaper weight scale requirements (WS)
 - Subgroup 1.4.7: Whitepaper pulse oximeter requirements (PLX)
 - Subgroup 1.4.8: Whitepaper continuous glucose monitoring requirements (CGM)
- Group 2: Personal Health Gateway (PHG)
- Group 2.1: Transport (TR)
 - Subgroup 2.1.1: Design guidelines: Common (DGC)
 - Subgroup 2.1.2: USB design guidelines (UDG)
 - Subgroup 2.1.3: Bluetooth design guidelines (BDG)
 - Subgroup 2.1.4: Cardiovascular design guidelines (CVDG)
 - Subgroup 2.1.5: Activity hub design guidelines (HUBDG)
 - Subgroup 2.1.6: ZigBee design guidelines (ZDG)
 - Subgroup 2.1.7: Bluetooth low energy design guidelines (BLEDDG)
 - Subgroup 2.1.8: NFC design guidelines (NDG)
 - Group 2.2: IEEE 20601 Optimized exchange protocol (OXP)
 - Subgroup 2.2.1: General (GEN)
 - Subgroup 2.2.2: PHD domain information model (DIM)
 - Subgroup 2.2.3: PHD service model (SER)
 - Subgroup 2.2.4: PHD communication model (COM)
 - Group 2.3: Devices class specializations (CLASS)
 - Subgroup 2.3.1: Weighing scales (WEG)
 - Subgroup 2.3.2: Glucose meter (GL)
 - Subgroup 2.3.3: Pulse oximeter (PO)
 - Subgroup 2.3.4: Blood pressure monitor (BPM)
 - Subgroup 2.3.5: Thermometer (TH)
 - Subgroup 2.3.6: Cardiovascular (CV)
 - Subgroup 2.3.7: Strength (ST)
 - Subgroup 2.3.8: Activity hub (HUB)
 - Subgroup 2.3.9: Adherence monitor (AM)
 - Subgroup 2.3.10: Insulin pump (IP)
 - Subgroup 2.3.11: Peak flow (PF)
 - Subgroup 2.3.12: Body composition analyser (BCA)
 - Subgroup 2.3.13: Basic electrocardiograph (ECG)
 - Subgroup 2.3.14: International normalized ratio (INR)
 - Subgroup 2.3.15: Sleep apnoea breathing therapy equipment (SABTE)
 - Subgroup 2.3.16: Continuous glucose monitor (CGM)
 - Group 2.4: Personal health device transcoding whitepaper (PHDTW)
 - Subgroup 2.4.1: Whitepaper general requirements (GEN)
 - Subgroup 2.4.2: Whitepaper thermometer requirements (TH)
 - **Subgroup 2.4.3: Whitepaper blood pressure requirements (BPM)**

- Subgroup 2.4.4: Whitepaper heart rate requirements (HR)
- Subgroup 2.4.5: Whitepaper glucose meter requirements (GL)
- Subgroup 2.4.6: Whitepaper weight scale requirements (WS)
- Subgroup 2.4.7: Whitepaper pulse oximeter requirements (PLX)
- Subgroup 2.4.8: Whitepaper continuous glucose monitoring requirements (CGM)

7 Electronic attachment

The protocol implementation conformance statements (PICS) and the protocol implementation extra information for testing (PIXIT) required for the implementation of Annex A can be downloaded from <http://handle.itu.int/11.1002/2000/12067>.

In the electronic attachment, letters "C" and "I" in the column labelled "Mandatory" are used to distinguish between "PICS" and "PIXIT" respectively during testing. If the cell is empty, the corresponding PICS is "independent". If the field contains a "C", the corresponding PICS is dependent on other PICS, and the logical expression is detailed in the "SCR_Expression" field. The static conformance review (SCR) is used in the test tool to assert whether the PICS selection is consistent.

Annex A

Test purposes

(This annex forms an integral part of this Recommendation.)

A.1 TP definition conventions

The test purposes (TPs) are defined according to the following rules:

- **TP Id:** This is a unique identifier (TP/<TT>/<DUT>/<GR>/<SGR>/<XX> – <NNN>). It is specified according to the naming convention defined below:
 - Each test purpose identifier is introduced by the prefix "TP".
 - <TT>: This is the test tool that will be used in the test case.
 - PAN: Personal area network (Bluetooth or USB)
 - LAN: Local area network (ZigBee)
 - PAN-LAN: Personal area network (Bluetooth or USB) – Local area network (ZigBee)
 - LP-PAN: Low power personal area network (Bluetooth low energy)
 - TAN: Touch area network (NFC)
 - PLT: Personal area network (Bluetooth or USB) – Local area network (ZigBee) – Touch area network (NFC)
 - <DUT>: This is the device under test.
 - PHD: Personal Health Device
 - PHG: Personal Health Gateway
 - <GR>: This identifies a group of test cases.
 - <SGR>: This identifies a subgroup of test cases.
 - <XX>: This identifies the type of testing.
 - BV: Valid behaviour test
 - BI: Invalid behaviour test
 - <NNN>: This is a sequential number that identifies a test purpose.
- **TP label:** This is the TP's title.
- **Coverage:** This contains the specification reference and clause to be checked by the TP.
 - Spec: This indicates the earliest version of the specification from which the testable items to be checked by the TP were included.
 - Testable item: This contains the testable items to be checked by the TP.
- **Test purpose:** This is a description of the requirements to be tested.
- **Applicability:** This contains the PICS items that define if the test case is applicable or not for a specific device. When a TP contains an "ALL" in this field it means that it applies to the device under test within that scope of the test (specialization, transport used, etc.).
- **Other PICS:** This contains additional PICS items (apart from the PICS specified in the Applicability row) which are used within the test case implementation and can modify the final verdict. When this row is empty, it means that only the PICS specified in the Applicability row are used within the test case implementation.
- **Initial condition:** This indicates the state to which the DUT needs to be moved at the beginning of TC execution.

- **Test procedure:** This describes the steps to be followed in order to execute the test case.
- **Pass/Fail criteria:** This provides criteria to decide whether the DUT passes or fails the test case.

A.2 Subgroup 2.4.3 – Whitepaper Blood pressure requirements (BP)

TP Id		TP/LP-PAN/PHG/PHDTW/BPM/BV-000		
TP label		Whitepaper. Blood Pressure MDS Object - System-Type Attribute		
Coverage	Spec	[b-Bluetooth PHDT v1.3]		
	Testable items	BPM Specific MDS 1; M		
Test purpose		Check that: PHG does not include MDS Object – System-Type attribute in transcoder output.		
Applicability		C_MAN_BLE_000 AND C_MAN_BLE_002 AND C_MAN_BLE_003		
Other PICS				
Initial condition		The PHG under test and the simulated PHD are in the Standby state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated PHD is configured with a Blood pressure profile (device specialization); it has a measurement ready to be sent and it is in the Advertising state (it is discoverable). 2. The PHG under test initiates a discovery process (Scanning state). It discovers the simulated PHD and it starts a pairing process with the simulated PHD (Initiating state). 3. When the pairing has been completed (Connection state), the simulated PHD sends the measurement to the PHG under test. 4. Check in PHG transcoder output for the MDS object – System-Type attribute. 		
Pass/Fail criteria		In step 4, the MDS object – System-Type attribute is not present.		
Notes		<p>Possible values in typical points of observation after transcoder output are:</p> <p>a) IEEE 11073 Objects and Attributes</p> <p>System-Type attribute is not present:</p> <ul style="list-style-type: none"> <input type="checkbox"/> Object: MDS object <input type="checkbox"/> Attribute-id: MDC_ATTR_SYS_TYPE (2438) <input type="checkbox"/> Attribute-type: TYPE <input type="checkbox"/> Attribute-value: <NOT PRESENT> <p>b) WAN PCD-01 message</p> <p>PCD-01 message does not include segments with a System-Type attribute value (67974^MDC_ATTR_SYS_TYPE^MDC).</p>		

TP Id		TP/LP-PAN/PHG/PHDTW/BPM/BV-001		
TP label		Whitepaper. Blood Pressure MDS Object - Dev-Configuration-Id Attribute		
Coverage	Spec	[b-Bluetooth PHDT v1.3]		
	Testable items	BPM Specific MDS 2; M		
Test purpose		Check that: PHG includes MDS Object – Dev-Configuration-Id attribute in transcoder output. [AND] Dev-Configuration-Id value is set to any value in range of 0x4000 to 0x7FFF (Extended Configuration)		
Applicability		C_MAN_BLE_000 AND C_MAN_BLE_002 AND C_MAN_BLE_003		
Other PICS				
Initial condition		The PHG under test and the simulated PHD are in the Standby state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated PHD is configured with a Blood pressure profile (device specialization); it has a measurement ready to be sent and it is in the Advertising state (it is discoverable). 		

	<ol style="list-style-type: none"> 2. The PHG under test initiates a discovery process (Scanning state). It discovers the simulated PHD and it starts a pairing process with the simulated PHD (Initiating state). 3. When the pairing has been completed (Connection state), the simulated PHD sends the measurement to the PHG under test. 4. Check in PHG transcoder output for the MDS object – Dev-Configuration-Id attribute.
Pass/Fail criteria	In step 4, the MDS object – Dev-Configuration-Id attribute is present and its value is inside the range 0x4000 - 0x7FFF.
Notes	<p>Possible values in typical points of observation after transcoder output are:</p> <ol style="list-style-type: none"> a) IEEE 11073 Objects and Attributes Dev-Configuration-Id attribute is present: <ul style="list-style-type: none"> <input type="checkbox"/> Object: MDS object <input type="checkbox"/> Attribute-id: MDC_ATTR_DEV_CONFIG_ID (2628) <input type="checkbox"/> Attribute-type: INT-U16 <input type="checkbox"/> Attribute-value: Any value inside the range 16384 - 32767 (dec) or 0x4000 – 0x7FFF (hex) b) WAN PCD-01 message According to [b-ITU-T H.810 (2013)], the Dev-Configuration-Id shall not be transmitted in the PCD-01 message; therefore it is not possible to check this attribute.

TP Id	TP/LP-PAN/PHG/PHDTW/BPM/BV-002		
TP label	Whitepaper. Blood Pressure MDS Object - System-Type-Spec-List Attribute		
Coverage	Spec	[b-Bluetooth PHDT v1.3]	
	Testable items	Common MDS 15; M	BPM Specific MDS 3; M
Test purpose	<p>Check that:</p> <p>PHG includes MDS Object – System-Type-Spec-List attribute in transcoder output.</p> <p>[AND]</p> <p>System-Type-Spec-List is set to (MDC_DEV_SPEC_PROFILE_BP, Version 1)</p>		
Applicability	C_MAN_BLE_000 AND C_MAN_BLE_002 AND C_MAN_BLE_003		
Other PICS			
Initial condition	The PHG under test and the simulated PHD are in the Standby state.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated PHD is configured with a Blood pressure profile (device specialization); it has a measurement ready to be sent and it is in the Advertising state (it is discoverable). 2. The PHG under test initiates a discovery process (Scanning state). It discovers the simulated PHD and it starts a pairing process with the simulated PHD (Initiating state). 3. When the pairing has been completed (Connection state), the simulated PHD sends the measurement to the PHG under test. 4. Check in PHG transcoder output for the MDS object – System-Type-Spec-List attribute. 		
Pass/Fail criteria	In step 4, the MDS object – System-Type-Spec-List attribute is present and its value is (MDC_DEV_SPEC_PROFILE_BP, Version 1).		
Notes	<p>Possible values in typical points of observation after transcoder output are:</p> <ol style="list-style-type: none"> a) IEEE 11073 Objects and Attributes System-Type-Spec-List attribute is present: <ul style="list-style-type: none"> <input type="checkbox"/> Object: MDS object <input type="checkbox"/> Attribute-id: MDC_ATTR_SYS_TYPE_SPEC_LIST (2650) <input type="checkbox"/> Attribute-type: SEQUENCE OF [{type (INT-U16), version (INT-U16)}] <input type="checkbox"/> Attribute-value: 		

	<ul style="list-style-type: none"> • type: MDC_DEV_SPEC_PROFILE_BP or 4103 (dec) or 10 07 (hex) • version: 1 (dec) or 00 01 (hex) <p>b) WAN PCD-01 message</p> <p>PCD-01 message includes a segment like this with System-Type-Spec-List attribute value (check OBX-5):</p> <p style="text-align: center;">OBX ? NM 68186^MDC_ATTR_SYS_TYPE_SPEC_LIST^MDC 1.0.0.a 528391^MDC_DEV_SPEC_PROFILE_BP ^MDC R</p>
--	--

TP Id		TP/LP-PAN/PHG/PHDTW/BPM/BV-003		
TP label		Whitepaper. Blood Pressure MDS Object - Reg-Cert-Data-List Attribute		
Coverage	Spec	[b-Bluetooth PHDT v1.3]		
	Testable items	Common MDS 14; M	Regulatory Conv 1; M	
Test purpose		<p>Check that:</p> <p>PHG transcodes IEEE 11073-20601 Regulatory Certification Data List characteristic into MDS Object – Reg-Cert-Data-List attribute</p>		
Applicability		C_MAN_BLE_000 AND C_MAN_BLE_002 AND C_MAN_BLE_003		
Other PICS				
Initial condition		The PHG under test and the simulated PHD are in the Standby state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated PHD is configured with a Blood pressure profile (device specialization) supported by the PHG under test; it has a measurement ready to be sent and it is in the Advertising state (it is discoverable). 2. The simulated PHD implements several BLE characteristics. The characteristic of interest for this test case is: <ol style="list-style-type: none"> a. IEEE 11073-20601 Regulatory Certification Data List (0x2A2A) <ul style="list-style-type: none"> • Format: reg-cert-data-list (opaque structure) • Value: 00 02 00 12 02 01 00 08 08 00 00 01 00 02 80 07 02 02 00 02 80 00 (hex) <ol style="list-style-type: none"> i. Element: <ul style="list-style-type: none"> • auth-body-and-struct-type: <ul style="list-style-type: none"> - auth-body: 02 (hex) auth-body-continua(2) - auth-body-struct-type: 01 (hex). continua-version-struct(1) • auth-body-data: <ul style="list-style-type: none"> - major-IG-version: 08 (hex) - minor-IG-version: 00 (hex) - certified-devices: 80 07 (hex). BLE Blood Pressure ii. Element: <ul style="list-style-type: none"> • auth-body-and-struct-type: <ul style="list-style-type: none"> - auth-body: 02 (hex). auth-body-continua(2) - auth-body-struct-type: 02 (hex). continua-reg-struct(2) • auth-body-data: <ul style="list-style-type: none"> - regulation-bit-field: 80 00 (hex). Unregulated device 3. The PHG under test initiates a discovery process (Scanning state). It discovers the simulated PHD and it starts a pairing process with the simulated PHD. 4. When the pairing has been completed (Connection state), force the PHG under test to read the IEEE 11073-20601 Regulatory Certification Data List characteristic. 5. The simulated PHD sends the measurement to the PHG under test. 		

	6. Check in PHG transcoder output for the MDS object – Reg-Cert-Data-List attribute.
Pass/Fail criteria	In step 6, the MDS object – Reg-Cert-Data-List attribute is present and its value matches with the IEEE 11073-20601 Regulatory Certification Data List characteristic value.
Notes	<p>Possible values in typical points of observation after transcoder output are:</p> <p>a) IEEE 11073 Objects and Attributes</p> <p>Reg-Cert-Data-List attribute is present:</p> <ul style="list-style-type: none"> ❑ Object: MDS object ❑ Attribute-id: MDC_ATTR_REG_CERT_DATA_LIST (2635) ❑ Attribute-type: SEQUENCE OF [{auth-body-and-struct-type, auth-body-data}, {...}] ❑ Attribute-value: 00 02 00 12 02 01 00 08 08 00 00 01 00 02 80 07 02 02 00 02 80 00 (hex) [Note that 0x00 0x02 is the number of elements in the sequence and 0x00 12 is the length of the sequence] <ul style="list-style-type: none"> i. Reg-Cert-Data Element: <ul style="list-style-type: none"> • auth-body-and-struct-type: <ul style="list-style-type: none"> - auth-body: 02 (hex) auth-body-continua(2) - auth-body-struct-type: 01 (hex). continua-version-struct(1) • auth-body-data: <ul style="list-style-type: none"> - major-IG-version: 08 (hex) - minor-IG-version: 00 (hex) - certified-devices: 80 07 (hex). BLE Blood Pressure ii. Reg-Cert-Data Element: <ul style="list-style-type: none"> • auth-body-and-struct-type: <ul style="list-style-type: none"> - auth-body: 02 (hex). auth-body-continua(2) - auth-body-struct-type: 02 (hex). continua-reg-struct(2) • auth-body-data: <ul style="list-style-type: none"> - regulation-bit-field: 80 00 (hex). Unregulated device <p>b) WAN PCD-01 message</p> <p>PCD-01 message includes five segments like these with Reg-Cert-Data-List attribute value (check OBX-5 in five segments):</p> <pre>OBX ? CWE 68218^MDC_REG_CERT_DATA_AUTH_BODY^MDC 1.0.0.a 2^auth-body-continua R</pre> <pre>OBX ? ST 532352^MDC_REG_CERT_DATA_CONTINUA_VERSION^MDC 1.0.0.a.x 6.1 R</pre> <pre>OBX ? NA 532353^MDC_REG_CERT_DATA_CONTINUA_CERT_DEV_LIST^MDC 1.0.0.a.y 32775 R</pre> <pre>OBX ? CWE 68218^MDC_REG_CERT_DATA_AUTH_BODY^MDC 1.0.0.b 2^auth-body-continua R</pre> <pre>OBX ? CWE 532354^MDC_REG_CERT_DATA_CONTINUA_REG_STATUS^MDC 1.0.0.b.z 1^unregulated-device(0) R</pre>

TP Id	TP/LP-PAN/PHG/PHDTW/BPM/BV-004		
TP label	Whitepaper. Systolic/Diastolic/Map Compound Numeric Object - Handle Attribute		
Coverage	Spec	[b-Bluetooth PHDT v1.3]	
	Testable items	BP Numeric 1; O	
Test purpose	<p>Check that:</p> <p>PHG does not include Systolic/Diastolic/Map Compound Numeric Object – Handle Attribute in transcoder output</p>		

	[OR] If PHG includes Systolic/Diastolic/Map Compound Numeric Object – Handle attribute in transcoder output, then its value shall be different than 0
Applicability	C_MAN_BLE_000 AND C_MAN_BLE_002 AND C_MAN_BLE_003
Other PICS	
Initial condition	The PHG under test and the simulated PHD are in the Standby state.
Test procedure	<ol style="list-style-type: none"> 1. The simulated PHD is configured with a Blood pressure profile (device specialization); it has a measurement ready to be sent and it is in the Advertising state (it is discoverable). 2. The simulated PHD implements several BLE characteristics. The characteristic of interest for this test case is: <ol style="list-style-type: none"> a. Blood pressure measurement (0x2A35) <ol style="list-style-type: none"> i. Field: Flags <ul style="list-style-type: none"> • Format: 8 bit • Value: 0000 0010 (MSB → LSB). Blood pressure measurement value in units of mmHg and Time Stamp fields are included, Pulse Rate, User ID and measurement Status fields are not included ii. Field: Blood Pressure Measurement Compound Value – Systolic (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant iii. Field: Blood Pressure Measurement Compound Value – Diastolic (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant iv. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant v. Field: Blood Pressure Measurement Compound Value – Systolic (kPa) <ul style="list-style-type: none"> • This field is not included vi. Field: Blood Pressure Measurement Compound Value – Diastolic (kPa) <ul style="list-style-type: none"> • This field is not included vii. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (kPa) <ul style="list-style-type: none"> • This field is not included viii. Field: Time Stamp <ul style="list-style-type: none"> • Format: Date and Time • Value: Not relevant ix. Field: Pulse Rate <ul style="list-style-type: none"> • This field is not included x. Field: User ID <ul style="list-style-type: none"> • This field is not included xi. Field: Measurement Status <ul style="list-style-type: none"> • This field is not included 3. The PHG under test initiates a discovery process (Scanning state). It discovers the simulated PHD and it starts a pairing process with the simulated PHD (Initiating state). 4. When the pairing has been completed (Connection state), the simulated PHD sends the measurement to the PHG under test. 5. Check in PHG transcoder output for the Systolic/Diastolic/Map compound numeric object

	– Handle attribute.
Pass/Fail criteria	In step 5, the Systolic/Diastolic/Map compound numeric object – Handle attribute is not present; however, if it is present then its value is different to 0.
Notes	<p>Possible values in typical points of observation after transcoder output are:</p> <p>a) IEEE 11073 Objects and Attributes</p> <p>Handle attribute is not present, or if it is present then:</p> <ul style="list-style-type: none"> <input type="checkbox"/> Object: Systolic/Diastolic/Map compound numeric object <input type="checkbox"/> Attribute-id: MDC_ATTR_ID_HANDLE (2337) <input type="checkbox"/> Attribute-type: INT-U16 <input type="checkbox"/> Attribute-value: Any value other than 0 <p>b) WAN PCD-01 message</p> <p>PCD-01 message does not include segments with a Handle attribute value.</p>

TP Id	TP/LP-PAN/PHG/PHDTW/BPM/BV-005		
TP label	Whitepaper. Systolic/Diastolic/Map Compound Numeric Object - Type Attribute		
Coverage	Spec	[b-Bluetooth PHDT v1.3]	
	Testable items	BP Numeric 2; M	
Test purpose	<p>Check that:</p> <p>PHG includes Systolic/Diastolic/Map Compound Numeric Object – Type attribute in transcoder output.</p> <p>[AND]</p> <p>Type is set to {MDC_PART_SCADA, MDC_PRESS_BLD_NONINV}</p>		
Applicability	C_MAN_BLE_000 AND C_MAN_BLE_002 AND C_MAN_BLE_003		
Other PICS			
Initial condition	The PHG under test and the simulated PHD are in the Standby state.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated PHD is configured with a Blood pressure profile (device specialization); it has a measurement ready to be sent and it is in the Advertising state (it is discoverable). 2. The simulated PHD implements several BLE characteristics. The characteristic of interest for this test case is: <ol style="list-style-type: none"> a. Blood pressure measurement (0x2A35) <ol style="list-style-type: none"> i. Field: Flags <ul style="list-style-type: none"> • Format: 8 bit • Value: 0000 0010 (MSB → LSB). Blood pressure measurement value in units of mmHg and Time Stamp fields are included, Pulse Rate, User ID and measurement Status fields are not included ii. Field: Blood Pressure Measurement Compound Value – Systolic (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant iii. Field: Blood Pressure Measurement Compound Value – Diastolic (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant iv. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant v. Field: Blood Pressure Measurement Compound Value – Systolic (kPa) 		

	<ul style="list-style-type: none"> • This field is not included <p>vi. Field: Blood Pressure Measurement Compound Value – Diastolic (kPa)</p> <ul style="list-style-type: none"> • This field is not included <p>vii. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (kPa)</p> <ul style="list-style-type: none"> • This field is not included <p>viii. Field: Time Stamp</p> <ul style="list-style-type: none"> • Format: Date and Time • Value: Not relevant <p>ix. Field: Pulse Rate</p> <ul style="list-style-type: none"> • This field is not included <p>x. Field: User ID</p> <ul style="list-style-type: none"> • This field is not included <p>xi. Field: Measurement Status</p> <ul style="list-style-type: none"> • This field is not included <p>3. The PHG under test initiates a discovery process (Scanning state). It discovers the simulated PHD and it starts a pairing process with the simulated PHD (Initiating state).</p> <p>4. When the pairing has been completed (Connection state), the simulated PHD sends the measurement to the PHG under test.</p> <p>5. Check in PHG transcoder output for the Systolic/Diastolic/Map compound numeric object – Type attribute.</p>
Pass/Fail criteria	In step 5, the Systolic/Diastolic/Map compound numeric object – Type attribute is present and its value is {MDC_PART_SCADA, MDC_PRESS_BLD_NONINV}.
Notes	<p>Possible values in typical points of observation after transcoder output are:</p> <p>a) IEEE 11073 Objects and Attributes</p> <p>Type attribute is present:</p> <ul style="list-style-type: none"> <input type="checkbox"/> Object: Systolic/Diastolic/Map compound numeric object <input type="checkbox"/> Attribute-id: MDC_ATTR_ID_TYPE (2351) <input type="checkbox"/> Attribute-type: SEQUENCE {partition (INT-U16), code (INT-U16)} <input type="checkbox"/> Attribute-value: <ul style="list-style-type: none"> • partition: MDC_PART_SCADA or 2 (dec) or 00 02 (hex) • code: MDC_PRESS_BLD_NONINV or 18948 (dec) or 4A 04 (hex) <p>b) WAN PCD-01 message</p> <p>PCD-01 message includes a segment like this with a Type attribute (check OBX-3):</p> <pre>OBX ? 150020^MDC_PRESS_BLD_NONINV^MDC 1.0.a X [current_date_time].</pre>

TP Id	TP/LP-PAN/PHG/PHDTW/BPM/BV-006		
TP label	Whitepaper. Systolic/Diastolic/Map Compound Numeric Object - Metric-Spec-Small Attribute		
Coverage	Spec	[b-Bluetooth PHDT v1.3]	
	Testable items	BP Numeric 3; M	
Test purpose	<p>Check that:</p> <p>PHG includes Systolic/Diastolic/Map Compound Numeric Object – Metric-Spec-Small attribute in transcoder output.</p> <p>[AND]</p>		

	Metric-Spec-Small is set to {0xF040} (mss-avail-intermittent mss-avail-stored-data mss-upd-aperiodic mss-msmt-aperiodic mss-acc-agent-initiated).
Applicability	C_MAN_BLE_000 AND C_MAN_BLE_002 AND C_MAN_BLE_003
Other PICS	
Initial condition	The PHG under test and the simulated PHD are in the Standby state.
Test procedure	<ol style="list-style-type: none"> 1. The simulated PHD is configured with a Blood pressure profile (device specialization); it has a measurement ready to be sent and it is in the Advertising state (it is discoverable). 2. The simulated PHD implements several BLE characteristics. The characteristic of interest for this test case is: <ol style="list-style-type: none"> a. Blood pressure measurement (0x2A35) <ol style="list-style-type: none"> i. Field: Flags <ul style="list-style-type: none"> • Format: 8 bit • Value: 0000 0010 (MSB → LSB). Blood pressure measurement value in units of mmHg and Time Stamp fields are included, Pulse Rate, User ID and measurement Status fields are not included ii. Field: Blood Pressure Measurement Compound Value – Systolic (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant iii. Field: Blood Pressure Measurement Compound Value – Diastolic (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant iv. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant v. Field: Blood Pressure Measurement Compound Value – Systolic (kPa) <ul style="list-style-type: none"> • This field is not included vi. Field: Blood Pressure Measurement Compound Value – Diastolic (kPa) <ul style="list-style-type: none"> • This field is not included vii. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (kPa) <ul style="list-style-type: none"> • This field is not included viii. Field: Time Stamp <ul style="list-style-type: none"> • Format: Date and Time • Value: Not relevant ix. Field: Pulse Rate <ul style="list-style-type: none"> • This field is not included x. Field: User ID <ul style="list-style-type: none"> • This field is not included xi. Field: Measurement Status <ul style="list-style-type: none"> • This field is not included 3. The PHG under test initiates a discovery process (Scanning state). It discovers the simulated PHD and it starts a pairing process with the simulated PHD (Initiating state). 4. When the pairing has been completed (Connection state), the simulated PHD sends the measurement to the PHG under test. 5. Check in PHG transcoder output for the Systolic/Diastolic/Map compound numeric object, Metric-Spec-Small attribute.

Pass/Fail criteria	In step 5, the Systolic/Diastolic/Map compound numeric object – Metric-Spec-Small attribute is present and its value is {0xF040} (mss-avail-intermittent mss-avail-stored-data mss-upd-aperiodic mss-msmt-aperiodic mss-acc-agent-initiated).
Notes	<p>Possible values in typical points of observation after transcoder output are:</p> <p>a) IEEE 11073 Objects and Attributes</p> <p>Metric-Spec-Small attribute is present:</p> <ul style="list-style-type: none"> <input type="checkbox"/> Object: Systolic/Diastolic/Map compound numeric object <input type="checkbox"/> Attribute-id: MDC_ATTR_METRIC_SPEC_SMALL (2630) <input type="checkbox"/> Attribute-type: BITS-16 <input type="checkbox"/> Attribute-value: F0 40 (hex) or BITS mss-avail-intermittent(0), mss-avail-stored-data(1), mss-upd-aperiodic(2), mss-msmt-aperiodic(3), mss-acc-agent-initiated(9) set to TRUE and remaining BITS set to FALSE <p>b) WAN PCD-01 message</p> <p>PCD-01 message does not include segments with a Metric-Spec-Small attribute value.</p>

TP Id	TP/LP-PAN/PHG/PHDTW/BPM/BV-007		
TP label	Whitepaper. Systolic/Diastolic/Map Compound Numeric Object - Metric-Structure-Small Attribute		
Coverage	Spec	[b-Bluetooth PHDT v1.3]	
	Testable items	BP Numeric 4; M	
Test purpose	<p>Check that:</p> <p>PHG includes Systolic/Diastolic/Map Compound Numeric Object – Metric-Structure-Small attribute in transcoder output.</p> <p>[AND]</p> <p>Metric-Structure-Small is set to {0x03, 0x03} (ms-struct-compound-fix, 3).</p>		
Applicability	C_MAN_BLE_000 AND C_MAN_BLE_002 AND C_MAN_BLE_003		
Other PICS			
Initial condition	The PHG under test and the simulated PHD are in the Standby state.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated PHD is configured with a Blood pressure profile (device specialization); it has a measurement ready to be sent and it is in the Advertising state (it is discoverable). 2. The simulated PHD implements several BLE characteristics. The characteristic of interest for this test case is: <ol style="list-style-type: none"> a. Blood pressure measurement (0x2A35) <ol style="list-style-type: none"> i. Field: Flags <ul style="list-style-type: none"> • Format: 8 bit • Value: 0000 0010 (MSB → LSB). Blood pressure measurement value in units of mmHg and Time Stamp fields are included, Pulse Rate, User ID and measurement Status fields are not included ii. Field: Blood Pressure Measurement Compound Value – Systolic (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant iii. Field: Blood Pressure Measurement Compound Value – Diastolic (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant iv. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT 		

	<ul style="list-style-type: none"> • Value: Not relevant <p>v. Field: Blood Pressure Measurement Compound Value – Systolic (kPa)</p> <ul style="list-style-type: none"> • This field is not included <p>vi. Field: Blood Pressure Measurement Compound Value – Diastolic (kPa)</p> <ul style="list-style-type: none"> • This field is not included <p>vii. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (kPa)</p> <ul style="list-style-type: none"> • This field is not included <p>viii. Field: Time Stamp</p> <ul style="list-style-type: none"> • Format: Date and Time • Value: Not relevant <p>ix. Field: Pulse Rate</p> <ul style="list-style-type: none"> • This field is not included <p>x. Field: User ID</p> <ul style="list-style-type: none"> • This field is not included <p>xi. Field: Measurement Status</p> <ul style="list-style-type: none"> • This field is not included <p>3. The PHG under test initiates a discovery process (Scanning state). It discovers the simulated PHD and it starts a pairing process with the simulated PHD (Initiating state).</p> <p>4. When the pairing has been completed (Connection state), the simulated PHD sends the measurement to the PHG under test.</p> <p>5. Check in PHG transcoder output for the Systolic/Diastolic/Map compound numeric object – Metric-Structure-Small attribute.</p>
Pass/Fail criteria	In step 5, the Systolic/Diastolic/Map compound numeric object – Metric-Structure-Small attribute is present and its value is {0x03, 0x03} (ms-struct-compound-fix, 3).
Notes	<p>Possible values in typical points of observation after transcoder output are:</p> <p>a) IEEE 11073 Objects and Attributes</p> <p>Metric-Spec-Small attribute is present:</p> <ul style="list-style-type: none"> <input type="checkbox"/> Object: Systolic/Diastolic/Map compound numeric object <input type="checkbox"/> Attribute-id: MDC_ATTR_METRIC_STRUCT_SMALL (2675) <input type="checkbox"/> Attribute-type: SEQUENCE {ms-struct (INT-U8), ms-comp-no (INT-U8)} <input type="checkbox"/> Attribute-value: <ul style="list-style-type: none"> • ms-struct Element: 03 (hex), ms-struct-compound-fix(3) • ms-comp-no Element: 03 (hex) <p>b) WAN PCD-01 message</p> <p>PCD-01 message does not include segments with a Metric-Structure-Small attribute value.</p>

TP Id	TP/LP-PAN/PHG/PHDTW/BPM/BV-008		
TP label	Whitepaper. Systolic/Diastolic/Map Compound Numeric Object - Metric-Id-List Attribute		
Coverage	Spec	[b-Bluetooth PHDT v1.3]	
	Testable items	BP Numeric 5; M	
Test purpose	Check that: PHG includes Systolic/Diastolic/Map Compound Numeric Object – Metric-Id-List attribute in transcoder output.		

	[AND] Metric-Id-List is set to {MDC_PRESS_BLD_NONINV_SYS, MDC_PRESS_BLD_NONINV_DIA, MDC_PRESS_BLD_NONINV_MEAN}.
Applicability	C_MAN_BLE_000 AND C_MAN_BLE_002 AND C_MAN_BLE_003
Other PICS	
Initial condition	The PHG under test and the simulated PHD are in the Standby state.
Test procedure	<ol style="list-style-type: none"> 1. The simulated PHD is configured with a Blood pressure profile (device specialization); it has a measurement ready to be sent and it is in the Advertising state (it is discoverable). 2. The simulated PHD implements several BLE characteristics. The characteristic of interest for this test case is: <ol style="list-style-type: none"> a. Blood pressure measurement (0x2A35) <ol style="list-style-type: none"> i. Field: Flags <ul style="list-style-type: none"> • Format: 8 bit • Value: 0000 0010 (MSB → LSB). Blood pressure measurement value in units of mmHg and Time Stamp fields are included, Pulse Rate, User ID and measurement Status fields are not included ii. Field: Blood Pressure Measurement Compound Value – Systolic (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant iii. Field: Blood Pressure Measurement Compound Value – Diastolic (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant iv. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant v. Field: Blood Pressure Measurement Compound Value – Systolic (kPa) <ul style="list-style-type: none"> • This field is not included vi. Field: Blood Pressure Measurement Compound Value – Diastolic (kPa) <ul style="list-style-type: none"> • This field is not included vii. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (kPa) <ul style="list-style-type: none"> • This field is not included viii. Field: Time Stamp <ul style="list-style-type: none"> • Format: Date and Time • Value: Not relevant ix. Field: Pulse Rate <ul style="list-style-type: none"> • This field is not included x. Field: User ID <ul style="list-style-type: none"> • This field is not included xi. Field: Measurement Status <ul style="list-style-type: none"> • This field is not included 3. The PHG under test initiates a discovery process (Scanning state). It discovers the simulated PHD and it starts a pairing process with the simulated PHD (Initiating state). 4. When the pairing has been completed (Connection state), the simulated PHD sends the measurement to the PHG under test. 5. Check in PHG transcoder output for the Systolic/Diastolic/Map compound numeric object

	– Metric-Id-List attribute.
Pass/Fail criteria	In step 5, the Systolic/Diastolic/Map compound numeric object – Metric-Id-List attribute is present and its value is {MDC_PRESS_BLD_NONINV_SYS, MDC_PRESS_BLD_NONINV_DIA, MDC_PRESS_BLD_NONINV_MEAN}.
Notes	<p>Possible values in typical points of observation after transcoder output are:</p> <p>a) IEEE 11073 Objects and Attributes</p> <p>Metric-Id-List attribute is present:</p> <ul style="list-style-type: none"> <input type="checkbox"/> Object: Systolic/Diastolic/Map compound numeric object <input type="checkbox"/> Attribute-id: MDC_ATTR_ID_PHYSIO_LIST (2678) <input type="checkbox"/> Attribute-type: SEQUENCE OF [{OID-Type(INT-U16)}] <input type="checkbox"/> Attribute-value: 00 03 00 06 4A 05 4A 06 4A 07 (hex) [Note that 0x00 0x03 is the number of elements in the sequence and 0x00 06 is the length of the sequence] <ul style="list-style-type: none"> i. First Element: 4A 05 (hex) or 18949 (dec) ii. Second Element: 4A 06 (hex) or 18950 (dec) iii. Third Element: 4A 07 (hex) or 18951 (dec) <p>b) WAN PCD-01 message</p> <p>PCD-01 message includes three segments like these with a Metric-Id-List attribute values (check OBX-3 in three segments):</p> <pre>OBX ? NM 150021^MDC_PRESS_BLD_NONINV_SYS^MDC 1.0.a.x 100 266016^MDC_DIM_MMHG^MDC R OBX ? NM 150022^MDC_PRESS_BLD_NONINV_DIA^MDC 1.0.a.y 70 266016^MDC_DIM_MMHG^MDC R OBX ? NM 150023^MDC_PRESS_BLD_NONINV_MEAN^MDC 1.0.a.z 80 266016^MDC_DIM_MMHG^MDC R</pre>

TP Id	TP/LP-PAN/PHG/PHDTW/BP/BV-009		
TP label	Whitepaper. Systolic/Diastolic/Map Compound Numeric Object - Unit-Code Attribute		
Coverage	Spec	[b-Bluetooth PHDT v1.3]	
	Testable items	BP Numeric 6; M	BP Numeric 7; M
Test purpose	<p>Check that:</p> <p>PHG includes Systolic/Diastolic/Map Compound Numeric Object – Unit-Code attribute in transcoder output.</p> <p>[AND]</p> <p>IF Blood Pressure Measurement Compound Value - Systolic (mmHg), Diastolic (mmHg) and Mean Arterial Pressure (mmHg) fields of Blood Pressure Measurement characteristic are present THEN Systolic/Diastolic/Map Compound Numeric Object – Unit-Code attribute is set to MDC_DIM_MMHG</p> <p>[AND]</p> <p>IF Blood Pressure Measurement Compound Value - Systolic (kPa), Diastolic (kPa) and Mean Arterial Pressure (kPa) fields of Blood Pressure Measurement characteristic are present THEN Systolic/Diastolic/Map Compound Numeric Object – Unit-Code attribute is set to MDC_DIM_KILO_PASCAL</p>		
Applicability	C_MAN_BLE_000 AND C_MAN_BLE_002 AND C_MAN_BLE_003		
Other PICS			
Initial condition	The PHG under test and the simulated PHD are in the Standby state		
Test procedure	<ol style="list-style-type: none"> 1. The simulated PHD is configured with a profile (device specialization) supported by the PHG under test; it has a measurement ready to be sent and it is in the Advertising state (it is discoverable). 2. The simulated PHD implements several BLE characteristics. The characteristic of interest 		

for this test case is:

- a. Blood pressure measurement (0x2A35)
3. The PHG under test initiates a discovery process (Scanning state). It discovers the simulated PHD and it starts a pairing process with the simulated PHD (Initiating state).
4. When the pairing has been completed (Connection state), the simulated PHD sends the measurement to the PHG under test with the following value:
 - a. Blood pressure measurement (0x2A35)
 - i. Field: Flags
 - Format: 8 bit
 - Value: 0000 0010 (MSB → LSB). Blood pressure measurement value in units of mmHg and Time Stamp fields are included, Pulse Rate, User ID and measurement Status fields are not included
 - ii. Field: Blood Pressure Measurement Compound Value – Systolic (mmHg)
 - Format: SFLOAT
 - Value: 100.0
 - iii. Field: Blood Pressure Measurement Compound Value – Diastolic (mmHg)
 - Format: SFLOAT
 - Value: 70.0
 - iv. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (mmHg)
 - Format: SFLOAT
 - Value: 80.0
 - v. Field: Blood Pressure Measurement Compound Value – Systolic (kPa)
 - This field is not included
 - vi. Field: Blood Pressure Measurement Compound Value – Diastolic (kPa)
 - This field is not included
 - vii. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (kPa)
 - This field is not included
 - viii. Field: Time Stamp
 - Format: Date and Time
 - Value: Not relevant
 - ix. Field: Pulse Rate
 - This field is not included
 - x. Field: User ID
 - This field is not included
 - xi. Field: Measurement Status
 - This field is not included
 5. Check in PHG transcoder output for the Systolic/Diastolic/Map compound numeric object – Unit-Code attribute.
 6. The simulated PHD sends the measurement to the PHG under test with the following value:
 - a. Blood pressure measurement (0x2A35)
 - i. Field: Flags
 - Format: 8 bit
 - Value: 0000 0011 (MSB → LSB). Blood pressure measurement value in units of kPa and Time Stamp fields are included, Pulse Rate, User ID and measurement Status fields are not included

	<ul style="list-style-type: none"> ii. Field: Blood Pressure Measurement Compound Value – Systolic (mmHg) <ul style="list-style-type: none"> • This field is not included iii. Field: Blood Pressure Measurement Compound Value – Diastolic (mmHg) <ul style="list-style-type: none"> • This field is not included iv. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (mmHg) <ul style="list-style-type: none"> • This field is not included v. Field: Blood Pressure Measurement Compound Value – Systolic (kPa) <ul style="list-style-type: none"> • Format: SFLOAT • Value: 13.33 vi. Field: Blood Pressure Measurement Compound Value – Diastolic (kPa) <ul style="list-style-type: none"> • Format: SFLOAT • Value: 9.33 vii. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (kPa) <ul style="list-style-type: none"> • Format: SFLOAT • Value: 10.67 viii. Field: Time Stamp <ul style="list-style-type: none"> • Format: Date and Time • Value: Not relevant ix. Field: Pulse Rate <ul style="list-style-type: none"> • This field is not included x. Field: User ID <ul style="list-style-type: none"> • This field is not included xi. Field: Measurement Status <ul style="list-style-type: none"> • This field is not included <p>7. Check in PHG transcoder output for the Systolic/Diastolic/Map compound numeric object – Unit-Code attribute.</p>
Pass/Fail criteria	<p>In step 5, the Systolic/Diastolic/Map compound numeric object – Unit-Code attribute is present and its value is MDC_DIM_MMHG.</p> <p>In step 7, the Systolic/Diastolic/Map compound numeric object – Unit-Code attribute is present and its value is MDC_DIM_KILO_PASCAL.</p>
Notes	<p>In step 5, possible values in typical points of observation after transcoder output are:</p> <ul style="list-style-type: none"> a) IEEE 11073 Objects and Attributes <ul style="list-style-type: none"> Unit-Code attribute is present: <ul style="list-style-type: none"> <input type="checkbox"/> Object: Systolic/Diastolic/Map compound numeric object <input type="checkbox"/> Attribute-id: MDC_ATTR_UNIT_CODE (2454) <input type="checkbox"/> Attribute-type: INT-U16 <input type="checkbox"/> Attribute-value: MDC_DIM_MMHG or 3872 (dec) or 0F 20 (hex) b) WAN PCD-01 message <ul style="list-style-type: none"> PCD-01 message includes three segments like these with Unit-Code attribute value (check OBX-6 in three segments): <pre>OBX ? NM 150021^MDC_PRESS_BLD_NONINV_SYS^MDC 1.0.a.x 100 266016^MDC_DIM_MMHG^MDC R [[current_date_time]</pre> <pre>OBX ? NM 150022^MDC_PRESS_BLD_NONINV_DIA^MDC 1.0.a.y 70 266016^MDC_DIM_MMHG^MDC R [[current_date_time]</pre>

	<p>OBX ? NM 150023^MDC_PRESS_BLD_NONINV_MEAN^MDC 1.0.a.z 80 266016^MDC_DIM_MMHG^MDC R [current_date_time]</p> <p>Note that " [current_date_time]" is optional at the end of each segment because the date and time can be specified in the parent segment (OBX ? 150020^MDC_PRESS_BLD_NONINV^MDC 1.0.a X [current_date_time])</p> <p>In step 7, possible values in typical points of observation after transcoder output are:</p> <p>a) IEEE 11073 Objects and Attributes</p> <p>Unit-Code attribute is present:</p> <ul style="list-style-type: none"> <input type="checkbox"/> Object: Systolic/Diastolic/Map compound numeric object <input type="checkbox"/> Attribute-id: MDC_ATTR_UNIT_CODE (2454) <input type="checkbox"/> Attribute-type: INT-U16 <input type="checkbox"/> Attribute-value: MDC_DIM_KILO_PASCAL or 3843 (dec) or 0F 03 (hex) <p>b) WAN PCD-01 message</p> <p>PCD-01 message includes three segments like these with Unit-Code attribute value (check OBX-6 in three segments):</p> <p>OBX ? NM 150021^MDC_PRESS_BLD_NONINV_SYS^MDC 1.0.a.x 13.33 265987^MDC_DIM_KILO_PASCAL^MDC R [current_date_time]</p> <p>OBX ? NM 150022^MDC_PRESS_BLD_NONINV_DIA^MDC 1.0.a.y 9.33 265987^MDC_DIM_KILO_PASCAL^MDC R [current_date_time]</p> <p>OBX ? NM 150023^MDC_PRESS_BLD_NONINV_MEAN^MDC 1.0.a.z 10.67 265987^MDC_DIM_KILO_PASCAL^MDC R [current_date_time]</p> <p>Note that " [current_date_time]" is optional at the end of each segment because the date and time can be specified in the parent segment (OBX ? 150020^MDC_PRESS_BLD_NONINV^MDC 1.0.a X [current_date_time])</p>
--	---

TP Id		TP/LP-PAN/PHG/PHDTW/BPM/BV-010		
TP label		Whitepaper. Systolic/Diastolic/Map Compound Numeric Object - Absolute-Time-Stamp Attribute		
Coverage	Spec	[b-Bluetooth PHDT v1.3]		
	Testable items	BP Numeric 9; M	Date-Time Conv 2; M	Date-Time Conv 3; M
		Date-Time Conv 4; M	Date-Time Conv 5; M	
Test purpose		<p>Check that:</p> <p>PHG transcodes Time Stamp field of Blood Pressure Measurement characteristic into Systolic/Diastolic/Map Compound Numeric Object - Absolute-Time-Stamp attribute</p> <p>[AND]</p> <p>PHG transcodes the Bluetooth Time Stamp field format to Absolute Time format</p> <p>[AND]</p> <p>The fraction of seconds in Absolute Time at transcoder output is 0</p>		
Applicability		C_MAN_BLE_000 AND C_MAN_BLE_002 AND C_MAN_BLE_003		
Other PICS				
Initial condition		The PHG under test and the simulated PHD are in the Standby state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated PHD is configured with a profile (device specialization) supported by the PHG under test; it has a measurement ready to be sent and it is in the Advertising state (it is discoverable). 2. The simulated PHD implements several BLE characteristics. The characteristic of interest for this test case is: <ol style="list-style-type: none"> a. Blood pressure measurement (0x2A35) 3. The PHG under test initiates a discovery process (Scanning state). It discovers the simulated PHD and it starts a pairing process with the simulated PHD (Initiating state) 		

	<p>4. When the pairing has been completed (Connection state), the simulated PHD sends the measurement to the PHG under test with the following value:</p> <ol style="list-style-type: none"> a. Blood pressure measurement (0x2A35) <ol style="list-style-type: none"> i. Field: Flags <ul style="list-style-type: none"> • Format: 8 bit • Value: 0000 0010 (MSB → LSB). Blood pressure measurement value in units of mmHg and Time Stamp fields are included, Pulse Rate, User ID and measurement Status fields are not included ii. Field: Blood Pressure Measurement Compound Value – Systolic (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: 100.0 iii. Field: Blood Pressure Measurement Compound Value – Diastolic (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: 70.0 iv. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: 80.0 v. Field: Blood Pressure Measurement Compound Value – Systolic (kPa) <ul style="list-style-type: none"> • This field is not included vi. Field: Blood Pressure Measurement Compound Value – Diastolic (kPa) <ul style="list-style-type: none"> • This field is not included vii. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (kPa) <ul style="list-style-type: none"> • This field is not included viii. Field: Time Stamp <ul style="list-style-type: none"> • Format: Date and Time • Value: August 2nd, 2012, 10:39:27 ix. Field: Pulse Rate <ul style="list-style-type: none"> • This field is not included x. Field: User ID <ul style="list-style-type: none"> • This field is not included xi. Field: Measurement Status <ul style="list-style-type: none"> • This field is not included <p>5. Check in PHG transcoder output for the Systolic/Diastolic/Map compound numeric object – Absolute-Time-Stamp attribute</p>
Pass/Fail criteria	<p>In step 5, the Systolic/Diastolic/Map compound numeric object – Absolute-Time-Stamp attribute is present, its value matches with the Time Stamp field of the Blood pressure measurement characteristic and the fraction of seconds is set to 0.</p>
Notes	<p>Possible values in typical points of observation after transcoder output are:</p> <ol style="list-style-type: none"> a) IEEE 11073 Objects and Attributes <ul style="list-style-type: none"> Absolute-Time-Stamp attribute is present: <ul style="list-style-type: none"> <input type="checkbox"/> Object: Systolic/Diastolic/Map compound numeric object <input type="checkbox"/> Attribute-id: MDC_ATTR_TIME_STAMP_ABS (2448) <input type="checkbox"/> Attribute-type: SEQUENCE {century (INT-U8), year (INT-U8), month (INT-U8), day (INT-U8), hour (INT-U8), minute (INT-U8), second (INT-U8), sec-fractions (INT-U8)} (BCD encoding)

	<ul style="list-style-type: none"> ❑ Attribute-value: <ul style="list-style-type: none"> • century: 20 (hex) or 32 (dec) • year: 12 (hex) or 18 (dec) • month: 08 (hex) or 8 (dec) • day: 02 (hex) or 2 (dec) • hour: 10 (hex) or 16 (dec) • minute: 39 (hex) or 57 (dec) • second: 27 (hex) or 39 (dec) • sec-fractions: 00 (hex) or 0 (dec) <p>b) WAN PCD-01 message</p> <p>PCD-01 message includes a segment like this with Absolute-Time-Stamp attribute value (check OBX-14):</p> <pre>OBX ? 150020^MDC_PRESS_BLD_NONINV^MDC 1.0.a X 20120802103927+0000</pre>
--	---

TP Id		TP/LP-PAN/PHG/PHDTW/BPM/BV-011		
TP label		Whitepaper. Systolic/Diastolic/Map Compound Numeric Object - Compound-Basic-Nu-Observed-Value Attribute 1		
Coverage	Spec	[b-Bluetooth PHDT v1.3]		
	Testable items	BP Numeric 10; M	Short Float Type 1; C	
Test purpose		<p>Check that:</p> <p>PHG transcodes Blood Pressure Systolic, Diastolic and Map Value fields of Blood Pressure Measurement characteristic into Systolic/Diastolic/Map Compound Numeric Object - Compound-Basic-Nu-Observed-Value attribute</p>		
Applicability		C_MAN_BLE_000 AND C_MAN_BLE_002 AND C_MAN_BLE_003		
Other PICS				
Initial condition		The PHG under test and the simulated PHD are in the Standby state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated PHD is configured with a profile (device specialization) supported by the PHG under test; it has a measurement ready to be sent and it is in the Advertising state (it is discoverable). 2. The simulated PHD implements several BLE characteristics. The characteristic of interest for this test case is: <ol style="list-style-type: none"> a. Blood pressure measurement (0x2A35) 3. The PHG under test initiates a discovery process (Scanning state). It discovers the simulated PHD and it starts a pairing process with the simulated PHD (Initiating state). 4. When the pairing has been completed (Connection state), the simulated PHD sends the measurement to the PHG under test with the following value: <ol style="list-style-type: none"> a. Blood pressure measurement (0x2A35) <ol style="list-style-type: none"> i. Field: Flags <ul style="list-style-type: none"> • Format: 8 bit • Value: 0000 0010 (MSB → LSB). Blood pressure measurement value in units of mmHg and Time Stamp fields are included, Pulse Rate, User ID and measurement Status fields are not included ii. Field: Blood Pressure Measurement Compound Value – Systolic (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: 100.0 iii. Field: Blood Pressure Measurement Compound Value – Diastolic (mmHg) 		

- Format: SFLOAT
 - Value: 70.0
 - iv. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (mmHg)
 - Format: SFLOAT
 - Value: 80.0
 - v. Field: Blood Pressure Measurement Compound Value – Systolic (kPa)
 - This field is not included
 - vi. Field: Blood Pressure Measurement Compound Value – Diastolic (kPa)
 - This field is not included
 - vii. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (kPa)
 - This field is not included
 - viii. Field: Time Stamp
 - Format: Date and Time
 - Value: Not relevant
 - ix. Field: Pulse Rate
 - This field is not included
 - x. Field: User ID
 - This field is not included
 - xi. Field: Measurement Status
 - This field is not included
5. Check in PHG transcoder output for the Systolic/Diastolic/Map compound numeric object, Compound-Basic-Nu-Observed-Value attribute.
6. The simulated PHD sends the measurement to the PHG under test with the following value:
- a. Blood pressure measurement (0x2A35)
 - i. Field: Flags
 - Format: 8 bit
 - Value: 0000 0011 (MSB → LSB). Blood pressure measurement value in units of kPa, Time Stamp field is included and Pulse Rate, User ID and measurement Status fields are not included
 - ii. Field: Blood Pressure Measurement Compound Value – Systolic (mmHg)
 - This field is not included
 - iii. Field: Blood Pressure Measurement Compound Value – Diastolic (mmHg)
 - This field is not included
 - iv. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (mmHg)
 - This field is not included
 - v. Field: Blood Pressure Measurement Compound Value – Systolic (kPa)
 - Format: SFLOAT
 - Value: 13.33
 - vi. Field: Blood Pressure Measurement Compound Value – Diastolic (kPa)
 - Format: SFLOAT
 - Value: 9.33
 - vii. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure

	<p>(kPa)</p> <p>viii. Format: SFLOAT</p> <ul style="list-style-type: none"> Value: 10.67 <p>ix. Field: Time Stamp</p> <ul style="list-style-type: none"> Format: Date and Time Value: Not relevant <p>x. Field: Pulse Rate</p> <ul style="list-style-type: none"> This field is not included <p>xi. Field: User ID</p> <ul style="list-style-type: none"> This field is not included <p>xii. Field: Measurement Status</p> <ul style="list-style-type: none"> This field is not included <p>7. Check in PHG transcoder output for the Systolic/Diastolic/Map compound numeric object, Compound-Basic-Nu-Observed-Value attribute.</p>
<p>Pass/Fail criteria</p>	<p>In step 5, the Systolic/Diastolic/Map compound numeric object – Compound-Basic-Nu-Observed-Value attribute is present and its value matches with the blood pressure measurement value (mmHg) fields of the Blood pressure measurement characteristic (Systolic: 100.0, Diastolic:70.0, MAP: 80.0).</p> <p>In step 7, the Systolic/Diastolic/Map compound numeric object – Compound-Basic-Nu-Observed-Value attribute is present and its value matches with the blood pressure measurement value (kPa) fields of the Blood pressure measurement characteristic (Systolic: 13.33, Diastolic:9.33, MAP: 10.67).</p>
<p>Notes</p>	<p>Possible values in typical points of observation after transcoder output are:</p> <p>a) IEEE 11073 Objects and Attributes</p> <p>Compound-Basic-Nu-Observed-Value attribute is present:</p> <ul style="list-style-type: none"> <input type="checkbox"/> Object: Systolic/Diastolic/Map compound numeric object <input type="checkbox"/> Attribute-id: MDC_ATTR_NU_CMPD_VAL_OBS_BASIC (2677) <input type="checkbox"/> Attribute-type: SEQUENCE OF [{SFLOAT}] <input type="checkbox"/> Attribute-value: 00 03 00 06 xx xx yy yy zz zz (hex), where 'xx xx' is the Systolic value, 'yy yy' is the Diastolic value and 'zz zz' is the MAP value [Note that 0x00 0x03 is the number of elements in the sequence and 0x00 06 is the length of the sequence] <ul style="list-style-type: none"> Systolic: F3 E8 (hex) or 00 64 (hex) or 10 0A (hex) or 20 01 (hex) or 100.0 (dec) Diastolic: F2 BC (hex) or 00 46 (hex) or 10 07 (hex) or 70.0 (dec) MAP: F3 20 (hex) or 00 50 (hex) or 10 08 (hex) or 80.0 (dec) <p>b) WAN PCD-01 message</p> <p>PCD-01 message includes three segments like these with a Simple-Nu-Observed-Value attribute value (check OBX-5):</p> <pre>OBX ? NM 150021^MDC_PRESS_BLD_NONINV_SYS^MDC 1.0.a.x 100 266016^MDC_DIM_MMHG^MDC R [[current_date_time]]</pre> <pre>OBX ? NM 150022^MDC_PRESS_BLD_NONINV_DIA^MDC 1.0.a.y 70 266016^MDC_DIM_MMHG^MDC R [[current_date_time]]</pre> <pre>OBX ? NM 150023^MDC_PRESS_BLD_NONINV_MEAN^MDC 1.0.a.z 80 266016^MDC_DIM_MMHG^MDC R [[current_date_time]]</pre> <p>Note that " [[current_date_time]]" is optional at the end of each segment because the date and time can be specified in the parent segment (OBX ? 150020^MDC_PRESS_BLD_NONINV^MDC 1.0.a X [[current_date_time]])</p> <p>In step 7, possible values in typical points of observation after transcoder output are:</p> <p>a) IEEE 11073 Objects and Attributes</p> <p>Simple-Nu-Observed-Value attribute is present:</p>

	<ul style="list-style-type: none"> ❑ Object: Systolic/Diastolic/Map compound numeric object ❑ Attribute-id: MDC_ATTR_NU_CMPD_VAL_OBS_BASIC (2677) ❑ Attribute-type: SEQUENCE OF [{SFLOAT}] ❑ Attribute-value: 00 03 00 06 E5 35 E3 A5 E4 2B (hex) [Note that 0x00 0x03 is the number of elements in the sequence and 0x00 06 is the length of the sequence] <ul style="list-style-type: none"> • Systolic: E5 35 (hex) or 13.33 (dec) • Diastolic: E3 A5 (hex) 9.33 (dec) • MAP: E4 2B (hex) 10.67 (dec) <p>b) WAN PCD-01 message</p> <p>PCD-01 message includes three segments like these with a Simple-Nu-Observed-Value attribute value (check OBX-5):</p> <pre>OBX ? NM 150021^MDC_PRESS_BLD_NONINV_SYS^MDC 1.0.a.x 13.33 265987^MDC_DIM_KILO_PASCAL^MDC R [[current_date_time]]</pre> <pre>OBX ? NM 150022^MDC_PRESS_BLD_NONINV_DIA^MDC 1.0.a.y 9.33 265987^MDC_DIM_KILO_PASCAL^MDC R [[current_date_time]]</pre> <pre>OBX ? NM 150023^MDC_PRESS_BLD_NONINV_MEAN^MDC 1.0.a.z 10.67 265987^MDC_DIM_KILO_PASCAL^MDC R [[current_date_time]]</pre> <p>Note that " [[current_date_time]]" is optional at the end of each segment because the date and time can be specified in the parent segment (OBX ? 150020^MDC_PRESS_BLD_NONINV^MDC 1.0.a X [[current_date_time]])</p>
--	--

TP Id		TP/LP-PAN/PHG/PHDTW/BPM/BV-012		
TP label		Whitepaper. Systolic/Diastolic/Map Compound Numeric Object - Compound-Basic-Nu-Observed-Value Attribute 2		
Coverage	Spec	[b-Bluetooth PHDT v1.3]		
	Testable items	BP Numeric 10; M	Short Float Type 1; C	Short Float Type 2; M
Test purpose		<p>Check that:</p> <p>PHG transcodes Blood Pressure Systolic, Diastolic and Map Value fields of Blood Pressure Measurement characteristic into Systolic/Diastolic/Map Compound Numeric Object - Compound-Basic-Nu-Observed-Value attribute</p> <p>[AND]</p> <p>PHG assigns the following special values: NaN (0x07FF), NRes (0x0800), +INFINITY (0x07FE) and -INFINITY (0x0802)</p>		
Applicability		C_MAN_BLE_000 AND C_MAN_BLE_002 AND C_MAN_BLE_003		
Other PICS				
Initial condition		The PHG under test and the simulated PHD are in the Standby state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated PHD is configured with a profile (device specialization) supported by the PHG under test; it has a measurement ready to be sent and it is in the Advertising state (it is discoverable). 2. The simulated PHD implements several BLE characteristics. The characteristic of interest for this test case is: <ol style="list-style-type: none"> a. Blood pressure measurement (0x2A35) 3. The PHG under test initiates a discovery process (Scanning state). It discovers the simulated PHD and it starts a pairing process with the simulated PHD (Initiating state). 4. When the pairing has been completed (Connection state), the simulated PHD sends the measurement to the PHG under test with the following value: <ol style="list-style-type: none"> a. Blood pressure measurement (0x2A35) <ol style="list-style-type: none"> i. Field: Flags <ul style="list-style-type: none"> • Format: 8 bit 		

- Value: 0000 0010 (MSB → LSB). Blood pressure measurement value in units of mmHg and Time Stamp fields are included, Pulse Rate, User ID and measurement Status fields are not included
 - ii. Field: Blood Pressure Measurement Compound Value – Systolic (mmHg)
 - Format: SFLOAT
 - Value: 100.0
 - iii. Field: Blood Pressure Measurement Compound Value – Diastolic (mmHg)
 - Format: SFLOAT
 - Value: 70.0
 - iv. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (mmHg)
 - Format: SFLOAT
 - Value: 80.0
 - v. Field: Blood Pressure Measurement Compound Value – Systolic (kPa)
 - This field is not included
 - vi. Field: Blood Pressure Measurement Compound Value – Diastolic (kPa)
 - This field is not included
 - vii. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (kPa)
 - This field is not included
 - viii. Field: Time Stamp
 - Format: Date and Time
 - Value: Not relevant
 - ix. Field: Pulse Rate
 - This field is not included
 - x. Field: User ID
 - This field is not included
 - xi. Field: Measurement Status
 - This field is not included
5. Check in PHG transcoder output for the Systolic/Diastolic/Map compound numeric object – Compound-Basic-Nu-Observed-Value attribute.
6. The simulated PHD sends the measurement to the PHG under test with the following value:
- a. Blood pressure measurement (0x2A35)
 - i. Field: Flags
 - Format: 8 bit
 - Value: 0000 0010 (MSB → LSB). Blood pressure measurement value in units of mmHg and Time Stamp fields are included, Pulse Rate, User ID and measurement Status fields are not included
 - ii. Field: Blood Pressure Measurement Compound Value – Systolic (mmHg)
 - Format: SFLOAT
 - Value: 07 FF (hex). Special value: NaN
 - iii. Field: Blood Pressure Measurement Compound Value – Diastolic (mmHg)
 - Format: SFLOAT
 - Value: 07 FF (hex). Special value: NaN
 - iv. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure

	<p>(mmHg)</p> <ul style="list-style-type: none"> • Format: SFLOAT • Value: 07 FF (hex). Special value: NaN <p>v. Field: Blood Pressure Measurement Compound Value – Systolic (kPa)</p> <ul style="list-style-type: none"> • This field is not included <p>vi. Field: Blood Pressure Measurement Compound Value – Diastolic (kPa)</p> <ul style="list-style-type: none"> • This field is not included <p>vii. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (kPa)</p> <ul style="list-style-type: none"> • This field is not included <p>viii. Field: Time Stamp</p> <ul style="list-style-type: none"> • Format: Date and Time • Value: Not relevant <p>ix. Field: Pulse Rate</p> <ul style="list-style-type: none"> • This field is not included <p>x. Field: User ID</p> <ul style="list-style-type: none"> • This field is not included <p>xi. Field: Measurement Status</p> <ul style="list-style-type: none"> • This field is not included <p>7. Check in PHG transcoder output for the Systolic/Diastolic/Map compound numeric object – Compound-Basic-Nu-Observed-Value attribute.</p> <p>8. The simulated PHD sends the measurement to the PHG under test with the following value:</p> <p>a. Blood pressure measurement (0x2A35)</p> <p>i. Field: Flags</p> <ul style="list-style-type: none"> • Format: 8 bit • Value: 0000 0010 (MSB → LSB). Blood pressure measurement value in units of mmHg and Time Stamp fields are included, Pulse Rate, User ID and measurement Status fields are not included <p>ii. Field: Blood Pressure Measurement Compound Value – Systolic (mmHg)</p> <ul style="list-style-type: none"> • Format: SFLOAT • Value: 08 00 (hex). Special value: NRes <p>iii. Field: Blood Pressure Measurement Compound Value – Diastolic (mmHg)</p> <ul style="list-style-type: none"> • Format: SFLOAT • Value: 08 00 (hex). Special value: NRes <p>iv. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (mmHg)</p> <ul style="list-style-type: none"> • Format: SFLOAT • Value: 08 00 (hex). Special value: NRes <p>v. Field: Blood Pressure Measurement Compound Value – Systolic (kPa)</p> <ul style="list-style-type: none"> • This field is not included <p>vi. Field: Blood Pressure Measurement Compound Value – Diastolic (kPa)</p> <ul style="list-style-type: none"> • This field is not included <p>vii. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (kPa)</p> <ul style="list-style-type: none"> • This field is not included
--	---

	<ul style="list-style-type: none"> viii. Field: Time Stamp <ul style="list-style-type: none"> • Format: Date and Time • Value: Not relevant ix. Field: Pulse Rate <ul style="list-style-type: none"> • This field is not included x. Field: User ID <ul style="list-style-type: none"> • This field is not included xi. Field: Measurement Status <ul style="list-style-type: none"> • This field is not included <p>9. Check in PHG transcoder output for the Systolic/Diastolic/Map compound numeric object, Compound-Basic-Nu-Observed-Value attribute.</p> <p>10. The simulated PHD sends the measurement to the PHG under test with the following value:</p> <ul style="list-style-type: none"> a. Blood pressure measurement (0x2A35) <ul style="list-style-type: none"> i. Field: Flags <ul style="list-style-type: none"> • Format: 8 bit • Value: 0000 0010 (MSB → LSB). Blood pressure measurement value in units of mmHg and Time Stamp fields are included, Pulse Rate, User ID and measurement Status fields are not included ii. Field: Blood Pressure Measurement Compound Value – Systolic (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: 07 FE (hex). Special value: +INFINITY iii. Field: Blood Pressure Measurement Compound Value – Diastolic (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: 07 FE (hex). Special value: +INFINITY iv. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: 07 FE (hex). Special value: +INFINITY v. Field: Blood Pressure Measurement Compound Value – Systolic (kPa) <ul style="list-style-type: none"> • This field is not included vi. Field: Blood Pressure Measurement Compound Value – Diastolic (kPa) <ul style="list-style-type: none"> • This field is not included vii. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (kPa) <ul style="list-style-type: none"> • This field is not included viii. Field: Time Stamp <ul style="list-style-type: none"> • Format: Date and Time • Value: Not relevant ix. Field: Pulse Rate <ul style="list-style-type: none"> • This field is not included x. Field: User ID <ul style="list-style-type: none"> • This field is not included xi. Field: Measurement Status <ul style="list-style-type: none"> • This field is not included
--	---

	<p>11. Check in PHG transcoder output for the Systolic/Diastolic/Map compound numeric object, Compound-Basic-Nu-Observed-Value attribute.</p> <p>12. The simulated PHD sends the measurement to the PHG under test with the following value:</p> <ol style="list-style-type: none"> a. Blood pressure measurement (0x2A35) <ol style="list-style-type: none"> i. Field: Flags <ul style="list-style-type: none"> • Format: 8 bit • Value: 0000 0010 (MSB → LSB). Blood pressure measurement value in units of mmHg and Time Stamp fields are included, Pulse Rate, User ID and measurement Status fields are not included ii. Field: Blood Pressure Measurement Compound Value – Systolic (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: 08 02 (hex). Special value: -INFINITY iii. Field: Blood Pressure Measurement Compound Value – Diastolic (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: 08 02 (hex). Special value: -INFINITY iv. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: 08 02 (hex). Special value: -INFINITY v. Field: Blood Pressure Measurement Compound Value – Systolic (kPa) <ul style="list-style-type: none"> • This field is not included vi. Field: Blood Pressure Measurement Compound Value – Diastolic (kPa) <ul style="list-style-type: none"> • This field is not included vii. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (kPa) <ul style="list-style-type: none"> • This field is not included viii. Field: Time Stamp <ul style="list-style-type: none"> • Format: Date and Time • Value: Not relevant ix. Field: Pulse Rate <ul style="list-style-type: none"> • This field is not included x. Field: User ID <ul style="list-style-type: none"> • This field is not included xi. Field: Measurement Status <ul style="list-style-type: none"> • This field is not included <p>13. Check in PHG transcoder output for the Systolic/Diastolic/Map compound numeric object, Compound-Basic-Nu-Observed-Value attribute.</p>
<p>Pass/Fail criteria</p>	<p>In step 5, the Systolic/Diastolic/Map compound numeric object, Compound-Basic-Nu-Observed-Value attribute is present and its value is 100.0 for Systolic, 70.0 for Diastolic and 80.0 for MAP.</p> <p>In step 7, the Systolic/Diastolic/Map compound numeric object, Compound-Basic-Nu-Observed-Value attribute is present and its value is 0x07FF for Systolic, Diastolic and MAP.</p> <p>In step 9, the Systolic/Diastolic/Map compound numeric object, Compound-Basic-Nu-Observed-Value attribute is present and its value is 0x0800 for Systolic, Diastolic and MAP.</p> <p>In step 11, the Systolic/Diastolic/Map compound numeric object, Compound-Basic-Nu-Observed-Value attribute is present and its value is 0x07FE for Systolic, Diastolic and MAP.</p> <p>In step 13, the Systolic/Diastolic/Map compound numeric object, Compound-Basic-Nu-</p>

<p>Notes</p>	<p>Observed-Value attribute is present and its value is 0x0802 for Systolic, Diastolic and MAP.</p> <p>In step 5, possible values in typical points of observation after transcoder output are:</p> <p>a) IEEE 11073 Objects and Attributes</p> <p>Compound-Basic-Nu-Observed-Value attribute is present:</p> <ul style="list-style-type: none"> <input type="checkbox"/> Systolic/Diastolic/Map compound numeric object <input type="checkbox"/> Attribute-id: MDC_ATTR_NU_CMPD_VAL_OBS_BASIC (2677) <input type="checkbox"/> Attribute-type: SEQUENCE OF [{SFLOAT}] <input type="checkbox"/> Attribute-value: 00 03 00 06 xx xx yy yy zz zz (hex), where 'xx xx' is the Systolic value, 'yy yy' is the Diastolic value and 'zz zz' is the MAP value [Note that 0x00 0x03 is the number of elements in the sequence and 0x00 06 is the length of the sequence] <ul style="list-style-type: none"> • Systolic: F3 E8 (hex) or 00 64 (hex) or 10 0A (hex) or 20 01 (hex) or 100.0 (dec) • Diastolic: F2 BC (hex) or 00 46 (hex) or 10 07 (hex) or 70.0 (dec) • MAP: F3 20 (hex) or 00 50 (hex) or 10 08 (hex) or 80.0 (dec) <p>b) WAN PCD-01 message</p> <p>PCD-01 message includes a segment like this with a Simple-Nu-Observed-Value attribute value (check OBX-5):</p> <pre>OBX ? NM 150021^MDC_PRESS_BLD_NONINV_SYS^MDC 1.0.a.x 100 266016^MDC_DIM_MMHG^MDC R [[current_date_time]] OBX ? NM 150022^MDC_PRESS_BLD_NONINV_DIA^MDC 1.0.a.y 70 266016^MDC_DIM_MMHG^MDC R [[current_date_time]] OBX ? NM 150023^MDC_PRESS_BLD_NONINV_MEAN^MDC 1.0.a.z 80 266016^MDC_DIM_MMHG^MDC R [[current_date_time]]</pre> <p>Note that "[[[current_date_time]]]" is optional at the end of each segment because the date and time can be specified in the parent segment (OBX ? 150020^MDC_PRESS_BLD_NONINV^MDC 1.0.a X [[current_date_time]])</p> <p>In step 7, possible values in typical points of observation after transcoder output are:</p> <p>a) IEEE 11073 Objects and Attributes</p> <p>Compound-Basic-Nu-Observed-Value attribute is present:</p> <ul style="list-style-type: none"> <input type="checkbox"/> Systolic/Diastolic/Map compound numeric object <input type="checkbox"/> Attribute-id: MDC_ATTR_NU_CMPD_VAL_OBS_BASIC (2677) <input type="checkbox"/> Attribute-type: SEQUENCE OF [{SFLOAT}] <input type="checkbox"/> Attribute-value: 00 03 00 06 07 FF 07 FF 07 FF (hex) [Note that 0x00 0x03 is the number of elements in the sequence and 0x00 06 is the length of the sequence] <ul style="list-style-type: none"> • Systolic: 07 FF (hex) or NaN (note that a decimal value is not allowed) • Diastolic: 07 FF (hex) or NaN (note that a decimal value is not allowed) • MAP: 07 FF (hex) or NaN (note that a decimal value is not allowed) <p>b) WAN PCD-01 message</p> <p>PCD-01 message does not include segments with a Simple-Nu-Observed-Value attribute value (150021^MDC_PRESS_BLD_NONINV_SYS^MDC, 150022^MDC_PRESS_BLD_NONINV_DIA^MDC and 150023^MDC_PRESS_BLD_NONINV_MEAN) because they have a special value and these values are not included in the PCD-01 message.</p> <p>In step 9, possible values in typical points of observation after transcoder output are:</p> <p>a) IEEE 11073 Objects and Attributes</p> <p>Compound-Basic-Nu-Observed-Value attribute is present:</p> <ul style="list-style-type: none"> <input type="checkbox"/> Systolic/Diastolic/Map compound numeric object <input type="checkbox"/> Attribute-id: MDC_ATTR_NU_CMPD_VAL_OBS_BASIC (2677) <input type="checkbox"/> Attribute-type: SEQUENCE OF [{SFLOAT}]
---------------------	---

	<ul style="list-style-type: none"> <input type="checkbox"/> Attribute-value: 00 03 00 06 08 00 08 00 08 00(hex) [Note that 0x00 0x03 is the number of elements in the sequence and 0x00 06 is the length of the sequence] <ul style="list-style-type: none"> • Systolic: 08 00 (hex) or NRes (note that a decimal value is not allowed) • Diastolic: 08 00 (hex) or NRes (note that a decimal value is not allowed) • MAP: 08 00 (hex) or NRes (note that a decimal value is not allowed) <p>b) WAN PCD-01 message</p> <p>PCD-01 message does not include segments with a Simple-Nu-Observed-Value attribute value (150021^MDC_PRESS_BLD_NONINV_SYS^MDC, 150022^MDC_PRESS_BLD_NONINV_DIA^MDC and 150023^MDC_PRESS_BLD_NONINV_MEAN) because they have a special value and these values are not included in the PCD-01 message.</p> <p>In step 11, possible values in typical points of observation after transcoder output are:</p> <p>a) IEEE 11073 Objects and Attributes</p> <p>Compound-Basic-Nu-Observed-Value attribute is present:</p> <ul style="list-style-type: none"> <input type="checkbox"/> Systolic/Diastolic/Map compound numeric object <input type="checkbox"/> Attribute-id: MDC_ATTR_NU_CMPD_VAL_OBS_BASIC (2677) <input type="checkbox"/> Attribute-type: SEQUENCE OF [{SFLOAT}] <input type="checkbox"/> Attribute-value: 00 03 00 06 07 FE 07 FE 07 FE (hex) [Note that 0x00 0x03 is the number of elements in the sequence and 0x00 06 is the length of the sequence] <ul style="list-style-type: none"> • Systolic: 07 FE (hex) or +INFINITY (note that a decimal value is not allowed) • Diastolic: 07 FE (hex) or +INFINITY (note that a decimal value is not allowed) • MAP: 07 FE (hex) or +INFINITY (note that a decimal value is not allowed) <p>b) WAN PCD-01 message</p> <p>PCD-01 message does not include segments with a Simple-Nu-Observed-Value attribute value (150021^MDC_PRESS_BLD_NONINV_SYS^MDC, 150022^MDC_PRESS_BLD_NONINV_DIA^MDC and 150023^MDC_PRESS_BLD_NONINV_MEAN) because they have a special value and these values are not included in the PCD-01 message.</p> <p>In step 13, possible values in typical points of observation after transcoder output are:</p> <p>a) IEEE 11073 Objects and Attributes</p> <p>Compound-Basic-Nu-Observed-Value attribute is present:</p> <ul style="list-style-type: none"> <input type="checkbox"/> Systolic/Diastolic/Map compound numeric object <input type="checkbox"/> Attribute-id: MDC_ATTR_NU_CMPD_VAL_OBS_BASIC (2677) <input type="checkbox"/> Attribute-type: SEQUENCE OF [{SFLOAT}] <input type="checkbox"/> Attribute-value: 00 03 00 06 08 02 08 02 08 02 (hex) [Note that 0x00 0x03 is the number of elements in the sequence and 0x00 06 is the length of the sequence] <ul style="list-style-type: none"> • Systolic: 08 02 (hex) or -INFINITY (note that a decimal value is not allowed) • Diastolic: 08 02 (hex) or -INFINITY (note that a decimal value is not allowed) • MAP: 08 02 (hex) or -INFINITY (note that a decimal value is not allowed) <p>b) WAN PCD-01 message</p> <p>PCD-01 message does not include segments with a Simple-Nu-Observed-Value attribute value (150021^MDC_PRESS_BLD_NONINV_SYS^MDC, 150022^MDC_PRESS_BLD_NONINV_DIA^MDC and 150023^MDC_PRESS_BLD_NONINV_MEAN) because they have a special value and these values are not included in the PCD-01 message.</p>
--	--

TP Id	TP/LP-PAN/PHG/PHDTW/BPM/BV-013
TP label	Whitepaper. Systolic/Diastolic/Map Compound Numeric Object - Blood Pressure measurement value

Coverage	Spec	[b-Bluetooth PHDT v1.3]		
	Testable items	Short Float Type 1; C	Date-Time Conv 1; M	BP Numeric 9; M
		BP Numeric 10; M		
Test purpose	<p>Check that:</p> <p>PHG processes correctly the Blood Pressure Measurement Compound Value (mmHg), Blood Pressure Measurement Compound Value (kPa) and Time Stamp fields of Blood Pressure Measurement</p>			
Applicability	C_MAN_BLE_000 AND C_MAN_BLE_003			
Other PICS				
Initial condition	The PHG under test and the simulated PHD are in the Standby state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated PHD is configured with a profile (device specialization) supported by the PHG under test; it has a measurement ready to be sent and it is in the Advertising state (it is discoverable). 2. The simulated PHD implements several BLE characteristics. The characteristic of interest for this test case is: <ol style="list-style-type: none"> a. Blood pressure measurement (0x2A35) 3. The PHG under test initiates a discovery process (Scanning state). It discovers the simulated PHD and it starts a pairing process with the simulated PHD (Initiating state). 4. When the pairing has been completed (Connection state), the simulated PHD sends the measurement to the PHG under test with the following value: <ol style="list-style-type: none"> a. Blood pressure measurement (0x2A35) <ol style="list-style-type: none"> i. Field: Flags <ul style="list-style-type: none"> • Format: 8 bit • Value: 0000 0010 (MSB → LSB). Blood pressure measurement value in units of mmHg and Time Stamp fields are included, Pulse Rate, User ID and measurement Status fields are not included ii. Field: Blood Pressure Measurement Compound Value – Systolic (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: 100.0 iii. Field: Blood Pressure Measurement Compound Value – Diastolic (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: 70.0 iv. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: 80.0 v. Field: Blood Pressure Measurement Compound Value – Systolic (kPa) <ul style="list-style-type: none"> • This field is not included vi. Field: Blood Pressure Measurement Compound Value – Diastolic (kPa) <ul style="list-style-type: none"> • This field is not included vii. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (kPa) <ul style="list-style-type: none"> • This field is not included viii. Field: Time Stamp <ul style="list-style-type: none"> • Format: Date and Time • Value: August 2nd, 2012, 11:08:25 ix. Field: Pulse Rate 			

	<ul style="list-style-type: none"> • This field is not included <p>x. Field: User ID</p> <ul style="list-style-type: none"> • This field is not included <p>xi. Field: Measurement Status</p> <ul style="list-style-type: none"> • This field is not included <p>5. Check that the PHG accepts the measurement and decodes its value properly (measurement values, units and time stamp).</p> <p>6. The simulated PHD sends the measurement to the PHG under test with the following value:</p> <p>a. Blood pressure measurement (0x2A35)</p> <p>i. Field: Flags</p> <ul style="list-style-type: none"> • Format: 8 bit • Value: 0000 0011 (MSB → LSB). Blood pressure measurement value in units of kPa and Time Stamp fields are included, Pulse Rate, User ID and measurement Status fields are not included <p>ii. Field: Blood Pressure Measurement Compound Value – Systolic (mmHg)</p> <ul style="list-style-type: none"> • This field is not included <p>iii. Field: Blood Pressure Measurement Compound Value – Diastolic (mmHg)</p> <ul style="list-style-type: none"> • This field is not included <p>iv. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (mmHg)</p> <ul style="list-style-type: none"> • This field is not included <p>v. Field: Blood Pressure Measurement Compound Value – Systolic (kPa)</p> <ul style="list-style-type: none"> • Format: SFLOAT • Value: 13.33 <p>vi. Field: Blood Pressure Measurement Compound Value – Diastolic (kPa)</p> <ul style="list-style-type: none"> • Format: SFLOAT • Value: 9.33 <p>vii. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (kPa)</p> <ul style="list-style-type: none"> • Format: SFLOAT • Value: 10.67 <p>viii. Field: Time Stamp</p> <ul style="list-style-type: none"> • Format: Date and Time • Value: August 2nd, 2012, 11:09:05 <p>ix. Field: Pulse Rate</p> <ul style="list-style-type: none"> • This field is not included <p>x. Field: User ID</p> <ul style="list-style-type: none"> • This field is not included <p>xi. Field: Measurement Status</p> <ul style="list-style-type: none"> • This field is not included <p>7. Check that the PHG under test accepts the measurement and decodes its value properly (measurement values, units and time stamp)</p>
Pass/Fail criteria	<p>In step 5, the PHG under test shows the following measurement: 100.0 mmHg for Systolic, 70.0 mmHg for Diastolic and 80.0 mmHg for MAP, with the time stamp '2012-08-02 11:08:25'.</p> <p>In step 7, the PHG under test shows the following measurement 13.33 kPa for Systolic, 9.33 kPa for Diastolic and 10.67 kPa for MAP, with the time stamp '2012-08-02 11:09:05'.</p>

Notes	
--------------	--

TP Id	TP/LP-PAN/PHG/PHDTW/BPM/BV-014		
TP label	Whitepaper. Pulse Rate Object - Handle Attribute		
Coverage	Spec	[b-Bluetooth PHDT v1.3]	
	Testable items	PR Numeric 1; O	
Test purpose	<p>Check that:</p> <p>PHG does not include Pulse Rate object, Handle Attribute in transcoder output</p> <p>[OR]</p> <p>If PHG includes Pulse Rate object, Handle attribute in transcoder output, then its value shall be different than 0</p>		
Applicability	C_MAN_BLE_000 AND C_MAN_BLE_002 AND C_MAN_BLE_003 AND C_MAN_BLE_005		
Other PICS			
Initial condition	The PHG under test and the simulated PHD are in the Standby state.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated PHD is configured with a Blood pressure profile (device specialization); it has a measurement ready to be sent and it is in the Advertising state (it is discoverable). 2. The simulated PHD implements several BLE characteristics. The characteristic of interest for this test case is: <ol style="list-style-type: none"> a. Blood pressure measurement (0x2A35) <ol style="list-style-type: none"> i. Field: Flags <ul style="list-style-type: none"> • Format: 8 bit • Value: 0000 0110 (MSB → LSB). Blood pressure measurement value in units of mmHg, Time Stamp and Pulse Rate fields are included, User ID and measurement Status fields are not included ii. Field: Blood Pressure Measurement Compound Value – Systolic (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant iii. Field: Blood Pressure Measurement Compound Value – Diastolic (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant iv. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant v. Field: Blood Pressure Measurement Compound Value – Systolic (kPa) <ul style="list-style-type: none"> • This field is not included vi. Field: Blood Pressure Measurement Compound Value – Diastolic (kPa) <ul style="list-style-type: none"> • This field is not included vii. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (kPa) <ul style="list-style-type: none"> • This field is not included viii. Field: Time Stamp <ul style="list-style-type: none"> • Format: Date and Time • Value: Not relevant ix. Field: Pulse Rate 		

	<ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant <p>x. Field: User ID</p> <ul style="list-style-type: none"> • This field is not included <p>xi. Field: Measurement Status</p> <ul style="list-style-type: none"> • This field is not included <p>3. The PHG under test initiates a discovery process (Scanning state). It discovers the simulated PHD and it starts a pairing process with the simulated PHD (Initiating state).</p> <p>4. When the pairing has been completed (Connection state), the simulated PHD sends the measurement to the PHG under test.</p> <p>5. Check in PHG transcoder output for the Pulse rate object, Handle attribute.</p>
Pass/Fail criteria	In step 5, the Pulse rate object, Handle attribute is not present; however, if it is present then its value is different to 0.
Notes	<p>Possible values in typical points of observation after transcoder output are:</p> <p>a) IEEE 11073 Objects and Attributes</p> <p>Handle attribute is not present, or if it is present then:</p> <ul style="list-style-type: none"> <input type="checkbox"/> Object: Pulse rate object <input type="checkbox"/> Attribute-id: MDC_ATTR_ID_HANDLE (2337) <input type="checkbox"/> Attribute-type: INT-U16 <input type="checkbox"/> Attribute-value: Any value other than 0 <p>b) WAN PCD-01 message</p> <p>PCD-01 message does not include segments with a Handle attribute value.</p>

TP Id	TP/LP-PAN/PHG/PHDTW/BPM/BV-015		
TP label	Whitepaper. Pulse Rate Object - Type Attribute		
Coverage	Spec	[b-Bluetooth PHDT v1.3]	
	Testable items	PR Numeric 2; M	
Test purpose	<p>Check that:</p> <p>PHG includes Systolic Pulse Rate object, Type attribute in transcoder output.</p> <p>[AND]</p> <p>Type is set to {MDC_PART_SCADA, MDC_PULS_RATE_NON_INV}</p>		
Applicability	C_MAN_BLE_000 AND C_MAN_BLE_002 AND C_MAN_BLE_003 AND C_MAN_BLE_005		
Other PICS			
Initial condition	The PHG under test and the simulated PHD are in the Standby state.		
Test procedure	<p>1. The simulated PHD is configured with a Blood pressure profile (device specialization); it has a measurement ready to be sent and it is in the Advertising state (it is discoverable).</p> <p>2. The simulated PHD implements several BLE characteristics. The characteristic of interest for this test case is:</p> <p>a. Blood pressure measurement (0x2A35)</p> <p>i. Field: Flags</p> <ul style="list-style-type: none"> • Format: 8 bit • Value: 0000 0110 (MSB → LSB). Blood pressure measurement value in units of mmHg, Time Stamp and Pulse Rate fields are included, User ID and measurement Status fields are not included <p>ii. Field: Blood Pressure Measurement Compound Value – Systolic (mmHg)</p>		

	<ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant <p>iii. Field: Blood Pressure Measurement Compound Value – Diastolic (mmHg)</p> <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant <p>iv. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (mmHg)</p> <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant <p>v. Field: Blood Pressure Measurement Compound Value – Systolic (kPa)</p> <ul style="list-style-type: none"> • This field is not included <p>vi. Field: Blood Pressure Measurement Compound Value – Diastolic (kPa)</p> <ul style="list-style-type: none"> • This field is not included <p>vii. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (kPa)</p> <ul style="list-style-type: none"> • This field is not included <p>viii. Field: Time Stamp</p> <ul style="list-style-type: none"> • Format: Date and Time • Value: Not relevant <p>ix. Field: Pulse Rate</p> <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant <p>x. Field: User ID</p> <ul style="list-style-type: none"> • This field is not included <p>xi. Field: Measurement Status</p> <ul style="list-style-type: none"> • This field is not included <p>3. The PHG under test initiates a discovery process (Scanning state). It discovers the simulated PHD and it starts a pairing process with the simulated PHD (Initiating state).</p> <p>4. When the pairing has been completed (Connection state), the simulated PHD sends the measurement to the PHG under test.</p> <p>4. Check in PHG transcoder output for the Pulse rate object, Type attribute.</p>
Pass/Fail criteria	In step 5, the Pulse rate object, Type attribute is present and its value is {MDC_PART_SCADA, MDC_PULS_RATE_NON_INV}
Notes	<p>Possible values in typical points of observation after transcoder output are:</p> <p>a) IEEE 11073 Objects and Attributes</p> <p>Type attribute is present:</p> <ul style="list-style-type: none"> <input type="checkbox"/> Object: Pulse rate object <input type="checkbox"/> Attribute-id: MDC_ATTR_ID_TYPE (2351) <input type="checkbox"/> Attribute-type: SEQUENCE {partition (INT-U16), code (INT-U16)} <input type="checkbox"/> Attribute-value: <ul style="list-style-type: none"> • partition: MDC_PART_SCADA or 2 (dec) or 00 02 (hex) • code: MDC_PULS_RATE_NON_INV or 18474 (dec) or 48 2A (hex) <p>b) WAN PCD-01 message</p> <p>PCD-01 message includes a segment like this with a Type attribute value (check OBX-3): OBX[?][NM]49546^MDC_PULS_RATE_NON_INV^MDC[1.0.0.a]</p>

TP Id		TP/LP-PAN/PHG/PHDTW/BPM/BV-016		
TP label		Whitepaper. Pulse Rate Object - Metric-Spec-Small Attribute		
Coverage	Spec	[b-Bluetooth PHDT v1.3]		
	Testable items	PR Numeric 3; M		
Test purpose		<p>Check that:</p> <p>PHG includes Pulse Rate object, Metric-Spec-Small attribute in transcoder output.</p> <p>[AND]</p> <p>Metric-Spec-Small is set to {0xF040} (mss-avail-intermittent mss-avail-stored-data mss-upd-aperiodic mss-msmt-aperiodic mss-acc-agent-initiated).</p>		
Applicability		C_MAN_BLE_000 AND C_MAN_BLE_002 AND C_MAN_BLE_003 AND C_MAN_BLE_005		
Other PICS				
Initial condition		The PHG under test and the simulated PHD are in the Standby state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated PHD is configured with a Blood pressure profile (device specialization); it has a measurement ready to be sent and it is in the Advertising state (it is discoverable). 2. The simulated PHD implements several BLE characteristics. The characteristic of interest for this test case is: <ol style="list-style-type: none"> a. Blood pressure measurement (0x2A35) <ol style="list-style-type: none"> i. Field: Flags <ul style="list-style-type: none"> • Format: 8 bit • Value: 0000 0110 (MSB → LSB). Blood pressure measurement value in units of mmHg, Time Stamp and Pulse Rate fields are included, User ID and measurement Status fields are not included ii. Field: Blood Pressure Measurement Compound Value – Systolic (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant iii. Field: Blood Pressure Measurement Compound Value – Diastolic (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant iv. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant v. Field: Blood Pressure Measurement Compound Value – Systolic (kPa) <ul style="list-style-type: none"> • This field is not included vi. Field: Blood Pressure Measurement Compound Value – Diastolic (kPa) <ul style="list-style-type: none"> • This field is not included vii. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (kPa) <ul style="list-style-type: none"> • This field is not included viii. Field: Time Stamp <ul style="list-style-type: none"> • Format: Date and Time • Value: Not relevant ix. Field: Pulse Rate 		

	<ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant <p>x. Field: User ID</p> <ul style="list-style-type: none"> • This field is not included <p>xi. Field: Measurement Status</p> <ul style="list-style-type: none"> • This field is not included <p>3. The PHG under test initiates a discovery process (Scanning state). It discovers the simulated PHD and it starts a pairing process with the simulated PHD (Initiating state).</p> <p>4. When the pairing has been completed (Connection state), the simulated PHD sends the measurement to the PHG under test.</p> <p>5. Check in PHG transcoder output for the Pulse rate object, Metric-Spec-Small attribute.</p>
Pass/Fail criteria	In step 4, the Pulse rate object, Metric-Spec-Small attribute is present and its value is {0xF040} (mss-avail-intermittent mss-avail-stored-data mss-upd-aperiodic mss-msmt-aperiodic mss-acc-agent-initiated).
Notes	<p>Possible values in typical points of observation after transcoder output are:</p> <p>a) IEEE 11073 Objects and Attributes</p> <p>Metric-Spec-Small attribute is present:</p> <ul style="list-style-type: none"> <input type="checkbox"/> Object: Pulse rate object <input type="checkbox"/> Attribute-id: MDC_ATTR_METRIC_SPEC_SMALL (2630) <input type="checkbox"/> Attribute-type: BITS-16 <input type="checkbox"/> Attribute-value: F0 40 (hex) or BITS mss-avail-intermittent(0), mss-avail-stored-data(1), mss-upd-aperiodic(2), mss-msmt-aperiodic(3), mss-acc-agent-initiated(9) set to TRUE and remaining BITS set to FALSE <p>b) WAN PCD-01 message</p> <p>PCD-01 message does not include segments with a Metric-Spec-Small attribute value.</p>

TP Id	TP/LP-PAN/PHG/PHDTW/BPM/BV-017		
TP label	Whitepaper. Pulse Rate Object - Unit-Code Attribute		
Coverage	Spec	[b-Bluetooth PHDT v1.3]	
	Testable items	PR Numeric 4; M	
Test purpose	<p>Check that:</p> <p>PHG includes Pulse Rate object, Unit-Code attribute in transcoder output.</p> <p>[AND]</p> <p>IF Blood Pressure Measurement – Pulse Rate field of Blood Pressure Measurement characteristic is present THEN Pulse Rate object, Unit-Code attribute is set to MDC_DIM_BEAT_PER_MIN</p>		
Applicability	C_MAN_BLE_000 AND C_MAN_BLE_002 AND C_MAN_BLE_003 AND C_MAN_BLE_005		
Other PICS			
Initial condition	The PHG under test and the simulated PHD are in the Standby state.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated PHD is configured with a Blood pressure profile (device specialization); it has a measurement ready to be sent and it is in the Advertising state (it is discoverable). 2. The simulated PHD implements several BLE characteristics. The characteristic of interest for this test case is: <ol style="list-style-type: none"> a. Blood pressure measurement (0x2A35) 3. The PHG under test initiates a discovery process (Scanning state). It discovers the simulated PHD and it starts a pairing process with the simulated PHD (Initiating state). 4. When the pairing has been completed (Connection state), the simulated PHD sends the 		

	<p>measurement to the PHG under test with the following value:</p> <ol style="list-style-type: none"> a. Blood pressure measurement (0x2A35) <ol style="list-style-type: none"> i. Field: Flags <ul style="list-style-type: none"> • Format: 8 bit • Value: 0000 0110 (MSB → LSB). Blood pressure measurement value in units of mmHg, Time Stamp field and Pulse Rate field are included, User ID and measurement Status fields are not included ii. Field: Blood Pressure Measurement Compound Value – Systolic (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant iii. Field: Blood Pressure Measurement Compound Value – Diastolic (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant iv. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant v. Field: Blood Pressure Measurement Compound Value – Systolic (kPa) <ul style="list-style-type: none"> • This field is not included vi. Field: Blood Pressure Measurement Compound Value – Diastolic (kPa) <ul style="list-style-type: none"> • This field is not included vii. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (kPa) <ul style="list-style-type: none"> • This field is not included viii. Field: Time Stamp <ul style="list-style-type: none"> • Format: Date and Time • Value: Not relevant ix. Field: Pulse Rate <ul style="list-style-type: none"> • Format: SFLOAT • Value: 110.0 x. Field: User ID <ul style="list-style-type: none"> • This field is not included xi. Field: Measurement Status <ul style="list-style-type: none"> • This field is not included <p>5. Check in PHG transcoder output for the Pulse rate object, Unit-Code attribute.</p>
Pass/Fail criteria	In step 5, the Pulse rate object, Unit-Code attribute is present and its value is MDC_DIM_BEAT_PER_MIN.
Notes	<p>In step 5, possible values in typical points of observation after transcoder output are:</p> <ol style="list-style-type: none"> a) IEEE 11073 Objects and Attributes <p>Unit-Code attribute is present:</p> <ul style="list-style-type: none"> <input type="checkbox"/> Object: Pulse rate object <input type="checkbox"/> Attribute-id: MDC_ATTR_UNIT_CODE (2454) <input type="checkbox"/> Attribute-type: INT-U16 <input type="checkbox"/> Attribute-value: MDC_DIM_BEAT_PER_MIN or 2720 (dec) or 0A A0 (hex) b) WAN PCD-01 message

	<p>PCD-01 message includes a segment like this with Unit-Code attribute value (check OBX-6):</p> <pre>OBX ? NM 149546^MDC_PULS_RATE_NON_INV^MDC 1.0.0.a 10 264864^MDC_DIM_BEAT_PER_MIN^MDC R current_date_time]</pre>
--	--

TP Id		TP/LP-PAN/PHG/PHDTW/BPM/BV-018		
TP label		Whitepaper. Pulse Rate Object - Absolute-Time-Stamp Attribute		
Coverage	Spec	[b-Bluetooth PHDT v1.3]		
	Testable items	PR Numeric 6; M	Date-Time Conv 2; M	Date-Time Conv 3; M
		Date-Time Conv 4; M	Date-Time Conv 5; M	
Test purpose		<p>Check that:</p> <p>PHG transcodes Time Stamp field of Blood Pressure Measurement characteristic into Pulse Rate Object - Absolute-Time-Stamp attribute</p> <p>[AND]</p> <p>PHG transcodes the Bluetooth Time Stamp field format to Absolute Time format</p> <p>[AND]</p> <p>The fraction of seconds in Absolute Time at transcoder output is 0</p>		
Applicability		C_MAN_BLE_000 AND C_MAN_BLE_002 AND C_MAN_BLE_003 AND C_MAN_BLE_005		
Other PICS				
Initial condition		The PHG under test and the simulated PHD are in the Standby state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated PHD is configured with a Blood pressure profile (device specialization); it has a measurement ready to be sent and it is in the Advertising state (it is discoverable). 2. The simulated PHD implements several BLE characteristics. The characteristic of interest for this test case is: <ol style="list-style-type: none"> a. Blood pressure measurement (0x2A35) 3. The PHG under test initiates a discovery process (Scanning state). It discovers the simulated PHD and it starts a pairing process with the simulated PHD (Initiating state). 4. When the pairing has been completed (Connection state), the simulated PHD sends the measurement to the PHG under test with the following value: <ol style="list-style-type: none"> a. Blood pressure measurement (0x2A35) <ol style="list-style-type: none"> i. Field: Flags <ul style="list-style-type: none"> • Format: 8 bit • Value: 0000 0110 (MSB → LSB). Blood pressure measurement value in units of mmHg, Time Stamp and Pulse Rate fields are included, User ID and measurement Status fields are not included ii. Field: Blood Pressure Measurement Compound Value – Systolic (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant iii. Field: Blood Pressure Measurement Compound Value – Diastolic (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant iv. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant v. Field: Blood Pressure Measurement Compound Value – Systolic (kPa) <ul style="list-style-type: none"> • This field is not included 		

	<ul style="list-style-type: none"> vi. Field: Blood Pressure Measurement Compound Value – Diastolic (kPa) <ul style="list-style-type: none"> • This field is not included vii. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (kPa) <ul style="list-style-type: none"> • This field is not included viii. Field: Time Stamp <ul style="list-style-type: none"> • Format: Date and Time • Value: August 2nd, 2012, 10:39:27 ix. Field: Pulse Rate <ul style="list-style-type: none"> • Format: SFLOAT • Value: 110.0 x. Field: User ID <ul style="list-style-type: none"> • This field is not included xi. Field: Measurement Status <ul style="list-style-type: none"> • This field is not included <p>5. Check in PHG transcoder output for the SystolicPulse rate object, Absolute-Time-Stamp attribute.</p>
Pass/Fail criteria	In step 5, the Pulse rate object, Absolute-Time-Stamp attribute is present, its value matches with the Time Stamp field of the Blood pressure measurement characteristic and the fraction of seconds is set to 0.
Notes	<p>Possible values in typical points of observation after transcoder output are:</p> <p>a) IEEE 11073 Objects and Attributes</p> <p>Absolute-Time-Stamp attribute is present:</p> <ul style="list-style-type: none"> <input type="checkbox"/> Object: Pulse rate object <input type="checkbox"/> Attribute-id: MDC_ATTR_TIME_STAMP_ABS (2448) <input type="checkbox"/> Attribute-type: SEQUENCE {century (INT-U8), year (INT-U8), month (INT-U8), day (INT-U8), hour (INT-U8), minute (INT-U8), second (INT-U8), sec-fractions (INT-U8)} (BCD encoding) <input type="checkbox"/> Attribute-value: <ul style="list-style-type: none"> • century: 20 (hex) or 32 (dec) • year: 12 (hex) or 18 (dec) • month: 08 (hex) or 8 (dec) • day: 02 (hex) or 2 (dec) • hour: 10 (hex) or 16 (dec) • minute: 39 (hex) or 57 (dec) • second: 27 (hex) or 39 (dec) • sec-fractions: 00 (hex) or 0 (dec) <p>b) WAN PCD-01 message</p> <p>PCD-01 message includes a segment like this with Absolute-Time-Stamp attribute value (check OBX-14):</p> <pre>OBX ? NM 149546^MDC_PULS_RATE_NON_INV^MDC 1.0.0.a 10 264864^MDC_DIM_BEAT_PER_MIN^MDC R 20120802103927+0000</pre>

TP Id	TP/LP-PAN/PHG/PHDTW/BPM/BV-019
TP label	Whitepaper. Pulse Rate Object - Basic-Nu-Observed-Value Attribute 1
Coverage	Spec [b-Bluetooth PHDT v1.3]

	Testable items	PR Numeric 7; M	Short Float Type 1; C	
Test purpose	Check that: PHG transcodes Blood Pressure Measurement – Heart Rate field of Blood Pressure Measurement characteristic into Pulse Rate Object - Simple-Nu-Observed-Value attribute			
Applicability	C_MAN_BLE_000 AND C_MAN_BLE_002 AND C_MAN_BLE_003 AND C_MAN_BLE_005			
Other PICS				
Initial condition	The PHG under test and the simulated PHD are in the Standby state.			
Test procedure	<ol style="list-style-type: none"> 1. The simulated PHD is configured with a Blood pressure profile (device specialization); it has a measurement ready to be sent and it is in the Advertising state (it is discoverable). 2. The simulated PHD implements several BLE characteristics. The characteristic of interest for this test case is: <ol style="list-style-type: none"> a. Blood Pressure Measurement (0x2A35) 3. The PHG under test initiates a discovery process (Scanning state). It discovers the simulated PHD and it starts a pairing process with the simulated PHD (Initiating state). 4. When the pairing has been completed (Connection state), the simulated PHD sends the measurement to the PHG under test with the following value: <ol style="list-style-type: none"> a. Blood pressure measurement (0x2A35) <ol style="list-style-type: none"> i. Field: Flags <ul style="list-style-type: none"> • Format: 8 bit • Value: 0000 0110 (MSB → LSB). Blood pressure measurement value in units of mmHg, Time Stamp and Pulse Rate fields are included, User ID and measurement Status fields are not included ii. Field: Blood Pressure Measurement Compound Value – Systolic (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant iii. Field: Blood Pressure Measurement Compound Value – Diastolic (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant iv. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant v. Field: Blood Pressure Measurement Compound Value – Systolic (kPa) <ul style="list-style-type: none"> • This field is not included vi. Field: Blood Pressure Measurement Compound Value – Diastolic (kPa) <ul style="list-style-type: none"> • This field is not included vii. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (kPa) <ul style="list-style-type: none"> • This field is not included viii. Field: Time Stamp <ul style="list-style-type: none"> • Format: Date and Time • Value: Not relevant ix. Field: Pulse Rate <ul style="list-style-type: none"> • Format: SFLOAT • Value: 110.0 x. Field: User ID 			

	<ul style="list-style-type: none"> • This field is not included xi. Field: Measurement Status • This field is not included <p>5. Check the output of the PHG transcoder for the Pulse rate object, Basic-Nu-Observed-Value attribute.</p>
Pass/Fail criteria	In step 5, the Heart rate object, Basic-Nu-Observed-Value attribute is present and its value matches with the Blood Pressure Measurement – Heart Rate Value (bpm) field of the Blood pressure measurement characteristic (110).
Notes	<p>Possible values in typical points of observation after transcoder output are:</p> <p>a) IEEE 11073 Objects and Attributes</p> <p>Basic-Nu-Observed-Value attribute is present:</p> <ul style="list-style-type: none"> <input type="checkbox"/> Object: Pulse rate object <input type="checkbox"/> Attribute-id: MDC_ATTR_NU_VAL_OBS_BASIC (2636) <input type="checkbox"/> Attribute-type: SFLOAT <input type="checkbox"/> Attribute-value: F4 4C (hex) or 00 6E (hex) or 01 0B (hex) or 110 (dec) <p>b) WAN PCD-01 message</p> <p>PCD-01 message includes a segment like this with a Basic-Nu-Observed-Value attribute value (check OBX-5):</p> <pre>OBX ? NM 149546^MDC_PULS_RATE_NON_INV^MDC 1.0.0.a 10 264864^MDC_DIM_BEAT_PER_MIN^MDC R [current_date_time]</pre>

TP Id	TP/LP-PAN/PHG/PHDTW/BPM/BV-020		
TP label	Whitepaper. Pulse Rate Object - Basic-Nu-Observed-Value Attribute 2		
Coverage	Spec	[b-Bluetooth PHDT v1.3]	
	Testable items	PR Numeric 7; M	Short Float Type 1; C
Test purpose	<p>Check that:</p> <p>PHG transcodes Blood Pressure Measurement- Heart rate field of Blood Pressure Measurement characteristic into Heart Rate Object - Basic-Nu-Observed-Value attribute [AND]</p> <p>PHG assigns the following special values: NaN (0x07FF), NRes (0x0800), +INFINITY (0x07FE) and -INFINITY (0x0802)</p>		
Applicability	C_MAN_BLE_000 AND C_MAN_BLE_002 AND C_MAN_BLE_003 AND C_MAN_BLE_005		
Other PICS			
Initial condition	The PHG under test and the simulated PHD are in the Standby state.		
Test procedure	<ol style="list-style-type: none"> 1. The simulated PHD is configured with a Blood pressure profile (device specialization); it has a measurement ready to be sent and it is in the Advertising state (it is discoverable). 2. The simulated PHD implements several BLE characteristics. The characteristic of interest for this test case is: <ol style="list-style-type: none"> a. Blood pressure measurement (0x2A35) 3. The PHG under test initiates a discovery process (Scanning state). It discovers the simulated PHD and it starts a pairing process with the simulated PHD (Initiating state). 4. When the pairing has been completed (Connection state), the simulated PHD sends the measurement to the PHG under test with the following value: <ol style="list-style-type: none"> a. Blood pressure measurement (0x2A35) <ol style="list-style-type: none"> i. Field: Flags <ul style="list-style-type: none"> • Format: 8 bit • Value: 0000 0110 (MSB → LSB). Blood pressure measurement value in 		

	<p>units of mmHg, Time Stamp and Pulse Rate fields are included, User ID and measurement Status fields are not included</p> <ul style="list-style-type: none"> ii. Field: Blood Pressure Measurement Compound Value – Systolic (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant iii. Field: Blood Pressure Measurement Compound Value – Diastolic (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant iv. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant v. Field: Blood Pressure Measurement Compound Value – Systolic (kPa) <ul style="list-style-type: none"> • This field is not included vi. Field: Blood Pressure Measurement Compound Value – Diastolic (kPa) <ul style="list-style-type: none"> • This field is not included vii. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (kPa) <ul style="list-style-type: none"> • This field is not included viii. Field: Time Stamp <ul style="list-style-type: none"> • Format: Date and Time • Value: Not relevant ix. Field: Pulse Rate <ul style="list-style-type: none"> • Format: SFLOAT • Value: 110.0 x. Field: User ID <ul style="list-style-type: none"> • This field is not included xi. Field: Measurement Status <ul style="list-style-type: none"> • This field is not included <p>5. Check in PHG transcoder output for the Pulse rate object, Basic-Nu-Observed-Value attribute.</p> <p>6. The simulated PHD sends the measurement to the PHG under test with the following value:</p> <ul style="list-style-type: none"> a. Blood pressure measurement (0x2A35) <ul style="list-style-type: none"> i. Field: Flags <ul style="list-style-type: none"> • Format: 8 bit • Value: 0000 0110 (MSB → LSB). Blood pressure measurement value in units of mmHg, Time Stamp and Pulse Rate fields are included, User ID and measurement Status fields are not included ii. Field: Blood Pressure Measurement Compound Value – Systolic (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant iii. Field: Blood Pressure Measurement Compound Value – Diastolic (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant iv. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure
--	--

	<p>(mmHg)</p> <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant <p>v. Field: Blood Pressure Measurement Compound Value – Systolic (kPa)</p> <ul style="list-style-type: none"> • This field is not included <p>vi. Field: Blood Pressure Measurement Compound Value – Diastolic (kPa)</p> <ul style="list-style-type: none"> • This field is not included <p>vii. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (kPa)</p> <ul style="list-style-type: none"> • This field is not included <p>viii. Field: Time Stamp</p> <ul style="list-style-type: none"> • Format: Date and Time • Value: Not relevant <p>ix. Field: Pulse Rate</p> <ul style="list-style-type: none"> • Format: SFLOAT • Value: 07 FF (hex). Special value: NaN <p>x. Field: User ID</p> <ul style="list-style-type: none"> • This field is not included <p>xi. Field: Measurement Status</p> <ul style="list-style-type: none"> • This field is not included <p>7. Check in PHG transcoder output for the Pulse rate object, Basic-Nu-Observed-Value attribute.</p> <p>8. The simulated PHD sends the measurement to the PHG under test with the following value:</p> <p>a. Blood pressure measurement (0x2A35)</p> <p>i. Field: Flags</p> <ul style="list-style-type: none"> • Format: 8 bit • Value: 0000 0110 (MSB → LSB). Blood pressure measurement value in units of mmHg, Time Stamp and Pulse Rate fields are included, User ID and measurement Status fields are not included <p>ii. Field: Blood Pressure Measurement Compound Value – Systolic (mmHg)</p> <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant <p>iii. Field: Blood Pressure Measurement Compound Value – Diastolic (mmHg)</p> <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant <p>iv. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (mmHg)</p> <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant <p>v. Field: Blood Pressure Measurement Compound Value – Systolic (kPa)</p> <ul style="list-style-type: none"> • This field is not included <p>vi. Field: Blood Pressure Measurement Compound Value – Diastolic (kPa)</p> <ul style="list-style-type: none"> • This field is not included <p>vii. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (kPa)</p>
--	--

	<ul style="list-style-type: none"> • This field is not included <p>viii. Field: Time Stamp</p> <ul style="list-style-type: none"> • Format: Date and Time • Value: Not relevant <p>ix. Field: Pulse Rate</p> <ul style="list-style-type: none"> • Format: SFLOAT • Value: 08 00 (hex). Special value: NRes <p>x. Field: User ID</p> <ul style="list-style-type: none"> • This field is not included <p>xi. Field: Measurement Status</p> <ul style="list-style-type: none"> • This field is not included <p>9. Check in PHG transcoder output for the Pulse rate object, Basic-Nu-Observed-Value attribute.</p> <p>10. The simulated PHD sends the measurement to the PHG under test with the following value:</p> <p>a. Blood pressure measurement (0x2A35)</p> <p>i. Field: Flags</p> <ul style="list-style-type: none"> • Format: 8 bit • Value: 0000 0110 (MSB → LSB). Blood pressure measurement value in units of mmHg, Time Stamp and Pulse Rate fields are included, User ID and measurement Status fields are not included <p>ii. Field: Blood Pressure Measurement Compound Value – Systolic (mmHg)</p> <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant <p>iii. Field: Blood Pressure Measurement Compound Value – Diastolic (mmHg)</p> <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant <p>iv. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (mmHg)</p> <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant <p>v. Field: Blood Pressure Measurement Compound Value – Systolic (kPa)</p> <ul style="list-style-type: none"> • This field is not included <p>vi. Field: Blood Pressure Measurement Compound Value – Diastolic (kPa)</p> <ul style="list-style-type: none"> • This field is not included <p>vii. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (kPa)</p> <ul style="list-style-type: none"> • This field is not included <p>viii. Field: Time Stamp</p> <ul style="list-style-type: none"> • Format: Date and Time • Value: Not relevant <p>ix. Field: Pulse Rate</p> <ul style="list-style-type: none"> • Format: SFLOAT • Value: 07 FE (hex). Special value: +INFINITY <p>x. Field: User ID</p>
--	--

	<ul style="list-style-type: none"> • This field is not included xi. Field: Measurement Status <ul style="list-style-type: none"> • This field is not included <p>11. Check in PHG transcoder output for the Pulse rate object, Basic-Nu-Observed-Value attribute.</p> <p>12. The simulated PHD sends the measurement to the PHG under test with the following value:</p> <p>a. Blood pressure measurement (0x2A35)</p> <p>i. Field: Flags</p> <ul style="list-style-type: none"> • Format: 8 bit • Value: 0000 0110 (MSB → LSB). Blood pressure measurement value in units of mmHg, Time Stamp and Pulse Rate fields are included, User ID and measurement Status fields are not included <p>ii. Field: Blood Pressure Measurement Compound Value – Systolic (mmHg)</p> <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant <p>iii. Field: Blood Pressure Measurement Compound Value – Diastolic (mmHg)</p> <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant <p>iv. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (mmHg)</p> <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant <p>v. Field: Blood Pressure Measurement Compound Value – Systolic (kPa)</p> <ul style="list-style-type: none"> • This field is not included <p>vi. Field: Blood Pressure Measurement Compound Value – Diastolic (kPa)</p> <ul style="list-style-type: none"> • This field is not included <p>vii. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (kPa)</p> <ul style="list-style-type: none"> • This field is not included <p>viii. Field: Time Stamp</p> <ul style="list-style-type: none"> • Format: Date and Time • Value: Not relevant <p>ix. Field: Pulse Rate</p> <ul style="list-style-type: none"> • Format: SFLOAT • Value: 08 02 (hex). Special value: -INFINITY <p>x. Field: User ID</p> <ul style="list-style-type: none"> • This field is not included <p>xi. Field: Measurement Status</p> <ul style="list-style-type: none"> • This field is not included <p>13. Check in PHG transcoder output for the Pulse rate object, Basic-Nu-Observed-Value attribute.</p>
Pass/Fail criteria	<p>In step 5, the Pulse rate object, Basic-Nu-Observed-Value attribute is present and its value is 110.</p> <p>In step 7, the Pulse rate object, Basic-Nu-Observed-Value attribute is present and its value is 0x07FF.</p> <p>In step 9, the Pulse rate object, Basic -Nu-Observed-Value attribute is present and its value is</p>

	<p>0x0800.</p> <p>In step 11, the Pulse rate object, Basic -Nu-Observed-Value attribute is present and its value is 0x07FE.</p> <p>In step 13, the Pulse rate object, Basic -Nu-Observed-Value attribute is present and its value is 0x0802.</p>
<p>Notes</p>	<p>In step 5, possible values in typical points of observation after transcoder output are:</p> <p>a) IEEE 11073 Objects and Attributes</p> <p>Basic-Nu-Observed-Value attribute is present:</p> <ul style="list-style-type: none"> <input type="checkbox"/> Object: Pulse rate object <input type="checkbox"/> Attribute-id: MDC_ATTR_NU_VAL_OBS_BASIC (2636) <input type="checkbox"/> Attribute-type: SFLOAT <input type="checkbox"/> Attribute-value: F4 4C (hex) or 00 6E (hex) or 01 0B (hex) or 110 (dec) <p>b) WAN PCD-01 message</p> <p>PCD-01 message includes a segment like this with a Simple-Nu-Observed-Value attribute value (check OBX-5):</p> <pre style="margin-left: 40px;">OBX ? NM 149546^MDC_PULS_RATE_NON_INV^MDC 1.0.0.a 10 264864^MDC_DIM_BEAT_PER_MIN^MDC R [[current_date_time]</pre> <p>In step 7, possible values in typical points of observation after transcoder output are:</p> <p>a) IEEE 11073 Objects and Attributes</p> <p>Basic-Nu-Observed-Value attribute is present:</p> <ul style="list-style-type: none"> <input type="checkbox"/> Object: Pulse rate object <input type="checkbox"/> Attribute-id: MDC_ATTR_NU_VAL_OBS_BASIC (2636) <input type="checkbox"/> Attribute-type: SFLOAT <input type="checkbox"/> Attribute-value: 07 FF (hex) or NaN (note that a decimal value is not allowed) <p>b) WAN PCD-01 message</p> <p>PCD-01 message does not include segments with a Simple-Nu-Observed-Value attribute value (149546^MDC_PULS_RATE_NON_INV^MDC) because it has a special value and these values are not included in the PCD-01 message.</p> <p>In step 9, possible values in typical points of observation after transcoder output are:</p> <p>a) IEEE 11073 Objects and Attributes</p> <p>Basic-Nu-Observed-Value attribute is present:</p> <ul style="list-style-type: none"> <input type="checkbox"/> Object: Pulse rate object <input type="checkbox"/> Attribute-id: MDC_ATTR_NU_VAL_OBS_BASIC (2636) <input type="checkbox"/> Attribute-type: SFLOAT <input type="checkbox"/> Attribute-value: 08 00 (hex) or NRes (note that a decimal value is not allowed) <p>b) WAN PCD-01 message</p> <p>PCD-01 message does not include segments with a Simple-Nu-Observed-Value attribute value (149546^MDC_PULS_RATE_NON_INV^MDC) because it has a special value and these values are not included in the PCD-01 message.</p> <p>In step 11, possible values in typical points of observation after transcoder output are:</p> <p>a) IEEE 11073 Objects and Attributes</p> <p>Basic-Nu-Observed-Value attribute is present:</p> <ul style="list-style-type: none"> <input type="checkbox"/> Object: Pulse rate object <input type="checkbox"/> Attribute-id: MDC_ATTR_NU_VAL_OBS_BASIC (2636) <input type="checkbox"/> Attribute-type: SFLOAT <input type="checkbox"/> Attribute-value: 07 FF (hex) or +INFINITY (note that a decimal value is not allowed) <p>b) WAN PCD-01 message</p>

	<p>PCD-01 message does not include segments with a Simple-Nu-Observed-Value attribute value (149546^MDC_PULS_RATE_NON_INV^MDC) because it has a special value and these values are not included in the PCD-01 message.</p> <p>In step 13, possible values in typical points of observation after transcoder output are:</p> <p>a) IEEE 11073 Objects and Attributes</p> <p>Basic-Nu-Observed-Value attribute is present:</p> <ul style="list-style-type: none"> <input type="checkbox"/> Object: Pulse rate object <input type="checkbox"/> Attribute-id: MDC_ATTR_NU_VAL_OBS_BASIC (2636) <input type="checkbox"/> Attribute-type: SFLOAT <input type="checkbox"/> Attribute-value: 08 02 (hex) or -INFINITY (note that a decimal value is not allowed) <p>b) WAN PCD-01 message</p> <p>PCD-01 message does not include segments with a Simple-Nu-Observed-Value attribute value (149546^MDC_PULS_RATE_NON_INV^MDC) because it has a special value and these values are not included in the PCD-01 message.</p>
--	--

TP Id		TP/LP-PAN/PHG/PHDTW/BPM/BV-021		
TP label		Whitepaper. Pulse Rate measurement value		
Coverage	Spec	[b-Bluetooth PHDT v1.3]		
	Testable items	Short Float Type 1; C	Date-Time Conv 1; M	PR Numeric 6; M
		PR Numeric 7; M		
Test purpose		<p>Check that:</p> <p>PHG processes correctly the Pulse Rate Value (bpm) and Time Stamp fields of Blood Pressure Measurement</p>		
Applicability		C_MAN_BLE_000 AND C_MAN_BLE_003 AND C_MAN_BLE_005		
Other PICS				
Initial condition		The PHG under test and the simulated PHD are in the Standby state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated PHD is configured with a Blood pressure profile (device specialization); it has a measurement ready to be sent and it is in the Advertising state (it is discoverable). 2. The simulated PHD implements several BLE characteristics. The characteristic of interest for this test case is: <ol style="list-style-type: none"> a. Blood pressure measurement (0x2A35) 3. The PHG under test initiates a discovery process (Scanning state). It discovers the simulated PHD and it starts a pairing process with the simulated PHD (Initiating state). 4. When the pairing has been completed (Connection state), the simulated PHD sends the measurement to the PHG under test with the following value: <ol style="list-style-type: none"> a. Blood pressure measurement (0x2A35) <ol style="list-style-type: none"> i. Field: Flags <ul style="list-style-type: none"> • Format: 8 bit • Value: 0000 0110 (MSB → LSB). Blood pressure measurement value in units of mmHg, Time Stamp and Pulse Rate fields are included, User ID and measurement Status fields are not included ii. Field: Blood Pressure Measurement Compound Value – Systolic (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant iii. Field: Blood Pressure Measurement Compound Value – Diastolic (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant 		

	<ul style="list-style-type: none"> iv. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant v. Field: Blood Pressure Measurement Compound Value – Systolic (kPa) <ul style="list-style-type: none"> • This field is not included vi. Field: Blood Pressure Measurement Compound Value – Diastolic (kPa) <ul style="list-style-type: none"> • This field is not included vii. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (kPa) <ul style="list-style-type: none"> • This field is not included viii. Field: Time Stamp <ul style="list-style-type: none"> • Format: Date and Time • Value: August 2nd, 2012, 10:39:27 ix. Field: Pulse Rate <ul style="list-style-type: none"> • Format: SFLOAT • Value: 110.0 x. Field: User ID <ul style="list-style-type: none"> • This field is not included xi. Field: Measurement Status <ul style="list-style-type: none"> • This field is not included <p>5. Check that the PHG accepts the measurement and decodes its value properly (pulse rate measurement value, pulse rate units and time stamp).</p>
Pass/Fail criteria	In step 5, the PHG under test shows the following measurement Pulse Rate = 110 beats per minute (bpm) with the time stamp '2012-08-02 10:39:27'.
Notes	

TP Id		TP/LP-PAN/PHG/PHDTW/BPM/BV-022		
TP label		Whitepaper. Systolic/Diastolic/Map Numeric object, Measurement-Status Attribute		
Coverage	Spec	[Bluetooth PHDT v1.5]		
	Testable items	BP Numeric 11; M		
Test purpose		Check that: PHG transcodes Blood Pressure Measurement – Measurement Status field of Blood Pressure Measurement characteristic into Systolic/Diastolic/Map Numeric object, Measurement-Status attribute		
Applicability		C_MAN_BLE_000 AND C_MAN_BLE_002 AND C_MAN_BLE_003		
Other PICS				
Initial condition		The PHG under test and the simulated PHD are in the Standby state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated PHD is configured with a Blood Pressure Profile (device specialization), it has a measurement ready to be sent and it is in the Advertising state (it is discoverable). 2. The simulated PHD implements several BTLE characteristics. The characteristics of interest for this Test Case are: <ol style="list-style-type: none"> a. Blood Pressure Feature (0x2A49) b. Blood Pressure Measurement (0x2A35) 3. The PHG under test initiates a discovery process (Scanning state), it discovers the simulated PHD and it starts a pairing process with the simulated PHD (Initiating state). 		

4. When the pairing has been completed (Connection state) the simulated PHD sends the Measurement to the PHG under test with the following value:
 - a. Blood Pressure Feature (0x2A49)
 - Format: 16 bit
 - Value: 0000 0000 0000 0000 (MSB → LSB).
 - b. Blood Pressure Measurement (0x2A35)
 - i. Field: Flags
 - Format: 8 bit
 - Value: 0001 0010 (MSB → LSB). Blood Pressure Measurement Value in units of mmHg, Time Stamp and Measurement Status fields are included, Pulse Rate and User ID fields are not included.
 - ii. Field: Blood Pressure Measurement Compound Value – Systolic (mmHg)
 - Format: SFLOAT
 - Value: Not relevant
 - iii. Field: Blood Pressure Measurement Compound Value – Diastolic (mmHg)
 - Format: SFLOAT
 - Value: Not relevant
 - iv. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (mmHg)
 - Format: SFLOAT
 - Value: Not relevant
 - v. Field: Blood Pressure Measurement Compound Value – Systolic (kPa)
 - This field is not included
 - vi. Field: Blood Pressure Measurement Compound Value – Diastolic (kPa)
 - This field is not included
 - vii. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (kPa)
 - This field is not included
 - viii. Field: Time Stamp
 - Format: Date and Time
 - Value: Not relevant
 - ix. Field: Pulse Rate
 - This field is not included.
 - x. Field: User ID
 - This field is not included
 - xi. Field: Measurement Status
 - Format: 16 bit
 - Value: 0000 0000 0000 0000
5. Check in the PHG transcoder output the Systolic/Diastolic/Map Numeric object, Measurement-Status attribute
6. The simulated PHD sends a Measurement to the PHG under test with the following value:
 - a. Blood Pressure Feature (0x2A49)
 - Format: 16 bit
 - Value: 0000 0000 0000 0100 (MSB → LSB). Irregular Pulse Detection supported.
 - b. Blood Pressure Measurement (0x2A35)

	<ul style="list-style-type: none"> i. Field: Flags <ul style="list-style-type: none"> • Format: 8 bit • Value: 0001 0010 (MSB → LSB). Blood Pressure Measurement Value in units of mmHg, Time Stamp and Measurement Status fields are included, Pulse Rate and User ID fields are not included. ii. Field: Blood Pressure Measurement Compound Value – Systolic (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant iii. Field: Blood Pressure Measurement Compound Value – Diastolic (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant iv. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant v. Field: Blood Pressure Measurement Compound Value – Systolic (kPa) <ul style="list-style-type: none"> • This field is not included vi. Field: Blood Pressure Measurement Compound Value – Diastolic (kPa) <ul style="list-style-type: none"> • This field is not included vii. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (kPa) <ul style="list-style-type: none"> • This field is not included viii. Field: Time Stamp <ul style="list-style-type: none"> • Format: Date and Time • Value: Not relevant ix. Field: Pulse Rate <ul style="list-style-type: none"> • This field is not included. x. Field: User ID <ul style="list-style-type: none"> • This field is not included xi. Field: Measurement Status <ul style="list-style-type: none"> • Format: 16 bit • Value: 0000 0000 0000 0100 <p>7. Check in the PHG transcoder output the Systolic/Diastolic/Map Numeric object, Measurement-Status attribute</p>
Pass/Fail criteria	<p>In Step 5, the Systolic/Diastolic/Map object, Measurement-Status attribute is not present.</p> <p>In Step 7, the Systolic/Diastolic/Map object, Measurement-Status attribute is present and set to “questionable” (0x4000).</p>
Notes	<p>In Step 7, possible values in typical points of observation after transcoder output are:</p> <ul style="list-style-type: none"> a) IEEE 11073 Objects and Attributes <ul style="list-style-type: none"> Measurement-Status attribute is present: <ul style="list-style-type: none"> <input type="checkbox"/> Object: Systolic/Diastolic/Map Object <input type="checkbox"/> Attribute-id: MDC_ATTR_MSMT_STAT (2375) <input type="checkbox"/> Attribute-type: BITS16 <input type="checkbox"/> Attribute-value: 40 00 (hex) b) WAN PCD-01 message

	PCD-01 message does not include segments with Metric-Spec-Small attribute value.
--	--

TP Id		TP/LP-PAN/PHG/PHDTW/BPM/BV-023		
TP label		Whitepaper. Pulse Rate object, Measurement-Status Attribute		
Coverage	Spec	[Bluetooth PHDT v1.5]		
	Testable items	PR Numeric 8; M		
Test purpose		<p>Check that:</p> <p>PHG transcodes Blood Pressure Measurement – Measurement Status field of Blood Pressure Measurement characteristic into Pulse Rate object, Measurement-Status attribute</p>		
Applicability		C_MAN_BLE_000 AND C_MAN_BLE_002 AND C_MAN_BLE_003 AND C_MAN_BLE_005		
Other PICS				
Initial condition		The PHG under test and the simulated PHD are in the Standby state.		
Test procedure		<ol style="list-style-type: none"> 1. The simulated PHD is configured with a Blood Pressure Profile (device specialization), it has a measurement ready to be sent and it is in Advertising state (it is discoverable). 2. The simulated PHD implements several BTLE characteristics. The characteristics of interest for this Test Case are: <ol style="list-style-type: none"> a. Blood Pressure Feature (0x2A49) b. Blood Pressure Measurement (0x2A35) 3. The PHG under test initiates a discovery process (Scanning state), it discovers the Simulated PHD and it starts a pairing process with simulated PHD (Initiating state). 4. When the pairing has been completed (Connection state) the simulated PHD sends the Measurement to the PHG under test with the following value: <ol style="list-style-type: none"> a. Blood Pressure Feature (0x2A49) <ul style="list-style-type: none"> • Format: 16 bit • Value: 0000 0000 0000 0000 (MSB → LSB). b. Blood Pressure Measurement (0x2A35) <ol style="list-style-type: none"> i. Field: Flags <ul style="list-style-type: none"> • Format: 8 bit • Value: 0001 0110 (MSB → LSB). Blood Pressure Measurement Value in units of mmHg, Time Stamp, Pulse Rate and Measurement Status fields are included, User ID field is not included ii. Field: Blood Pressure Measurement Compound Value – Systolic (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant iii. Field: Blood Pressure Measurement Compound Value – Diastolic (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant iv. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant v. Field: Blood Pressure Measurement Compound Value – Systolic (kPa) <ul style="list-style-type: none"> • This field is not included vi. Field: Blood Pressure Measurement Compound Value – Diastolic (kPa) <ul style="list-style-type: none"> • This field is not included 		

	<ul style="list-style-type: none"> vii. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (kPa) <ul style="list-style-type: none"> • This field is not included viii. Field: Time Stamp <ul style="list-style-type: none"> • Format: Date and Time • Value: Not relevant ix. Field: Pulse Rate <ul style="list-style-type: none"> • Format: SFLOAT • Value: 110.0 x. Field: User ID <ul style="list-style-type: none"> • This field is not included xi. Field: Measurement Status <ul style="list-style-type: none"> • Format: 16 bit • Value: 0000 0000 0000 0000 <p>5. Check in the PHG transcoder output the Pulse Rate object, Measurement-Status attribute</p> <p>6. The simulated PHD sends a Measurement to the PHG under test with the following value:</p> <ul style="list-style-type: none"> a. Blood Pressure Feature (0x2A49) <ul style="list-style-type: none"> • Format: 16 bit • Value: 0000 0000 0000 0100 (MSB → LSB). Irregular Pulse Detection supported. b. Blood Pressure Measurement (0x2A35) <ul style="list-style-type: none"> i. Field: Flags <ul style="list-style-type: none"> • Format: 8 bit • Value: 0001 0110 (MSB → LSB). Blood Pressure Measurement Value in units of mmHg, Time Stamp, Pulse Rate and Measurement Status fields are included, User ID field is not included ii. Field: Blood Pressure Measurement Compound Value – Systolic (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant iii. Field: Blood Pressure Measurement Compound Value – Diastolic (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant iv. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (mmHg) <ul style="list-style-type: none"> • Format: SFLOAT • Value: Not relevant v. Field: Blood Pressure Measurement Compound Value – Systolic (kPa) <ul style="list-style-type: none"> • This field is not included vi. Field: Blood Pressure Measurement Compound Value – Diastolic (kPa) <ul style="list-style-type: none"> • This field is not included vii. Field: Blood Pressure Measurement Compound Value – Mean Arterial Pressure (kPa) <ul style="list-style-type: none"> • This field is not included viii. Field: Time Stamp <ul style="list-style-type: none"> • Format: Date and Time • Value: Not relevant
--	--

	<ul style="list-style-type: none"> ix. Field: Pulse Rate <ul style="list-style-type: none"> • Format: SFLOAT • Value: 110.0 x. Field: User ID <ul style="list-style-type: none"> • This field is not included xi. Field: Measurement Status <ul style="list-style-type: none"> • Format: 16 bit • Value: 0000 0000 0000 0100 <p>7. Check in the PHG transcoder output the Pulse Rate object, Measurement-Status attribute</p>
Pass/Fail criteria	<p>In Step 5, the Pulse Rate object, Measurement-Status attribute is not present.</p> <p>In Step 7, the Pulse Rate object, Measurement-Status attribute is present and set to "questionable" (0x4000).</p>
Notes	<p>In Step 7, possible values in typical points of observation after transcoder output are:</p> <ul style="list-style-type: none"> a) IEEE 11073 Objects and Attributes <p>Measurement-Status attribute is present:</p> <ul style="list-style-type: none"> <input type="checkbox"/> Object: Pulse Rate Object <input type="checkbox"/> Attribute-id: MDC_ATTR_MSMT_STAT (2375) <input type="checkbox"/> Attribute-type: BITS16 <input type="checkbox"/> Attribute-value: 40 00 (hex) b) WAN PCD-01 message <p>PCD-01 message does not include segments with Metric-Spec-Small attribute value.</p>

Bibliography

- [b-ITU-T H.810 (2013)] Recommendation ITU-T H.810 (2013), *Interoperability design guidelines for personal health systems*.
- [b-ITU-T H.810 (2015)] Recommendation ITU-T H.810 (2015), *Interoperability design guidelines for personal health systems*.
- [b-Bluetooth PHDT v1.3] Bluetooth SIG (2012), *Personal Health Devices Transcoding White Paper* (version 1.3)
https://www.bluetooth.org/docman/handlers/downloaddoc.ashx?doc_id=294540
- [b-CDG 1.0] Continua Health Alliance, Continua Design Guidelines v1.0 (2008), *Continua Design Guidelines*.
- [b-CDG 2010] Continua Health Alliance, Continua Design Guidelines v1.5 (2010), *Continua Design Guidelines*.
- [b-CDG 2011] Continua Health Alliance, Continua Design Guidelines (2011), "Adrenaline", *Continua Design Guidelines*.
- [b-CDG 2012] Continua Health Alliance, Continua Design Guidelines (2012), "Catalyst", *Continua Design Guidelines*.
- [b-CDG 2013] Continua Health Alliance, Continua Design Guidelines (2013), "Endorphin", *Continua Design Guidelines*.
- [b-CDG 2015] Continua Health Alliance, Continua Design Guidelines (2015), "Genome", *Continua Design Guidelines*.
- [b-CDG 2016] Personal Connected Health Alliance, Continua Design Guidelines (2016), "Iris", *Continua Design Guidelines*.
- [b-CDG 2017] Personal Connected Health Alliance, Continua Design Guidelines (2017), "Keratin", *Continua Design Guidelines*.
- [b-ETSI SR 001 262] ETSI SR 001 262 v1.8.1 (2003-12), *ETSI drafting rules*
<https://docbox.etsi.org/MTS/MTS/10-PromotionalMaterial/MBS-20111118/Referenced%20Documents/Drafting%20Rules.pdf>
- [b-PHD PICS & PIXIT] PHD PICS and PIXIT Test Tool v8.0.0.0 – Excel sheet v1.13.
<http://handle.itu.int/11.1002/2000/12067>
- [b-PHG PICS & PIXIT] PHG PICS and PIXIT Test Tool v8.0.0.0 – Excel sheet v1.11.
<http://handle.itu.int/11.1002/2000/12067>
- [b-TI] PHD Testable items. Test Tool v8.0.0.0 – Excel sheet v1.10.
<http://handle.itu.int/11.1002/2000/12067>

SERIES OF ITU-T RECOMMENDATIONS

Series A	Organization of the work of ITU-T
Series D	Tariff and accounting principles and international telecommunication/ICT economic and policy issues
Series E	Overall network operation, telephone service, service operation and human factors
Series F	Non-telephone telecommunication services
Series G	Transmission systems and media, digital systems and networks
Series H	Audiovisual and multimedia systems
Series I	Integrated services digital network
Series J	Cable networks and transmission of television, sound programme and other multimedia signals
Series K	Protection against interference
Series L	Environment and ICTs, climate change, e-waste, energy efficiency; construction, installation and protection of cables and other elements of outside plant
Series M	Telecommunication management, including TMN and network maintenance
Series N	Maintenance: international sound programme and television transmission circuits
Series O	Specifications of measuring equipment
Series P	Telephone transmission quality, telephone installations, local line networks
Series Q	Switching and signalling, and associated measurements and tests
Series R	Telegraph transmission
Series S	Telegraph services terminal equipment
Series T	Terminals for telematic services
Series U	Telegraph switching
Series V	Data communication over the telephone network
Series X	Data networks, open system communications and security
Series Y	Global information infrastructure, Internet protocol aspects, next-generation networks, Internet of Things and smart cities
Series Z	Languages and general software aspects for telecommunication systems